



## Toilet Support Instructions

**Models #4200, 4300, 4400 (Lo-Back)**  
**4600, 4700, 4800, 4900 (Hi-Back)**

### Thank you for Choosing Columbia

Your Columbia Toilet Support has been built to the highest standards of quality to ensure years of superior service. Please take the time to read this informative instruction guide and pay particular attention to the safety instructions provided. Thank you again for choosing Columbia!

### Important Safety Instructions

When using your Toilet Support, basic safety precautions should always be followed to reduce the risk of injury to persons.

1. Read and understand all instructions.
2. Determine, with a physician or therapist, the suitability of this support for the intended user.
3. Users with limited physical capabilities should be supervised or assisted when using this product.
4. Check that all locking knobs, buckles, and other parts are tight at all times.

## Save These Instructions

### Installing your Toilet Support

#### Step 1: Remove your Toilet Seat

Remove your toilet seat from your toilet bowl. This is done by unscrewing the nuts on the two bolts that hold your seat onto your toilet bowl. If the bolts are rusted, you may need to use new nuts and bolts when you replace your toilet seat.

(See Fig. 1)

#### Step 2: Mount the Base Plate

Set the base plate directly on your toilet bowl, as in **Figure 1**, so the holes in the plate line up with those in your toilet bowl. Note: you have a choice of using the front or rear pair of holes in the plate. You may need to use the rear pair in order to clear the water tank on your toilet. Otherwise you may use your choice of holes, to position the support more forward or rearward, depending on the seat depth that would be most comfortable for the user of the support.




Figure 1

Replace your toilet seat, on top of the base plate, so the two bolts you removed in **Step 1** go through the holes in the plate as well as into the holes in your toilet bowl. Tighten the nuts firmly on the two bolts.

### Step 3: Attach the Support Back

For use, slide the support back over the two upright posts on the base plate. When the back height is where you would like it, tighten the plastic locking knobs. To open the safety buckle, squeeze the top and bottom tabs together.

When not in use, remove the support back by loosening the plastic locking knobs and sliding the back off the posts.

### Cleaning Your Toilet Support

Scouring powders (Ajax, Comet, etc.) are recommended for cleaning the chrome and metal parts of your support. Cleaning sprays and liquids are best for cleaning the plastic back. Always keep the chrome parts dry to prevent deterioration of the finish.

### Recommended Accessories

#### Padded Toilet Seat Reducer Ring

**Models #4960, 4965**

If you have a slender or young child who may require extended toileting time, our Padded Toilet Seat Reducer Ring offers comfort and non-slip security. The padded reducer has a large abductor/splashguard that is padded and removable. The padded reducer fits any toilet size. There are two sizes available: Model #4960, which has a small hole size (7" x 8") for young, slender children, and Model #4965, which has an extra-small hole size (5" x 6") for toddler age children.

#### Padded Back Support Cushion

**Models #4205, 4305, 4405 (Lo-Back)  
4605, 4705, 4805, 4905 (Hi-Back)**

For users that require a soft support surface, or simply desire more comfort, padded back cushions are available. These cushions are made of seamless, contoured, waterproof urethane. Each cushion attaches to your toilet support back with self-stick Velcro® and is removable. When using a Padded Back, the inside width of each support back is reduced by one inch. Specify your Toilet Support Model # when ordering, as there is a cushion molded exactly for each size of support.

Columbia Medical  
13577 Larwin Circle  
Santa Fe Springs, CA 90670  
☎ (562) 282-0244  
Fax: (310) 305-1718  
E-mail: [info@columbiamedical.com](mailto:info@columbiamedical.com)  
Website: [www.columbiamedical.com](http://www.columbiamedical.com)  
© 2007 Columbia Medical Mfg., LLC

**COLUMBIA MEDICAL MFG., LLC warrants this product to be free of defects in material and workmanship. Our obligation under this warranty is limited to repair or replacement of any part or entire unit at our option for a period of one year from date of delivery to the original purchaser.**

**The warranty does not include cost of inconvenience, property damage, misuse, abuse, accident or similar incidents. The warranty will not apply if product has been tampered with or repaired by unauthorized individuals. This warranty gives you specific rights which vary from state to state.**