



Model **NE-U22V**  
**MicroAir Vibrating Mesh Nebulizer**

**CLEANING INSTRUCTIONS – EACH USE**

**⚠ WARNING**

**FAILURE TO PROPERLY CLEAN AND DISINFECT THE MICROAIR COMPONENTS WILL RESULT IN THE RISK OF BACTERIAL GROWTH, ILLNESS, AND SERIOUS INJURY FROM CONTAMINATION, COMPROMISED MEDICATION DELIVERY, AND POSSIBLE DAMAGE TO THE DEVICE.**

**⚠ CAUTION**

- Do **NOT** operate the device if not properly assembled.
- Do **NOT** operate the device if not working correctly.
- Do **NOT** rinse or immerse the Main Unit in any liquid.
- Do **NOT** wash or rinse any of the parts under the strong running water.
- Do **NOT** touch the mesh with your hand or any other object.

Following the cleaning instructions after each use will prevent any remaining medication in the bottle from drying, adhering to the mesh cap, and resulting in the device not nebulizing effectively.

**Wash the Medication Bottle, Mesh Cap, Mask, Mouthpiece, and Mask Adapter after each use.**

**It is important to clean your unit using distilled water after each use.**

1. Remove the Mouthpiece or Mask and the Mask Adapter from the Main Unit.
2. Remove the Medication Bottle from the Main Unit.
3. Open the Medication Bottle and discard any remaining medication.
4. Attach the Medication Bottle to the Main Unit. Open the Mesh Cap.
5. Pour a small amount of distilled water into the Medication Bottle and close the Mesh Cap.
6. Turn on the device to nebulize the distilled water for 1 to 2 minutes to remove residual medication from the mesh holes.
7. Turn off the device and remove the Medication Bottle from the Main Unit.
8. Remove the Mesh Cap from the Medication Bottle and discard any remaining distilled water from the Medication Bottle.
9. Rinse the Medication Bottle, Mesh Cap, Mask, Mouthpiece, and Mask Adapter with distilled water.
10. Gently wipe off excess water with a soft clean cloth or allow the parts to air dry in a clean environment.
11. Assemble the device. Store the device in the Storage Case or in a clean environment.

**OVER**



Model **NE-U22V**  
**MicroAir Vibrating Mesh Nebulizer**

**DISINFECTING INSTRUCTIONS – DAILY**

**⚠ WARNING**

**FAILURE TO PROPERLY CLEAN AND DISINFECT THE MICROAIR COMPONENTS WILL RESULT IN THE RISK OF BACTERIAL GROWTH, ILLNESS, AND SERIOUS INJURY FROM CONTAMINATION, COMPROMISED MEDICATION DELIVERY, AND POSSIBLE DAMAGE TO THE DEVICE.**

**⚠ CAUTION**

- Do **NOT** operate the device if not properly assembled.
- Do **NOT** operate the device if not working correctly.
- Do **NOT** rinse or immerse the Main Unit in any liquid.
- Do **NOT** wash or rinse any of the parts under the strong running water.
- Do **NOT** touch the mesh with your hand or any other object.

**Disinfect the Medication Bottle, Mesh Cap, Mask, Mouthpiece, and Mask Adapter after the last treatment of each day.**

**It is important to disinfect your unit on a daily basis.**

1. Make disinfecting solution of 1 part white vinegar and 3 parts distilled water.  
**⚠ Caution** Do **NOT** use household bleach. The mesh will rust.
2. Lift open the Mesh Cap and pour a small amount of disinfecting solution into the Medication Bottle.
3. Turn on the device to nebulize the disinfecting solution for 1 to 2 minutes.
4. Turn off the device and remove the Medication Bottle from the Main Unit.
5. Remove the Mesh Cap from the Medication Bottle and discard any remaining disinfecting solution from the Medication Bottle.
6. Soak the Medication Bottle, Mesh Cap, Mask, Mouthpiece, and Mask Adapter in the disinfecting solution for 10 to 15 minutes.
7. Rinse the Medication Bottle, Mesh Cap, Mask, Mouthpiece, and Mask Adapter with distilled water.
8. Gently wipe off excess water with a soft clean cloth or allow the parts to air dry in a clean environment.
9. Assemble the device. Store the device in the Storage Case or in a clean environment.