

Manual Supplement

Manual Title:	1652C/1653B/1654B Users Supplement Issue:	1
Print Date:	September 2010	Issue Date: 12/10
Revision/Date:		Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	1652C/1653B/1654B
CD Rev. & Date:	9/2010
CD PN:	3798444

Change #1, 56800

On page 3, add the following bullet:

- **Do not use in CAT III or CAT IV environments without the protective cap installed. The protective cap decreases the possibility of arc flash caused by short circuits.**

On page 6, under *Operating the Tester Using the Rotary Switch*, add the following:

⚠Warning

Do not use in CAT III or CAT IV environments without the protective cap installed. The protective cap decreases the exposed probe metal to <4mm. This decreases the possibility of arc flash from short circuits.

On page 45, under *General Specifications*, add the following to the Safety section:

Complies with EN/IEC 61010-031:2002+A1:2008.