

USER GUIDE

for

Automated purification of DNA from cells or tissue samples with

**KingFisher 96/ KingFisher mL instrument
and
MACHERY-NAGEL NucleoMag 96 Tissue kit**

10.11.2006



Description

Purification of DNA from cells or tissue (e.g. mouse tails) samples with MACHEREY-NAGEL's NucleoMag 96 Tissue kit can easily be automated using KingFisher® instruments (Thermo Electron Corporation). The KingFisher platforms utilize patented technology where magnetic rods move particles through the processing steps. KingFisher 96 instrument operates on microplates and can process up to 96 samples per run, whereas KingFisher mL can handle 15 samples per run. With both instruments the volumes handled can be up to 1 ml.

Typically, DNA isolation from 20 mg of mouse tail clippings using KingFisher instruments results up to 20 µg DNA. Generally, DNA yields vary according to sample type and storage conditions.

The protocol described here is designed for general use and can be modified according to customer individual needs using KingFisher® Software provided with the instrument.

Important notes

- See MACHEREY-NAGEL NucleoMag 96 Tissue kit user manual for reagent storage, product use limitations, safety information etc.
- Resuspend magnetic beads (DNA Binding Beads) thoroughly before use.

KingFisher 96 protocol

Importing protocols from the web

KingFisher Software protocol for MACHEREY-NAGEL NucleoMag 96 Tissue kit can be downloaded from the website (www.thermo.com/kingfisher). First you have to save the file "**NucleoMagTissueKF96**" to your computer.

1. Open KingFisher Software.
2. Select **Protocol** → **Import/Export data**.
3. Click **Read file**.
4. Select the database (*.KF2) by browsing in the **Open** dialog and click **Open**.
5. Select the protocol(s) you wish to import from the *Protocols in file* list. Use the SHIFT key together with the mouse button to select protocols between two clicked protocols and the CTRL key to select only the clicked protocols.
6. Tick **Update existing** if you wish to overwrite the protocols with identical protocol name(s) in the target database.
7. Click **Import**.

If there are protocols with identical names and you have not ticked the **Update existing** tick box, you will be prompted to change the name of the protocol that is being imported:

- Type in a new name and click **OK**.
 - o **Note:** Check that the name of the protocol does not exceed 17 characters.
- You will receive a message stating whether the database updating procedure was successful or not.

Sample preparation

- KingFisher DNA protocol **NucleoMag TissueKF96** is designed to purify DNA from tissue samples, cells or bacteria).
- Use Microtiter 96 DW plate (Catalog No. 95040450), KingFisher 96 tip comp for DW magnets (Catalog No. 97002534) and KingFisher 96 KF plate (Catalog No. 97002540) with NucleoMag 96 Tissue protocol.
- Add sample and other reagents except Elution Buffer supplied by NucleoMag 96 Tissue kit to KingFisher deepwell microplate according to table 1 and instructions below. Add the Elution Buffer to KingFisher 96 plate and start the NucleoMagTissue_KF96 process.

KingFisher 96 process

Table 1 Pipetting instructions for KingFisher 96 and NucleoMag 96 Tissue protocol.

Plate *	Plate	Content	Sample/ Reagent volume
A	1	Lysed Sample	225 µl
		Magnetic Beads (added after lysis step)	24 µl
		Binding buffer MB2 (added after lysis step)	360 µl
A	2	Wash buffer MB3	600 µl
A	3	Wash buffer MB4	600 µl
A	4	Wash buffer MB5	800 µl
B	5	Elution buffer MB6	100 µl

*A = Microtiter 96 DW plate, B = KingFisher 96 KF plate

1. Add 205 µl of buffer T1 and 20 µl of proteinase K solution to sample. Perform lysis at 56°C for 1h -16h depending on samples. Following lysis load lysed sample, 24 µl of resuspended Magnetic Beads and 360 µl Binding Buffer MB2 to plate 1.
2. Add 600 µl of Wash Buffer MB3 to plate 2.
3. Add 600 µl of Wash Buffer MB4 to plate 3.
4. Add 800 µl of Wash Buffer MB5 to plate 4.
5. Add 100 µl of Elution Buffer MB6 to plate 5.
6. Combine the tip comb and The KingFisher plate. See KingFisher 96 User manual.
7. Select the NucleoMag 96 Tissue protocol using arrow keys and press START button.
8. Load the plates according to protocol request and press START after every plate to confirm the action.
9. **Note!** Confirm that the plates are placed in correct orientation: A1 well to be pointed to upper right corner of the plate holder in turntable. A1 row of the plate is then always located in the inner circle of the turntable.
10. The purification protocol will start when the last plate is loaded and START button is pressed.
11. After lysis-step add 24 µl of resuspended Magnetic Beads and 360 µl of Binding Buffer MB2 to plate 1. Beads and Binding buffer may be premixed before addition to plate 1.
12. After the purification process is completed the plates are removed according to instructions shown in instrument screen.

Press START after each plate removal to confirm the action.

13. When the last plate is removed text End_of _run will appear. Press STOP to complete the run.

Description of NucleoMag 96 Tissue protocol with KingFisher 96

1. Samples are lysed with buffer T1 and proteinase K for 1-16 h at 56°C. Lysis incubation time depends on sample type. Refer NucleoMag 96 Tissue kit protocol for details. Following Lysis all further steps are done on the KingFisher 96 instrument.
2. Lysed samples are incubated first with magnetic beads and Binding Buffer MB2 in plate 1 for 5 minutes. Magnetic bead/DNA complexes are formed.
3. Magnetic beads are washed with Wash Buffer MB3 and MB4 in plates 2 and 3 respectively.
4. A final short wash step of magnetic beads with Wash buffer MB5 in plate 4 removes ethanol from previous wash buffers.
5. DNA is released to Elution Buffer in plate 5 for 10 minutes with heating.
6. Beads are discarded into plate 3.

KingFisher mL protocol

Importing protocol from the web

KingFisher mL Software protocol for NucleoMag 96 Tissue kit can also be downloaded from the website (www.thermo.com/kingfisher). Save the file "**NucleoMagTissueKF_mL**" to your computer.

Sample preparation

- KingFisher mL DNA protocol **NucleoMag TissueKFmL** is designed to purify DNA from tissue samples, cells or bacteria). Use KingFisher mL tubestrips and tip combs (Catalog No. 97002141) with **NucleoMagTissueKFmL** protocol.
- Add sample and other reagents supplied by **NucleoMag 96 Tissue** kit to KingFisher mL tubestrips according to table 2 and instructions below.

KingFisher mL process

Table 2 Pipetting instructions for KingFisher mL and NucleoMag 96 Tissue protocol.

Tube	Content	Sample/ Reagent volume
A	Lysed Sample	225 µl
	Magnetic Beads	24 µl
	Binding Buffer MB2	360 µl
B	Wash buffer MB3	600 µl
C	Wash buffer MB4	600 µl
D	Elution buffer MB5	800 µl
E	Elution buffer MB6	100 µl

1. Place an appropriate number of tube strips needed for the samples (one tube strip per sample) into removable tube strip tray.
2. Add 225 µl of lysed sample 24 µl of Magnetic Beads and 360 µl of Binding Buffer MB2 to tube strip **A**.
3. Add 600 µl of Wash Buffer MB3 and Wash Buffer MB4 to tube strip **B** and **C** respectively.
4. Add 800 µl of Wash Buffer MB5 to tube strip **D**.
5. Add 100 µ of Elution Buffer MB6 to tube strip **E**.

6. Close the front lid and start the process by selecting intended protocol **NucleoMagTissue_KFmL** using arrow keys and by pressing START.
7. Remove the tube strip tray from the KingFisher mL after program has completed.

Description of NucleoMag 96 Tissue protocol with KingFisher mL

1. Samples are lysed with buffer T1 and proteinase K for 1-16 h at 56°C. Lysis incubation time depends on sample type. Refer NucleoMag 96 Tissue kit protocol for details. Following lysis all further steps are done on the KingFisher ml instrument.
2. Lysed samples are incubated first with magnetic beads and Binding Buffer MB2 for 5 minutes in the tube strips A. Magnetic bead/DNA complexes are formed.
3. Magnetic beads are washed with Wash Buffer MB3 and MB4 in tube strips B and C respectively.
4. A final short wash step of magnetic beads with Wash buffer MB5 in tube strip D removes ethanol from previous wash buffers.
5. Sample is released to Elution Buffer MB6 in tube strip E. Optionally: The sample is then moved manually to 1.5 ml reaction tube and incubated in a heat block for 10 minutes to release the DNA. The sample is then moved back to tube strip E.
6. Beads are discarded into tube strip C.

Trouble shooting

1. Low yield
 - Elution buffer volume insufficient
 - Beads pellet must be covered completely with elution buffer
 - Partial elution in Wash Buffer MB5 already
 - Keep the beads on the magnet while washing in Wash Buffer MB5. Do not resuspend beads in this buffer, and do not incubate beads in this buffer for more than 2 min, as this buffer is water-based and might elute the DNA already.

2. Low purity
 - Poor sample quality
 - Check sample storage. Samples can be stored at 2 – 8°C for some days. Freeze samples if stored for longer periods.
 - Insufficient washing procedure
 - Use only recommended Deep-well blocks (see KingFisher User Manual).
 - Carry-over of ethanol from wash buffer MB4
 - Increase time for washing step with buffer MB5 to 2 min. Alternatively add additional air dry step (5 min). Be sure to remove all of the ethanolic Wash Buffer MB4, as residual ethanol interferes with downstream applications.

3. Any steps of the protocol (e.g. sample incubation and elution times) and the reagent volumes can be modified with KingFisher® software.

4. Tip comb was forgotten
 - Clean the magnetic rods using a soft cloth or tissue paper soaked in mild detergent solution, soap or alcohol.

5. The processor is not working properly
 - Refer to Kingfisher User Manual

Ordering Information

Product no.	Product Description
Thermo Electron	
540 05 00	KingFisher 96, 110V-240V, Magnetic particle processor
24073430	Magnet head for Deep Well plate
97002534	KingFisher 96 tip comb for DW magnets (10 x 10 pcs/box)
97002540	KingFisher 96 plate (200 µl), 48 plates/box
95040450	Deep Well 96 plate, V-bottom, Polypropylene
540 00 50	KingFisher mL, 110-240 V, Magnetic particle processor
97002131	KingFisher mL Combi 60 (tubes and tips for 60 samples)
97002141	KingFisher mL Combi 240 (tubes and tips for 240 samples)
97002111	KingFisher mL tip comb, 800 pcs
97002121	KingFisher mL tube, 900 pcs (20X45 pcs)
MACHEREY-NAGEL	
744 300.1	NucleoMag 96 Tissue kit (1x96 preps)
744 300.4	NucleoMag 96 Tissue kit (4x96 preps)
744 300.24	NucleoMag 96 Tissue kit (24x96 preps)

Contact information

ThermoFisher
SCIENTIFIC

Ratastie 2, P.O. Box 100
FIN-01621 Vantaa
Finland
Tel. +358-9-329100
Fax. +358-9-32910415
www.thermo.com



Neumann-Neander Str. 6-8
D-52355 Düren
Germany
T. +49 (0) 2421 969-0
F. +49 (0) 2421 969 199
www.mn-net.com, email: sales@mn-net.com