



PCR clean-up

User manual

NucleoMag[®] 96 PCR

May 2014 / Rev. 03

MACHEREY-NAGEL

www.mn-net.com



Table of contents

1	Components	4
1.1	Kit contents	4
1.2	Equipment and consumables to be supplied by user	4
2	Product description	5
2.1	The basic principle	5
2.2	Kit specifications	5
2.3	Magnetic separation systems	6
2.4	Adjusting the shaker settings	7
2.5	Handling of beads	7
2.6	Elution procedures	8
3	Storage conditions and preparation of working solutions	9
4	Safety instructions	10
5	Protocol for clean-up of 50 µL PCR reactions	11
6	Appendix	17
6.1	Troubleshooting	17
6.2	Ordering information	18
6.3	Product use restriction/warranty	19

1 Components

1.1 Kit contents

NucleoMag® 96 PCR			
REF	1x 96 preps 744100.1	4 x 96 preps 744100.4	24 x 96 preps 744100.24
NucleoMag® P-Beads	1.4 mL	5.6 mL	33.6 mL
Binding Buffer MP1	20 mL	80 mL	500 mL
Wash Buffer MP2	40 mL	160 mL	1000 mL
Wash Buffer MP3 (Concentrate)*	25 mL	100 mL	2 x 200 mL
Elution Buffer MP4	15 mL	30 mL	250 mL
Elution Plate U-bottom (including Self-adhering PE Foil)	1	4	24
User manual	1	1	1

1.2 Equipment and consumables to be supplied by user

Product	REF	Pack of
<ul style="list-style-type: none"> Magnetic separation system e.g., NucleoMag® SEP (see section 2.3) 	744900	1
<ul style="list-style-type: none"> Separation plate for magnetic beads separation, e.g., Square-well Block (96-well block with 2.1 mL square-wells) 	740481 740481.24	4 24

* For preparation of working solutions and storage conditions see section 3.

2 Product description

2.1 The basic principle

The **NucleoMag® 96 PCR** procedure is based on reversible adsorption of nucleic acids to paramagnetic beads under appropriate buffer conditions. Adjusting the PCR reaction to binding conditions and addition of paramagnetic beads can be carried out simultaneously. After magnetic separation and removal of supernatant, the beads are washed to remove contaminants and salt. A short drying step is necessary to remove ethanol from previous washing steps. Finally, highly purified DNA is eluted with low salt elution buffer or water and can be used directly for downstream applications. The **NucleoMag® 96 PCR** kit can be used either manually or automated on standard liquid handling instruments.

2.2 Kit specifications

- **NucleoMag® 96 PCR** is designed for rapid manual and automated clean-up of PCR fragments using the NucleoMag® SEP Magnetic Separator (see ordering information) other magnetic separation systems (see section 2.3). Manual processing time for 96 samples is about 45 min.
- **NucleoMag® 96 PCR** is easily adapted to common liquid handling instruments. The actual processing time and sample volume to be processed depends on the configuration of your instrument and the used magnetic separation system. Typically, 96 samples can be purified in about 30–45 min.
- The kit provides reagents for the purification of 3–5 µg DNA from 50 µL PCR samples.
- Typical concentration of the purified PCR samples of 75–200 ng/µL can be achieved.
- The purity of recovered PCR products is $A_{260}/A_{280} \geq 1.7$ –1.9.
- The kit is designed for use with or without detergent containing PCR buffers.
- Purified PCR products are ready-to-use for downstream applications, like automated fluorescent sequencing, labeling, microarray analysis, cloning, or restriction digestion.
- **NucleoMag® 96 PCR** can be processed completely at room temperature. Elution at 55 °C will increase the recovery by about 10–15 %.
- NucleoMag® P-Beads are highly reactive, superparamagnetic beads. The binding capacity is 0.3 µg of DNA per 1 µL of NucleoMag® P-Bead suspension, 1 µL of suspension contains 150 µg of beads.

2.3 Magnetic separation systems

For use of **NucleoMag® 96 PCR**, the use of the magnetic separator NucleoMag® SEP is recommended. Separation is carried out in a Square-well Block (see ordering information). The kit can also be used with other common separators.

Magnetic separator	Separation plate or tube
NucleoMag® SEP (MN REF 744900)	Square-well Block (MN REF 740481)
Tecan Te-MagS™	1.5 mL tubes without lid (Sarstedt)

Static magnetic pins

Separators with static magnetic pins, for example, NucleoMag® SEP (for manual use and for use on liquid handling workstations): This type of separator is recommended in combination with a suitable microplate shaker for optimal resuspension of the beads during the washing and elution steps. Alternatively, beads can be resuspended in the buffer by pipetting up and down several times. For fully-automated use on liquid handling workstations, a gripper tool is required, the plate is transferred to the magnetic separator for separation of the beads and transferred to the shaker module for resuspension of the beads.

Movable magnetic systems

Separators with moving magnetic pins: Magnetic pins/rods are moved from one side of the well to the other and vice versa. Beads follow this movement and are thus pulled through the buffer during the wash and elution steps. Separation takes place when the system stops.

Automated separators

Separators with moving magnets: Magnetic beads are transferred into suitable plates or tubes. Beads are resuspended from the rod-covered magnets. Following binding, washing or elution beads are collected again with the rod-covered magnets and transferred to the next plate or tube.

2.4 Adjusting the shaker settings

When using a plate shaker for the washing and elution steps, the speed settings have to be adjusted carefully for each specific separation plate and shaker to prevent cross-contamination from well to well. Proceed as follows:

Adjusting shaker speed for binding and wash steps:

- Load 300 µL dyed water (select to the wells of the separation plate. Place the plate on the shaker and start shaking with a moderate speed setting for 30 seconds. Turn off the shaker and check the plate surface for small droplets of dyed water.
- Increase speed setting, shake for an additional 30 seconds, and check the plate surface for droplets again.
- Continue increasing the speed setting until you observe droplets on top of the separation plate. Reduce speed setting, check again, and use this setting for the washing step.

Adjusting shaker speed for the elution step:

- Load 100 µL dyed water to the wells of the collection plate and proceed as described above.

2.5 Handling of beads

Distribution of beads

A homogeneous distribution of the magnetic beads to the individual wells of the separation plate is essential for a high well-to-well consistency. Therefore, before distributing the beads, make sure that the beads are completely resuspended. Shake the storage bottle well or place it on a vortexer shortly. Premixing magnetic beads with the binding buffer allows easier homogenous distribution of the beads to the individual wells of the separation plate. During automation, a premix step before aspirating the beads/binding buffer mixture from the reservoir is recommended to keep the beads resuspended.

Magnetic separation time

Attraction of the magnetic beads to the magnetic pins depends on the magnetic strength of the magnetic pins, the selected separation plate, distance of the separation plate from the magnetic pins, and the volume to be processed. The individual times for complete attraction of the beads to the magnetic pins should be checked and adjusted on each system. It is recommended using the separation plates or tubes specified by the supplier of the magnetic separator.

Washing the beads

Washing the beads can be achieved by shaking or mixing. In contrast to mixing by pipetting up and down, mixing by shaker or magnetic mixing allows simultaneous mixing of all samples. This reduces the time and number of tips needed for the preparation.

Resuspension by pipetting up and down, however, is more efficient than mixing by a shaker or magnetic mix.

Method	Resuspension efficiency	Speed	Number of tips needed
Magnetic mix	+	++	Low
Shaker	++	++	Low
Pipetting	+++	+*	High

+: acceptable, ++: good, +++: excellent, * 8-channel pipetting device

2.6 Elution procedures

Purified DNA can be eluted directly with the supplied Elution Buffer MP4 or water (pH 7.5–8.5). Elution can be carried out in a volume of $\geq 25 \mu\text{L}$ (per $12 \mu\text{L}$ bead suspension). It is essential to cover the NucleoMag® P-Beads completely with elution buffer during the elution step. The volume of dispensed elution buffer depends on the magnetic separation system (e.g., the position of the pellet inside the separation plate). For efficient elution, the magnetic bead pellet should be resuspended completely in the elution buffer. For some separators, high elution volumes might be necessary to cover the whole pellet.

Elution is possible at room temperature. Yield can be increased by 10–15 % if elution is performed at 55 °C.

* 8-channel pipetting device

3 Storage conditions and preparation of working solutions

Attention: Buffers MP1 and MP2 contain chaotropic salt! Wear gloves and goggles!

CAUTION: Buffers MP1 and MP2 contain guanidine hydrochloride which can form highly reactive compounds when combined with bleach (sodium hypochlorite). DO NOT add bleach or acidic solutions directly to the sample-preparation waste.

Storage conditions:

- All components of the **NucleoMag® 96 PCR** kit should be stored at room temperature (18–25 °C) and are stable for up to one year.

Before starting any **NucleoMag® 96 PCR** protocol, prepare the following:

- **Wash Buffer MP3:** Add the indicated volume of ethanol (96–100%) to Buffer MP3 Concentrate. Mark the label of the bottle to indicate that ethanol was added. Wash Buffer MP3 is stable at room temperature (18–25 °C) for at least one year.

NucleoMag® 96 PCR			
	1 x 96 preps	4 x 96 preps	24 x 96 preps
REF	744100.1	744100.4	744100.24
Buffer MP3 (Concentrate)	25 mL Add 100 mL ethanol	100 mL Add 400 mL ethanol	2 x 200 mL Add 800 mL ethanol to each vial

4 Safety instructions



The following components of the **NucleoMag® 96 PCR** kits contain hazardous contents.

Wear gloves and goggles and follow the safety instructions given in this section.

GHS classification

Only harmful features do not need to be labeled with H and P phrases up to 125 mL or 125 g.

Mindergefährliche Eigenschaften müssen bis 125 mL oder 125 g nicht mit H- und P-Sätzen gekennzeichnet werden.

Component	Hazard contents	GHS symbol	Hazard phrases	Precaution phrases
Inhalt	Gefahrstoff	GHS Symbol	H-Sätze	P-Sätze
MP1	Guanidine hydrochloride 36–50 % <i>Guanidinhydrochlorid 36–50 %</i>	 Warning <i>Achtung</i>	302, 319	280, 301+312, 305+351+338, 330, 337+313
MP2	Guanidine hydrochloride 24–36 % <i>Guanidinhydrochlorid 24–36 %</i>	 Warning <i>Achtung</i>	302	301+312, 330

Hazard phrases

- H 302 Harmful if swallowed.
Gesundheitsschädlich bei Verschlucken.
- H 319 Causes serious eye irritation.
Verursacht schwere Augenreizung.

Precaution phrases

- P 280 Wear protective gloves / eye protection.
Schutzhandschuhe / Augenschutz tragen.
- P 301+312 IF SWALLOWED: Call a POISON CENTER/ doctor/.../if you feel unwell.
BEI VERSCHLUCKEN: Bei Unwohlsein GIFTINFORMATIONSZENTRUM / Arzt /... anrufen.
- P 305+351+313 IF IN EYES: Rinse continuously with water for several minutes. Remove contact lenses if present and easy to do – continue rinsing
BEI KONTAKT MIT DEN AUGEN: Einige Minuten lang behutsam mit Wasser spülen. Vorhandene Kontaktlinsen nach Möglichkeit entfernen. Weiter ausspülen.
- P 330 Rinse mouth.
Mund ausspülen.
- P 337+313 Get medical advice / attention.
Bei anhaltender Augenreizung: Ärztliche Rat einholen / ärztliche Hilfe hinzuziehen.

For further information please see Material Safety Data Sheets (www.mn-net.com).
Weiterführende Informationen finden Sie in den Sicherheitsdatenblättern (www.mn-net.com).







5 Protocol for clean-up of 50 µL PCR reactions









Protocol-at-a-glance

- For additional equipment and hardware requirements, refer to section 1.2 and 2.3, respectively.
- For detailed information on each step, see page 17.

Before starting the preparation:

- Check if Buffer MP3 was prepared according to section 3.


1 Transfer PCR reaction mixture to appropriate 96-well plate	50 µL PCR reaction For PCR reactions < 50 µL, adjust the volume to 50 µL using sterile water	
2 Bind DNA to NucleoMag® P-Beads	12 µL NucleoMag® P-Beads 138 µL MP1	
	Mix by shaking for 5 min at RT <i>(Optional: Mix by pipetting up and down)</i>	
	Remove supernatant after 2 min separation	
3 Wash with MP2	Remove Square-well Block from NucleoMag® SEP 300 µL MP2	
	Resuspend: Shake 5 min at RT <i>(Optional: Mix by pipetting up and down)</i>	

		Remove supernatant after 2 min separation	
4	1st wash with MP3	Remove Square-well Block from NucleoMag® SEP 300 µL MP3	
		Resuspend: Shake 5 min at RT <i>(Optional: Mix by pipetting up and down)</i>	
		Remove supernatant after 2 min separation	
5	2nd wash with MP3	Remove Square-well Block from NucleoMag® SEP 300 µL MP3	
		Resuspend: Shake 5 min at RT <i>(Optional: Mix by pipetting up and down)</i>	
		Remove supernatant after 2 min separation	
6	Dry the beads	10 min at RT	
7	Elute DNA	Remove Square-well Block from NucleoMag® SEP 25–100 µL MP4 <i>(Optional: Elute at 55 °C)</i>	

Shake 5 min at RT
*(Optional: Mix by pipetting
up and down)*



**Separate 2 min and transfer
DNA into Elution Plate
U-bottom**



Detailed protocol

This protocol is designed for magnetic separators with static pins (e.g., NucleoMag® SEP) and suitable plate shakers (see section 2.3). It is recommended using a Square-well Block for separation (see section 1.2). Alternatively, isolation of DNA can be performed in reaction tubes with suitable magnetic separators. This protocol is for manual use and serves as a guideline for adapting the kit to robotic instruments.

Before starting the preparation:

- Check if Buffer MP3 was prepared according to section 3.
-

1 Transfer PCR reaction mixture

Transfer PCR reaction mixture to an appropriate 96-well plate.

For PCR reaction volumes < 50 µL, adjust the volume to 50 µL using sterile water.

Note: See recommendations for suitable plates (e.g., Square-well Block; not included in the kit) and compatible magnetic separators section 2.3.

2 Bind DNA to NucleoMag® P-Beads

Add **12 µL NucleoMag® P-Beads** and **138 µL Binding Buffer MP1** to each well of the separation plate. Mix by pipetting up and down 6 times and **shake** for **5 min** at **room temperature**. Alternatively, when processing the kit without a shaker, pipette up and down 10 times and incubate for 5 min at room temperature.

Note: NucleoMag® P-Beads and Binding Buffer MP1 may be premixed. For 96 samples, premix 1248 µL of NucleoMag® P-Beads with 14.35 mL of Buffer MP1, mix by vortexing. Use 150 µL of the suspension per well. Be sure to resuspend the NucleoMag® P-Beads before removing them from the storage bottle. Vortex storage bottle briefly until a homogenous suspension has been formed.

Separate the magnetic beads against the side of the wells by placing the 96-well plate on the NucleoMag® SEP magnetic separator. Wait at least **2 min** until all the beads have been attracted to the magnets. Remove and discard supernatant by pipetting.

Note: Do not disturb the attracted beads while aspirating the supernatant. The magnetic pellet is not visible in this step. Remove supernatant from the opposite side of the well.

3 Wash with MP2

Remove the Square-well Block from the NucleoMag® SEP magnetic separator.

Add **300 µL Buffer MP2** to each well and resuspend the beads by shaking until the beads are resuspended completely (**5 min**). Alternatively, resuspend beads completely by repeated pipetting up and down (15 times).

Separate the magnetic beads by placing the Square-well Block on the NucleoMag® SEP magnetic separator. Wait at least **2 min** until all the beads have been attracted to the magnet. Remove and discard supernatant by pipetting.

4 1st wash with MP3

Remove the Square-well Block from the NucleoMag® SEP magnetic separator.

Add **300 µL Buffer MP3** to each well and resuspend the beads by shaking until the beads are resuspended completely (**5 min**). Alternatively, resuspend beads completely by repeated pipetting up and down (15 times).

Separate the magnetic beads by placing the Square-well Block on the NucleoMag® SEP magnetic separator. Wait at least **2 min** until all the beads have been attracted to the magnet. Remove and discard supernatant by pipetting.

5 2nd wash with MP3

Remove the Square-well Block from the NucleoMag® SEP magnetic separator.

Add **300 µL Buffer MP3** to each well and resuspend the beads by shaking until the beads are resuspended completely (**5 min**). Alternatively, resuspend beads completely by repeated pipetting up and down (15 times).

Separate the magnetic beads by placing the Square-well Block on the NucleoMag® SEP magnetic separator. Wait at least **2 min** until all the beads have been attracted to the magnet. Remove and discard supernatant by pipetting.

6 Dry the beads

Dry the beads by incubating the Square-well Block **10 min** at **room temperature** with the particles held against the magnet in order to allow the remaining traces of alcohol to evaporate.

Note: Allow the pellet to dry sufficiently so that there is no visible droplets of buffer in the bottom of the tube. Allowing the pellet to dry completely, indicated by visible cracking. Do not overdry beads (e.g., by prolonged drying at 55 °C). This will reduce yield.

7 Elute DNA

Remove the Square-well Block from the NucleoMag® SEP magnetic separator.

Add desired volume of **Elution Buffer MP4 (25–100 µL)** to each well and resuspend the by shaking **5–10 min** at **56 °C**. Alternatively, resuspend beads completely by repeated pipetting up and down and incubate for **5–10 min** at room temperature.

Separate the magnetic beads by placing the Square-well Block on the NucleoMag® SEP magnetic separator. Wait at least **2 min** until all the beads have been attracted to the magnets. Transfer the supernatant containing the purified PCR products to the Elution Plate U-bottom.

Note: Yield can be increased by 15–20 % by using pre-warmed elution buffer (55 °C) or by incubating the bead/elution buffer suspension at 55 °C for 5 min.

6 Appendix

6.1 Troubleshooting

Problem	Possible cause and suggestions
Poor DNA yield	<p><i>Wash Buffer MP3 did not contain ethanol</i></p> <ul style="list-style-type: none"> Addition of the the indicated volume of 96–100% ethanol to Buffer MP3 Concentrate is required before use.
	<p><i>Elution buffer volume insufficient</i></p> <ul style="list-style-type: none"> Bead pellet must be covered completely with elution buffer.
	<p><i>Insufficient performance of elution buffer during elution step</i></p> <ul style="list-style-type: none"> Remove residual wash buffers during the separation steps completely. Remaining buffers decrease efficiency of following wash steps and elution step.
	<p><i>Beads overdried</i></p> <ul style="list-style-type: none"> Do not dry beads longer than 20 min at room temperature. Overdrying of beads may result in lower elution efficiencies.
Suboptimal performance of DNA in downstream applications	<p><i>Carry-over of ethanol from Wash Buffer MP3</i></p> <ul style="list-style-type: none"> Be sure to remove all of the ethanolic Wash Buffer MP3 after the final wash step. Dry beads 10–15 min at room temperature.
	<p><i>Elution of DNA with TE buffer</i></p> <ul style="list-style-type: none"> Use supplied elution buffer or sterile water. Do not use TE buffer. EDTA may inhibit sequencing reactions. Repurify or precipitate DNA by ethanol and elute/redissolve in Elution Buffer MP 4 buffer or water.
	<p><i>Eluted DNA contains residual primers/primer dimers</i></p> <ul style="list-style-type: none"> Minimize amount of primers in PCR reaction mixture. Do not use higher volumes of binding buffer than specified.

Problem	Possible cause and suggestions
Carry-over of beads	<i>Time for magnetic separation too short</i>
	<ul style="list-style-type: none"> • Increase separation time to allow the beads to be attracted to the magnetic pins completely.
	<i>Aspiration speed too high (elution step)</i>
	<ul style="list-style-type: none"> • High aspiration speeds during the elution step may cause bead carry-over. Reduce aspiration speed for elution step.

6.2 Ordering information

Product	REF	Pack of
NucleoMag® 96 PCR	744100.1	1 x 96 preps
	744100.4	4 x 96 preps
	744100.24	24 x 96 preps
NucleoMag® SEP	744900	1
Square-well Blocks	740481	4
	740481.24	24
Elution Plate U-bottom	740486.24	24
Self-adhering PE Foil	740676	50 sheets

Visit www.mn-net.com for more detailed product information.

6.3 Product use restriction/warranty

NucleoMag® 96 PCR kit components are intended, developed, designed, and sold FOR RESEARCH PURPOSES ONLY, except, however, any other function of the product being expressly described in original MACHEREY-NAGEL product leaflets.

MACHEREY-NAGEL products are intended for GENERAL LABORATORY USE ONLY! MACHEREY-NAGEL products are suited for QUALIFIED PERSONNEL ONLY! MACHEREY-NAGEL products shall in any event only be used wearing adequate PROTECTIVE CLOTHING. For detailed information please refer to the respective Material Safety Data Sheet of the product! MACHEREY-NAGEL products shall exclusively be used in an ADEQUATE TEST ENVIRONMENT. MACHEREY-NAGEL does not assume any responsibility for damages due to improper application of our products in other fields of application. Application on the human body is STRICTLY FORBIDDEN. The respective user is liable for any and all damages resulting from such application.

DNA/RNA/PROTEIN purification products of MACHEREY-NAGEL are suitable for *IN VITRO*-USES ONLY!

ONLY MACHEREY-NAGEL products specially labeled as IVD are also suitable for *IN VITRO*-diagnostic use. Please pay attention to the package of the product. *IN VITRO*-diagnostic products are expressly marked as IVD on the packaging.

IF THERE IS NO IVD SIGN, THE PRODUCT SHALL NOT BE SUITABLE FOR *IN VITRO*-DIAGNOSTIC USE!

ALL OTHER PRODUCTS NOT LABELED AS IVD ARE NOT SUITED FOR ANY CLINICAL USE (INCLUDING, BUT NOT LIMITED TO DIAGNOSTIC, THERAPEUTIC AND/OR PROGNOSTIC USE).

No claim or representations is intended for its use to identify any specific organism or for clinical use (included, but not limited to diagnostic, prognostic, therapeutic, or blood banking). It is rather in the responsibility of the user or - in any case of resale of the products - in the responsibility of the reseller to inspect and assure the use of the DNA/RNA/protein purification products of MACHEREY-NAGEL for a well-defined and specific application.

MACHEREY-NAGEL shall only be responsible for the product specifications and the performance range of MN products according to the specifications of in-house quality control, product documentation and marketing material.

This MACHEREY-NAGEL product is shipped with documentation stating specifications and other technical information. MACHEREY-NAGEL warrants to meet the stated specifications. MACHEREY-NAGEL's sole obligation and the customer's sole remedy is limited to replacement of products free of charge in the event products fail to perform as warranted. Supplementary reference is made to the general business terms and conditions of MACHEREY-NAGEL, which are printed on the price list. Please contact us if you wish to get an extra copy.

There is no warranty for and MACHEREY-NAGEL is not liable for damages or defects arising in shipping and handling (transport insurance for customers excluded), or out of accident or improper or abnormal use of this product; defects in products or

components not manufactured by MACHEREY-NAGEL, or damages resulting from such non-MACHEREY-NAGEL components or products.

MACHEREY-NAGEL makes no other warranty of any kind whatsoever, and SPECIFICALLY DISCLAIMS AND EXCLUDES ALL OTHER WARRANTIES OF ANY KIND OR NATURE WHATSOEVER, DIRECTLY OR INDIRECTLY, EXPRESS OR IMPLIED, INCLUDING, WITHOUT LIMITATION, AS TO THE SUITABILITY, REPRODUCTIVITY, DURABILITY, FITNESS FOR A PARTICULAR PURPOSE OR USE, MERCHANTABILITY, CONDITION, OR ANY OTHER MATTER WITH RESPECT TO MACHEREY-NAGEL PRODUCTS.

In no event shall MACHEREY-NAGEL be liable for claims for any other damages, whether direct, indirect, incidental, compensatory, foreseeable, consequential, or special (including but not limited to loss of use, revenue or profit), whether based upon warranty, contract, tort (including negligence) or strict liability arising in connection with the sale or the failure of MACHEREY-NAGEL products to perform in accordance with the stated specifications. This warranty is exclusive and MACHEREY-NAGEL makes no other warranty expressed or implied.

The warranty provided herein and the data, specifications and descriptions of this MACHEREY-NAGEL product appearing in MACHEREY-NAGEL published catalogues and product literature are MACHEREY-NAGEL's sole representations concerning the product and warranty. No other statements or representations, written or oral, by MACHEREY-NAGEL's employees, agent or representatives, except written statements signed by a duly authorized officer of MACHEREY-NAGEL are authorized; they should not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

Product claims are subject to change. Therefore please contact our Technical Service Team for the most up-to-date information on MACHEREY-NAGEL products. You may also contact your local distributor for general scientific information. Applications mentioned in MACHEREY-NAGEL literature are provided for informational purposes only. MACHEREY-NAGEL does not warrant that all applications have been tested in MACHEREY-NAGEL laboratories using MACHEREY-NAGEL products. MACHEREY-NAGEL does not warrant the correctness of any of those applications.

Last updated: 07/2010, Rev. 03

Please contact:
MACHEREY-NAGEL GmbH & Co. KG
Tel.: +49 24 21 969-270
tech-bio@mn-net.com

Trademarks:

KingFisher is a registered trademark of Thermo Fisher Scientific
NucleoMag is a registered trademark of MACHEREY-NAGEL GmbH & Co KG
Te-MagS is a trademark of Tecan Group Ltd., Switzerland

All used names and denotations can be brands, trademarks, or registered labels of their respective owner – also if they are not special denotation. To mention products and brands is only a kind of information (i.e., it does not offend against trademarks and brands and can not be seen as a kind of recommendation or assessment). Regarding these products or services we can not grant any guarantees regarding selection, efficiency, or operation.