

Office of The Chief General Manager-Materials Management NEYVELI LIGNITE CORPORATION LIMITED

(A Government of India Enterprise)



MATERIALS MANAGEMENT COMPLEX
Dr. Rajendra Prasad Road , Block - 27, Neyveli - 607 807. Tamilnadu, India.
Phone: 0091-4142-268268 Fax: 269 458/269197Grams: EMCEEPUR



TENDER ENQUIRY NO: ENQ/14-15/000730/MM15(01)

DATE: 20/05/2014

Page - 1

Due Date of Cover-I Opening: 02/06/2014 15:00:00 hrs. IST (Check NLC eproc website for any change in due date of opening.) Note: Quotation should be valid for 60 days from the date of opening.

Gentlemen,

We shall be glad to receive your lowest online offer submitted through http://eproc.nlcindia.com for the material mentioned below, subject to our terms and conditions. Any other mode of offer will be rejected. Please indicate the enquiry number in all future correspondence. Offer details of the participating bidders will be available online after the tender is opened.

bidders will be available online after the tender is opened.								
Sl No.			Unit of Measurement	Quantity				
1	31 015 0	056 : DIGITAL CLAM	Number	10.0000				
1	SL No.	OSO: DIGITAL CLAW	Number	10.0000				
	1	Make and model for the	Technical Factor(s) coffered item may be mentioned.					
	2	Digital clamp meter 3 1						
	3	Range selection : Auto						
	4 DC Voltage : up to 1000 V or Better							
	5 Accuracy: +/- (1.0 % rdg +3 dgt) or better 6 AC Voltage: up to 750 V or Better							
	7	Accuracy: +/- (1.5 % r	dg + 3 dgt) or better					
	8 DC Current :up to 1000 A or Better							
	9 Accuracy: +/- (2%rdg + 8 dgt) or better							
	10 AC Current :up to 1000 A or Better							
	11 Accuracy: +/- (3% rdg + 5 dgt) or better. 12 Ressistance: 0.1 Ohm to 4 K Ohm or Better							
	13	Accuracy : +/- (1.5 % ro	dg + 5 dgt) or better.					
	14	The following features sa) Continuty tester						
	15 b) True RMS 16 c) Data Hold.							
	17	Accessories : 1) Carrying case.						
	18 2) User Manual							
	19	3) Test Lead						
	20	4) Battery						
	21	Bidder should attach the technical leaflet for the offered Make and Model along with the offer. otherwise offer will be liable for rejection.						
	22	ADC: In case of being Certificate from the m	rrent Valid Authorized Dealership e bid offer.					
	Delivery Type		Period From PO Date (Days)	Quantity				
		bulk	0 to 90	10.0000				
	Manufact	urer's test certificate shall						
			- ***					
					I	I		

YOUR OFFER SHOULD BE ON DOOR DELIVERY BASIS AT NLC STORES NEYVELI

Inspection Clause: Inspection at Neyveli

Guarantee Requirement: General Guarantee for Warranty as per NLC format Warranty Period: 12 months from the date of commissioning or 18 months from the date of receipt of material, whichever is earlier.

Attachemnts: Terms and Conditions of Enquiry



Office of The Chief General Manager-Materials Management NEYVELI LIGNITE CORPORATION LIMITED (A Government of India Enterprise)



MATERIALS MANAGEMENT COMPLEX
Dr. Rajendra Prasad Road, Block - 27, Neyveli - 607 807. Tamilnadu, India.
Phone: 0091-4142-268268 Fax: 269 458/269197Grams: EMCEEPUR

DATE: 20/05/2014 TENDER ENQUIRY NO: ENQ/14-15/000730/MM15(01) Page - 2

RINA (Net

Sl No.		Details	Unit of Measurement	Quantity
		Specification		
2	Specificat	061 : 4 1/2 DIGIT - DIGITAL MULTIMETER.	Number	15.0000
	1).Basic I	DC Voltage Accuracy: 0.1 % rdg. or better.		
	2).AC/DC	C Voltage: upto 1000 V .		
	3).AC/DC	C Current: upto 10 A.		
	5). Freque	nnce Range: upto 40 M . ohm ency: upto 50 KHz. Auto Ranging.		
	7).Diode	Test: Yes		
	8). Contir	nuity Test: Yes.		
	9). Data	Hold: Yes		
	10). Low	Battery Indication: Yes		
		load Warning: Yes.		
		Power Off: Yes.		
		Lit Display: Yes.		
	14). Acces i).Carryin ii). Fuses	ssories: ng Case with Probe(test leads 1 pair)		
	SL No.	Technical Factor(s)		
	1	4 1/2 Digital Display		
	2	AC/DC Voltage : Up to 1000V		
	3	AC/DC Current: upto 10A		
	4	Basic DC Voltage accuracy: 0.1 % rdg or better		
	5	Resistance range: upto 40M ohm		
	6	Frequency: upto 50 KHz		
	7	Range: Auto Ranging		
	8	THe following features shall be provided. Diode Test		
	9	Continuity Test		
	10	Data Hold Low Potters Indication		
	12	Low Battery Indication Overload warning		
	14	Auto power off		
	15	Back Lit Display		
	16	Accessories: i) carrying case with Probe (test leads 1 pair) ii) Fuses: 10 nos iii) Instruction manual (English)		
	16	AUTHORISED DEALERSHIP CERTIFICATE FOR THE OFFERED MAKE SHALL BE FURNISHED ALONG WITH THE OFFER		
		Delivery Schedule		
		Delivery Type Period From PO Date (Days) Quantity		
		bulk 0 to 90 15.0000		
		Test Certificate		
	Manufacti	arer's test certificate shall be furnished along with supply at free of cost.		
		and a summer many was supply to not on a soun		