Contents

Chapter 1	The	Acquire Menu	A-1
-	1.1	General	A-1
	1.2	Preparing for data acquisition	A-2
	1.3	The configuration suite [config]	A-2
	1.4	Defining the acquisition data set [new, edc]	A-35
	1.5	Setting up acquisition parameters	A-36
	1.6	Interface control commands	A-79
	1.7	Starting and stopping data acquisition	A-90
Chapter 2	The	Windows Menu	. A-97
-	2.1	Interactive plot editor [xwinplot]	A-98
	2.2	Routine Spectrometer operation [iconnmr]	A-98
	2.3	Lock [lockdisp]	A-98
	2.4	Amplifier control [acbdisp]	A-99
	2.5	Pulse program [pulsdisp]	. A-101
	2.6	BSMS panel [bsmsdisp]	. A-105
	2.7	Temperature monitor [temon]	. A-118
	2.8	MAS rate monitor [masrmon]	. A-118
Chapter 3	The S	Shape Tool	A-119
F	3.1	Introduction	. A-119
	3.2	Generate a new shape	. A-120
	3.3	Manipulate existing Shape	. A-125
	3.4	Analyze existing Shape	. A-129
	3.5	Add 2 existing shapes to form a new shape	. A-132
	3.6	Merge 2 existing shapes to form a new shape	. A-132
	3.7	The Interactive Display Command stdisp.	. A-133
Chapter 4	Writ	ing Pulse Programs	A-139
-	4.1	Introduction	. A-139
	4.2	Pulse program library	. A-140
	4.3	Pulse program display	. A-140
	4.4	Basic syntax rules	. A-140
	4.5	Pulse generation commands	. A-144
	4.6	Delay generation commands	. A-161
	4.7	Simultaneous pulses and delays	. A-166
	4.8	Decoupling	. A-169
	4.9	Composite pulse decoupling (cpd)	. A-171
	4.10	Loop commands	. A-177
	4.11	Conditional pulse program execution	. A-179
	4.12	Commands to suspend the pulse program execution	. A-184
	4.13	Commands to start data acquisition	. A-185
	4.14	Working with acquisition memory buffers	. A-192

4.15	Writing memory buffers to disk A	\-194
4.16	Multiple receivers	\-196
4.17	Real time outputs A	\-197
4.18	Gradients	\-198
4.19	Miscellaneous commands A	4-201
Index		I-1

XWIN-NMR Comment Form

Chapter 1 The *Acquire* Menu

1.1 General

XWIN-NMR provides the following data acquisition commands:

- 1. <u>zg</u> and <u>go</u>. Used to execute a NMR experiment based on acquisition parameters set up with the command <u>eda</u>, The commands <u>gs</u> helps with the interactive adjustment of the acquisition parameters by real-time displaying the fid or spectrum. <u>zg</u> and <u>go</u> may be invoked in AU programs to control several experiments.
- 2. <u>iconnmr</u>. You should use ICON-NMR for routine spectroscopy based on standard experiments, and for automation using a sample changer. ICON-NMR is described in its own manual. The following two commands serve the same purpose, but are historically older. They are just maintained for compatibility reasons:
 - a) <u>run</u>. Used to execute a series of NMR experiments defined with the command <u>set</u>. A typical application of <u>run</u> is automated spectrometer operation with a sample changer.
 - b) **<u>quicknmr</u>**. Execution of experiments based on standard parameter sets provided by Bruker or defined by the spectrometer administrator for routine

applications. Requires only solvent and experiment to be specified before execution can start.

1.2 Preparing for data acquisition

Before you can start with data acquisition, the following preparations are necessary:

- 1. *Execute the configuration suite (command <u>config</u>). This is a sequence of commands which allows you to define your spectrometer environment. It is required once after program installation, or if the environment changes.*
- Set up sample/experiment specific acquisition parameters (commands <u>eda</u>, <u>wobb</u>, <u>gs</u>, <u>rga</u>, <u>set</u>, <u>ased</u>). Required for every experiment. Set up depends on acquisition command to be used (<u>zg/go</u>, <u>run</u>, <u>quicknmr</u>). Display the lock and fid windows. Required to observe fid signal and lock.
- 3. Start acquisition.

- <u>zg</u>: Starts an experiment based on free parameter set up with <u>eda</u> <u>-iconnmr</u>: routine execution of standard experiments Older commands:

- quicknmr: Executes experiments based on standard parameter sets
- <u>run</u>: Executes a series of experiments defined with the command <u>set</u>, based on standard parameter sets (may use a sample changer).

1.3 The configuration suite [config]

Please execute the following sequence of commands to prepare XWIN-NMR for data acquisition. Some of the commands will request the *NMR superuser* password. XWIN-NMR uses the concept of an NMR superuser. This is a normal system user, who was defined to be the NMR superuser by the system administrator.

If an NMR superuser is defined, commands such as \underline{cf} will prompt for the password of the NMR superuser, otherwise for the root password.

An NMR superuser may be set up or changed using the NMR user manager of ICON-NMR. Please check the ICON-NMR manual. In addition, the NMR superuser may also be entered at installation time of XWIN-NMR.

The command cfpp requires the root password since it modifies the system file /

etc/inittab.

- 1. Create a user id on your computer for anyone who should be allowed to start up XWIN-NMR. Invoke the respective operating system user manager tool for this purpose.
- 2. Define a data set. You may either select an existing data set using the command <u>search</u>, the commands of the *File->Open* menu, or create a new one with the *File->New* command.
- **3.** Some of the configuration commands display a text file requiring changes. Execute the command <u>setres</u> and set the system variable *Editor* to the name of your preferred text editor. The default editor is *xedit*, provided by the X Windows system. On SGI computers, the editor *jot* is preferred by many users.
- **4.** Execute the command <u>config</u> (type it in or call it from the *Acquire->Spectrometer setup* menu).

<u>config</u> will display a menu of the required configuration steps (Figure 1.1). By default, all steps except for configuring a MAS or a BPSU unit are enabled. If you click on the *Start* button, the configuration commands will be executed in this sequence. Whenever a step is complete, you are invited to click on *Continue* to proceed, or on *Cancel* to stop execution of the suite. You may restart execution by clicking on *Start*. Execution will begin with the first enabled command of the suite, and skip all disabled commands. Each step of the suite has a command name assigned, given in brackets. You may invoke a command directly from the keyboard.

1.3.1 Spectrometer configuration [cf]

The purpose of \underline{cf} is to make your spectrometer type and its hardware equipment known to XWIN-NMR. The program is capable of recognizing certain hardware components automatically, others are asked for by \underline{cf} . The result of this configuration process is saved in the files *uxnmr.conf* and *uxnmr.par*, both located in the directory *XWINNMRHOME/conf/instr/<name>/*. <name> is the instrument name defined during \underline{cf} . Acquisition commands will read the configuration parameters from these files. It is therefore not necessary to invoke \underline{cf} again if you terminate XWIN-NMR and restart it. However, after installation of a new XWIN-NMR version \underline{cf} is mandatory. If your hardware configuration did not change, you may simply type Return to any question posed during \underline{cf} .

Correct execution of \underline{cf} is a prerequisite for the acquisition commands to work, and should only be performed by the administrator of the spectrometer. We strongly

a configure	•
XWIN-NMR config	guration suite
Spectrometer configuration	[cf]
Temperature unit configuration	[cfte]
Sample changer configuration	[cfbacs]
Solvent table setup	[edsolv]
Probehead table setup	[edhead]
Solvent dependent parameter setup	[solvloop]
📕 User permission setup	[eduser]
Lock parameter setup	[locnuc, edlock]
📕 Plotter / printer setup	[cfpp]
🔲 Configure MAS pneumatic unit	[cfmas]
Configure BPSU unit for LC-NMR	[cfbpsu]
Avance spectrometer constants	[edscon]
Avance frequency routing	[edsp]
Installation of standard experiments	[expinstall]
Start Continue	Cancel Quit

Figure 1.1 Configuration suite

recommend to save the files *uxnmr.conf* and *uxnmr.par* on tape, diskette, or elsewhere after a successful configuration. If needed, they may be restored into the *XWINNMRHOME/conf/instr/<name>/* directory. <u>cf</u> executed thereafter will only require Returns to be entered.

1.3.1.1 Reconfiguration

If at any ealier time \underline{cf} was executed successfully, configuration is an easy procedure since the answers to all questions are stored in a file and are prompted as default. As long as the spectrometer hardware has not changed the operator may simply answer Return to all questions until \underline{cf} is finished. At the end a window will appear giving an overview of the spectrometer configuration. You should now check whether the configuration is correct and repeat \underline{cf} if required.

1.3.1.2 Configuration from scratch

If there was no configuration directory on disk (e.g. on a replaced disk drive) \underline{cf} will start a configuration from scratch.

Please note: If the spectrometer requires a *hardware_list* file (to be described later in this section) it is necessary to create the configuration directory manually and put the *hardware_list* file there. Open a unix shell, retrieve the *hardware_list* from the backup device and proceed as follows:

```
su
cd XWINNMRHOME/conf/instr
mkdir <instrument name>
cp <any-dir>/hardware_list <instrument name>
exit
```

Now start XWIN-NMR and configure the spectrometer. Enter the name of the instrument (usually *spect*):

Enter new instrument name: spect

If you chose a name different from *spect* be sure to have this name set as alias in the file */etc/hosts* (see Troubleshooting page 10). Otherwise later \underline{cf} is not able to contact the spectrometer when it wants to check the spectrometer hardware.

cf offers a selection of spectrometer types:

What type of spectrometer? (Datast AMX ARX ASX DMX DRX DPX DSX APEX-A APEX-D AMX3) DMX

cf needs the 1H frequency of the magnet:

```
Basic 1H frequency (with offset o1=0) in MHz: 500.13
```

Please wait for some seconds for \underline{cf} to check the spectrometer hardware. On Avance spectrometers it will examine the FCU's, the RCU and all its connected digitizers, while on AMX/ARX systems it will detect the Aspect 3001 configuration and the digitizers.

Afterwards all units controlled via serial port are configured. On Avance systems a



window (see Figure 1.2) appears which contains all installed units and allows to

Figure 1.2 Configuration of RS 232 devices

set the device name of each unit. If an unit is not in use or has not been used before, the device name shows *no*.

On non Avance spectrometers each unit is prompted for. Enter the name of the device they are connected to. Usually a default device name is offered, e.g.

```
Which device is used for <unit>: tty03
```

You will be asked whether the spectrometer is equipped with a triple housing

preamplifier. If it is equipped with a HPPR instead, you must enter *n*, then you are asked for the device used by the HPPR.

As soon as all configurator questions have been answered, the nucleus table stored in the file *XWINNMRHOME/exp/stan/nmr/lists/nuclei.all* is displayed in a dialog window (see Table 1.1). This table may now be modified according to your needs.

nuclei table						
11	H	500.13				
2	Н	76.77				
3	H	533.46				
4H	łe	380.55				
SAVE	ADD	RESTORE	QUIT			

Table 1.1 Nuclei table

XWIN-NMR offers the following possibilities of editing the nucleus table:

Changing the frequency

Move the cursor onto a frequency value and depress the left mouse button. A new number may be entered now.

Deleting a nucleus

Move the cursor onto the name of a nucleus and depress the left mouse button.

ADD (insert nucleus)

Move the cursor onto this command field and depress the left mouse button. The file *nuclei.all* appears on the screen, and a nucleus can be selected. It is inserted according to its mass number.

RESTORE

Activate this command field if something failed during the edit session. The nucleus table is restored entirely from the file *nuclei.all*.

SAVE

All changes are saved, and the nuclei table disappears.

QUIT

All changes are discarded, and the nuclei table disappears.

The result ist stored in the file *XWINNMRHOME/conf/instr/<instrument name>/ nuclei*, and is required by the commands <u>edasp/edsp</u>, where the table is displayed to select a nucleus .

At the end a window appears presenting an overview of the spectrometer configuration. This overview is stored in the text file *XWINNMRHOME/conf/instr/<instrument name>/uxnmr.info*, and may be viewed or printed like any text file.

1.3.1.3 Related Files

The text file *XWINNMRHOME/conf/instr/curinst* contains the instrument name as entered in <u>cf</u>. All files used and created by <u>cf</u> reside in the directory *XWIN-NMRHOME/conf/instr/<instrument name>*. A description may be found in Table 1.2. There are several subdirectories whose descriptions are shown in Table 1.3.

The hardware_list file

All AMX, ARX and ASX spectrometers and all non standard Avance spectrometers need the text file

XWINNMRHOME/conf/instr/<Instrument Name>/hardware_list,

which is set up by the service engineer at installation time of the instrument, and contains information about the hardware equipment of the spectrometer. The file must not be modified. A safe copy should always be available on magnetic tape and in form of a printout.

If a spectrometer is equipped with a 4-phase modulator, but not with a HPCU it is possible to define this hardware equipment in the hardware list file. \underline{cf} is required afterwards. The unit may be connected to *tty10* or *tty20* on the CCU.

DMX spectrometers which are equipped with HRD16 and FADC digitizers can use the analog filters on the HRD16 external box during measurements with the FADC. For this kind of application, the RXC must *not* be defined in the hardware list file, and the acquisition parameter SEOUT must be set to BB.

1.3.1.4 Layout of serial device connectors

All serial devices on the spectrometer are located at the CCU¹:

• The RS-232 device *tty00* is located on the CCU frontpanel and may be used by a console terminal. During the boot procedure all messages are printed to this

acqu.conf	text file created by <u>cf</u> containing information about the Aspect 3001 hardware configuration (not for Avance spectrometers).
acqu.conf_info	text file created by <u>cf</u> containing an explanation of the numbers found in <i>acqu.conf</i> (not for Avance spectrometers).
bbis_fcu	text file created by <u>cf</u> containing information about the FCU's installed in the spectrometer (Avance only).
bbis_rcu	text file created by <u>cf</u> containing information about the RCU and other boards connected to the RCU (Avance only).
bsmsdisp.calibr	text file created as empty file by <u>cf</u> and filled in or used by <u>bsmsdisp</u> (Avance only).
bsmsdisp.calibr.bak	backup file for <i>bsmsdisp.calibr</i> (Avance only).
hardware_list	text file in a special format containing a list of hard- ware components
nuclei	text file containing the table of nuclei selected either in \underline{cf} or by \underline{ednuc} .
scon	parameter file containing spectrometer parameters; is created by <u>edscon</u> .
uxnmr.conf	data file created by <u>cf</u> , and used by several commands; its content may be examined by the <i>XWINNMRHOME/</i> <i>prog/bin/printconf</i> program.
uxnmr.info	text file created by \underline{cf} containing useful information about the spectrometer hardware. It is displayed when \underline{cf} is finished.
uxnmr.par	standard parameter file created by <u>cf</u> containing the answers of the operator to the <u>cf</u> questions

 Table 1.2 Configuration files

1. Communication Control Unit

rs232_device	Is created by <u>cf</u> and contains text files for the device configu- ration of several units (e.g. HPPR, BSMS, RX22).
rs232_lock	Is created by <u>cf</u> and contains empty files used to lock devices (to prevent that a device from being used simultaneously by more than one unix process).
cortab	Is created by the automatic spectrometer adjustment.
prosol	Is created by <u>prosol</u> .
users	Is created by <u>eduser</u> .

Table 1.3 Configuration subdirectories

device.

It is therefore not allowed to connect any spectrometer unit to this device!

- There is a connector panel above the CCU equipped with 9 RS-232 and 2 RS-485 devices. The 9 RS-232 devices correspond to *tty01-tty09*, the RS-485 devices correspond to *tty10* and *tty20*. Note: tty08 and tty10 (and respectively tty09 and tty20) may only be used alternatively as these devices are connected to the same port!
- If more devices are needed it is possible to add up to 4 SIB's¹. Each SIB is equipped with 6 RS-232 devices (CH1 CH6) which correspond to *tty11-tty16* on the 1st SIB, *tty21-tty26* on the 2nd SIB, *tty31-tty36* on the 3rd SIB and *tty41-tty46* on the 4th SIB.

1.3.1.5 Trouble shooting: part 1

While <u>cf</u> is checking the spectrometer hardware (after entering the 1H frequency) the following error message appears:

----- < iiconf > -----connection to <Instrument Name> (aqport0) failed
('startd' demon active on CCU ?)

This error message may be caused by several reasons:

1. Serial Interface Board

• There is no *startd* demon running on the CCU. To check this execute the following commands in a unix shell:

```
telnet spect
login: root
ps -ef | grep startd | grep -v grep
```

If the last commands does not list any running *startd* it can be started manually with:

/etc/startd

or by rebooting the CCU with:

```
/etc/init 6
```

• The operating system on the host does not know the name of the spectrometer. This happens if <Spectrometer Name> is a newly chosen name which has never been used before.

To solve this problem the superuser must add the new name to the file */etc/hosts* directly behind the name spect, e.g. for the name drx300:

149.236.99.99 spect drx300

1.3.1.6 Trouble shooting: part 2

Some of the units controlled via serial port (e.g. ACB-board, HPPR, 3-channel-SE451) are addressed during <u>cf</u> to read their internal configuration. During this procedure the following error message appears:

----- < cf > -----Error during open of <unit>:
no such device or address

There are several reasons leading to this error:

- The <unit> is connected to a serial port other than the one entered by the operator.
- The <unit> is switched off.

- The <unit> is connected with the wrong cable.
- The <unit> is defective.

1.3.2 Temperature unit configuration [cfte]

This command must be used to inform XWIN-NMR about the channel to which the temperature is connected, e.g. *tty03*. The configured channel is stored in the file *XWINNMRHOME/conf/instr/<instrument name>/rs232_device/temp*.

1.3.3 Sample changer configuration [cfbacs]

This command must be used to inform XWIN-NMR about the channel to which the sample changer is connected, e.g. *tty12*. The configured channel is stored in the file *XWINNMRHOME/conf/instr/<instrument name>/rs232_device/bacs*. The second question of <u>cfbacs</u> asks you whether to use the bar code reader. If your answer is *yes*, you are invited to define the RS232 channel of the bar code printer. Finally, the delay between the change sample command <u>sx</u> and the next command (frequently <u>ro</u>, turn sample rotation on) must be specified to allow for the sample to settle in its position. Typical values are 5 to 30 seconds, depending on the magnet.

If you want to make use of XWIN-NMR's automation features (setting up a series of experiments with the commands <u>set</u>) without a sample changer being available, you must call <u>cfbacs</u> anyway. In this case you are asked to enter the *set up capacity*, a number from 1 to 120 defining the number of experiment entry fields displayed by the <u>set</u> command. This number is automatically set to the number of sample holders if a sample changer is installed. It is stored in the file *XWINNMRHOME/conf/ instr/bacs_capacity*.

For spectrometers equipped with BSMS the question "Airflow for Sample Lift from BACS connected?" allowes one to configure the BACS/BSMS so that the LIFT function is not performed by the sample changer, but by the BSMS Lift only. In this case the connection of the airflow from the sample changer to the Sample lift must be cut off.

1.3.4 Setting up the solvent table [edsolv]

Use this command to set up a table of the solvents you intend to use for your NMR experiments. It is a configuration command which should only be executed by the administrator. <u>edsolv</u> opens a dialog window where you may add, change, or delete

lines by clicking on the corresponding button. New entries must get assigned an arbitrary, but unique reference number. A typical entry in the table looks like the following:

The reference number must be specified in brackets. It is used for identifying the solvent on a barcode label when a barcode controlled experiment is performed, e.g. using an automatic sample changer equipped with a barcode reader. In order for such experiments to be executed properly, we recommend not to change the numbers any more once assigned to a particular solvent.

Initially, edsolv displays a solvent table suggested by Bruker from the text file

XWINNMRHOME/exp/stan/nmr/lists/solvents.all.

If you apply modifications to the table and then exit via the SAVE button, the modified table is stored in the file

```
XWINNMRHOME/exp/stan/nmr/lists/solvents.
```

Now the file *solvents* has been created, and future invocations of <u>edsolv</u> will display this file containing your personal settings.

In order to define a solvent for an experiment, type <u>eda</u> (<u>eda</u> is described further below in this chapter), and in the up-coming dialog window click on the downarrow button right of the SOLVENT parameter. The *solvents* file is displayed, and you may select a solvent from the table. SOLVENT is evaluated by the commands <u>prosol</u>, <u>lock -acqu</u>, <u>sref</u> and <u>lopo</u>, and during <u>quicknmr</u> and <u>run</u>. For the latter two applications the acquisition parameters are obtained in the following way: They are initialized with the parameters of the specified experiment. The probe head and solvent dependent parameters (see command <u>prosol</u>) are then inserted according to the setting of SOLVENT and the current probe head (see next section). Finally, any parameter changes the user possibly requested are applied.

Table 1.4 describes the available command buttons.

1.3.5 Setting up the probe head table [edhead]

Use this command to set up a table of the probe heads you intend to use for your NMR experiments, and to define the current probehead installed in the magnet. It is a configuration command which should only be executed by the administrator.

SAVE	Store modifications in the file <i>XWINNMRHOME/exp/stan/nmr/lists/solvents</i> and quit.
ADD/CHANGE	Opens a new entry field at the end of the table.
DELETE	After enabling this button, clicking on a probe- head entry will delete it.
CTRL/K key	undo last change.
ABORT	close edsolv window without storing changes.

Table 1.4 Command buttons in edsolv dialog window

<u>edhead</u> opens a dialog window where you may add, change, or delete lines by clicking on the corresponding button. New entries must get assigned an arbitrary, but unique reference number. A typical entry in the table looks like the following:

```
5 mm Dual 13C/1H [03]
```

The reference number must be specified in brackets. It is used for identifying the probe head on a barcode label when a barcode controlled experiment is performed, e.g. using an automatic sample changer equipped with a barcode reader. In order for such experiments to be executed properly, we recommend not to change the numbers any more once assigned to a particular solvent.

Initially, edhead displays a probe head table suggested by Bruker from the text file

XWINNMRHOME/exp/stan/nmr/lists/probeheads.all.

If you apply modifications to the table and then exit via the SAVE button, the modified table is stored in the file

XWINNMRHOME/exp/stan/nmr/lists/probeheads.

Now the file *probeheads* has been created, and future invocations of <u>edhead</u> will display this file containing your personal settings.

In order to inform XWIN-NMR which probe head of the table is currently installed in the magnet, click on the *Define current* button, and then on the desired table entry. It will be stored in the file *XWINNMRHOME/conf/instr/probehead*. The current probe head is evaluated by the commands <u>prosol</u>, <u>edlock</u>, <u>lock -acqu</u>, and <u>lopo</u>, and during <u>quicknmr</u> and <u>run</u>. For the latter two applications the acquisition parameters are obtained as described in the previous section (edsolv).

The XWIN-NMR acquisition commands (zg, go) store the current probe head in the acquisition parameter PROBHD of the acquired data set for future reference.

Table 1.5 describes the available command buttons.

SAVE	Store modifications in the file <i>XWINNMRHOME/exp/stan/nmr/lists/probeheads</i> and quit.
ADD/CHANGE	Opens a new entry field at the end of the table.
DELETE	After enabling this button, clicking on a probe- head entry will delete it.
CTRL/K key	undo last change.
ABORT	close edhead window without storing changes.

Table 1.5 Command buttons in edhead dialog window

1.3.6 Solvent/probehead dependent parameters [prosol, solvloop]

These parameters are required to run series of experiments using the commands <u>run</u> or <u>quicknmr</u>, or by the PROSOL function of the command <u>eda</u>. If the values are not known at this time, you may skip the command for now, and invoke it at a later time.

The commands <u>run</u> and <u>quicknmr</u> are designed to run experiments based on standard parameter sets (provided by Bruker, or set up by your system administrator). If one of these commands is active, the acquisition parameters for an experiment are obtained in the following sequence:

- 1. They are initialized with the standard parameters of the selected experiment.
- **2.** Parameters depending on the probe head and solvent are inserted according to the setting of SOLVENT and the current probe head.
- 3. Finally, any parameter changes the user requested are applied.

The purpose of <u>prosol</u> and <u>solvloop</u> is to set up the parameters of step 2. <u>prosol</u> lets you select a solvent and a probehead by displaying the tables you have defined earlier with <u>edsolv</u> and <u>edhead</u>. Then, for a selected solvent/nucleus combination,

power levels and pulse lengths may be specified (e.g. *Proton high power level, transmitter pulse, power level for pre-saturation* etc.). If you click on the SAVE button of the respective dialog box, the parameters will be saved in the file

XWINNMRHOME/conf/instr/<instrument name>/prosol/<solvent.probehead>

e.g.

XWINNMRHOME/conf/instr/dmx300/prosol/acetic.01,

where *probehead* is the id number assigned to a probehead during execution of *edhead*.

In contrast to <u>prosol</u>, <u>solvloop</u> immediately starts up with the probehead table. If you click on the SAVE button after having defined the parameters, they are automatically stored for all solvents (e.g. the files *.01 are created). Afterwards you may change special parameters of any solvent/probehead combination using the command <u>update</u>. It displays a list of existing *<solvent.probehead>* files, from where you can open the desired one by clicking on it.

1.3.7 Setting up user permissions [eduser]

Execute the command <u>eduser</u> to define which experiments a XWIN-NMR user may execute. You may skip this command if you do not intend to use the acquisition commands <u>run</u> or <u>quicknmr</u>. You may also skip <u>eduser</u> for now, and execute it at a later time (but before invoking <u>run</u> or <u>quicknmr</u>).

First, a list of installed users is displayed. After selecting the desired one, a permission file

XWINNMRHOME/conf/instr/<instrument name>/users/<login id>

is created for this user (if not yet existing), e.g.

XWINNMRHOME/conf/instr/dmx300/users/guest.

It is a copy of the default permission file

XWINNMRHOME/exp/stan/nmr/lists/sam_users_exam

provided by XWIN-NMR. The file is displayed by a text editor, and you may modify it. Figure 1.3 shows an example. Comment lines begin with a '#' character.

```
#Example user permission file
#
---Data set names-----
$DATE
Name1
Name2
---Experiments:-----
0 PROTON128 - 1H experiment 128 scans
0 PROTON
               - 1H experiment 16 scans
0 C13CPD - C13 exp. comp. pulse dec. 1024 scans
0 N15IG - 15N exp. inverse gated
1 INV4SW - sw opt. inv. 4 pulse MC
1 INV4NDLRSW - sw opt. inv. 4 pulse MC long range
2 HCCOSW - sw opt. CH-correlation
2 HCCOLOCSW - sw opt. COLOC
---Permissions (urgent editPAR composite exit(qnmr))---
no yes no no
```

Figure 1.3 Example of a user permission file

Data set names

In the section ---*Data set names*--- you may predefine a list of data set names. These will be the only names which may be given a data set during experiment set up with the command <u>set</u>. The special name *\$DATE* will be converted to the current date during <u>set</u>. Please note that the header text of a section may be arbitrary. The leading '---' characters indicate the start of a section.

Experiments

The section *---Experiments---* is the list of experiments this user is permitted to run. Each entry consists of a number, a name, and a comment. The number specifies the experiment type according to Table 1.6. The name must denote a parameter

0	normal experiment
1	experiment depending on 1 preparation experiment
2	experiment depending on 2 preparation experiments
Т	variable temperature experiment

Table 1.6 Experiment types in a user permission file

set in the directory *XWINNMRHOME/exp/stan/nmr/par/*. Experiments of type 1 or 2 require one or two preparation experiments to be performed before the experiment itself can start (e.g. a 1D preparation experiment determining the optimized sweep width for a subsequent 2D experiment). T characterizes a variable temperature experiment. If such an experiment is selected during <u>set</u>, the initial temperature, the temperature increment, and the number of increments are requested. The corresponding number of measurements are performed with this sample.

Permissions

The section ---Permissions--- contains 4 flags *urgent*, *editPAR*, *composite*, and *exit(qnmr)* used to enable or disable certain features for this user. Legal values of the flags are *no* or *yes*, to be specified in the next line in the correct sequence.

- *urgent* flag = *yes* This user may classify a sample as *urgent* during the set procedure. In a sample changer run, it will get priority.
- *editPAR* flag = yes
 This user is allowed to change acquisition or processing parameters during <u>set</u>.
- composite flag = yes
 This user may define composite experiments during set. A composite experiment is a sequence of up to 9 standard experiments. After a new composite experiment was defined, it will be added to the experiment table of the set dialog window.
- *exit(qnmr)* flag = yes This user gets the permission to terminate the <u>quicknmr</u> command.

1.3.8 Setting up the lock parameter table [edlock]

The purpose of <u>edlock</u> is to define the lock parameters for the solvents to be used, for a particular lock nucleus, and store them on disk. Before you call <u>edlock</u>, enter the command <u>locnuc</u> and type 2H or 19F to define the lock nucleus. <u>edlock</u> opens a dialog window according to Figure 1.4. Table 1.7 describes its command buttons.

For instruments equipped with a BSMS unit, the lock parameters are stored in two files. The first one contains those parameters depending on the lock nucleus (acquisition parameter LOCNUC) and the solvent. For deuterium, its file path name is

XWINNMRHOME/conf/instr/<instrument name>/2Hlock.

edlock										
	Edit 2H	lock File	• CC	Curhead	11: 5	mm Mult	inuclear i	nverse 2	Z-grad]	
Lockfreq:	46.03	340585		FIELI	D: 481			BFRE	Q: 299.	88 [MHz]
Solvent	LPower	LGain	LTime	LFilt	LPhase	Nucl.	Distance	Ref.	Width	RShift
Acetic	-40.0	-10.0	0,100	100	-1.0	2H	2,03	0.0	0.5	0.000
Aceton	-40.0	-10.0	0.100	100	-1.0	2H	2.04	0.0	0.5	0.000
CDC13	-25.0	-15.0	0,200	100	-1.0	2H	7.24	0.0	0.5	0.000
CD2C12	-30.0	-15.0	0.100	100	-1.0	2H	5,32	0.0	0.5	0.000
CD3CN	-40.0	-10.0	0.100	100	-1.0	2H	1.93	0.0	0.5	0.000
Cede	-26.0	-15.0	0,200	100	-1.0	2H	7.28	0.0	0.5	0.000
D2O	-20.0	-5.0	0.200	200	-1.0	2H	4.70	0.0	0.5	0.000
H20	-23.0	-15.0	0,200	100	-1.0	2H	4.70	0.0	0.5	0.000
DEE	-30.0	-15.0	0,200	100	-1.0	2H	1.07	0.0	0.5	0.000
DME	-35.0	-10.0	0.100	100	-1.0	2H	3,30	0.0	0.5	0.000
DMF	-25.0	-8.0	0,200	200	-1.0	2H	2,91	0.0	0.5	0.000
DMSO	-25.0	-8.0	0.200	200	-1.0	2H	2,49	0.0	0.5	0.000
Dioxan	-30.0	-15.0	0,200	100	-1.0	2H	3.53	0.0	0.5	0.000
EtOH	-30.0	-15.0	0.200	100	-1.0	2H	1.11	0.0	0.5	0.000
MeOH	-35.0	-10.0	0.100	100	-1.0	2H	3,30	0.0	0.5	0.000
KI KI										
SAVE	SAVE BSMS_FIELD NUCLEUS NEW SOLVENT DELETE ABORT									
LOADSTAN +/- POWER COPY_VALUE LIST HELP										

Figure 1.4 edlock parameter dialog box (for a BSMS unit)

For a fluorine lock (i.e. the acquisition parameter LOCNUC is set to *19F* before calling <u>edlock</u>), the file name is *19Flock* in the same directory. The second file contains parameters depending on the solvent and on the probehead. The name of the directory where it is located, is the solvent name, with the probe head identification

SAVE	Store current setting in their files and quit.
BSMS-FIELD	Read current value from BSMS and update top line.
NUCLEUS	Select new nucleus. Equivalent to clicking on a nucleus button in the dialog window.
NEW SOLVENT	Insert new entry at table end. Initialize it with the last selected solvent (click on a solvent to select it).
DELETE	Click on this button to activate delete mode. Then click on a solvent to remove an entire solvent entry, or click on a nucleus to remove only this nucleus for the corresponding solvent.
ABORT	Discard any changes and quit.
LOADSTAN	Discard the current settings and load standard values from the file <i>XWINNMRHOME/exp/stan/nmr/list/2Hlock</i> .
+/- POWER	Increment the LPower values of all solvents by the specified amount.
LIST	Print the lock parameters on the device given by the parameter CURPRIN. Set it up with <u>edo</u> (e.g. CURPRIN=\$hplj4p).
COPY_VALUE	automatically set the selected value for all solvents
HELP	Print help messages.

Table 1.7 Command buttons in edlock dialog window

number (see <u>edhead</u> command) as a name extension, e.g. *Acetic.03*. The file name itself is *param*:

XWINNMRHOME/conf/instr/<instrument name>/prosol/<solvent.probeID>/ param.

The first line of the <u>edlock</u> dialog window shows the *2Hlock* or *19F* lock file names, and the current probe head including is ID number as defined with <u>edhead</u>. The second line shows the lock frequency, the field value and the basic spectrometer frequency BFREQ. The lock frequency is calculated by the software from the BFREQ and the nucleus table.

In the following lines, for a desired solvent you may enter the BSMS lock parame-

ters *LPower*, *LGain*, *LTime*, *LFilt*, and *LPhase* valid for the current probehead. These are the parameters stored in the *param* file. The parameters are loaded into the hardware by the commands <u>lock</u>, <u>lopo</u>. and <u>lopoi</u>. XWIN-NMR also provides the special commands <u>ltime</u>, <u>lgain</u>, <u>lfilter</u> from which you may load the hardware directly, by ignoring the values of the <u>edlock</u> table. These commands (also available in the *Acquire->Interface control->BSMS unit* submenu) request the numbers to be entered on the keyboard, or they may be specified on the command line.

For SCM units, only *LPower* is available, stored in the *2Hlock* or *19Flock* file.

In the right half of the dialog window, for each solvent and a selected nucleus the shift of the lock from 0 ppm (*distance*), the chemical shift value of the reference signal (*Ref.*, 0 ppm for TMS), and a *width* value may be specified. They are stored in the 2*Hlock* or 19*Flock* file.

The Bruker standard *2Hlock* file contains default values for lock power, loop gain, loop time and loop filter for each solvent. If a spectrometer installation is started from scratch, these values are automatically read when you do <u>edlock</u> the first time. If the *2Hlock* file already exists and the lock/loop parameters were already defined, then these values are displayed and the values from the default file are not used. The LOADSTAN button in <u>edlock</u> will read the default file and overwrite all previous settings. Please make sure, before you use LOADSTAN, that you have a print-out or disk copy of the original settings.

The table shown with the command <u>edlock</u> contains an additional column R.Shift. The value (in ppm) entered here is added to the default calibration done by the <u>sref</u> command. This allows chemical shift corrections where, for whatever reason, the reference shift is not calculated accurately enough.

The parameters described last are used by the command <u>sref</u> for automatic calibration (referencing) of a spectrum. *width* (in ppm) defines the range that a reference signal is searched for by <u>sref</u> in order to determine the exact origin.

1.3.9 Printer/plotter configuration [cfpp]

Execute the command <u>cfpp</u> to inform XWIN-NMR of which types of plotters or printers are connected to your computer, and to which channels. During this procedure you will also assign names to these devices. They are used to define where a spectrum plot or text printout is to be sent. All details of <u>cfpp</u> are explained in the chapter *The 1D Output Menu*.

1.3.10 Configure MAS unit [cfmas]

Execute <u>cfmas</u> if your spectrometer is equipped with a MAS pneumatic unit. You must specify the RS232 channel number to which the unit is connected. You may also specify the pressure, the air on time (in seconds) for sample insertion and ejection, and the sample diameter (normal or wide bore).

1.3.11 Configure BPSU unit for LC-NMR [cfbpsu]

Execute <u>cfbpsu</u> if your spectrometer is equipped with a BPSU accessory designed to run LC-NMR experiments. This unit is connected to a RS232 channel of the spectrometer whose channel number must be specified, e.g. *tty13*.

1.3.12 Avance spectrometer constants [edscon]

The command <u>edscon</u> is only available for Avance type spectrometers. It opens a dialog window, where you may change the default value of certain spectrometer constants. Unlike the normal acquisition parameters, these constants are not part of a particular data set. If changed, the modifications will influence all further measurements on this spectrometer. The constants are stored in the file

XWINNMRHOME/conf/instr/<instrument name>/scon.

They define the timing of *transmitter blanking*, *ASU blanking*, *preamplifier blanking*, *phase presetting*, and *shaped pulse presetting* with respect to the transmitter gating pulses. Furthermore, the pre-scan delays DE1 and DE2 may be changed. The pulse blanking constants are explained in Figure 1.5.

TGPCH[1..8] (transmitter gating pulses)

They are generated on the TCU (word3, Bits 24-31) with the pulse program commands $\underline{p1:f(1..8)}$ or $\underline{cw:f(1..8)}$. Routing is done according to

TGPCH[FCUCHAN[channel]].

BLKTR[1..8] (transmitter blanking pulses)

The transmitter blanking may be switched on a time t1 before the pulse. Up to 8 different times can be used for the 8 different ASU channels. The same blanking is applied to the amplifiers.

They are generated on the TCU (word 3, Bits 0-7 and NMR2, Bits 5 and 6) together with the transmitter gating pulses TGPCH[1..8], but prolonged at the



Figure 1.5 Presetting of blanking pulses

beginning by the times BLKTR[1..9]. Routing is done according to

BLKTR[RSEL[FCUCHAN[channel]]].

The routing according to RSEL is done by setting setting the RSEL parameters on the digital routers. This routing can be switched at runtime by setting the corresponding NMR-control-words:

NMR1, Bits 0-3 for RSEL[1] NMR1, Bits 4-7 for RSEL[2] NMR1, Bits 8-11 for RSEL[3]

For more than 3 FCUs:

NMR1, Bits 12-15 for RSEL[3] and if RSEL[3] < 6 NMR1, Bits 8-11 NMR7, Bits 0-3 for RSEL[4] " " 4 " NMR7, Bits 4-7 for RSEL[5] " " 5 " For more than 5 FCUs:

NMR7, Bits 8-11 for RSEL[5] and if RSEL[5] < 10 NMR7, Bits 4-7 NMR7, Bits 12-15 for RSEL[6] " " 6 " NMR8, Bits 0-3 for RSEL[8] " " 7 "

The corresponding routing of BLKTR, BLKPA and BPCH will follow these settings. Examples:

 $\frac{\text{setnmr1} ^{3} ^{2} |1|0}{\text{setnmr1} ^{7} ^{6} |5 ^{4}} ; \text{set RSEL[1]} = 3$

BPCH[1..8] (ASU blanking pulses)

They are generated on the TCU (word 3, Bits 16-23) together with the transmitter gating pulses TGPCH[1..8], but prolonged at the beginning by the times BLKTR[1..5]. Routing is done according to

BLKTR[RSEL[FCUCHAN[channel]]].

TGPPA[1..5] (*Preamplifier blanking pulses*)

They are generated on the TCU (word 3, Bits 8-12) together with the transmitter gating pulses TGPCH[1..8], prolonged at the beginning by the times BLKPA[1..5]. Routing is done according to

BLKPA[PRECHAN[SWIBOX[RSEL[FCUCHAN[channel]]]]]

The routing according to PRECHAN is done by selecting the preamplifier module (HPPR) via RS232 at the beginning of the experiment and cannot be changed afterwards. The routing according to SWIBOX is done with the switches SELH!H/ F (NMR2, Bit 2) and SELX!X/F (NMR2, Bit3). This can be changed at runtime.

BLKPA[1..5]

The blanking of the 5 preamplifier modules may be switched on a time t2 before the pulse.

PHASPR[1..8] (phase presetting)

The switching of the phase programs may be done a time t3 before the pulse in order to ensure a stable phase when the pulse begins.

SHAPPR[1..8] (shaped pulse presetting)

The propagation time of the phase versus the amplitude may be taken into account by setting SHAPPR (t5). This value should be equal or larger than $1.4 \,\mu$ sec. Other-

wise the shaped pulse itself will be delayed by this time.

Please note: At the end of each shaped pulse the power is reset to its default value (e.g. for channel 1 pl1) and as consequence of this power setting the phase correction corresponding to this power is set as well. If the shaped pulse is terminated before this setting has taken place because there was not enough time to execute the whole shaped pulse, the phase is still loaded as required for the shaped pulse, namely with the phase cycle of the shaped pulse and not with the phase correction value for the power pl[channel].

DE1, DE2

The acquisition parameter DE is the waiting period between the last pulse and the begin of data acquisition (digitizer start) to avoid pulse feed through. You may change DE by entering a new value. DE includes two sections, DE1 and DE2 (Figure 1.6). After DE1 the receiver gate is opened and the frequency is switched to receive mode. If DE1 = -1, a default value of 4.0 μ sec is used.

After DE2, the phase of the observe channel is set to 0. DE1 and DE2 both start with DE and may range from 0 to DE. DE1 may be larger or smaller than DE2. If DE2 = -1, a default value of 1.0 µsec will be used. The parameter PHASPR, which will pull forward the phase setting, is taken into account automatically such that the true phase setting occurs after DE2. This may lead to the situation where the program prints the error message *DE too small*... even if DE2 is smaller than DE, because DE2 + PHASPR[FCUCHAN[1]] is larger than DE..



Figure 1.6 The DE delay

1.3.13 Avance frequency routing [edsp, edasp]

<u>edasp</u> is usually called from the NUCLEI button of the <u>eda</u> window, but may also be invoked as a keyboard command. In addition to nuclei setup, the <u>edasp</u> window also shows the connections between the hardware parts of the spectrometer, namely FCUs, amplifiers, routers, high power modules and preamplifier modules. This feature is only available for Avance type spectrometers. Table 1.8 shows the available command buttons.

SAVE	The aquisition parameters will be saved on the disk
SWITCH F1/F2	Exchanges the F1 nucleus with the F2 nucleus
SWITCH F1/F3	Exchanges the F1 nucleus with the F3 nucleus. With the SWITCH but- tons you may easily swap between observed and decoupled nuclei wit- hout having to reenter any frequen- cies
DEFAULT	sets the default routing for the sel- ected set of nuclei
CANCEL	Exit without saving the parameters
PARAM	Shows the aquisition parameters as they would be stored by SAVE

Table 1.8 Command buttons of edasp/edsp window

The Avance <u>edasp/edsp</u> display is logically divided into several vertical parts. Some elements may be connected with the elements of the next part by clicking on the corresponding two buttons. The rules to be obeyed are described below.

Frequency

The basic frequencies BF1-BF8 cannot be modified, they correspond to the selected nucleus. For spectrum referencing the parameter SR is derived from SF and BF1:

SR = SF - BF1

(SR is specified in Hertz). Note that for spectrometers equipped with a BSMS unit this number should be set to a value near 0, since the basic frequency of the nucleus is chosen such that the frequency reference standard (e.g. TMS) will have the frequency BF1 if the lock frequency has been set properly by the BSMS.

Logical channels

There are two buttons for each logical channel. The buttons F1....F8 are used to show and to alter the assignment of the physical FCU to the logical channel, e.g. if F1 is connected to FCU2 this means that FCU number 2 will be used for the logical channel F1. If the nucleus for channel i is selected (not set to *off*) the button F with the number i must be connected to one of FCU1...FCU8. The selected FCU button must be connected to one of the amplifier buttons.

The lower button in NUC1...NUC8 in the logical channel group allows you to select the nucleus. The displayed table is taken from the file

XWINNMRHOME/conf/instr/<instrument name>/nuclei,

which is set up during the configuration command cf.

The parameters NUC1-NUC8, and the frequency offsets OFSxx are global parameters in XWIN-NMR: They are not only stored in the *acqu* file of the current data set as acquisition parameters, but also in the file

XWINNMRHOME/conf/instr/ < instrument name>/specpar

at the time the SAVE command is executed in <u>edsp</u> or <u>edasp</u>. They may be imported from there into any dataset with the command <u>edsp</u> (this is actually the difference between <u>edasp</u> and <u>edsp</u>).

Usage of edsp and edasp

Invoke <u>edasp</u> if the parameters of the current data set should be used as default. <u>edsp</u> should be used, if the frequencies and nuclei of another experiment, previously defined with <u>edasp</u> or <u>edsp</u>, but in a different dataset, should be transferred to the current data set. A typical example is the setup of a ¹H experiment and a ¹³C experiment with ¹H-decoupling. For the first experiment a dataset 1 is created. Within this dataset the nucleus ¹H and the frequency SFO1 are chosen with the command <u>edasp</u>. Then within a second data set the nucleus ¹³C for F1 and ¹H for F2 are chosen using the command edsp. Automatically the frequency SFO1 from experiment 1 is transferred to SFO2 of experiment 2, because the same nucleus is used for channel 2 as that one which was used in experiment 1 for channel F1. If an inverse experiment would be done next, it is possible to interchange the nuclei for F1 and F2 from experiment 2 with the command edsp.

Amplifiers

The frequency output of each FCU is hardwired to a router input and each router output is hardwired to a specific amplifier. The edsp display shows in the first column the logical assignment of channels (F1-F8) to the FCUs, in the second column the connections of the FCUs to the amplifiers which is done by the router. The third column of connections shows the so-called switchbox, which can connect the output of the first X-amplifier and of the 1H-amplifier to different preamplifiers by means of relays or diode-switches.

Preamplifier

Up to 5 preamplifier modules can be installed in the HPPR. They are connected directly to switch box outputs X, 19F, ¹H or to optional High Power Transmitters which may also be connected to these outputs.

Note and rules for manual routing

After each nucleus selection the default routing will be set. It can be accepted or, may be changed by the user. This should be done only after the complete set of nuclei has been selected. All changes in the routing are made by moving from the left to the right through the display. Connections between two units may be created or cut by two mouse clicks on the corresponding two buttons. The following rules apply:

- Only one F1...F8 logical channel must be connected to each FCU.
- Several FCU's may be connected to a single amplifier.
- *Router restrictions*: Router input 1 may be connected only to router output 1,2, or 3, router input 2 only to 1, 2, 3, and 4. Router inputs of the second or third router may be connected to the output of the first router only if no other input channel of this router goes to a different output channel of the first router.
- In the switch box each input button may be connected to any output button but this connection must be *one to one*. Double connections from the same input button are not allowed.
- Each preamplifier may only be connected to one amplifier, either through the switch box or directly. Note however that the displayed connection to a preamplifier is of no physical influence as far as the connections between the amplifiers and the preamplifier modules are concerned. It is up to the operator to

ensure that the wiring is correct. The same is true for the correct connection to the probehead.

At the time the program creates the default routing, the amplifier is chosen such that the nucleus type is considered. For ¹H or ¹⁹F nuclei an amplifier of type H is selected, for other nuclei an X type amplifier.

1.3.14 Install standard parameter sets [expinstall]

Execute the command <u>expinstall</u> (also available in the *Acquire->Spectrometer setup* menu). XWIN-NMR is delivered with sets of acquisition, processing, and plot parameters for many types of NMR experiments. They were compiled and tested on various instruments in Bruker laboratories. <u>expinstall</u> lets you select your instrument type, and stores the corresponding parameter sets, pulse programs, etc. in their working directories *XWINNMRHOME/exp/stan/nmr/par/, XWINNMRHOME/exp/ stan/nmr/lists/pp/*, etc. respectively. During this process, it adapts certain acquisition parameters such as the observe frequency to the basic frequency of your spectrometer. <u>cf</u> must therefore have been executed before.

Since <u>expinstall</u> may overwrite existing files, it should only be invoked by the administrator. In order to exclude such conflicts, we recommend not to modify parameter sets, pulse programs, etc. provided by Bruker. Instead, before applying changes, create a copy with a different name. <u>expinstall</u> needs only be executed once after installation of a new XWIN-NMR version, or, if at a later time additonal items are to be installed omitted initially.

On XWIN-NMR release media, in addition to the software modules Bruker provides pulse program libraries, parameter sets etc. as listed in Table 1.9. The purpose of <u>expinstall</u> is to install all desired items in their working directories for your type of instrument, and to update certain parameters accordingly. Since this is a critical configuration procedure, it should only be excuted by the system administrator. <u>expinstall</u> starts up with a table of spectrometers. Select the correct one (the instrument specified during <u>cf</u> configuration is enabled by default) and click on the Proceed button. A new table comes up showing the possible actions that may be performed. The highlighted buttons define the minimum required for data acquisition with Bruker pulse programs, and automated measurements based on Bruker standard parameter sets. You may click on a highlighted button to disable installation, and on other buttons to enable installation. As soon as you click on the Proceed button of the dialog box, all highlighted features will be installed. <u>expinstall</u> may be executed again at any later time to install items not selected earlier.

Pulse program library	
Composite pulse decoupling library	
AU program library	
Gradient file library	
Shape file library	
Standard experiments library	
Standard composite experiments library	
Scaling region files libary	

Table 1.9 Libraries to be installed with expinstall

Please do not enter other commands while <u>expinstall</u> is in progress. It will print a message when complete. Without compilation of AU programs, <u>expinstall</u> terminates within a few minutes. Compilation may take a few hours, depending on the number of AU programs and computer speed.

The following sections describe the features that may be installed.

1.3.14.1 Pulse programs

On the release media, pulse programs for many NMR experiments are delivered for various types of instruments in own directories according to Table 1.10. <u>expin-</u>

XWINNMRHOME/exp/stan/nmr/lists/pp.exam	AMX high resolution
XWINNMRHOME/exp/stan/nmr/lists/pp.rexam	ARX high resolution
XWINNMRHOME/exp/stan/nmr/lists/pp.dexam	AVANCE
XWINNMRHOME/exp/stan/nmr/lists/pp.solids	AMX/ASX solids
XWINNMRHOME/exp/stan/nmr/lists/pp.imag	micro-imaging
XWINNMRHOME/exp/stan/nmr/lists/pp.tomo	tomography

Table 1.10 Sample pulse program directories

stall copies them into the working directory XWINNMRHOME/exp/stan/nmr/lists/pp/, where they are searched for by pulse program manipulation commands such as

<u>edpul</u>, and by the acquisition commands. If the item *Enable Define Statements in Pulse Programs* is higlighted, <u>expinstall</u> opens each pulse program prior to installation, and removes all double semicolons (;;) occurring at the beginning of a pulse program line. Definitions of pulse program parameters such as d11=30m or d12=20u are now enabled as suggested by Bruker for certain experiments.

1.3.14.2 Composite pulse decoupling programs

expinstall copies them from the directories of Table 1.10. to the directory

XWINNMRHOME/exp/stan/nmr/lists/cpd.rexam	AMX/ARX/ASX
XWINNMRHOME/exp/stan/nmr/lists/cpd.dexam	AVANCE

Table 1.11 Sample cpd program directories

XWINNMRHOME/exp/stan/nmr/lists/cpd/.

1.3.14.3 Install and compile AU programs/modules

The XWIN-NMR release media include AU programs and modules for a number of special applications. You may inspect the source code of AU programs after their installation using the <u>edau</u> command. Usually the function of an AU program is described in its header. AU programs and modules are written in the C language and must be compiled before they can be executed with one of the command <u>xau</u>, <u>xaup</u>, <u>xaua</u>. <u>expinstall</u> copies the C source files from the directories

XWINNMRHOME/prog/au/src.exam/ XWINNMRHOME/prog/au/modsrc.exam/

to the working directories

XWINNMRHOME/exp/stan/nmr/au/src/ XWINNMRHOME/exp/stan/nmr/au/modsrc/.

Then they are compiled, and the excutable files are stored in the directories

XWINNMRHOME/prog/au/bin/ XWINNMRHOME/prog/au/modbin/.

1.3.14.4 Recompile all user AU programs

AU programs a user wrote himself using the <u>edau</u> command are also stored in the directory *XWINNMRHOME/exp/stan/nmr/au/src/*. They are not deleted if a new XWIN-NMR version is installed (unless they happen to have the name of a Bruker AU program; please avoid this). You must, however, recompile your own AU programs in order to make them compatible with the new program version.

1.3.14.5 Install library gradient files

On the release media, gradient files for a number of applications are delivered for various types of instruments in own directories according to Table 1.10. expinstall

XWINNMRHOME/exp/stan/nmr/lists/gp.exam	AMX high resolution
XWINNMRHOME/exp/stan/nmr/lists/gp.dexam	AVANCE
XWINNMRHOME/exp/stan/nmr/lists/gp.solids	AMX/ASX solids
XWINNMRHOME/exp/stan/nmr/lists/gp.imag	micro-imaging
XWINNMRHOME/exp/stan/nmr/lists/gp.tomo	tomography

Table 1.12 Sample gradient file directories

copies them into the working directory *XWINNMRHOME/exp/stan/nmr/lists/gp/*, where they are searched for by commands such as <u>edgp</u>, and by the acquisition commands.

1.3.14.6 Install library shape files

On the release media, shape file for a number of NMR experiments are delivered for various types of instruments in own directories according to Table 1.10. <u>expin-stall</u> copies them into the working directory *XWINNMRHOME/exp/stan/nmr/lists/ wave/*, where they are searched for by the acquisition commands.

1.3.14.7 Convert standard parameters sets

After installation of XWIN-NMR from the release media, the directory

XWINNMRHOME/exp/stan/nmr/par.300/

XWINNMRHOME/exp/stan/nmr/lists/wave.exam	AMX high resolution
XWINNMRHOME/exp/stan/nmr/lists/wave.solids	AMX/ASX solids
XWINNMRHOME/exp/stan/nmr/lists/wave.imag	micro-imaging
XWINNMRHOME/exp/stan/nmr/lists/wave.tomo	tomography

Table 1.13 Sample shape files directories

contains a collection of so-called standard experiments. An *experiment* is a directory with a complete set of acquisition, processing, and plot parameters prepared e.g. for a proton measurement, a 13C decoupling measurement, a COSY experiment, etc. These parameter sets were compiled and tested on a 300 MHz spectrometer in the Bruker application laboratory. Before you can make use of them, they must be adapted to your local requirements. Afterwards, they are stored in the directory

XWINNMRHOME/exp/stan/nmr/par/,

where they may be accessed by commands such as <u>rpar</u> and <u>dirpar</u>.

The conversion utility first requests the logical name of your printer and plotter, e.g. *\$hplh4p*. See command <u>cfpp</u> on printer/plotter installation. These names are inserted in the CURPRIN and CURPLOT parameters defining the output devices. At any later time you may overwrite these default settings if required using the command <u>edo</u>.

The next question is about the paper format of your plotter. The plot parameters of the standard parameter sets are adjusted for A3 size. You may leave A3, or change it to A, A4, or B. A3 and B will take over the default settings, A4 and A will change parameters according to the contents of the text file

XWINNMRHOME/exp/stan/nmr/lists/plotconvpar.

If you want to produce your own default values, you may edit this file according to the description in its header, and run <u>expinstall</u> again, enabling only the *Convert standard parameter sets* function.

The next question asks for the type of digitizer installed in your instrument. The answer is stored in the DIGTYP acquisition parameter of the standard acquisition parameter files.

Finally, for AMX instruments you may initialize the parameters XL and YL with the BSV10 attenuator setting 3 to prevent probe head damage.

Parameter set conversion will now start and take a few minutes. During this process, frequencies are adapted to your spectrometer, the sweep width and 2D increments are corrected, and, for Avance type spectrometers, the default frequency routing is initiated.

1.3.14.8 Update user permission files

Enable this item if you intend to use XWIN-NMR's automated spectrometer operation features (commands <u>set</u>, <u>run</u>), or the commands <u>quicknmr</u>, <u>setexp</u>, and <u>runexp</u>. For any XWIN-NMR user, a permission file may be created by the system administrator using the <u>eduser</u> command. Such a file contains the standard experiments which may be carried out by this user, and other possibilies (see <u>eduser</u> command; an example file is <u>XWINNMRHOME/exp/stan/nmr/lists/sam_users_exam</u>). Usually, a new XWIN-NMR version includes new standard experiments stored in the file

XWINNMRHOME/exp/stan/nmr/par.300/.News.

This part of <u>expinstall</u> displays a dialog box of all users. If you click on a user, <u>expinstall</u> merges the new experiments into his (her) permission file. If you click on the *Select all* button, all users will be updated.

1.3.14.9 Install standard composite experiments

A sequence of standard experiments may be composed to build a new standard experiment, called a *composite experiment* (see also <u>set</u> command). Such experiments may be selected for execution in the dialog windows opened by <u>set</u> or <u>quick-nmr</u>. These routines locate composite experiments in the file

XWINNMRHOME/conf/instr/<Instrument Name>/users/.pool.

If you enable this <u>expinstall</u> item, this file is created by making a copy of the standard composite experiments provided by Bruker in the file

XWINNMRHOME/exp/stan/nmr/pp.300/.pool.

The *.pool* file of composite experiments is common to all users. With the <u>set</u> command, you may define own composite experiments, and define which user is allowed to execute a particular experiment.
1.3.14.10 Install standard scaling region files

The directory XWINNMRHOME/exp/stan/nmr/lists/scl.exam/ includes files whose names are composed of a nucleus and a solvent name, e.g. 13C.Acetic. The files contain spectral regions, usually around the solvent and reference, which are to be excluded when the program scales a plot, or during sweep width optimization with commands such as getlim. If this expinstall item is enabled, the sample region files will be copied to their working directory

XWINNMRHOME/exp/stan/nmr/lists/scl/,

where they are accessed by the respective commands. If you need solvent/nucleus combinations other than those provided by Bruker, you may create own suitable files in this directory.

1.4 Defining the acquisition data set [new, edc]

Before you can set up the parameters for a data acquisition, you must define a data set where the acquired data (fid), and later on the spectrum are to be stored. In XWIN-NMR, data set a characterized by the 5 parameters

DU, USER, NAME, EXPNO, PROCNO.

From these parameters, two directories are derived:

/DU/data/USER/nmr/NAME/EXPNO/ /DU/data/USER/nmr/NAME/EXPNO/pdata/PROCNO/.

Examples:

/u/data/guest/nmr/sucrose/1/ /u/data/guest/nmr/sucrose/1/pdata/1/

In the first directory, acquisition data are stored by the acquisition commands in the files *fid* or *ser*. A *ser* file contains multiple fids and is the result of a two-, three, or higher dimensional experiment.

Defining a data set for a single experiment acquisition with zg

In order to define a new data set, call the command <u>New</u> from the *File* menu, or enter <u>new</u> or <u>edc</u> on the keyboard. A dialog box is displayed where you may enter the data set parameters. DU is the disk partition where the data are to be located,

and USER is the user's own or another legal login name. NAME is an arbitrary name assigned to the data set. EXPNO is a number, allowing you to run experiments under the same NAME, and differentiating them by the EXPNO. Similarly, PROCNO is a number, and may be used to store several processed data sets derived from the same acquisition data. If you enter data set parameters for which a data set already exists, it will be display as soon as you click on the OK button of the dialog box. This is the same as if you had selected the data set using one of the dir commands, or the search command (see *The File Menu*). Any acquisition command issued now would overwrite existing acquisition data with the new *fid* or *ser* file. XWIN-NMR will output a warning prior to acquisition start only if the system variable *ZGsafety* is set to *yes*. Use the command <u>setres</u> to set *ZGsafety*, which is equivalent to invoking <u>User interface</u> from the *Display->Options* menu.

If no data set exists corresponding to the specified parameters, a new one is created, and is initialized with the same set of acquisition, processing, and plot parameters valid for the currently displayed data. The program display switches to the new data set (i.e. makes it the *current* data set), but the data area of the XWIN-NMR window remains blank since no data are present yet.

Defining the data set for an experiment to be executed from <u>quicknmr</u> Use the *sample name* entry field in the <u>quicknmr</u> dialog window.

Defining the data sets for a series of experiment to be executed by <u>run</u> Use the *name* entry field in the <u>set</u> dialog window.

1.5 Setting up acquisition parameters

This section will discuss the following set up commands for data acquisition:

- <u>edinfo</u> (edit sample information)
- <u>eda</u>, <u>rpar</u> (general acquisition parameter set up for the commands <u>zg</u>, <u>go</u>, <u>gs</u>)
- <u>ased</u>, <u>as</u> (acquisition parameter set up for the commands <u>zg</u>, <u>go</u>, <u>gs</u> restricted to parameters referred to in the current pulse program)
- <u>wobb</u> (probehead tuning)
- **gs** (interactive adjustment of acquisition parameters)
- <u>set</u> (parameter set up for a series of experiments started with <u>run</u>)
- <u>extset</u> (external parameter set up for <u>run</u> using ASCII files)

1.5.1 Setting up the sample information file [edinfo]

This command is *not* required for data acquisition to perform properly. However, it allows you to store sample or company specific information of your choice such as sample id, order number etc. to be entered and stored along with acquisition data. You may also append this information to the spectrum plot. Please refer to the description of <u>edinfo</u> in the chapter *The File Menu*.

1.5.2 Acquisition parameter setup with eda

<u>eda</u> is the most general parameter setup for a data acquisition to be started with <u>zg</u>. With <u>eda</u>, you get the settings for the current data set displayed, and you may modify them. <u>eda</u> may also be called from the <u>set</u> dialog window, which is the parameter set up for experiments executed by the <u>run</u> command. Likewise, <u>eda</u> may be called from the <u>quicknmr</u> dialog window. Before you call <u>eda</u>, you may use <u>rpar</u> (see chapter *The File Menu*) to initialize the current acquisition parameters with a standard parameter set from the directory *XWINNMRHOME/exp/stan/nmr/par/*. *Please note:*

Certain buttons in the <u>eda</u> dialog window, such as PULPROG, will open a box with a list of items, e.g. pulse program names. In order to select an item, you must click on it. If you happen to click outside the box, the box will be *disabled*: If you move the cursor back into the box, no mouse clicks will be accepted until you *re-enable* it by hitting the ESC key.

1.5.2.1 Invoking eda from the keyboard or from the Acquire menu

The acquisition parameters displayed in this case are a copy of those of the data set where the <u>new</u> command was given. You may overwrite them with an arbitrary predefined parameter set using the command <u>rpar</u>. <u>rpar</u> displays the table of parameter sets available in the directory *XWINNMRHOME/exp/stan/nmr/par/*. It contains the Bruker standard experiments (see <u>expinstall</u>) as well as others you or your system administrator copied there with the <u>wpar</u> command. <u>rpar</u> allows you to selectively overwrite either your current acquisition, plotting, or processing parameters (or all of them). After <u>rpar</u>, you may change individual parameters with <u>eda</u>.

The command <u>eda</u> displays all acquisition parameters of the current data set at once in a table, and allows you to apply modifications. The parameters are loaded from the parameter file (a text file)

/DU/data/USER/nmr/NAME/EXPNO/acqu.

Upon exit from <u>eda</u> via the SAVE button of the <u>eda</u> dialog window, all changes are written into this file. Any parameter may also be entered from the keyboard: Type in the parameter name, followed by Return. This will print the current value and waits for change or confirmation. As a second possibility, enter the parameter name followed by a space character, then type in the new value, followed by Return.

If the current data set is a 2D or 3D data set, <u>eda</u> shows two or three columns labelled F2 and F1, or F3, F2, and F1, respectively. The leftmost column (F2 for 2D, F3 for 3D) defines the acquisition dimension, whose parameters are stored in the file *acqu*. Table 1.14 shows which parameter files correspond to the different dimensions. You may access the parameters of any dimension also from the key-

	F3		F2		F1	
Data set type	Param. Files	Com- mand	Param. Files	Com- mand	Param. Files	Com- mand
1D	-	-	-	-	acqu	<u>td</u>
2D	-	-	acqu	<u>td</u>	acqu2	<u>2 td</u>
3D	acqu	<u>td</u>	acqu2	<u>2 td</u>	асqиЗ	<u>1 td</u>

 Table 1.14 eda columns, corresponding parameter files and command example

board. Table 1.14 also shows the keyboard commands to access the time domain size parameter TD, which exists for all dimensions of a 2D or 3D data set. Please note that typing the parameter name (in lower case characters) always accesses the acquisition dimension if not preceded by a dimension number.

Acquisition commands such as <u>gs</u>, <u>zg</u>, <u>go</u>, <u>wobb</u>, or <u>rga</u> read in the acquisition parameters from the files *acqu (acqu2, acqu3)*, compile them into a form suitable for the spectrometer hardware, and load them into the hardware for execution of the experiment.

The next section describes all the acquisition parameters in detail. Some of them depend on the spectrometer type and are required e.g. for an AMX, but not for a DMX. The contents of the <u>eda</u> dialog box may therefore differ accordingly. Actually, the parameters appearing in this window are taken from the so-called format file

XWINNMRHOME/exp/stan/nmr/form/acqu.e,

which is linked (made identical) to a format file suitable for your instrument at the time of spectrometer configuration with \underline{cf} . Format files are text files, which you may adapt to your requirements, e.g. remove parameters you will never use or change, or change their output format. If you re-arrange the sequence of parameters in a format file, the <u>eda</u> display will change accordingly.

Please note: Before modifying a format file, make a copy!

Format files do not only define the arrangement of parameters in the <u>eda</u> window. Some parameters, such as DW, AQ, FIDRES, SWH are defined in the format file by means of an equation, expressing the relation to other parameters (e.g. SWH=SW*SFO1). Parameters defined in a format file in this way are called *temporary parameters*, since they are not stored in an acquisition parameter file *acqu*, which only contains parameters not depending on each other. You may insert own temporary parameters into a format file by generating a corresponding new entry in the format file similar to the entries of DW, AQ, etc.

1.5.2.2 Invoking eda from the keyboard or from set or quicknmr

Now, the parameters displayed are not those contained in the $acqu^*$ files of the currently displayed data set. Instead, they correspond to the experiment selected in the <u>set</u> or <u>quicknmr</u> dialog window, and are valid only for the data set currently selected in these dialog windows. See also command <u>set</u> for more details.

1.5.2.3 eda command buttons

Table 1.4 lists the command buttons available in the eda dialog box.

1.5.2.4 The acquisition parameters

The acquisition parameters are described in the standard eda order.

PULPROG (pulse program)

The pulse program to be executed by the acquisition commands <u>gs</u>, <u>zg</u>, <u>go</u>, and <u>rga</u>. Type in a name, or click on the down arrow button right of the entry field to get a list of available pulse programs. Click on the desired one to store it in PULPROG. The first list entry is EDIT CURRENT, a command which displays the text of the current pulse program. The pulse program list is generated according to the con-

SAVE	Save modifications and quit.	
1-COL (only 1D)	Change <u>eda</u> display such that the parameters are arranged in a single column, including their description.	
2-COL (only 1D)	Change <u>eda</u> display such that the parameters are arranged in two columns, omitting their description.	
Parameter	Search. An entry field where you may specify the name or some initial characters of a parameter. After typing Return, the cursor is auto-positioned to the parameter.	
NEXT	Continue search to next parameter matching the string.	
ABORT	Discard any changes and quit.	

Table 1.15 Command buttons in eda dialog window

tents of the directory

XWINNMRHOME/exp/stan/nmr/lists/pp/.

It contains the Bruker pulse programs according to the <u>expinstall</u> installation, and the pulse programs created by yourself or your system administrator by means of the command <u>edpul</u>.

A special chapter in this manual (*Writing Pulse Programs*) describes the pulse program language. The command <u>pulsdisp</u> displays a graphical representation of the current pulse program. You will find a description of <u>pulsdisp</u> in the chapter *The Windows Menu*.

AQ_mod (acquisition mode)

The acquisition mode with the possible settings *qf*, *qseq*, *qsim* (=*standard setting*), *DQD*.

- qf = single channel detection. Only one detector is used.
- *qseq* = quadrature detection, sequential mode. Two detectors with a reference phase shifted by 90 degrees are used. In the resulting fid, two successive data points originate from different detectors. Their acquisition time difference is given by the dwell time parameter DW.
- qsim = quadrature detection, simultaneous mode. Two detectors with a reference phase shifted by 90 degrees are used. In the resulting fid, two successive data points originate from different detectors, but were sampled at the same

time. The acquisition time difference of two pairs of data points is two times the dwell time parameter DW.

• *DQD*= digital quadrature detection (only available on suitably equpipped Avance type spectrometers). Requires the acquisition parameter DIGMOD to be set to *digital* or *homodecoupling digital*. Digital quadrature detection has the advantage over the previously described analog modes that the two receiver channels are generated mathematically by the digital filters of such instruments, and for this reason are not offset against each other, leading to a cleaner fid signal.

AQSEQ (acquisition sequence)

This parameter AQSEQ which take on one of the values 312 or 321. It is evaluated by the 3D Fourier transform and describes the sequence of fids acquired in a *ser* file during a 3D experiment (3 = the acquisition dimension, 1 and 2 = the orthogonal dimensions). The processing parameter AQORDER is not evaluated if AQSEQ is set.

On Avance type systems, 3D pulse programs will set AQSEQ automatically if td and td1 are used consistently within the pulse program. However, you may explicitly define AQSEQ in the pulse program. For this purpose, insert one of the following statements in the pulse program header: aqseq 321 or aqseq 312.

On both, AMX/ARX and Avance type systems, you can set or modify AQSEQ using the command <u>3s AQSEQ</u> before starting the transform.

TD (time domain size)

Defines the number of data points of the fid. For data set dimensions bigger than 1, TD for F1 and F2 usually specifies the number of increments. It may be used in pulse programs, e.g. by the pulse program commands <u>lo to x times td1</u> or <u>lo to x</u> times td2. Enlarging TD will give a better resolution of the fid (see parameter FID-RES) and hence the spectrum at the expense of a longer scan time AQ.

PARMODE (dimension of acquisition data)

1D, *2D*, or *3D*. Specifies the type of data to be acquired. If you change PARMODE to a smaller dimension, unnecessary parameter files are deleted. If you switch to a higher dimension, the required additional parameter files are fetched from the directories

XWINNMRHOME/exp/stan/nmr/par/standard2D/, or *XWINNMRHOME/exp/stan/nmr/par/standard3D/*.

Existing acquisition data are deleted after approval by the user.

NS (*number of scans*)

Defines the number of accumulation loops the pulse program commands $\underline{go=n}$ and $\underline{rcyc=n}$ should perform.

DS (number of dummy scans)

Defines the number of loops the pulse program command $\underline{go=n}$ should perform without digitizing and accumulating data, before continuing with NS true scans. Used to get steady state conditions.

D0-D31 (delay parameters in seconds)

32 duration parameters, to be specified in seconds. They are executed as delays without any further actions by the pulse program commands d0-d31. In Bruker pulse programs, D1 is most often used as a relaxation delay, and D0 and D10 as incrementable delays for 2D and 3D experiments, although any delay may be incremented or decremented using the pulse program commands <u>id0-id31</u> or <u>dd0-dd31</u>, respectively. The changes are given by the parameters IN0-IN31. The pulse program commands <u>rd0-rd31</u> reset a respective delay to its original value. On Avance systems, delays may also be changed by means of arithmetic expressions during pulse program execution.

P0-P31 (pulse length parameters in microseconds)

32 duration parameters, to be specified in microseconds. They are executed by the pulse program commands <u>p0-p31</u>, which execute a pulse on the specified channel during the respective duration parameter. In Bruker pulse programs, P1 is most often used as a 90 degree pulse. Pulses may be incremented or decremented by the pulse program commands <u>ipu0-ipu31</u> or <u>dpu0-dpu31</u>, respectively. The changes are given by the parameters INP0-INP31. The pulse program commands <u>rpu0-rpu31</u> reset a respective pulse length to its original value. On Avance systems, pulses may also be changed by means of arithmetic expressions during pulse program execution.

ND0, ND10 (number of delays D0 or D10)

Number of d0 and d10 commands in the increment loops of a pulse program for 2D or 3D experiments. Used to calculate the sweep width for the F1 and F2 dimensions according to

SW(F1)=1/(SFO1*ND0*IN0), SW(F2)=1/(SFO1*ND10*IN10),

respectively.

IN0, IN10 (increments for delays D0 and D10 in seconds)

The pulse program commands <u>id0</u> and <u>id10</u> increment the delays D0 and D10 by IN0 and IN10, respectively. See also description of D0-D31. If you change IN0 during the setup of a 2D experiment, the sweep width in the dimension F1 is recalculated according to SW=10e6/(SFO1*DR*IN0). If you change IN0 during the setup of a 3D experiment, the sweep width in the dimension F1 is recalculated according to SW=10e6/(SFO1*DR*IN0). If you change IN10 during the setup of a 3D experiment, the dimension F2 is recalculated according to SW=10e6/(SFO1*DR*IN0).

Features:

a) changing ND10 on a 3D data set in <u>eda</u> results in a correct calculation of IN10, SW and SWH.

b) if the parameter IN0 or IN10 is changed by typing, for instance, in0 or in10 from the XWIN-NMR command line, then the sweep width in ppm is updated correctly, as well as the sweep width in Hz.

c) changing the sweep width by typing 1 SWH or 3 SWH from the XWIN-NMR command line will change the increment.

d) Setting the parameters in an AU program with STOREPAR is possible.

If you want to change the parameters from the XWIN-NMR command line, type:

 $\underline{nd0}$: to change the sweep width in F1 of a 2D or 3D data set (the increment IN0 is NOT changed)

nd10: to change the sweep width in F2 of a 3D data set (the increment IN10 is NOT changed)

 $\underline{in0}$: to change the increment IN0 in F1 of a 2D or 3D data set (the sweep width will be adjusted, but not ND0)

 $\underline{in10}$: to change the increment IN10 in F2 of a 3D data set (the sweep width will be adjusted, but not ND10)

1 SWH (1 SW): to change the sweep width in Hz (ppm) in F1 of a 2D or 3D data set (the increment IN0 will be adjusted, but not ND0)

2 SWH (1SW): to change the sweep width in Hz (ppm) in F2 of a 3D data set (the increment IN10 will be adjusted, but not ND10)

Note : <u>2 SWH</u> (<u>2 SW</u>) on a 2D data set will change the sweep width in F2 and <u>3</u> <u>SWH</u> (<u>3 SW</u>) on a 3D data set will change the sweep width in F3! Therefore, increments and ND0 or ND10 values will NOT be effected, because the changes are made in the acquisition dimension and not in one of the indirect detected dimension.

If you want to change the parameters in an AU program, use :

storepar1("ND0", value) : changes the sweep width in F1 of a 2D data set, if ND0 is changed. *value* must be a INTEGER variable.

storepar3("ND0", value) : changes the sweep width in F1 of a 3D data set, if ND0 is changed. *value* must be a INTEGER variable.

<u>storepar1("ND10", value</u>) : changes the sweep width in F2 of a 3D data set, if ND10 is changed. *value* must be a INTEGER variable.

storepar1("INO", value) : to change the increment IN0 in F1 of a 2D data set. *value* must be a FLOAT variable.

storepar3("IN0", value) : to change the increment IN0 in F1 of a 3D data set. *value* must be a FLOAT variable.

storepar1("IN10", value) : to change the increment IN10 in F2 of a 3D data set. *value* must be a FLOAT variable.

storepar1("SWH", value) : to change the sweep width in Hz in F1 of a 2D data set. *value* must be a FLOAT variable.

storepar3("SWH", value) : to change the sweep width in Hz in F1 of a 3D data set. *value* must be a FLOAT variable.

storepar1("SWH", value) : to change the sweep width in Hz in F2 of a 3D data set. *value* must be a FLOAT variable.

storepar1("SW", value) : to change the sweep width in ppm in F1 of a 2D data set. *value* must be a DOUBLE variable.

storepar3("SW", value) : to change the sweep width in ppm in F1 of a 3D data set. *value* must be a DOUBLE variable.

storepar1("SW", value) : to change the sweep width in ppm in F2 of a 3D data set. *value* must be a DOUBLE variable.

SW, SWH (sweep width in ppm, Hz)

The sweep width may be specified in ppm or Hertz units by setting one of the parameters SW or SWH. The other parameter is calculated automatically (SWH=SW*SFO1). The sweep width is the spectral range to be observed. SW in conjunction with TD or the desired scan time AQ define the dwell time DW, which triggers the digitizer to sample the individual data points. The digitizer hardware only allows for discrete values of DW, and therefore for SW. This explains the observation that the program often changes SW slightly if you enter an arbitrary value. SW may also be set interactively from the *sw-sfo1* button in the 1D *utilities* menu. It ajusts SW so that it has the same value as the expanded region, and also adjusts SFO1 so that the carrier frequency lies in the center of the expanded region.

For 2D and 3D experiments, SW as described above corresponds to the width in the acquisition dimension, while the widths of the other dimensions are calculated from the parameters IN0, IN10, ND0, and ND10.

FIDRES (fid resolution in Hertz)

A temporary parameter calculated according to FIDRES=SW*SFO1/TD. If you modify FIDRES to get a desired resolution, the time domain size is recalculated according to TD=nextpower of 2 of ((SW*SFO1)/FIDRES).

FW (filter width in Hertz)

The analog filters are set according to this value when acquisition is started, or with the <u>ii</u> command. If you change the sweep width, FW is recalculated according to FW=1.25*SWH or FW=0.3125/DWOV. The latter equation is valid for Avance systems, if digital filtering is enabled (DIGMOD=*digital* or *homodecoupling digital*). DWOV is the oversampling dwell time (see also parameter DIGMOD).

AQ (acquisition time in seconds)

The time to record one scan. Doubles if you increase TD or SW by a factor of two. If you change AQ, TD is recalculated accordingly. It will *not* be rounded to the next power of two, but be set to an even number giving the closest match to the entered AQ value. Then AQ is recalculated from TD and SW.

RG (receiver gain)

The receiver gain controls the amplitude of the fid signal before it is fed into the digitzer. It must be adjusted to maximum amplitude, without causing overflow of the digitizer (in this case signal cutoff would occur, generating artefacts in the transformed spectrum). The values depend on the receiver system of your spectrometer. You may use the command <u>rga</u> for automatic determination of the optimum RG value. <u>rga</u> performs some scans using the pulse program specified in PULPROG and sets RG to the found value. You may then further adjust RG to fit your special requirements.

DW (dwell time in microseconds)

The dwell time is the time difference between two data points of the fid. It is calculated from the sweep width according to DW=10e6/(2*SW*SFO1). If you change DW, SW is recalculated according to SW=10e6/(2*(0.05+DW)*SFO1). The minimum dwell time (corresponding to the maximum possible sweep width) depends on the digitizer type installed in the spectrometer (see also: DIGTYP parameter, Table 1.16).

DWOV (oversampling dwell time)

For spectrometers equipped with digital filters (Avance) only. It is displayed as an information for the user, and is related to the dwell time DW according to DW=DWOV*DECIM. DWOV is the actual sample rate of the digitzer if digital filtering is enabled. The decimation factor DECIM is chosen so that after the Fourier transform the spectrum will have a spectral width as close as possible to the chosen SW parameter. See also: DIGMOD parameter.

DECIM (decimation rate of digital filter)

For spectrometers equipped with digital filters (Avance) only. For details see: DWOV and DIGMOD parameters.

DSPFIRM (firmware used for digital filtering)

For spectrometers equipped with digital filters (Avance) only. For details see: DIG-MOD parameter.

DIGTYP (*digitizer type*)

XWIN-NMR supports the digitizer types of Table 1.16. The SADC is not prepared

DIGTYP	dynamic range	minimum DW
slow	12 bits	3 microsec
16bit	16 bits	4 microsec
fast	9 bits	0.1/0.05 microsec
BC132-12	12 bits	0.1 microsec
BC132-16	16 bits	2 microsec
FADC (=BC133)	12 bits	0.05 microsec
HADC (=HRD16)	16 bits	2.5 microsec
SADC	16 bits	3.325 microsec
HADC+	16 bits	3.325 microsec
SADC+	16 bits	3.325 microsec
IADC	16 bits	0.1/0.05 microsec

Table 1.16 Digitizer types

for sequential acquisition, i.e. AQ_mod must not be set to *qseq*. DIGTYP must be set to a type installed in your spectrometer (more than one may be present). When

executing the <u>expinstall</u> command, you may enter a desired type. In this case, the DIGTYP parameter of all standard parameters sets are updated accordingly.

DIGMOD (*digitization mode*)

For spectrometers equipped with digital filters (Avance) only. The following three modes are available: *analog*, *digital*, *homodecoupling-digital*. *Analog* disables digital filtering and uses the analog filters in the conventional way. *Digital* turns on digital filtering. *Homodecoupling-digital* also enables digital filtering, but limits the sampling rate so that homodecoupling is still possible.

The receiver control unit (RCU) of all Avance instruments is equipped with two digital signal processors (DSP) which can perform digital filtering during the acquisition. The digitizer will sample with a dwell time DWOV which is related to the conventional dwell time by

DW = DWOV * DECIM

The DSP will decimate the sampled points by a factor DECIM, and at the same time perform digital filtering. The oversampling dwell time DWOV must be in the range 2.5-5 µsec for HADC/HRD16 and FADC, and in the range 3.3-6.6 µsec for a SADC. For homodecoupling the oversampling dwell time must be between 25 and 50 µsec. The decimation factor DECIM is chosen so that after the Fourier transform the spectrum will have a spectral width as close as possible to the chosen SW parameter. The factor DECIM is stored as an acquisition status parameter (in the file *acqus*). It is required to calculate a correct baseline correction of the fid, and the first order phase correction during the Fourier transform (for the standard DSP firmware; the parameter DSPFIRM may be set to *used-defined* if non-standard firmware should be used. In this case the firmware file (in ASCII) must be located in *XWINNMRHOME/exp/stan/nmr/lists/DSPFIRM*).

Digital filtering enhances the digitizer resolution by a factor given by the parameter DDR (digital digitizer resolution). The total digitizer resolution, which is the sum of the hardware resolution according to Table 1.16 and DDR is still characterized by DR. DDR is related to DECIM according to(.

$$DDR = 2 \log \sqrt{DECIM} + 1$$

With DIGMOD=*digital* or *homodecoupling digital* it is possible to use AQ_mod = qsim or DQD (digital quadrature detection), but not AQ_mod = qseq (sequential data acquisition).

DR (digitizer resolution)

Acquisition commands will normally set the digitzer to its maximum resolution (the default of DR). If you set DR to a smaller value, acquisition will start with the modified setting.

DDR (digital digitizer resolution)

For spectrometers equipped with digital filters (Avance) only. For details see: DIG-MOD parameter.

DE (pre-scan delay in microseconds)

A waiting period between the last pulse and the begin of data acquisition (digitizer start) to avoid pulse feed through. You may change DE by entering a new value. DE includes the sections, DE1 and DE2 (Figure 1.6). After DE1 the receiver gate is opened.

For AMX/ARX type spectrometers, the program calculates DE according to DE=1.4288*DW (or half this value if AQ_mod=*qf*). DE1 and DE2 are constants.

For Avance type spectrometers, you may change DE1 and DE2 using the command <u>edscon</u>. See <u>edscon</u> for details.

PHP (preamplifier module selection)

AMX/ARX type instruments only. 0=X-BB(HR), 2=X-QNP(HR), 8=X-BB(HP), 10=X-QNP(HP).

QNP (*QNP selection*) An integer number (1-4) is required.

HL1-HL4, (ecoupler high power levels in db) FL1-FL4 (F19 decoupler high power levels in db) XL, YL (BSV!0-X, BSV!0-X power level) ZL1-ZL4 (4th channel low power levels in db)

AMX/ARX type instruments only. The pulse program commands <u>hl1-hl4</u> and <u>fl1-fl4</u> may be used to enable the respective high power level of the ecoupler or 19F decoupler. The default power (which is set if power is not explicitly switched with <u>hl1-hl4</u>) is HL1 (and FL1 for F19).

If the spectrometer is equipped with a 4th channel, the pulse program commands $\underline{zl1}$ - $\underline{zl4}$ may be used to set the low power according to the parameters ZL1-ZL4. The command \underline{zlo} switches the 4th channel decoupler to low power mode, \underline{zhi} to high power mode. There is no possibility of controlling the high power.

PL0-PL31 (32 power levels in db)

Avance type instruments only. The pulse program commands <u>pl0-pl31</u> may be used to set power levels. <u>pl0</u> takes the power value from the PL0 parameter, etc. The frequency channel must be specified behind the command. E.g., the pulse program command <u>pl1:f2</u> sets the power level of channel f2 to the value PL1. Channels f1 to f8 are legal, and require the corresponding hardware equipment. The default power setting for channel n (which is used if power is not explicitely switched with <u>pl0-pl31</u>) is PLn. The parameters may also be accessed from the keyboard. Enter the command <u>pl0</u> to set PL0, etc. PL0-PL31 may aslo be accessed from CPD programs to control the power of hard and shaped pulses. For example, the CPD program command <u>p31:sp1:180 pl=pl1</u> uses PL1 for the shaped pulse sp1.

CNST (array of constants)

32 floating point constants. For free use in pulse programs, usually in arithmentic expressions. Example: p2=p1*cnst1. This statement would use the array element CNST[1].

NUCLEI (set up nuclei), NUC1, BF1, SFO1, O1

The *edit* NUCLEI button is used to assign nuclei to the frequency channels of the spectrometer. The program will set the basic frequeny of a channel according to the selected nucleus. For channel 1, this frequency is designated as BF1, for channel 2 BF2, etc. The corresponding actual irradiation frequencies are SFO1, SFO2, etc. They are calculated from the basic frequencies by adding a *frequency offset* which may also be specified during nucleus setup. Nucleus, basic frequency, irradiation frequency and frequency offset are displayed in the <u>eda</u> window using the parameter names NUC1, BF1, SFO1, O1/O1P; NUC2, BF2, SFO2, O2/O2P, etc. O1P corresponds to O1, but is displayed in ppm rather than Hertz (channels 5-8 for Avance systems are located at the end of the <u>eda</u> table). Only the offsets may be changed in the <u>eda</u> window, the other parameters are there just to inform you about their current setting.

For Avance systems, while editing nuclei, you may also change the default routing (i.e. assign a different amplifier or preamplifier to a channel).

The *edit* NUCLEI button in fact invokes the command <u>edasp</u>. You may therefore set up nuclei not only from <u>eda</u>, but directly from the keyboard by entering this command. See also the description of <u>edasp</u> to find out further details.

For AMX/ARX type spectrometers, the nucleus names for the channels 1-3 are

designated NUCLEUS, DECNUC, and DECNUCB, respectively, rather than NUC1-NUC3. Channel 4, which is a special hardware accessory for these instruments, has no nucleus parameter assigned.

TL, DL, DBL (low power levels in db for square pulse transmitters)

AMX/ARX type instruments with 3-channel interface (MCI) only. These are 3 parameter arrays of size 8, TL[0-7], DL[0-7], and DBL[0-7], one array for each channel. The square pulse low power level may be set in a pulse program with the commands <u>tl0-tl7</u>, <u>dl0-dl7</u>, <u>dbl0-dbl7</u>. For example, <u>tl3</u> sets the power value TL[3] in channel 1, <u>dbl6</u> sets the value DBL[6] in channel 3. Low power mode for the 3 channels is activated with the pulse program commands <u>tl0</u>, <u>dbl0</u>. Correspondingly, high power mode is activated with the commands <u>thi</u>, <u>dhi</u>, <u>dbhi</u> (channel 1 = transmitter, channel 2 = decoupler, channel 3 = second decoupler).

CPDPRG1-CPDPRG8 (cpd programs)

Avance type instruments only. These parameters must contain the names of composite pulse decoupling (cpd) programs if a cpd experiment is to be carried out, i.e. if the pulse program contains commands such as <u>cpd1-cpd8</u> and <u>cpds1-cpds8</u>. They execute the cpd program given by the parameters CPDPRG1-CPDPRG8 (the <u>cpds</u> commands differ from the <u>cpd</u> commands in that they will execute the cpd program synchronously with the pulse program). The channel where a <u>cpd</u> command is to execute its associated cpd program must be specified behind the <u>cpd</u> command. For example, <u>cpds5:f2</u> executes CPDPRG5 on channel f2 in synchronous mode.

If you click on the down-arrow button right of the CPDPRG parameters, a list of available cpd programs is displayed, from which you may select one. The list is created according to the contents of the directory

XWINNMRHOME/exp/stan/nmr/lists/cpd/.

Initially (after configuration of XWIN-NMR is complete) it contains Bruker's cpd programs. You may add own ones with the command <u>edcpd</u>.

CPDPRG, CPDPRGT, CPDPRGB, CPDPRG4 (cpd programs)

AMX/ARX type instruments with 3-channel interface (MCI) only. These parameters must contain the names of composite pulse decoupling (cpd) programs if a cpd experiment is to be carried out, i.e. if the pulse program contains <u>cpd</u> type commands. The pulse program commands <u>cpd</u> and <u>cpds</u> execute the program CPDPRG (<u>cpds</u> synchronously with the pulse program) on the decoupler. <u>cpdt</u> and <u>cpdts</u> execute CPDPRGT on the transmitter. <u>cpdb</u> and <u>cpdbs</u> execute CPDPRGB on the second decoupler. <u>cpd4</u> executes CPDPRG4 on the fourth channel (requires a special hardware equipment).

If you click on the down-arrow button right of the CPDPRG parameters, a list of available cpd programs is displayed, from which you may select one. The list is created according to the contents of the directory

XWINNMRHOME/exp/stan/nmr/lists/cpd/.

Initially (after configuration of XWIN-NMR is complete) it contains Bruker's cpd programs. You may add own ones with the command <u>edcpd</u>.

GRDPROG (gradient program)

Requires Avance or AMX/ARX type instruments with 3-channel interface (MCI). This is the name of a gradient program which must be located in the directory

XWINNMRHOME/exp/stan/nmr/lists/gp/.

It is executed by the pulse program command <u>ngrad</u>. See also <u>edgp</u> command. For Avance type spectrometers, we recommend to use the pulse program commands <u>gron/groff</u> for gradient control, and the :gp pulse option for shaped gradients, which are more convenient.

LOCNUC (*lock nucleus*)

Must be set to the desired lock nucleus before invoking the <u>edlock</u> command, which sets up the lock parameters for this nucleus.

SOLVENT (*solvent*)

Must be set to the solvent used in the current sample. SOLVENT is evaluated by the commands <u>prosol</u>, <u>lock -acqu</u>, <u>sref</u> and <u>lopo</u>, and during <u>quicknmr</u> and <u>run</u>. For the latter two applications the acquisition parameters are obtained in the following way: They are initialized with the parameters of the specified experiment. The probe head and solvent dependent parameters (see command <u>prosol</u>) are then inserted according to the setting of SOLVENT and the current probe head. Finally, any parameter changes the user possibly requested are applied.

PROBHD (probehead)

A status parameter stored at the end of an experiment according to the current probehead set with <u>edhead</u>.

PROSOL (update probehead/solvent dependent parameters)

This parameter may take on the values true or false. If set to true, the following

will happen in <u>eda</u>: The pulse lengths and power levels of the current data set depending on the solvent and probehead will be overwritten by the values defined with <u>prosol</u> or <u>solvloop</u>. The pulses or powers (e.g. P1, PL0) concerned may be taken from the AU program *pulsesort* which describes the prosol parameters and their associated <u>eda</u> parameters.

EXP (experiment performed)

Reserved to be updated by <u>quicknmr</u> and <u>run</u> with the experiment peformed.

RO (spinner rotation frequency in Hz)

The AU command \underline{ro} will set sample rotation to the frequency RO. The command will wait for 15 seconds and check if the desired rate could be achieved. If this is not the case, an error message is printed. If you enter the command \underline{ro} on the keyboard, you may disable or enable sample rotation, and specify the desired rate.

TE (temperature in Kelvin)

The command <u>teset</u> sets the Eurotherm temperature unit according to TE. teset may be entered on the keyboard, or be used in an AU program. See also: Eurotherm unit, command <u>edte</u>.

NBL (number of buffers)

Used by pulse programs acquiring NBL fids in the memory of the acquisition processor before writing them to disk. NBL is evaluated by the pulse program commands <u>lo to x times nbl</u> (which performs this many loops), <u>ze</u>, <u>zd</u>, <u>wr</u>, <u>if</u>, <u>df</u>, <u>st</u>, and <u>st0</u>. Normally, NBL is 1. If you set NBL to a bigger value, you may acquire up to NBL fids in memory (limited by the equipment of the acquisition processor). A buffer of the required size is reserved (NBL*TD). <u>ze</u> and <u>zd</u> will clear the buffer. <u>st0</u> will set the memory pointer where the next fid should be placed during acquisition to buffer start. <u>st</u> will increment the pointer by TD. <u>wr</u> will copy NBL fids from acquisition memory to disk in a *ser* file, starting at buffer begin. <u>if</u> and <u>df</u> will increment the file pointer in the *ser* file by NBL*TD. If TD is not a multiple of 256 (corresponding to 1024 bytes), the buffer space for an fid is extended to become a multiple of 1024 bytes. The pulse program commands decribed above will all work with the corrected buffer size. The effect is that an fid always begins at a 1K byte block (in acquisition memory as well as on disk).

PAPS (phase cycling mode)

Only for AMX/ARX type spectrometers. The values cp, ap, and qp are legal. They define in which way consecutive scans of an fid are handled. cp adds them, ap adds two scans and subtracts the next two, qp uses a more complicated procedure. If the

transmitter pulses in the pulse program do not have phases assigned explicitly, their phases are shifted automatically during the corresponding scan to ensure accumulation of the scans.

WBST (*number of wobble steps*), **WBSW** (*wobble sweep width in Mhz*) Parameters required by the command <u>wobb</u> (probehead tuning). See <u>wobb</u> for more information.

V9 (variation width of random delay in per cent)

In a pulse program, any of the delay commands $\underline{d0}$ - $\underline{d31}$ may get appended the option :*r*, e.g. $\underline{d1:r}$. The result will be that each time this command is encountered, the delay will get a different, randomized value . The maximum variation with respect to the original value may be specified in V9.

AUNM (acquisition AU program)

The AU program specified in this parameter is executed by the command <u>xaua</u>. <u>xaua</u> is most often used in an AU program to start another AU program, which may depend on the current acquisition parameters.

POWMOD (power mode)

Possible settings: *low*, *high*, and *linear*. Power mode selection for spectrometers equipped with a high power accessory.

HPPRGN (gain for HPPR preamplifier)

Possible settings: *normal* and *plus*. Amplifier selection for spectrometers equipped with the respective accessory.

PRGAIN (high power preamplifier gain)

Possible settings: *low* and *high*. Gain selection for spectrometers equipped with a high power accessory.

IN0-IN31 (*increment values for the delays D0-D31 in seconds*) See earlier description of D0-D31, and of IN0 and IN10.

INP0-INP31 (*increment values for the pulses P0-P31 in microseconds*) See earlier description of P0-P31.

L0-L31 (loop counter parameters)

The pulse program commands <u>lo to x times 10</u>, ..., <u>lo to x times 131</u> execute a loop to label *x* L0, ..., L31 times, respectively. Outside the loop, you may use the pulse program commands <u>iu0-iu31</u> and <u>du0-du31</u>, which increase or decrease the values given by L0-L31 by 1. Assume the command <u>iu0</u> is the next statement after <u>lo to x</u>

<u>times 10</u> and both, <u>lo to x times 10</u> and <u>iu0</u> are contained in larger loop. Then, the first time the <u>lo to x times 10</u> command is executed it will loop L0 times, the second time L0+1 times, etc.

SEOUT (SE 451 receiver unit output to be used)

Possible settings: *HR* and *BB*. The *BB* channel is used for special experiments only.

S0-S7 (ecoupler power 1 in db)

Only for AMX/ARX type spectrometers. The pulse program commands <u>s0-s7</u> set the Ecoupler power values according to the parameters S0-S7. We recommend, however, to use <u><u>hl1-hl4</u> instead, which provide a faster power switching using the 4 fast ecoupler registers (parameters HL1-HL4).</u>

PH_ref (reference phase in degrees)

The pulse program command <u>go=n</u> may have a phase program for the receiver phase appended, e.g. <u>go=n ph1:r ph2</u>. The option :r indicates that the phase program ph1 should control the reference phase. The value PH_ref is added to all phases of such a phase program. For Avance systems the :r option may also be appended to the <u>adc</u> pulse program command, for AMX/ARX systems to the <u>rcyc=n</u> command. The example also has a second receiver phase program *ph2*. Since it has not a :r appended, it may only contain phases corresponding to 0, 90, 180, or 270 degrees. Such phases are realized in such a way that the fids A and B of the 2 quadrature channels are added to (or subtracted from) the previous scans as follows:

0: +A, +B. 90: -B +A. 180: -A, -B. 270: +B -A.

The total receiver phase is the sum of the phases of the two phase programs specified.

PHCOR0-PHCOR31 (correction angles for phase programs in degrees)

In pulse programs, any phase program may have appended the option :r, e.g. <u>p1:ph2:r</u>. The value PHCOR2 is added to any phase of phase program ph2 before execution. See PH_ref for the special case of receiver phases.

ROUTWD1, ROUTWD2 (router control words)

For AMX type spectrometers only. Do *not* modify these parameters unless you have a detailed hardware knowledge of your spectrometer. The control words are set automatically at the time you define the nuclei (NUCLEI button in <u>eda</u>, or with <u>edsp</u>).

SP07 (shaped pulse parameter table)

For Avance type spectrometers only. The pulse program command <u>p1:sp2:f1</u> would execute a shaped pulse of length P1 on channel f1. The shaped pulse parameters are taken from entry 2 of this table, indicated by the :sp2 option. The table has totally 16 entries (with index 0-15), so that the options :sp0, ..., :sp15 may be used with a pulse command. Each table entry has 3 parameters assigned: a power level, a frequency offset, and a file name. A file with the specified name must exist in the directory

XWINNMRHOME/exp/stan/nmr/lists/wave/.

It must contain the shape of the pulse, generated for example with Bruker's *shape* or *xshape* programs. Shaped pulse parameters may also be set from the keyboard. Example: The commands <u>spnam2</u>, <u>spoffs2</u>, and <u>sp2</u> allow you to set the file name, frequency offset, and power level for entry 2 of the table.

TP07, DP07, DBP07 (shaped pulse parameter tables)

For AMX/ARX type spectrometers only. Each table has 8 entries. The :*spx:fy* type pulse options for Avance systems (see SP07 parameter table above) must be replaced by the options :*tp0*, ..., *tp7* to execute a shaped pulse on the transmitter, :*dp0*, ..., *dp7* to execute a shaped pulse on the decoupler, and :*dbp0*, ..., *dbp7* to execute a shaped pulse on decoupler B.

GP031 (gradient parameter table)

For Avance type spectrometers only. The pulse program commands <u>gron0-gron31</u> are used to switch on static gradients, <u>groff</u> to turn them off. Example to enable static gradients for the duration D1+D2:

d1 gron2 d2 d3 groff

The gradient parameters for this example are taken from entry 2 of the gradient parameter table. Other entries are accessed using the respective gron command. The table has totally 32 entries (with index 0-31). Each table entry has 4 parameters assigned: a gradient strength for each dimension (in the range -100%...100%), and a file name. The file name does not play a role for static gradients. However, you may also generate shaped gradient pulses by means of the pulse options :*gp0*, ..., :*gp31*. For example, the pulse program command <u>p1:gp5</u> would generate a gradient pulse of length P1. Gradient strength and pulse shape would be taken from entry 5 of the gradient parameter table. The specified file must be located in the directory

XWINNMRHOME/exp/stan/nmr/lists/gp/.

It must contain the shape of the pulse, generated for example with Bruker's *shape* or *xshape* programs. Gradient parameters may also be set from the keyboard. Example: The commands <u>gpx2</u>, <u>gpy2</u>, <u>gpz2</u>, and <u>gpnam2</u> allow you to set the gradient strength in the 3 dimensions and the file name for entry 2 of the table.

F1LIST-F3LIST (frequency list files)

For AMX/ARX type spectrometers only. The pulse program commands $\underline{o1}$ - $\underline{o3}$ set the frequency of the transmitter, decoupler, or second decoupler, respectively, from the current position of the respective frequency list file. The next time such a command is encountered, the next frequency is taken from the list. Example: $\underline{d1 o1}$ (set transmitter frequency during a delay of length D1 from the list defined by F1LIST). See command <u>edlist</u> (in *The File Menu*) how to set up frequency lists.

FQ1LIST-FQ8LIST (frequency list files)

For Avance type spectrometers only. The pulse program commands $\underline{fq1}$ - $\underline{fq8}$ set the frequency of a spectrometer channel from the current position of the respective frequency list file. You must append one of the options : $f1, \ldots, :f8$ to these commands to define the channel. For example, the command $\underline{d1}$ <u>fq2:f1</u> takes the frequency value from the current entry of FQ2LIST, and loads it to channel f1. Execution is performed during the delay D1. The next time the command is encountered, the next frequency is taken from the list. See command <u>edlist</u> (in *The File Menu*) how to set up frequency lists,.

DSLIST (data set list file)

The pulse program command $\underline{\mathrm{wr \#0}}$ stores acquisition data in the files *fid* or *ser* of the current data set. In contrast, the pulse program command $\underline{\mathrm{wr \#1}}$ stores these files in the first data set specified in the list file, $\underline{\mathrm{wr \#2}}$ in the second data set, etc. You may also access the data set list via a *pointer* (pointing to the first data set in the list when the pulse program is started). The command $\underline{\mathrm{wr \##}}$ stores the acquisition data in the data set corresponding to the current pointer position. The commands <u>ifp</u> and <u>dfp</u> increment or decrement the pointer by 1, <u>rfp</u> resets it to the beginning of the list. See command <u>edlist</u> (in *The File Menu*) how to set up data set lists.

VCLIST (variable loop counter list file)

The pulse program command lo to x times c executes a loop to the label x in the pulse program. The number of times the loop is performed is taken from the current position in the loop counter list file. In order to proceed to the next list posi-

tion, the command <u>ivc</u> (for Avance, <u>vc</u> for AMX/ARX) must be used (e.g. <u>d1 ivc</u>). See command <u>edlist</u> (in *The File Menu*) how to set up loop counter lists.

VDLIST (variable delay list file)

The pulse program command \underline{vd} executes a delay whose length is taken from the current position in the variable delay list file. In order to proceed to the next list position, the command \underline{ivd} (for Avance, \underline{vd} for AMX/ARX) must be used (e.g. $\underline{d1}$ \underline{ivd} , or \underline{vd} ivd). See command <u>edlist</u> (in *The File Menu*) how to set up variable delay lists.

VPLIST (variable pulse list file)

The pulse program command \underline{vp} executes a pulse whose length is taken from the current position in the variable pulse list file. In order to proceed to the next list position, the command \underline{ivp} (for Avance, \underline{ip} for AMX/ARX) must be used (e.g. $\underline{d1}$ \underline{ivp}). You may append the usual pulse and phase program options to \underline{vp} , e.g. $(\underline{vp}$ <u>ph1):f2</u>. See command <u>edlist</u> (in *The File Menu*) how to set up variable pulse lists.

VTLIST (variable temperature list file)

The AU program command <u>rvtlist</u> makes the specified list file available to the AU program. The AU program command <u>vt</u> sets the Eurotherm temperature according to the current position in the list. <u>teready(<seconds>, <accuracy>)</u> waits for the temperature to settle: The AU program continues if the temperature is reached with the specified accuracy (0.0-1.0), or the specified time has elapsed. In order to proceed to the next list position, the command <u>ivtlist</u> must be used (<u>dvtlist</u> goes to the previous position). If you replace <u>rvtlist</u> by <u>gvtlist</u>, the temperature list file is not taken from the VTLIST parameter, but the AU program will prompt for it. See command <u>edlist</u> (in *The File Menu*) how to set up variable temperature lists. The command <u>edte</u> is provided to control the Eurotherm parameter settings

FCUCHAN (*FCU for channel fx*)

Avance only. This is an array describing the usage of the 8 possible FCUs for the 8 logical channels (:f1 - :f8). FCUCHAN[1] = 2 means that FCU no. 2 is used for channel f1. FCUCHAN[x] = 0 if the channel is not used. FCUCHAN[0] is always unused. This routing is done completely by software.

RSEL (transmitter select)

Avance only. This is an array describing the connections of the 8 possible FCUs to the amplifiers. The routing is done by the digital routers. RSEL[x] = 0 means *no connection*. RSEL[1] = 2 means that FCU no. 1 is connected to amplifier no. 2 (1H 100 W standard). RSEL[0] is unused. If there are more than 3 FCUs, a second

router must be installed and the output 1 of the second router must be connected to input 3 of the first one. Router 2 is used for FCUs no. 3-5. If there are more than 5 FCUs a third router must be installed and output 1 of the third router must be connected to input 3 of the second. In this case RSEL[5] = 1 means that input 1 of the third router goes to output 1 of the first router (via output 1 of 3, input 3 of 2, output 1 of 2 and input 3 of 1). The graphical interface in edsp and edasp shows these connections in a simplified manner.

SWIBOX (channel switching)

Avance only. Some of the amplifiers contain fast switches which direct the output to the 1H module, or to the X-BB module or to the 19F-module of the preamplifier. These switches can be set by software with the parameter SWIBOX[1..15]. If no such switch is installed, SWIBOX[n] = n. XWIN-NMR cannot guarantee, however, the correct connections between these outputs and the preamplifier modules.

PRECHAN (preamplifier channel)

Avance only. The connection to the preamplifier modules is described by PRE-CHAN[1..15]. If there is no preamplifier involved, PRECHAN[n] = 5. The order of the preamplifier modules is determined by their soft addresses:

0 = 2H 1 = X-BB 2 = 1H 3 = UserBox (QNP) 4 = 19F

OBSCHAN (observation channel)

Avance only. The parameter array OBSCHAN is used to define the channels which are used for observation. By default always channel 1 is used as the default observation channel (OBSCHAN[1] = 1). With the HPPR it is possible to switch between up to 3 channels for observation during the experiment. To define the second and third channel which can be used for interleaved acquisition, set OBSCHAN[n] to 2 or 3 if the n-th channel should be used as the 2nd or 3rd observe channel in an interleaved acquisition experiment. The value of OBSCHAN[1] is ignored by the program and the program behaves always as if OBSCHAN[1] = 1.

With pulses from NMR control word 2, bit 7, the HPPR can be switched from one module to the next in the sequence defined by OBSCHAN. The pulses must be 1 μ sec to switch to the next module and 4 μ sec to reset to the first one:

The modules which are chosen on the HPPR are defined by the complete routing in the <u>edasp</u> or <u>edsp</u> window.

OVERFLW (overflow handling)

Avance type systems only. In certain applications small signals must be detected in the presence of very strong ones, and many accumulating scans are required. This can lead to an overflow of the 32 bit maximum dynamic range of an fid, resulting in a distorted spectrum after Fourier transformation. XWIN-NMR does not normally check for such an event since such situations only occur in special experiments, and there are possiblities to work around those. However, an acquisition parameter OVERFLW is available that allows for user control of overflow detection. By default, OVERFLW is set to *ignore*, instructing the acquisition software not to care about overflows. Alternatively, you may set OVERFLW to *check*, and the acquisition software will check before each scan whether an overflow would most probably occur during the next transient. In this case the experiment is halted. This allows you to specify a large NS whithout the inherent danger of overflowing. When the experiment is done, you can examine the actual number of scans performed using the command <u>dpa</u> or <u>2s ns</u>.

Please note:

Overflow checking is a time consuming procedure carried out by the RCU. If enabled, the minimum recycle time between scans is noticeably enlarged, which may cause certain experiments to fail.

DQDMODE

Avance type systems only. The acquisition parameter DQDMODE defines the frequency shift applied in Digital Quadrature Detection mode as positive (*add*) or negative (*subtract*).

1.5.3 Acquisition parameter setup with ased or as.

<u>ased</u> opens a dialog window of acquisition parameters similar to <u>eda</u>. While the parameters appearing in the <u>eda</u> window are defined by the corresponding format file, <u>ased</u> calls the pulse program compiler for the current pulse program PUL-PROG and lets it determine which acquisition parameters are required. Only those are displayed.

A description of the parameters is contained at the end of the pulse program delivered with XWIN-NMR in the following way:

;p1 : f1 channel - 90 degree pulse

Such a description is taken over into the text field of the editor <u>ased</u> to facilitate the parameter editing. For this purpose the description must follow the above convention. The following parameters can be described in this way:

AVANCE-Series:

• P0-31, D0-31, PL0-31, IN0-31, INP0-31, L0-31, SP0-15, PCPD1-8, CNST0-31, GPX0-31, GPY0-31, GPZ0-31, NBL.

AMX/ARX/ASX:

• P0-31, D0-31, IN0-31, INP0-31, L0-31, HL1-4, S0-7, FS0-7, TP0-7, DP0-7, DBP0-7, TL0-15, SP0-15, NBL, CNST0-31.

<u>as</u> works similarly, but the parameters required by the pulse program are requested in form of a dialog.

1.5.4 Tuning the probehead [wobb]

1.5.4.1 Introduction

During the acquisition electric energy in form of pulses is transferred from the transmitter (source) to the probe (drain). As these pulses are in the radio frequency range it is vital that the output impedance of the transmitter is equal to the input impedance of the probe. If the impedances don't match a part of (or in worst case all) the energy is reflected back to the transmitter. The results are

- a bad performance like long 90 degree pulses and bad signal-to-noise and
- a high risk of destroying the transmitter by reflecting too much power into its output stage¹.

All probes used in spectrometers behave like a resonance circuit consisting of the coil and one or more capacitors. The impedance of this circuit is frequency dependent and the nominal impedance is given for the resonance frequency only. Therefore it is necessary

- to tune the probe with the tune knob to set the resonance frequency of the curcuit to the same value as the frequency of the transmitter pulses and
- to match the probe with the match knob to set the impedance of the probe to the same value as the output impedance of the transmitter, which is generally 50 Ω .

Additionally, only special cables with 50 Ω impedance (as provided with the spectrometer) may be used for all radio frequency cables.

1.5.4.2 How to tune and match a probe

The <u>wobb</u> command allows you to tune and match a probe in an easy way even when the coil is heavily mistuned and mismatched. Firstly, we should set the parameters which define the wobble curve:

• SFO1..8: nuclei frequencies

These parameters are setup within <u>edsp</u>, <u>edasp</u> or <u>eda</u>. All nuclei specified in the experiment whose rf signals pass through a HPPR¹ preamplifier module can be wobbled. If more than one nucleus is defined, <u>wobb</u> starts with the nucleus having the lowest frequency. The selected frequency (SFO*x*) specifies the center of the wobble window.

- WBSW: wobble sweep width in MHz This parameter sets the frequency range of the wobble sweep, which is from SFOx-WBSW/2 to SFOx+WBSW/2. The lower limit for WBSW is at 0.001 MHz, the default value being 4 MHz. It can be set within <u>eda</u> or by typing <u>WBSW</u> via the keyboard.
- WBST: number of wobble steps This parameter determines the number of single frequencies measured by wobb

^{1.} The <u>acbdisp</u> command (on Avance) or the router display (on AMX/ASX) allow you to watch the reflected power as well as the forward power (the part of the energy which really is transferred to the probe).

^{1.} High Performance PReamplifier with up to five preamplifier modules. The <u>wobb</u> command is only available on spectrometers equipped with a HPPR preamplifier.

and can be set within <u>eda</u> or by typing <u>WBST</u> via the keyboard. The allowed range is between 256 and 4096, the default being 256. However, as the precision of the wobble curve displayed on screen is limited by the number of pixels it is of no use to set WBST higher than the horizontal pixel number. On the other hand the display refresh rate decreases with higher WBST.

Now enter the acquisition window by clicking *Acquire -> Observe fid window* and start the wobble routine by clicking on *Acquire -> Acquisition parameter setup -> Tune probehead*. Alternatively enter acqu and then wobb via the keyboard. The acquisition is started and after a few seconds the wobble curve (see Figure 1.7) is displayed and refreshed continuously. A vertical line is drawn at the center frequency (SFOx) to provide optical information on the frequency which is to be tuned. The horizontal axis of the coordinate system is scaled in MHz and labeled accordingly. Useful information like nucleus, tuning frequency, frequency of the minimum of the wobble curve and wobble width, is displayed in the information window. Simultaneously the LED display on the preamplifier is set and refreshed accordingly (See "Preamplifier operating panel" on page 64.).

The wobble curve shows a dip downwards which changes while you turn the tune or match knobs of the probe.

- 1. Turn the tune knob so that the dip moves towards the center of the screen. Keep on turning until the dip is exactly in the center across the vertical line.
- **2.** Turn the match knob so that the dip becomes deeper. Keep on turning until the base of the dip is at a minimum. This occurs at the zero level line for most probes.
- **3.** On most probes the matching influences the tuning and vice versa, so repeat steps 1 and 2 until the dip is exactly in the center of the screen and its base at minimum level. Then the probe is tuned and matched.

Now you are finished with tuning and matching and may stop <u>wobb</u> by activating the *stop* button or by entering <u>stop</u> via the keyboard.

Change wobble sweep width and center frequency during <u>wobb</u>

Activate the button *wobb-SW* and confirm the question Change the nucleus? with the ENTER key. Then enter the desired center frequency and wobble sweep width (here the minimum value for WBSW is also 0.001 MHz).



Figure 1.7 Examples of wobble curves with different matching and tuning

After that, the wobble routine stops the acquisition and restarts with the new parameter setting.

Select another nucleus

If more than one nucleus is defined which may be wobbled, you can switch to another nucleus either

- by pressing the CHANNEL SELECT button on top of the preamplifier or
- by activating the button *wobb-SW* and typing y to the question Change the nucleus? Alternatively you can enter <u>wbchan</u> via the keyboard.

wobb then selects the nucleus with the next higher frequency and restarts the acquisition. After a few seconds the wobble curve is displayed again.

1.5.4.3 Preamplifier operating panel

Depending on the location of the magnet you may not be able to see the display screen when tuning the probe. Therefore the operating panel of the preamplifier has been equipped with an LED display which, in the case of wobbling, shows the interpreted wobble curve. The vertical component displays the deviation in the direction matching, the horizontal component displays the deviation in the direction tuning. Matching represents a measure for the impedance value at the currently tuned frequency, one lit matching LED meaning zero and all LEDs lit representing the curve maximum. The tuning display corresponds to the deviation of the frequency of the curve minimum from the center frequency. If the difference between the two frequencies is zero the two middle LEDs are lit, if the difference is plus or minus half the wobble sweep width the entire left side or the entire right side of the LEDs is lit, respectively.

The goal of the tuning process is therefore to keep the number of matching LEDs minimal (optimal: 1) and to bring the lit tuning LEDs into the middle (optimal: the middle two).

For accurate tuning, the deviations are weighted with a square root function. This means that the sensitivity of the LEDs around zero becomes greater. However, in order to get an idea for the direction of tuning for greatly mistuned impedances, tendency LEDs have also been integrated into the operating panel. If the matching tendency LED is lit upwards the deviation from the optimal matching setting is becoming greater, and conversely. If the tuning tendency LED is lit leftwards, the curve minimum moves to the left (towards smaller frequencies), and vice-versa.

A tendency of direction X in the wobble curve causes LED X to be lit.

1.5.4.4 How does wobble work?

Since the output impedance of the transmitter and the impedance of the cables is 50 Ω , the probe should, if possible, also have this impedance in order to eliminate reflections. The wobb routine therefore compares the probe impedance with a 50 Ω reference resistor which is built into each preamplifier module. The preamplifier offers the possibility of internally switching between this 50 Ω resistor and the probe.

The heart of the wobble software is an endless loop which steps through the frequencies in the wobble window. The command <u>wobb</u> is very similar to the command gs (go setup) and is therefore used similarly in many cases.

After startup the preparations needed for the acquisition, such as reading and defining acquisition parameters¹, setting up a frequency list, compiling a pulse program², etc., are made.

Now the 50 Ω reference is measured first. The receiver gain is adjusted so that the highest point of the aquired data falls within the range of 18% to 45% of the digitizer resolution. If a complete measurement with the reference resistor has been acquired, the preamplifier switches to the probe. After that, measurement of the impedance of the probe is continuously repeated. After a completed scan the complex magnitude of the difference between probe impedance and 50 Ω reference is calculated, scaled for the optimal display region and then displayed. The LED display of the preamplifier is set accordingly.

1.5.4.5 Trouble shooting guide

- a) The wobble curve shows no dip. There are two possible reasons:
 - •The probe is heavily mistuned and mismatched. Just turn the tune knob and check the wobble curve for changes. This works best when watching the screen instead of the preamplifier display. Often a tiny change somewhere in the curve indicates the dip.
 - •The dip is outside the wobble sweep range. This happens especially on broadband probes which previously were tuned for another X-nucleus. Increase WBSW (e.g. 10 to 20 MHz) and try again.
- b) The wobble curve moves periodically although the tune and match knobs are not touched.

The sample is spinning: either ignore the movement or switch off the spinning.

^{1. &}lt;u>wobb</u> reads most parameters from the current data set with the exception of few parameters needed by the wobble pulse program which are read from either *acqu_go4* (Avance) or *acqu_go2* (AMX,ARX, ASX) located in *XWINNMRHOME/prog/wobble*.

^{2. &}lt;u>wobb</u> uses the pulse program *XWINNMRHOME/prog/wobble/pulsprog_X*, which is linked to any of the pulse programs *pp_amx_X*, *pp_arx_X*, *pp_dmx_X* or *pp_drx_X* by <u>cf</u> depending on the type of instrument.

c) After starting acquisition one of these error messages appears:

```
wobble signal too weak with RG=...
or
Error changing RG ... : maximum RG reached
```

The rf signal received from the preamplifier is too low to be used for the wobble curve. In most cases the rf signal path has been interrupted somewhere. Checking the rf signal path usually helps.

d) After starting acquisition the error message appears:

```
wobble signal too strong despite of smallest receiver gain
```

The rf signal received from the preamplifier is too high to be used for the wobble curve. Unplug the TUNE-IN cable at the preamplifier, put in a 10dB attenuator and reconnect the cable. If this does not fix the problem a hard-ware fault may be the reason:

- The DC voltage of one or both of the receiver outputs is so high that the acquired data always fall outside the valid range of the automatic receiver gain adjustment. Get Bruker service to check and adjust the DC voltages of both receiver outputs.
- •There are unwanted signals at the digitizer input, e.g. spikes or oscillating signals from the stages before. Get Bruker service to check the receiver outputs or the digitizer inputs and fix the problem.
- e) Errors in the preamplifier-controller:

If an error occurs in the preamplifier-controller¹ an error message appears on screen showing the type of error. After confirmation the error is cleared in the preamplifier-controller and <u>wobb</u> is cancelled. Depending on the type of error <u>wobb</u> can be restarted without errors. However, in most cases a thorough investigation and correction of the error is necessary.

^{1.} The preamplifier-controller is a microcontroller located in the top of the HPPR housing which is connected to the CCU via RS-232. It controls the different preamplifier modules and their signal routing.

1.5.5 Interactive adjustment of acquisition parameters [gs]

The command <u>gs</u> executes, like <u>zg</u>, the pulse program defined by the acquisition parameter PULPROG in order to perform a data acquisition. In contrast to <u>zg</u> only the first acquisition loop in the pulse program plays a role, e.g. up to the first <u>go=n</u> or <u>rcyc=n</u> pulse program command. Data is not accumulated by <u>gs</u>, but each new scan replaces the data of the previous one. Phase programs are also not executed: every scan uses the first phase of a referred list.

The purpose of <u>gs</u> is to observe the effect of a changed acquisition parameter on the signal, thereby optimizing this parameter. <u>gs</u> is terminated immediately either by <u>stop</u> or <u>halt</u>, and in contrast to <u>go</u> no data is saved on disc with <u>halt</u>.

1.5.5.1 gs on AVANCE spectrometers

As soon as <u>gs</u> has been activated, the acquisition starts and the fid is displayed in the acquisition window, while the fid integral is displayed in the information window. The iconified *XWinNmr-gs* (gsdisp) appears on the desktop. After klicking on it or activating <u>modify</u> gsdisp appears allowing the manipulation of parameters with sliders (see Figure 1.8). On startup the frequency offset parameter group (O1-8) is active allowing all frequencies used by the current pulse program to be varied. On 13C observe / 1H decoupling experiments for example, O1 and O2 are displayed.

Adjust parameter:

There are several possibilities to adjust a parameter:

- Move the cursor onto the slider, keep the left mouse button depressed and move the mouse along the scale. The slider follows the cursor while changing the parameter value.
- Move the cursor onto the scale and press the left mouse button. The parameter value is incremented or decremented (depending on the position of the cursor in relation to the slider position) and the slider moves closer to the cursor. Keeping the mouse button depressed results in a continuous parameter change. In order to increase the stepsize of the increment/decrement by a factor of 10 activate the *sens* *10 button. The stepsize is decreased accordingly by a factor of 10 when the *sens* /10 button is activated.
- Move the cursor onto the scale and press the middle mouse button. The slider moves to the cursor and the parameter value changes accordingly.





• Enter the parameter via the keyboard. If a parameter is entered that may not be varied within <u>gs</u>, it is saved in its parameter file. If it is an acquisition parameter, it will only become effective after <u>gs</u> has been stopped and restarted.

All changed parameters except RG will become active at the earliest with the next scan while RG itself will become active immediately.

Select new parameter group to modify:

In the left part of the gsdisp-window all selectable parameter groups are listed, so simply select one of these groups by klicking onto it. Klicking on *Stop acquisition* will stop the acquisition.

Note: it is not necessary to click exactly on the button, klicking onto the name of the parameter group works just as well.

Save adjusted parameter:

In order to save the last touched parameter for a later data accumulation with \underline{go} or \underline{zg} , activate the *save* button in the gsdisp-window. A status message gives informa-

tion about the parameter and its saved value.

Save all adjusted parameters:

In order to save all ,,touched" parameters for a later data accumulation with <u>go</u> or <u>zg</u>, activate the *save all* button in the gsdisp-window. A status message gives information about which parameters have been saved.

Reset adjusted parameter:

In order to reset the last touched parameter to its last saved value, activate the *reset* button. A status message gives information about the parameter and its value.

Reset all parameters:

In order to reset all parameters to their last saved value, activate the *reset all* button. A status message gives information about which parameters have been reset.

Caveats:

• Changing RG with the slider may cause problems because the slider changes RG linearly although RG has an exponential effect. With small RG values especially the slider is too sensitive.

Use any of the other methods mentioned above to change RG by a small value.

1.5.5.2 gs on AMX/ARX/ASX spectrometers

As soon as <u>gs</u> has been activated, the information window displays the parameter O1, its current value along with its range of values, the mouse sensitivity and the fid integral.

Adjust parameter:

Move the cursor onto the *modify* button, keep a mouse button depressed and move the mouse up or down. Fine ajustment may be achieved via the left mouse button, coarse adjustment via the righthand button. The middle button has an intervening sensitivity.

Alternatively all parameters which may be varied within <u>gs</u> (as displayed in the select window) can be entered via the keyboard. If a parameter is entered that may not be varied within <u>gs</u>, it is saved in its parameter file. If it is an acquisition parameter, it will only become effective after <u>gs</u> has been stopped and restarted.

All changed parameters except RG will become active at the earliest with the next scan while RG itself will become active immediately.

Select new parameter to modify:

Activate the *select* button. A dialog window is opened containing all parameters that may be varied within <u>gs</u>. Use the mouse cursor to select the desired one. Certain parameters (pulses, delays, ...) require an index (e.g. 5 to adjust P5). It is asked for by the program. Alternately this parameter may be typed in after clicking the PARAMETER field (e.g. type P5).

Alternatively enter <u>gschan <Parameter></u> via the keyboard to select this new parameter for manipulation by mouse.

Changing the mouse sensitivity:

Below the label *mouse*: are the four buttons *2, /2, *8 and /8. The sensitivity of the mouse for the variation of parameters is increased or decreased by the appropriate factor by activating these buttons.

Save adjusted parameter:

In order to save the current parameter adjusted by mouse for a later data accumulation with g_0 or \underline{zg} , activate the *save* button.

Alternatively enter gsstore via the keyboard.

Reset adjusted parameter:

In order to reset the parameter currently adjusted by mouse to the last saved value, activate the DEF button.

Alternatively enter gsres via the keyboard.

1.5.6 Acquisition parameter setup with set

Please note that you should use ICON-NMR for routine spectroscopy based on standard experiments, and for automation using a sample changer rather than the command <u>set/run/quicknmr</u>, which are historically older and are just maintained for compatibility reasons.

The acquisition command <u>run</u> allows you to start a series of experiments. <u>run</u> is most often used to work with an automatic sample changer, but may also be used if samples are changed manually. Please refer to the description of the <u>run</u> command for details. The command <u>set</u> must be used to set up the experiments to be executed by <u>run</u>. All experiments you define in the <u>set</u> dialog window are stored in a file, from where they are retrieved by <u>run</u> for execution. <u>set</u> requests the name of the file upon start up, as does <u>run</u>. This feature, that you may specify the name of the file yourself, allows you to set up and maintain several such experiment files, and run the experiments at a desired time.
An experiment file is stored in the directory

XWINNMRHOME/prog/curdir/changer/.

The file name extension .0 will be appended for internal reasons.

<u>set</u> comes up with a dialog window containing up to 120 entries for experiments. If a sample changer is to be used, the number of entries is 60 or 120, depending on its capacity (number of *holder* positions). Otherwise, you my configure the number of entries using the <u>cfbacs</u> command.

1.5.6.1 The experiment entry fields

Each entry consists of the following fields:

STAT HOLDER NAME EXPNO SOLVENT EXPERIMENT Priority

STAT

Sample status.

U=unused. This entry is available for a new experiment.

R=ready for acquisition. Parameter set up is complete for this holder. You may still apply changes as long as acquisition is not yet in progess. Click on this field to cancel R mode, and to return to U mode. Everything set up for this holder will be lost.

nR=*ready for multiple experiments*. Like R, but indicates that more than one experiments have been set up for this sample.

F=*finished*. All experiments with this sample are complete, the holedr may be used for a new one. Click on the HOLDER field if you want to re-activate the last parameters used.

HOLDER

You must fill in the entry field for a particular holder in order to define the experiment for the sample in this holder. Click on the holder number to anable this entry for editing. Clicking on the next holder position will terminate the parameter set up for the current one, and switch its status to R or nR. In non-sample changer mode, after the experiments for a holder are terminated by <u>run</u>, the operator is invited to insert the next sample manually.

NAME

Define data set NAME parameter for this experiment (up to 10 characters). See <u>eduser</u> how to set up predefined names. You may use the same NAME for all sam-

ples.

EXPNO

Define data set EXPNO parameter for this experiment, a number with possible values, 10, 20, Used to differentiate data sets with the same NAME parameter. The increment must be 10 to allow for multiple (up to 10) experiments defined on the same holder position.

SOLVENT

Define solvent for this sample. The solvent table was set up with edsolv.

EXPERIMENT

Define experiment for this sample. Bruker standard experiments were installed with the <u>expinstall</u> command. The system administrator may have installed own experiments and copied them to the experiment directory with <u>wpar</u>. When you click on the experiment field, a table of the available experiments is displayed. The first column shows one of the character 0, 1, 2, or, C, referring to experiments requiring 0, 1, or 2 preparation experiments, or indicating *composite* (C) experiments. Experiments of type 1 or 2 require one or two preparation experiments to be performed before the experiment itself can start (e.g. a 1D preparation experiment determining the optimized sweep width for a subsequent 2D experiment).

Priority

Clicking on this field will toggle on or off *urgent* mode. During sample changer operation, such a sample will get priority. Up to 10 samples may be declared urgent, and will be processed in the order they were set up. You must have the permission to use this field (see <u>eduser</u>).

1.5.6.2 The title entry field

Enter the plot title for the experiment you are currently setting up. You may seperate lines in a title text by the 2-character sequence n.

Plots generated by composite experiments may have two titles, one common to all plots, and a particular one for each component experiment. The common title is the one entered in the main <u>set</u> dialog window, the other titles are taken from the title entries of the component experiments. See *EditPar* button.

1.5.6.3 Command buttons at the bottom of the set window

n EXPERIMENTS

Opens a dialog window where you may specify up to 9 additional experiments for this sample. If you use the same NAME parameter for the data sets, the program will automatically assign EXPNO data set parameters, e.g. 10, 11, 12,

Defining a new composite experiment

If the experiment you have selected for a holder number has the name COMPOS-ITE (corresponding to the last entry of the experiment table), the n EXPERI-MENTS button is used to create a new composite experiment. This requires a permission to be granted with eduser. A composite experiment is a sequence of up to 9 standard experiments (i.e. non-composite experiments), which, once defined, will appear under the name it was given in the experiment table. After clicking on the *n* EXPERIMENTS button, a table is show where you can define the sequence of experiments that should constitute the composite experiment. With the define EXP button you start a dialog which allows you to enter the name of the composite experiment, a comment to appear in the experiment table together with the experiment name, and a list of users that should get the permission to execute the composite experiment (separate the list of users by a space character). Before you define the new experiment in this way, you may modify selected parameters of the component experiments via the EditPar button (see below). In order to delete a composite experiment, click on a holder number. Select the desired composite experiment. Click on *n EXPERIMENTS*. Click on *define EXP*. Answer *Return* to all question until the list of allowed users is shown. Remove all users with the backspace key and type Return. Now you will be asked whether the experiment should be deleted.

n SAMPLES

Click on this button and enter a number n. The program will automatically set up the next n free holder entries with the same parameters. The data set NAME parameters are assigned in the following way. Assume the NAME parameter of the current holder is *June15*. The next holders will get the names *June15.1*, *June15.2*, ... assigned. If NAME of the current holder is *June15.1*, the next holders will get the names *June15.2*, *June15.2*, ... assigned. You may further modify the parameters of a particular holder by clicking on its number.

Change User

The data set parameters for a XWIN-NMR data set are NAME, EXPNO, PROCNO, DU, and USER. NAME and EXPNO may be specified in the corrsponding holder

entry fields, and PROCNO is usually 1. USER is initilized with the current login name. You may, however, change user by clicking on this button. You will get a table of users displayed set up with <u>eduser</u>. These are the users owning permission files. Select the desired table entry enter the requested password. From now on, the entries of this user's permission file (experiments, data set names, various permissions) are valid, and the USER parameter for the data sets will be set to the login name of this user.

Holder

Click in this field and enter a number. The set window will scroll to the corresponding holder.

EditPar

Clicking on this button will open a list box containing parameter editing commands (acquisition parameters <u>eda</u>, processing parameters <u>edp</u>, plot parameters <u>edg</u>, output device parameters <u>edo</u>, information file <u>edinfo</u>). Their execution requires a permission, to be set with <u>eduser</u>. Acquisition parameters are normally defined by the selected experiment. At the time the experiment is started, the solvent and probehead dependent parameters are inserted (see <u>prosol</u>, <u>solvloop</u>). Finally, the acquisition parameters you have defined with <u>eda</u> are applied, and therefore obtain the highest priority.

If the experiment you have selected is a composite experiment (indicated by the letter C in the first column of the experiment table), and you want to modify parameters of its component experiments, you must click on the *n* EXPERIMENTS button, select the component experiment by clicking on its holder number, and then invoke *EditPar*. Leave the dialog window with the component experiment via the *Return* button.

Quit

Terminate <u>set</u> (requires a permission, to be set with <u>eduser</u>). You may terminate set at any time. Everything you have set up will be preserved in the set up file. If you re-enter <u>set</u> at a later time by specifying the same set up file name, the dialog box will be filled in with the correct information. You may also leave <u>set</u> while <u>run</u> is in progress.

1.5.7 Acquisition parameter setup with extset

The command <u>extset</u> was designed to set up experiments to be executed with the command <u>run</u>, like the command <u>set</u> described in the previous section. While <u>set</u>

provides a dialog box to be filled in by the user, <u>extset</u> looks for ASCII type text files containing the required information, and converts the contents of such files to a form suitable for <u>run</u>.

Laboratories often employ a centralized sample management, which requires that the experiments are not defined locally on the NMR spectrometer with <u>set</u>, but rather on PCs or a central laboratory computer, from where they must be transferred into the spectrometer via the network or via magnetic storage media.

XWIN-NMR offers the following means for accomplishing such tasks:

- The experiments for each sample may be defined in a text file in ASCII format, e.g. on a remote computer. The structure of such a file is described below. The file name must be of the form Name.NNN. NNN is a sequence of 3 digits, e.g. 001, and with Name may consist of up to 8 characters.
- If <u>run</u> should execute such an experiment, the file must be copied into the directory *XWINNMRHOME/prog/tmp/* of the spectrometer computer (e.g. using <u>rcp</u> or <u>ftp</u> in an Ethernet network).
- Start <u>extset</u>. <u>run</u> may also be active, but it does not have to be. The command <u>extset</u> will remain active in the background as long as XWIN-NMR is running, or until it is called again, which will terminate it. <u>extset</u> periodically (every 60 seconds) checks the directory *XWINNMRHOME/prog/tmp/* for new ASCII experiment files. Any new file will automatically be inserted into the desired setup file. The result is the same as if the user had entered the information directly on the spectrometer with the command <u>set</u>. Once entered into the setup file, the ASCII file is deleted from the temporary directory *XWINNMRHOME/prog/tmp/*. It is therefore the user's responsibility to keep a backup copy of this file, if needed. If several ASCII experiment files are found by <u>extset</u>, they are processed in the order of their file name extensions NNN.

When <u>extset</u> is called, a dialog window appears first with the question *Allow permanent access to setup*?. If the OK field is clicked <u>extset</u> will become permanently active in the background, and will enter all arriving ASCII experiment files into the desired setup file. If CANCEL is selected, a single ASCII experiment file may be specified for insertion into the setup file, and <u>extset</u> will terminate.

extset may be entered on the keyboard with arguments:

extset <path of an ASCII file>

Example:

extset /usr/people/guest/expdat.234

In this example <u>extset</u> will only insert the ASCII experiment file *expdat.234* into the desired setup file, and performs no other task. In particular, <u>extset</u> will not be active in the background afterwards.

Figure 1.9 illustrates the major properties of ASCII experiment files. Lines start-

SETUPNAME	set20			
USER	guest			
#				
SAMPLES				
#				
HOLDER	19			
NAME	Feb25			
EXPNO	50			
SOLVENT	CDC13			
EXPTITLE	created	by	setd	
#				
HOLDER	7			
SOVENT	Aceton			
EXPERIMENT	double			
END				

Figure 1.9 Example of an ASCII experiment file

ing with the character # are comment lines, which are ignored. The file consists of a series of keywords (one per line) combined with information.

SETUPNAME

Specifies the setup file into which <u>extset</u> should enter the following experiments. This is the file from which <u>run</u> extracts the experiments to be executed.

USER

Defines the USER parameter of the data set specification (a data set is defined by NAME, EXPNO, PROCNO, USER, DU).

SAMPLES

Between this keyword and the keyword END experiment names for the holder positions may be listed. The holder positions must be empty at the time <u>extset</u> tries to enter the experiments into the setup file. Otherwise an error situation occurs. <u>extset</u> logs errors in the protocol file generated by <u>run</u>.

An experiment is defined by the keywords

HOLDER, NAME, EXPNO, SOLVENT, EXPERIMENT, TITLE.

They correspond exactly to the fields of the <u>set</u> dialog window. NAME, EXPNO, and TITLE are optional, i.e. they may be omitted. In that case XWIN-NMR uses standard default values: EXPNO=10; no title; NAME is formed from the name of the experiment file, the extension being replaced by the HOLDER number.

Only experiments of type 0 (no preparation experiments required) or C (composite experiments) are allowed.

Table 1.17 summarizes all legal keywords.

#	comment line
SETUPNAME	name of destination setup file
USER	data set USER parameter
SAMPLES	start of holder definitions section
HOLDER	holder number
NAME	data set NAME parameter (optional)
EXPNO	data set EXPNO parameter (optional)
SOLVENT	solvent
EXPERIMENT	experiment name (of type 0 or C)
TITLE	rest of line is the plot title (optional)
END	end of holder definitions section

Table 1.17 extset ASCII file keywords

Features

1. Up to 9 different experiments per holder.

- 2. The first experiment on a holder can be a composite experiment.
- 3. Each experiment can have an individual title. This applies also for the individual experiments of a composite experiment (see the example below). Titles can have more than one line. The line seperator is the n sequence.

If several experiments are defined for one holder, each experiment can have an individual name (NAME) and experiment number (EXPNO). The following example ASCII text file summarizes the above points.

```
Example:
```

```
# in 1. column means : comment
SETUPNAME mike1
USER
          eng
SAMPLES
HOLDER
            8
#
NAME
           June15
EXPNO
           10
SOLVENT
             CDCl3
EXPERIMENT sw_cosy45
TITLE
          This is a title\nwith two lines
TITLE
          This is a second title for the first exp. of the comp.
TITLE
          This is a third title for the second exp. of the comp.
NAME
           June16
EXPERIMENT PROTON
TITLE
          This is a title\nwith two lines
EXPERIMENT C13CPD
TITLE
          This is a title\nwith three lines\nThird line
#
HOLDER
            9
NAME
           June17
SOLVENT
             CDCl3
EXPERIMENT PROTON
TITLE
          This is a default title
EXPERIMENT C13CPD
TITLE
          This is a default title
EXPERIMENT C13DEPT45
TITLE
          This is a default title
#
       Only one END statement per complete file !
```

END

1.6 Interface control commands

1.6.1 Initialize interface

The command \underline{ii} loads acquisition parameters subg as frequencies, power levels, etc. into the spectrometer hardware just as it would do it before a data acquisition, but without starting the acquisition. \underline{ii} can, for example, be used to check whether the software has access to all spectrometer components.

1.6.2 Shim control

1.6.2.1 Writing, reading, deleting, viewing and setting shim values

Shim values are stored in files located in the directory

XWINNMRHOME/exp/stan/nmr/lists/scm .

wsh, wsh <filename> Save the shim values (which are read from the shim unit) in the specified file.

<u>rsh</u>, <u>rsh <filename></u> Read the shim values from the specified file and loads them into the shim unit.

<u>delsh</u>, <u>delsh <filename></u> Delete a shim file.

<u>vish</u>, <u>vish <filename></u> View the shim values stored in the specified file.

<u>lish</u>, <u>lish <filename></u> List the shim values stored in the specified file on the current printer CURPRIN, which can be set using the command <u>edo</u>.

<u>setsh</u>

Display a table of shim gradients. In this table you can set shim values which are immediately loaded into the shim hardware.

1.6.2.2 Automatic shimming

XWIN-NMR provides the command <u>tune</u> to start autoshimming. Its syntax is described in Table 1.18. The autoshim procedure is controlled by the parameters

tune	Select tune file from a dialog box and start autoshim proce- dure using the parameters stored in the tune file.
<u>tune <filename></filename></u>	Start autoshim procedure using the parameters stored in the specified tune file. In AU programs, use the syntax <u>tune("filename")</u> .
<u>tune <gradient></gradient></u>	Start autoshim procedure for the specified gradient only (e.g. <u>tune z1</u>). Uses XWIN-NMR-internal parameters. In AU programs, use the syntax <u>tune("gradientname")</u> .

Table 1.18 tune command

contained in a *tune* file, which can be set up using the command <u>edtune</u>. <u>edtune</u> may also be invoked with a tune file as an argument. A tune file is a text file stored in the directory

XWINNMRHOME/exp/stan/nmr/lists/group.

It contains two example tune files, *example* and *example_bsms*, suitable for BSN18 and BSMS hardware, respectively, and show how to control auto-shimming using special commands and parameters that are discussed now. Note that lines in a tune file starting with a # character are treated as comments.

USE_FIDAREA

The auto-shim procedure will maximize the area under the fid.

USE_LOCKLEVEL

The auto-shim procedure will maximize the lock level.

LOCKDWELL n

Determines a mean value of the lock level by measuring it n times (3 < n < 32, default: n=5). Maximizing the lock level is based on mean values to suppress noise effects. Has no effect if USE_FIDAREA was set.

LOCKPHASE s i

The autoshim module is capable of optimizing the lock phase or shim gradient. The shim optimization commands are GRADIENT and SIMPLEX (see below), while LOCKPHASE adjusts the lock phase. s=maximum step width, i=maximum number of iterations.

MAXLOCK m

Limits the maximum value of the lock level to the specified value. Avoids that the lock signal move out of the display during the shimming procedure. Only required for testing.

ROTATION ON (OFF)

Enables shimming with or without sample rotation.

GRADIENT s i

Example: z1 1000 3. All gradients that should be optimized during auto-shimming must be specified in this way. s=maximum step width, i=maximum number of iterations. Each time a GRADIENT is encountered, this shim will be optimized. This assumes that different gradients can be adjusted independently. See the SIMPLEX command on how to adjust several shims simultaneously.

DELAY n

In order to give the lock enough time to settle when new shim values are loaded, the program will wait n seconds before it starts reading the lock level.

TIMES n, END

This is a loop construct. Everything between TIMES n and END will be executed n times. Nested looping up to a depth of 5 is possible.

RSH, RSH <filename>

The command <u>rsh</u> is executed. If no name is specified, the program will assume the current solvent name as defined by the acquisition parameter SOLVENT.

AUTOSHIM ON (OFF) <gradient list>

When the <u>tune</u> command has terminated, the shim unit itself will continuously auto-shim the gradients specified by this command.

Example for BSMS: AUTOSHIM ON Z1=2 Z2 Example for BSN18: AUTOSHIM ON Z1 Z2

For the BSMS, a step width may be set for each shim individually. For the BSN18, the maximum number of shims is 4.

AUTOSHIM OFF will disable auto-shimming by the shim unit.

The AUTOSHIM command may be written anywhere in a tune file. It will not become effective before <u>tune</u> has terminated.

SIMPLEX <list of gradients>, SET <gradient> w c

The commands SET and SIMPLEX provide a simplex-based optimization procedure for several shims. SET sets the maximum step width w and the convergence limit c for a gradient, e.g. SET Z1 20 3. SIMPLEX starts the simplex optimization for the specified gradients, e.g. SIMPLEX Z1 Z2.

1.6.2.3 Gradient shimming (FOCUS) [gradshim]

Gradient shimming is described in its own manual. Please open the *Help -> Other topics -> Gradient shimming* menu item to view or print it.

1.6.3 BSMS unit

Please refer to the description of the command edlock.

1.6.4 HPCU High power control unit

The HPCU handling (inlcuding commands such as <u>edhpcu</u>) is described in a separate manual, which can be viewed by selecting the

Help --> Other topics --> Solids manual

menu item.

1.6.5 MAS Magic angle spinning unit

The MAS handling (including commands such as \underline{mas}) is described in a separate manual, which can be viewed by selecting the

Help --> Other topics --> SB/MAS manual

menu item.

1.6.6 Amplifier protection limits

The command <u>edacb</u> opens a dialog windows where you can define the amplifier hardware protection limits for pulse power, pulse width, and duty cycle.

1.6.7 Sample rotation, insertion and injection

The command <u>ro</u> allows one to turn sample rotation on or off, and to enter the spinning rate (in Hertz). Furthermore, *variable* rotation may be selected. In this case the rotation frequency is randomly varied by a few Hertz in order to eliminate solvent side bands. In AU programs, the command <u>ro</u> will set the rotation speed according to the acquisition parameter RO, without variation. After turning on rotation, the command will wait for 15 sec and check whether the selected rate was reached. Otherwise an error message is sent.

The command <u>ij</u> inserts the sample into the magnet by switching off the pneumatics, <u>ej</u> ejects it from the magnet.

1.6.8 Set preemphasis [setpre]

1.6.8.1 Introduction

Using this standard XWIN-NMR control feature it is possible to modify or reload the current settings of gradient preemphasis in any plane on the fly. The <u>setpre</u> command is the standard adjustment tool required for the various different types of preemphasis units which Bruker currently offers. Previously installed values may also be loaded and stored as required.

The preemphasis values adjusted and stored with the ParaVision preemphasis adjustment tool and with the <u>setpre</u> command are interchangeable with each other. It is also possible to read the preemphasis parameters stored by preceding version of <u>setpre</u> and convert them to the new native format. However, it is no longer possible to write parameters in the old format.

1.6.8.2 Features

GREAT, BGU-II and Acustar features, such as Gradients generation, Preemphasis bypass, B0 compensation, Amplifier modules and Protection switches, if available for the actual hardware equipment, may be accessed from the setpre control window. For the BGU-II and Acustar units the Gradient coil temperature is also displayed and updated at a user defined interval.

Each gradient may be independently selected by means of the Channel pulldown

📥 setpre: BGL	I-II ON, Gradient coil temperature: 0 degrees
<u>File Edit</u>	Channel Enable Disable (Jew)
Slow Base Preemp. Z 0.2 ms =	Time [ms] 0.00000 Gain [%] 0.00
Mid Base Preemp. Z 0.2 ms =	Time [ms] 0.00100 ▲ Gain [%] 0.888
Fast Base Preemp. Z	Time [ms] 0.00100 ▲ Gain [%] 0.001
Store	Recall Exchange Undo Clear

Figure 1.10 setpre adjusting of pre-emphasis parameters

menu.

It is possible to set some parameters important for microimaging, such as Gradient calibration constant and scaling factor.

The slider box consists of six sliders controlling the time and gain values for the respective phase of the preemphasis action (slow, mid, fast) and three option menus controlling the time base for which the time sliders are responsible; either 0.2 msec, 2 msec, 20 msec or 200 msec. The slider values themselves are shown on the right hand side next to the slider. Should you wish to enter new values directly from the keyboard simply click on the slider value to move text cursor into it and type the new value in. The time values are shown in milliseconds, the time slider ranges and sensitivity depend on the time base. The gain values are shown on a scale from -100% to +100%. The two arrow buttons with the numeric value to the right from the slider value are used to change mouse sensitivity for each slider.

For the GREAT and GREAT-3 units there exist one more slider to adjust Amplifier DC offset between -100% and +100%.

The amplifier matching parameters on the GREAT and GREAT-3, such as control loop resistors, capacitors and impedance may also be adjusted by qualified personnel if enabled via special commands in the setpre menu.

There are five pushbuttons under the slider box: Store, Recall, Exchange, Undo, Clear with the functions similar to a pocket calculator to store currently displayed slider values in memory, recall them back, exchange memory and display, undo all changes, and clear the gain preemphasis values.

At the bottom of the setpre window is the scrollable error message field.

On startup the currently set values contained within the preemphasis unit are read by the program and displayed accordingly. The sliders are then on-line and changes in their values are written directly and instantly to the preemphasis unit. Also, the type of the preemphasis hardware is automatically detected, and the features unavailable for this particular hardware are excluded from the menus and from the main window.

If the <u>setpre</u> command is given while an instance of <u>setpre</u> is already started, then the preceding instance of <u>setpre</u> is deiconified if necessary and its window is raised to the top of the screen.

If the user exits <u>setpre</u> via the *Close* or *Exit* command of the window manager menu, then the module exits leaving the preemphasis unit in its current state.

1.6.8.3 The sliders and option menus

As with all X sliders there are three possible ways of employment. The first and easiest way is by placing the mouse on the slider knob, pressing the middle mouse button and moving the mouse sideways to achieve the desired result. The sliders have the rolling feature. It means that if the slider reaches its rightmost (leftmost) side but not the numerical limit of its value, then the slider rolls over to the other side and continues moving to the given direction. Another possibility is to move the mouse to the position on the slider where you want to be, and then by clicking once on the middle button, the chosen value is set. The third possibility is is by placing the mouse on either side of the slider button and holding down the left mouse button. This has the effect of moving the slider in the desired direction in steps according to the step value shown at the rightmost end of the slider, which can be changed via the two arrow buttons on either side of the step value.

The input field between the slider itself and the step control field is used to enter

directly the desired value of the slider.

The option menu to the left of the slider is used to set the time base for the corresponding time silder. Changing time base changes also the time slider limits and step size, but not the time slider value in milliseconds, except in the case where the time value becomes larger than the maximum limit - then it is truncated.

1.6.8.4 The pushbuttons

There are five pushbuttons under the sliders with functions equivalent to popular pocket calculators.

Store - store current slider and option menu values for the current preemphasis channel in memory for later use. The offset, impedance and loop parameters of GREAT and GREAT-3 are also put in memory.

Recall - recall previously stored values from memory, set the sliders and hardware accordingly.

Exchange - swap the values stored in memory and in the hardware to see which set is better.

Undo - undo slider changes, i.e. reset back the values which were just after switching to this preemphasis channel or after startup of setpre if no channel was switched yet.

Clear - set all three Gain sliders to zero to begin adjusting the channel from a well defined state.

Important notes.

1. Neither Recall nor Exchange is possible if nothing has been Stored after switching to this channel or after startup.

2. Undo is possible only after something has been changed for this channel.

3. Important: One has to distinguish between the *Store* and *Recall* actions, which operate only on the current preemphasis channel and in the computer memory (NOT on the disk), and the *File->Write*... and *File->Read*... which influence all the preemphasis parameters for all the channels and operate on the disk files also clearing the *Store/Recall* and *Undo* states.

4. Very important: Note that these five buttons only influence the current preem-

phasis channel and are immediately forgotten after switching to another one. The *File->Cancel* and *Edit->Clear* commands influence all preemphasis values and channels and are always in effect.

1.6.8.5 The File pulldown

Read default - read preemphasis parameters from the default file and load them into the preemphasis unit. The default file is .../*exp/stan/nmr/parx/preemp/<CUR-HEAD>/default*, where <CURHEAD> is current probehead number defined by the edhead command. If no probehead is defined, the word CURHEAD is used instead. Therefore, each probehead has its own default preemphasis file. The offset, loop and impedance parameters of the GREAT and GREAT-3, as well as the gradient calibration constant, scaling factors and rates to measure temperature and to check status are also contained in these preemphasis parameter files.

Read from... - read preemphasis parameters from the file defined by the user via the standard file selection dialogue and load them into the preemphasis unit. The file *.../exp/stan/nmr/parx/preemp/<CURHEAD>/default* is proposed as the default.

Convert... - read preemphasis parameters in the old format, convert them to the native format and load into the preemphasis unit. The file to convert is selected by the user via the file selection dialogue, default being *.../exp/stan/nmr/lists/preemp/ default*. The offset, loop and impedance parameters of the GREAT and GREAT-3 can not be converted and therefore will not be changed.

Write default - write current preemphasis parameters to the default file *.../exp/stan/ nmr/parx/preemp/<CURHEAD>/default*.

Write to... - write current preemphasis parameters to the user defined file, the default filename is proposed.

Exit - exit <u>setpre</u>. If the parameters have been changed and not yet written to disk, a warning message is shown asking if the user really wants to exit.

Cancel - cancel adjustment. The user is asked via the warning dialogue to confirm the cancel request. If confirmed, then all preemphasis parameters which were in the preemphasis unit before starting setpre will be reloaded back into the unit, and then setpre will exit.

1.6.8.6 The *Edit* pulldown

Grad. calib. const. [Hz/cm] - enter the gradient calibration constant in Hz/cm. Hz/cm are the primary units for this constant, same as in the ParaVision package, and this constant is read and saved along with the preemphasis parameters by the Read... and Write... commands.

Grad. calib. const. [G/mm] - enter the same gradient calibration constant in G/mm. These units are supported for compatibility with microimaging AU programs written by Dr. Dieter Gross. This constant is automatically and instantly written to the file .../conf/instr/gradient_calib, required by these AU programs, and it is also read from there on startup.

Gradient scaling factor - enter the gradient scaling factor for the current gradient direction. Ideally, all gradients should have scaling factors of 1.0. Using these parameters it is possible to correct inaccuracies in the gradients. The scaling factors are read and saved by the Read... and Write... commands in the .../conf/instr/ gradient_calib file.

Rate to measure temperature - enter how often (in seconds) the <u>setpre</u> module should measure the gradient coil temperature and update the <u>setpre</u> window title bar. It cannot be guaranteed that rapid changes in the coil temperature will be shown on the title bar earlier than in that time. However, if a dangerous increase of the temperature appears, the preemphasis unit hardware should automatically switch the gradients off. Entering the value of zero disables temperature measurement. The temperature measuring feature is unavailable for GREAT and GREAT-3.

Rate to check error status - enter how often (in seconds) the <u>setpre</u> module should check whether some error has occured in the preemphasis hardware. It cannot be guaranteed that any error message from the preemphasis unit be shown in the error message field earlier than that time. It follows that, if several error messages come very rapidly after each other, some of them may be lost, while at least one of them (namely the latter one) is guaranteed to be displayed. Entering the value of zero for this parameter disables the error checking feature.

Clear all preemphasis values - set all preemphasis parameters for all channels to zero. This could be used at the very start of the adjustment procedure to put the preemphasis unit in a well defined state. The offset, loop and impedance parameters of the GREAT and GREAT-3 units are not cleared by this command.

1.6.8.7 The Channel pulldown

The *Channel* pulldown is used to switch the <u>setpre</u> module to adjust another preemphasis channel. All preemphasis parameters of the preceding channel are frosen before switching and cannot be recalled from memory or undone (except from exiting the setpre module via Cancel).

On startup of <u>setpre</u> the Z channel is chosen for adjustment as the default.

1.6.8.8 The Enable pulldown

Gradients generation - enable gradient generation. If disabled, no gradient reaches the gradients amplifiers and the probe. This feature only exists for the BGU-II.

Preemphasis bypass - enable preemphasis bypass. If enabled, preemphasis will not be applied to the gradients, and they will be directly connected to the amplifiers as they come from the gradient controller.

B0 compensation - enable B0 compensation. This feature only exists for the BGU-II.

Amplifier modules - enable gradient amplifiers. If disabled, no gradient reaches the probe.

Reset protection - reset protection circuit after dangerous rising of the gradient coil temperature or some other erroneous state of preemphasis.

Impedance&loop editing - enable editing of impedance and loop parameters for GREAT and GREAT-3. Wrong settings of these important parameters can damage the hardware, therefore editing is disabled by default. To enable it the NMR superuser password must be known. If enabled, the corresponding sliders and option menu appear in the <u>setpre</u> window.

Important notes

1. Preemphasis bypass and Amplifier modules can be enabled or disabled only for all channels at once for the BGU-II and Acustar preemphasis units, while they are enabled or disabled for the current channel if you are using the GREAT-3 preemphasis unit.

2. Important: while it is possible to enable or disable some preemphasis feature explicitly, there is no way to detect in which state it currently is. (Check the LEDs

on the preemphasis panel).

3. Very important: it is not possible to load anything into the hardware while the Preemphasis bypass is enabled. The setpre module memorizes all changes on the sliders (if any) and tries to actualize them in hardware. After disabling bypass the changed values will be written into the unit.

1.6.8.9 The Disable pulldown

All the commands of the *Enable* pulldown have their counterparts in the *Disable* pulldown menu. Exceptions: the *Reset* protection command is simply duplicated here to simplify reaching it, and knowing NMR superuser password is unnecessary to disable Impedance&loop editing.

See also the notes for the *Enable* pulldown.

1.6.8.10 The Help pulldown

The *Help* pulldown, as well as the *Help* buttons in various dialogues are reserved for online help feature but have not yet been implemented.

1.7 Starting and stopping data acquisition

This section covers the following commands:

- <u>zg</u>, <u>go</u> Start 1 experiment based on parameter set up with eda, ased, as.
- 4. <u>run</u> Start a series of experiments based on parameter set up with <u>set</u>, <u>extset</u>, or by means of bar code labels. Must be used for automatic sample changer operation (Please be aware that you should use ICON-NMR for routine spectroscopy based on standard experiments, and for automation using a sample changer. ICON-NMR is described in its own manual. <u>set</u>, <u>run</u> and <u>quicknmr</u> serve the same purpose, but are historically older. They are just maintained for compatibility reasons).
- **<u>quicknmr</u>** Easy, routine execution of experiments.
- Displaying the lock and fid window

1.7.1 The commands zg, go, tr, halt, stop, suspend and resume

Both the \underline{zg} and the \underline{go} command compile the current pulse program PULPROG and the acquisition parameters set up with <u>eda</u> (or <u>ased</u>, <u>as</u>, <u>gs</u>). The compiled output is loaded into the acquisition processor, and the pulse program is started. The acquired data are written to disk according to the <u>wr</u> statements in the pulse program. The most frequently used <u>wr</u> statement is <u>wr #0</u>. Each time it is encountered in the pulse program, the acquisition data are stored on disk in the file *fid* or *ser* (depending on the type of pulse program) of the data set from where <u>zg</u> or <u>go</u> was started. Acquisition runs in background, and you may switch to another data set and process or plot it while the experiment is in progress. Alternatively, you may enable the *Observe fid* window, and look at the fid or (for 1D experiments) at the transformed fid in real time. Acquisition may also be started from the *Observe fid* window. In 1D pulse programs, the <u>wr #0</u> statement is often placed at the end, after NS scans are complete. Using the command <u>tr</u> (*transfer*) you may force the software to write the accumulated data on disk for processing or plotting before acquisition terminates.

And this is the difference between \underline{zg} and \underline{go} : If your current data set already contains acquisition data (a file *fid* or *ser*) from a previous acquisition, \underline{zg} will overwrite them, and they are lost. This is done silently if the system variable *ZGsafety* is set to *no*. Otherwise a warning is printed, and you may prevent acquisition start. Enter the command <u>setres</u> (or call <u>User interface</u> from the *Display->Options* submenu) to set this variable accordingly. In contrast to \underline{zg} , the command <u>go</u> will add the acquired data to an already existing *fid* or *ser* file. It is dangerous, however, to continue an experiment with <u>go</u> interrupted with <u>stop</u> or <u>halt</u>, which terminate acquisition immediately or after the current scan, respectively. This could occur in the middle of a phase cycle or a nD delay increment loop. In order to interrupt acquisition at a well defined position of a pulse program, the pulse program command <u>suspend</u> (see Chapter *Writing Pulse Programs*) is provided. The keyboard command <u>suspend</u> will halt the pulse program as soon as it encounters the pulse program <u>suspend</u> statement, and resume acquisition there when the command <u>resume</u> is typed in.

1.7.2 The command <u>run</u>

<u>Run</u> starts a series of experiments. You may enter <u>run</u> as soon as you have defined at least 1 experiment in <u>set</u>. <u>run</u> will invite you to answer the following questions:

1. Enter set up file name:

<u>run</u> will execute the experiments stored in the specified file. The file is created and filled with experiments either by the commands <u>set</u> or <u>extset</u>, or from the information read from bar code labels. It is stored in the directory

XWINNMRHOME/prog/curdir/changer/.

2. Enter sample changer AU program:

<u>run</u> will start up an AU program which takes over control of data acquisition, processing, and plotting. Please enter:

- a) *single_sx* for spectrometer operation with manual sample handling. Before <u>run</u> will execute the next experiment(s) corresponding to the next holder number in the set up file, it will invite you to insert the next sample. See b) how to terminate <u>run</u>.
- b) stan_sx for sample changer operation. run will move the sample carousel to the position corresponding to the next holder in the set up file (taking into account samples with priority), change the sample, execute the experiment(s) defined for this sample, and carry on with the next holder number in the set up file. If no more experiments are defined, run will not terminate, but wait until new experiments are inserted into the set up file with set or extset. In order to terminate the run command, type in run again. run will realize that it is already active, and ask you whether to terminate.
- c) *barcode_sx* for sample changer operation with bar code reader. This AU program moves the sample carousel to the next sample and reads the sample information from the bar code label. The information is entered into the set up file, the sample is inserted and the experiment executed. A bar code consists of the following components:

Experiment (2 digits), Solvent (2), UserID (3), LabelID (5), Checksum (1)

Example (in decimal form): 01 01 002 00001 9

Experiment, Solvent, UserID, LabelID also appear in readable for on the label.

Experiment

Code for the experiment to be performed (0-99). The code assigned to a particular experiment may be viewed with <u>edexp</u>. This command displays a table of experiments allowed for bar code operation. It may be adapted to

your requirements. In order to restore the original table provided by Bruker type <u>edexp new</u>. edexp displays its information from the file *XWINNMRHOME/ conf/instr/barcode.exp*.

Solvent

The solvent.

UserID

Defines the USER parameter of the data set. Remember that a the data set path is /DU/data/USER/nmr/NAME/EXPNO/pdata/PROCNO/. NAME is constructed from the first 5 characters of the name of the set up file, followed by the *LabelID*. EXPNO is set to 10 for the first experiment with a sample, 11 for the next experiment, etc. (if any). PROCNO ist set to 1. DU is defined in *barcode_sx* and may be changed there, if required.

LabelID

A label identification number, created automatically when printing labels with the command <u>prlabel</u> (described in the chapter *The 1D Output Menu*).

Checksum

Provided to be able to detect read errors.

While <u>run</u> based on *barcode_sx* is active, you may invoke <u>set</u> at any time and view which sample is in progress, and what has been measured so far. You may even insert a sample without bar code label in an empty holder of the sample changer, and define an experiment in <u>set</u>, also with priority.

- d) An own AU program (which must be stored in the directory *XWINNMRHOME/ exp/stan/nmr/au/modsrc/*), e.g. to realize centralized sample management. An example AU program *remote_sx* is part of the XWIN-NMR release. It fetches the experiment file via Ethernet *ftp* from a remote computer. Please note that such solutions depend on the particular laboratory environment.
- **3.** *First holder position (0=continue):*

<u>run</u> will start with the specified holder position. It remembers the last one and allows you to continue there if a run was aborted.

4. Enter name of list file:

While <u>run</u> proceeds, it builds up a list of data sets successfully acquired and stores it in the specified file. At any a later time, you may process or re-process the acquisition data using the command <u>run proc</u>. It will ask you to enter the list file name (the whole path, or only the file name if located in your home directory), and the command to apply to all or to the selected data sets of the list. You may specify any XWIN-NMR command such as <u>efp</u>, or <u>xau <process</u>>,

where <process> is a suitable AU program. The data set list is created by the command <u>mkcurdatlist</u> in the AU program started by <u>run</u>.

- 5. *First sample already in magnet and locked:* Answer *y* or *n* to inform <u>run</u> in which way to start.
- 6. Processing all samples:

Answer *y* or *n*. *y* enables processing. After a data set has been acquired, its corresponding processing AU program will be started (given by the processing parameter AUNMP of this data set). The AU program usually contains Fourier transform and plot commands. Processing is carried out in background, i.e. the next sample will be measured while processing or plotting of the previous ones are in progress. If you answer *n*, processing will be omitted. At any later time, you may process the data with <u>run proc</u> (see 4.)

Each time you start run, a protocol file is created (overwriting an existing one):

XWINNMRHOME/prog/curdir/protocol.

The file contains the starting times of the experiments. It is a text file you may view or print with the respective operating system utilities. For error tracking or other purposes, you may generate additional entries in the protocol file. You must add the AU command PROTOCOL(text) at the desired place of the AU program. *text* must be a *char* * variable (a text string) according to the C language.

1.7.3 The command <u>quicknmr</u>

<u>quicknmr</u> is an easy to use tool for routine spectroscopy. It is a dialog window where you define the data set for the acquired data, the solvent, the experiment, and a plot title. Then acquisition may be started. Please note that you should now use <u>iconnmr</u> rather than <u>quicknmr</u>, which is a newer product.

It is the responsibility of the spectrometer administrator to assign permissions to the users of <u>quicknmr</u> with the command <u>eduser</u>. The allowed data set names, experiments, right to modify parameters, and to exit from <u>quicknmr</u>.

1.7.4 Displaying the lock and fid window

• Open the acquisition window by selecting the command <u>Observe fid window</u> from the *Acquire* menu. In this window, the acquisition data (fid) may be observed in real time either in the time domain (as an fid) or in the frequency

domain (as a spectrum). This window will also show the wobble curve during the probehead tuning procedure.

• Execute the command <u>lockdisp</u> (type it in or call it from the *Windows* menu). The lock signal will be displayed for locking, shimming, or vising.

Chapter 2 The *Windows* Menu

The Windows menu contains XWIN-NMR commands which generate new windows

📥 Windows	
Plot-editor	[xwinplot]
ICON-NMR	[iconnmr]
Lock	[lockdisp]
Amplifier control	[acbdisp]
Pulse program	[pulsdisp]
Shape Tool	[stdisp]
BSMS panel	[bsmsdisp]
Temperature Monitor	[temon]
MAS Rate Monitor	[masrmon]
Command panel	[cpan]
Plot preview	4

Figure 2.1 The Windows menu

independent of the main XWIN-NMR window. They may remain on screen simultaneously with other windows of this type, and with the main XWIN-NMR window. Since XWIN-NMR is a multi-tasking program, the commands are executed in parallel. For example, if a data acquisition is in progress, and the lock and the amplifier control windows are open, the contents of all three windows are refreshed at the same time, and the user may at any time move the mouse into one of the windows and execute a command.

2.1 Interactive plot editor [xwinplot]

The command <u>xwinplot</u> opens an interactive plot editor which allows you to set up plots interactively on the screen and plot data on a variety of printers, or store data in ecapsulated PostScript or other formats for inclusion in documents. The layouts created with <u>xwinplot</u> may be used for plotting of other data sets. Particularly, the XWIN-NMR command <u>autoplot</u> plots the current data set using the layout file whose path name is defined by the parameter LAYOUT. This parameter must be set up using the command <u>edo</u>. The following path name conventions are valid: If the path name begins with a "~" character, the specified path will be relative the the login user's home directory. If it begins with a "+" sign, the path will be relative to *\$WXINNMRHOME/plot/layouts/*.

xwinplot is described in detail in its own manual, which is also accessible on line.

2.2 Routine Spectrometer operation [iconnmr]

The command <u>iconnmr</u> opens a special user interface designed for easy spectrometer operation in a routine environment. Please refer to the ICON-NMR manual for details.

2.3 Lock [lockdisp]

The command <u>lockdisp</u> opens a new window and displays the spectrometer's lock signal. Parameters, such as shim values, influencing the lock signal may be adjusted from the SCM or BSMS keyboards, or from the BSMS panel window (see below). The lock signal enters an RS232 channel on the acquisition rack of the spectrometer, and is transferred to the workstation over a network connection. The RS232 channel number must be specified when executing the configuration command <u>cf</u>.

The lock display window provides a number of commands which may be activated

by clicking on the bottom row command buttons.

grid This is a toggle command displaying a grid, a vertical bar, a horizontal bar, or nothing in addition to the lock signal. The vertical bar divides the window in a left and a right part of equal size. The horizontal bar divides the window in an upper and a lower part at 85% of the window's height.

mode This command toggles between two color modes. In the first mode, both forward and back sweeps of the lock are displayed in the same color. In the second mode, the color of the back sweep is different.

store This command allows you to save the current grid and mode settings, the size, and the position of the lock window in a file. When <u>lockdisp</u> is executed, and no such file exists, the lock window will come up with default settings programmed in XWIN-NMR. If several save files exist, and one of the files is called *default*, the lock window will appear according the settings stored there. The files are stored in the directory */u/exp/stan/nmr/loc_win/*.

read The purpose of this command is to read in the lock window settings saved in a file generated with the store command described above. A list of available files is displayed, from which you may select the desired one.

quit Terminates the lockdisp command and closes the lock window.

2.4 Amplifier control [acbdisp]

2.4.1 Introduction

The <u>acbdisp</u> (Amplifier Control Board Display) XWIN-NMR command represents another new digital control approach, improving the functionality of Bruker Spectrometers. It is available for the Avance spectrometer series. The program allows all amplifier activity to be monitored at the workstation terminal.

2.4.2 Functionality

Decide which amplifiers to display with the aid of the *Transmitters* pulldown menu. It contains a list of all currently activated amplifiers. An amplifier is said to be active when it is used in a particular routing arrangement which has been activated by means of the <u>edasp</u> or <u>edsp</u> command. The system is completely flexible

and able to monitor any of the 48 possible amplifiers in any given spectrometer. Amplifier type and maximum output power is also displayed. Not only that but depending on the type of experiment the correct set of amplifiers will automatically be monitored by the program.

2.4.3 Initialization

The aquisition startup routine informs <u>acbdisp</u> which amplifiers are being used before the first pulse hits the probe. For those of you still fond of the *Led System* <u>acbdisp</u> also configures the top part of the BSMS keyboard display where the chosen amplifiers forward and reflected power values, pulse type, observed channel and mismatch leds may still be found. The set of leds on the BSMS keyboard may also be configured using the <u>acbbsms</u> command which updates the display silently without starting up the acb display.

2.4.4 Amplifier Sampling

The amplifiers are sampled at regular intervals, between five and fifteen times per second (user selectable from the *Controls* window). This ensures that the workstation need only handle that amount of sampling necessary to achieve an accurate picture of the current experiment. Should a fault condition develop the user is notified by means of the status display line showing the first major fault plus a list of all faulting amplifiers. The XWIN-NMR status line also gives notice of the fault. In *paused* mode the amplifiers are still scanned once a second to check the respective amplifier status bits for a possible fault condition. If everything is functioning correctly the Status line shows a *standby status* and may occasionly show *blank* meaning that specified transmitter is currently blanked. The *blank* status is not a failure condition.

2.4.5 Controls

The *Controls* pulldown menu provides access to <u>acbdisp</u> program itself as well as the ACB directly. The *scan rate* push buttons control the rate at which the ACB is sampled for transmission information. The push buttons may be activated by clicking on the appropriate button after having pulled down the window by clicking on the *Controls* part of the menu bar. If you wish to pause the display at any time use the *pause display* feature. *ACB Reset* resets the ACB and initializes all observed channels afresh. This is necessary to free all amplifiers for action after the Transmission Power Down Button has been activated on the BSMS Keyboard, it takes about 10 seconds. This also provides a useful way of testing all amplifiers currently connected to the system as the reset command searches for all amplifiers on the system and displays all of them (up to a total of 6).

2.4.6 Motif Resources

Various colours and modes including the chosen font for the display may be changed by editing the file */u/prog/version/app-defaults/AcbDisp*. This file is delivered with your release and sets up a default session's characteristics. It may be changed at any time according to your personal preferences.

2.5 Pulse program [pulsdisp]

The command <u>pulsdisp</u> displays pulse programs written for Avance type spectrometers. It requires that the <u>cf</u> configuration command was executed for an Avance. The spectrometer executes the timing of pulses on a processor called TCU (timing control unit), while the frequencies, phases and power levels of pulses are handled by the frequency control units (FCUs). For each frequency channel one FCU is responsible. The pulse program display simulates the behaviour of the TCU and the FCUs using a TCU simulator and a FCU simulator (Figure 2.2). The TCU simulator uses the same source code that is used by the embedded controller on the TCU, and the FCU simulator uses that same memory image that is also used by the FCU. That means that everything that comes out of this two simulation programs reflect exactly what happens on the spectrometer hardware. You may specify the duration of the simulation in terms of seconds or number of scans. The pulse program output is recorded during this time and visualized on the display. You will find any fine detail of a timing sequence, even of implicite delays certain pulse program macro commands such as <u>go=n</u> apply.

The command <u>pulsdisp</u> uses the current pulse program defined by the parameter PULPROG. In order to set or examine it, type in <u>pulprog</u> or <u>eda</u>. Figure 2.3 shows the main window of the pulse program display routine. First click on the *Observe setup* button. A table of possible channels will appear. Enable the channels you want to observe by clicking on the checkbuttons. *CH1* etc. denote the timing channels, *FCU1* etc. the output of the frequency units, and *GCU-X/Y/Z* the outputs of the gradient unit. FCU channels need only be enabled if phases or powers (shapes) are to be displayed. Click on *Apply* or *OK* to make the current checkbutton settings effective. *OK* will, in addition to *Apply*, close the table. *Cancel* will close it without



Figure 2.2 TCU and FCU simulator

📥 Pulse Pi	rogram Display	•
<u>File V</u> iev	w <u>O</u> ptions	<u>H</u> elp
TCU simula	ator setup	Undo
Start: 0	Stop: 1 Unit: Time (se	:c) 🚸 Scans
Observe set	up Run simulation New display Show program Quit	

Figure 2.3 Pulse Progam Display

changing the last settings. If the *RCUGO* channel is enabled, the pulse program display window will contain a section with the *RCUGO* pulse, whose purpose is to start the receiver control unit RCU. This is equivalent to the start pulse given to the

digitizer. If you enable the *Receiver* checkbox, a receiver field will be displayed showing the on/off times of the receiver, and its phase. Finally, enabling the *Prog* checkbox will cause the pulse program line responsible for the generation of a certain duration to be displayed in the same column as the graphical representation of the duration.

Next, you must specify the duration of the simulation. Click on *Time* or *Scans* in the main pulse program display window, depending on whether you want to specify a time or a scan interval to be simulated. Insert the time or the scan number where simulation should start in the *Start* entry field, and the stopping time or the last scan in the *Stop* entry field. Please note that if any FCU channel is enabled in the *Observe setup* table, the contents of the *Start* entry will be ignored, and simulation will always start at time zero, equivalent to the first scan number. You may use negative numbers for the scan number. This will cause the simulation of dummy scans if specified in the acquisition parameter set. If you want to simulate just the first scan of an experiment, enter 0 for *Start*, and 1 for *Stop*.

You are now ready to click on the *Run simulation* button. Note that if you change acquisition parameters or the pulse program you have to quit the pulse program display and restart it. Just running the simulation again is not sufficient. Simulation time depends on the selected time or scan interval, usually a few seconds for a small number of scans. If you are doing FCU simulation and if you are using a pulse program that changes the phase very frequently (e.g. a pulse program with a CPD sequence), the simulation time can be very long, and it is also possible that the memory limit is exceeded or the graphic gets too large and will be truncated (an error message will appear if this happens).

After running the simulation the actual pulse program display window (Figure 2.4) is opened containing all the channels selected in the *Observe setup*. You may disable or re-enable selected channels from the *Setup display* button, without modifying the *Observe setup*. The *Zoom in* button allows you to investigate the fine details of the pulse program. Particularly text labels indicating the length of durations, power levels in db or per cent units, phases in degrees, or pulse program text will not appear before you zoom in deep enough so as to create space for the text. The *Zoom out* button reverses *Zoom in* and allows you get an overview of the entire simulation. You may create an additional pulse program window (a copy of the current one) from the *New display* button. You can use the one window as an overview, and the second one to zoom in. The procedure is not limited to two windows. Use the *Kill display* button to close a window. If you want to get a print out of the

A-104	1
-------	---

Display								1	•
CH 1 (F1)	0.	00505300s		2.02507720	s 2.0	4510890s			
	4995.75µs	50.00µs	2000000.00µs	20000.00µs	25.00µs	19997.OOµs	2497. OOµs	122.00µs	2.
CH 2 (F2)	i.			2.02507050s	1				
	4995.75µs	50.00µs	::2000000.00µs;::	20000.00µs	: :::25. 00µs ::: :	19997.OOµs ¦¦	2497.00µs	122.00µs	2
FCU 1 (F1)	0.00db	0.00db	27.01db	0.00db	0.00db	30.00db	30.00db	30.00db	1 3
	359.99°	359.99°	359.99°	359. 99°	359.99°	359.99°	89.99°	359.99°	
ECIL2 (E2)	4.83db	4.83db	ii 4.83db iii	4,83db	iii4.83db iii	4.83db ii	4.83db	ii 4.83db	1
	9.06°	9.06°	9.06°	9.06°	9.06°	189.06°	189.06°	189.06°	
Receiver									
Prog	1 ze	2 d11			111 111			4 (p6 pł	122 p
Setup displa	y Align panes	Zoom In	Zoom Out N	ew display Ki	II display Prir	nt			

Figure 2.4 Display Window

current display, click on the *Print* button. The print dialog box allows you to start printing, or to preview the output. The printer type, paper format, layout parameters, the preview command and the plot command must be specified in the print setup dialog box. It is not always possible to output a long simulation on a single sheet of paper with enough resolution. The print setup dialog therefore allows you to expand the output in x and y direction across several pages. This is achieved by filling the *Split to X*Y pages* entry with the desired numbers.

Finally, the *Options* menu of the main <u>pulsdisp</u> window contains the commands *Scale setup* and *FCU simulator setup*. The *Scale setup* allows you to select logarithmic scaling of the time axis versus linear scaling. Linear scaling will cause a duration to be drawn in dashed lines if it exceeds a certain limit, which may be specified. The length of the drawing corresponds to the specified limit. The reason for this feature is the fact that pulse programs usually contain microsecond pulses and millisecond or second delays. A reasonable display of the pulse sequence can only be obtained by artificially limiting the width of long durations on the screen. Logarithmic scaling, although distorted, often results in an appealing presentation of a sequence. The *FCU simulator setup* allows you select the pulse power units (db, a logarithmic unit, or %, a linear unit).

2.6 BSMS panel [bsmsdisp]

2.6.1 Introduction

Prerequisite Skills:

This description to the BSMS display tool version 1.0 is intended for users which are already familiar with the operation of the Bruker Smart Magnet Control System (BSMS) accessed by either the BSMS keyboard version HR-20 (BOSS 1) or version BOSS 2. For any questions to the general operation of the BSMS, please refer to the BSMS User Manual.

About the BSMS Display Tool:

The BSMS display tool is a user interface to the BSMS hardware. Functions controlled by the BSMS can be accessed and manipulated from this application. The primary user interface has been up to now an additional hardware control unit: a keyboard attached directly to the BSMS. The tool provides an alternative user interface with the additional capability to access the BSMS from any remote host connected by a network.

The BOSS 2 keyboard is an extension of BOSS 1; its complete functionality is not supported by the display tool version 1.0. Spectrometers equipped with the newer BOSS 2 keyboard cannot fully be controlled by this software, as some of the BSMS functions are missing (some shim currents cannot be accessed).

2.6.2 Getting Started

2.6.2.1 How to Start the BSMS Display Tool

The BSMS display tool can be started from either the XWIN-NMR-package, or

from a UNIX[®] command shell by entering the command <u>bsmsdisp</u>. An "&" typed after the command in a UNIX[®] shell allows to continue operation with the shell immediately.

Starting the Display Tool after a New Installation:

When the tool is started up the first time after a new installation, an automatic calibration process is invoked, checking the BSMS hardware for the limits of the functions supported by the display tool. The values displayed on the BSMS keyboard (if connected) during this calibration process may be neglected. BSMS parameters are left unchanged.

Start-up Options:

Start-up options are not recognized by the BSMS display tool directly; they are passed to the X Window System[®]. To specify another font, add a font resource on the command line as for example <u>bsmsdisp -xrm "XBsmsDisp*fontList: *-helvet-ica-bold-r-normal-*-10-*-*-*-*-*</u>"

2.6.2.2 The Main Panel

When starting the BSMS display tool, a first window, the so-called "main panel" comes up, a window with the following layout:



Figure 2.5 Layout of the Main Panel

In the upper part of the panel, a menu bar with pulldown menus labelled "Display", "Options", "Service", as well as a "Help" menu provide commands controlling the general behaviour of the display tool. The lower part offers 5 so-called
"control buttons" to open the "control panels" (refer to Chapter 2.6.3.1: "Opening and Closing Control Panels").

The main panel is intended to control the general tool operation. Therefore, this panel is always open when working with any one of the control panels. The latter can, however, be opened and closed individually to provide just the functionality currently needed.

2.6.3 The Control Panels

Control panels are used to invoke commands on the BSMS unit. For each subsystem of the BSMS, a separate control panel exists offering functions referring to that subsystem. Due to the extensive functionality provided by this tool, the BSMS functions are distributed to 5 different panels. Normally, only one or two control panels are needed at one time.

2.6.3.1 Opening and Closing Control Panels

Control panels can be opened (only) from within the main panel using the control buttons "Level", "Lock", "Sample", etc. of the lower button row in the main panel. A click on the "Level" button for example, will open the "Level" panel:



Figure 2.6 Opening a Control Panel

When a panel is opened, an entry like "Close Level Panel" is appended to the "Display" pulldown menu to close the corresponding panel again (Figure 2.6).

A control panel currently not needed can be closed either by using the entry "Close Panel" from the "Display" pulldown menu (refer to Figure 2.6) or by using the window-managers "Close" or "Quit" entry.

When a control panel is closed, the additional menu entry "Close ... Panel" in the "Display" pulldown menu of the main panel disappears again. The "Display" pulldown menu always contains the list of all currently opened control panels.

2.6.3.2 Control Panel Design

A closer look at the control panels reveals one design used throughout the display tool. The upper area of the control panel contains a set of buttons, the so called "Function Buttons", which refer to functions on the BSMS device. They are used either to switch functions on the device on or off, or are used to select a so-called "value function", a function representing a value displayed in the lower area of the control panel, which can be changed with buttons, sliders or text/display fields. The following figure describes the general layout of a control panel:



Figure 2.7 Layout of a Control Panel

2.6.3.3 Toggle and Value Functions

Most of the function buttons in the control panels can be assigned to one of two function classes: toggle or value functions.

The BSMS device has many functions that can be controlled by the display tool. Some of them represent a simple state like "being turned on" or "being turned off". The "LIFT ON-OFF" command is one example. These functions are referred to as "toggle functions". A click on a function button representing a toggle function has an immediate effect on the BSMS device: when the button is highlighted, the function is currently turned on. Clicking that button will turn off the corresponding function on the BSMS, visualized by changing the button color to the background color¹. These buttons show the state of the function on the BSMS unit directly.

The other class of functions represent a value, which is not displayed on the button itself (there's no space left on the button), but in two value fields in the lower part of the control panel. The two display fields show the (current and old) function value. With additional control elements, this value can be changed (refer to chapter 2.6.3.7 for changing function values).

Toggle and value functions cannot be distinguished on the panel directly; there's no visual difference between these two classes of buttons. This "knowledge" must come out of your experience.



Figure 2.8 Toggle and Value Functions

^{1.} Depending on the current color setup

Figure 2.8 shows the "Sample" panel with its five function buttons. "LIFT ON-OFF", "SPIN ON-OFF" and "SPIN CALIB" are toggle functions, whereas the first and third currently is turned off and the second ("SPIN ON-OFF") turned on (indicated by a highlighted button). "SPIN RATE" and "SPIN MEAS" in turn are value functions; "SPIN RATE" currently is highlighted as its value is displayed in the two value fields.

2.6.3.4 Toggle (On/Off-) Functions

A toggle function is a function with two states: turned on and turned off.

The state of a toggle function is indicated on the corresponding button itself: a button with a bright color represents a function "turned on", whereas a dark colored button a function that is "turned off"¹. For example in Figure 2.8, the "SPIN ON-OFF" function is turned on (highlighted) and the "LIFT ON-OFF" and "SPIN CALIB" functions turned off.

To change a toggle function, just click on the corresponding button and it's state will toggle; clicking on a button turned on will turn it off and vice versa, unless the BSMS currently does not allow a change of that function, in which case a short "beep" is emitted.

2.6.3.5 Value Functions

Value functions have assigned a value, which is displayed in the two display fields in the lower part of the control panel (see "left display field" and "right display field" in Figure 2.7). This means, only one value function can be displayed and controlled in a panel at a time.

To change a functions value, "select" that function by clicking on the corresponding function button. The function values are going to be displayed in the two value fields.

In Figure 2.8, the value function "SPIN RATE" has been selected as the "current function" (indicated above the two display fields); the function button "SPIN RATE" is highlighted. The current function values can be changed with the control elements in the lower area, described in the following paragraphs.

^{1.} Depending on the current color setup.

Each control panel can have a current function, once there is clicked on a value function button. The current function name appears above the two display fields and its corresponding function button is highlighted.

2.6.3.6 The "Left / Right Display Field"

The "left display field" contains as a reminder the initial value of the current function, e.g. the value at the time the current function was selected. The "right display field" contains the actual value as currently set and confirmed by the BSMS device.

Once a value function has been selected, their values are going to be displayed in the "left display field" and "right display field". Both value fields show the current function value set on the BSMS device. When changing now that value (as described below), the new (current) value is displayed in the right value field and the left field remains unchanged. The actual value set on the BSMS device is **always** displayed in the right value field, emphasized by the label "ACTUAL" above. The left value field only changes when switching to another function; so you can find there the "old" function value, indicated by the label "PREVIOUS" above.

Changing a functions value is allowed only within the functions range, which is displayed above the "slider": The "function minimum" and "function maximum" (Figure 2.7) show the functions smallest resp. largest value allowed to be set.

2.6.3.7 Methods to Change the Current Value of a Function

There are different methods provided to change the current function value:

- 7. Using the "-/+" button
- 8. Using the "slider"
- 9. Using the "right display field"

In the following paragraphs, each method is discussed.

2.6.3.8 Changing the Current Value Using the "-/+" Button

By clicking on the "-/+" button with the **left** mouse button, the current value displayed in the right value field is decreased (unless it has already reached its lower limit):



Figure 2.9 Usage of the "-/+" Button

If you click on the "-/+" button with the **middle** mouse button, the button label will toggle from "-/+" to "+" and the value will be increased instead of decreased.

The size by which the value is in-/decreased is displayed next to the label "Step:". It is currently set to two units in the figure above, but can be changed by using the "2/*2" button (see paragraph 2.6.3.11: "Function Sensitivity").

When holding down the mouse button on the "-/+" key for a longer time, continuous adjustment of the function value is possible. This button has an auto repeat behaviour.

2.6.3.9 Changing the Current Value Using the "Slider"

Setting Values by "Dragging" the Slider Button:

The slider provides a method to change a value by using the "slider button". Push (hold down) the left mouse button with the mouse pointer onto the slider button and move the mouse to the left or to the right. You will see that the slider button follows your mouse pointer and the actual value (ACTUAL) will change accordingly.

When the slider button has reached the left edge of the slider, the function is set to the minimal value allowed (compare right display field with function minimum; they should be equal). At the right sliders edge, the function is set to its maximum, where function maximum and the right display field should show the same value.

Setting Values by Positioning the Slider Button Directly:

The slider button can be placed immediately by a click with the middle mouse button on the estimated spot within the slider. The button jumps to that position without the dragging behaviour:



Figure 2.10 Usage of the Slider

Continuous Adjustment of Values:

The slider can be used for continuous adjustment of the current value in a way that maps to the behaviour of the "-/+" key: using the left mouse button, a click in the slider area on the right of the slider button (not on the button itself) will increment the value by the displayed step size. Holding down the button for a while will adjust the value continuously. In turn, when clicking on the left of the slider button (but still within the slider), the value is going to be decremented.

2.6.3.10 Changing the Current Value Using the "Right Display Field"

The right display field can accept values entered from the (computer-)keyboard. Click with the left mouse button in the right display field. A frame is drawn around that field and a cursor is placed after the value. Clear the current value, enter the new value and press <return> ("Enter" key). The new value is set now.

Values out of the function range will not be accepted and the current value is redisplayed in the right value field.

2.6.3.11 Function Sensitivity

Two of the three methods to change the current function value use an "offset" for incrementation / decrementation (the "-/+" button and the slider method). This offset is displayed next to the label "Step:".

When clicked on the ":2/*2" button with the left mouse button, the current step size is approximately halved; when clicked with the middle mouse button, approximately doubled. Change that step size and the current value, either with the slider or the "-/+" button. Incrementation resp. decrementation will follow the new step size now.

This size is useful for constant adjustment of function values and can be seen as some kind of "function sensitivity": a large step size will change the value faster when using for example the "-/+" button held down for a certain time, while a lower step size will be "less sensitive".

Each value function has assigned an own step size! Switching to another function will also switch to the other functions private step size. These function sensitivities can be saved with the "Save Function Sensitivity" command.

2.6.3.12 The Differential Mode (Diff.Mode)

For normal operation, the panels display absolute numbers for value functions in the left and right display fields.

In differential mode, the PREVIOUS field displays the change made the last time the function was accessed. The ACTUAL field contains the total change made since the function was selected to be the current function.

2.6.4 Display Tool Commands

2.6.4.1 Terminating the Display Tool

... from within the BSMS Display Tool:

Exit the BSMS display tool with the "Exit" command in the "Display" pulldown menu, or by using the window-managers "Quit" or "Close" command. It is recom-

mended to terminate the BSMS display tool always this way. If this method does not work (the tool does not react on inputs any more), the application must be killed "from outside" by a UNIX[®] shell (see below).

... from XWIN-NMR

When the BSMS display tool has been started from XWIN-NMR, next to the builtin exit- and quit-command, the tool can also be terminated with the "kill cmd" command entered on the XWIN-NMR command line.

... from a $\text{UNIX}^{(\!R\!)}$ Shell

This method should be used only, when the methods explained above fail.

Terminate the tool from a command shell with the command "kill" and the correct process-ID. The latter can be found with the UNIX[®] - "ps" and "grep" command: <u>ps | grep bsmsdisp</u> will display something like "2010 ttyq6 0:00 bsmsdisp". The first number (here: 2010) is the process-ID. The tool can be killed with the command <u>kill 2010</u>. If that doesn't terminate the application, enter <u>kill -9 2010</u> and the BSMS display tool disappears.

2.6.4.2 Turn the Display Tool On/Off

The display tool can be turned off when access to the BSMS is not needed for a while. Use the menu entry "Turn Off" from the "Display" pulldown menu in the main panel to switch the display tool off. This menu entry changes to "Turn On", when the tool has been turned off, and all control panels are closed (as they are not needed when the display tool is off-line). You will see that the control panel buttons are disabled (dimmed) in the off-line state, when reopening any control panel. Accessing the BSMS is not possible, until the tool is turned on again with the same menu entry (which has changed now to) "Turn On".

If the tool (turned on) is iconified, it turns off automatically during the iconified state.

To prevent any interference of the BSMS display tool with other applications, it is

recommended to turn the BSMS display tool off if not in use.

As long as a dialog with a BSMS error message is open, no other application can access the BSMS!

2.6.4.3 Save Panel Layout

When the tool is started the first time (after a new display tool installation), the panels pop up at default positions on the screen, which can be customized individually. If you wish the display tool to place the panels at your customized positions on the screen on start-up, use the "Save Layout" command from the "Options" pulldown menu of the main panel. Calling this command saves the current panel positions; so the panels must be placed correctly before the command is issued.

2.6.4.4 Save Function Sensitivities

Using the "-/+" button to change the current function value needs the step size, a value which is added / subtracted, when this button is pushed (the same is true for the slider). We saw earlier, that the step size can be compared to a kind of sensitivity, by which the function value is changed. The "-/+" button held down with a larger step size changes the function value faster (in bigger steps).



Figure 2.11 Step-Size

2.6.4.5 Setup Colors

5 different colors are used by the display tool, which can be set up individually: A color for (1) the background of all panels, (2) the font used for the buttons and labels of all panels, exclusive the font of the function buttons, (3) the function buttons in the control panel in the "turned off" state, (4) the function buttons in the control panel in the "turned on" state, and (5) the function button font.

To change the colors of the application, use the "Color" command in the "Options" pulldown menu of the main panel. Issuing the color command pops up a color dialog. Select the type of color to change from the left box in the color dialog (e.g. "Panel background" to set a new panel background color). Select now one of the colors from the color palette on the right by clicking on the corresponding color button (e.g. click on the red color on the color palette; the panel background will change to red). The panels will show the new color immediately¹.

2.6.4.6 Display Tool Calibration

The panel can be calibrated with the "Calibrate Display" command in the "Display" pulldown menu.

When value functions are changed with the control panels, the valid function range is displayed above the slider ("function minimum" and "function maximum" in Figure 2.7). A value can be set only within this range; other values are not accepted by the BSMS device. The BSMS display tool does not know, however, these function ranges in advance; they depend on the currently installed BSMS hardware and must be queried from the BSMS device by the display tool. This is done with the display tool calibration; the ranges are saved into a defaults-file. If you think, function maxima / minima are not set correctly in the control panels due to hardware exchange, just start a calibration.

^{1.} Only the control panels show an immediate effect. The main panels colors change when you restart the application.

2.7 Temperature monitor [temon]

The command <u>temon</u> opens a window and displays the sample temperature in Kelvin or Celsius, depending on your selection. The last setting is saved in the text file

\$HOME/.xwinnmr-<host>/temon

where \$HOME is the XWIN-NMR user's home directory, and <host> the computer's network name. If the feature *Write log file* is enabled, the temperature is written into a text file in regular, selectable time intervals. The log file name is

/DU/data/<user>/nmr/NAME/EXPNO/temperatureLog

i.e. in the acquisition directory of the data set from where temon was started.

The <u>temon</u> window has two modes. The first one displays the temperature along with the time it was measured. In addition, it shows the minimum and the maximum value measured during the period the <u>temon</u> window was open. The second mode only displays the temperature in a larger font.

2.8 MAS rate monitor [masrmon]

The command <u>masrmon</u> opens a window and displays the Magic Angle Spinning Rate in Hertz. Its behaviour is similar to the <u>temon</u> window of the previous section. The respective log file name is *masRateLog*.

Chapter 3 The Shape Tool

3.1 Introduction

XWIN-NMR 2.0 and later provides a number of utilities for the creation of pulse shapes:

- The command <u>stdisp</u> opens a new display that allows you to view shapes, and to interactively create and manipulate shapes using various parameters.
- The command <u>st <parameters></u> allows you to generate or manipulate shapes from the XWIN-NMR command line. The main purpose of this syntax is that the command can easily be included in AU programs to produce shapes automatically.
- Once created or changed, a shape is stored in a text file in JCAMP-DX format suitable to be displayed with stdisp, and to be executed by the Xwin-nmr acquisition commands.

The <u>st</u> command is available in the following forms (<u>stdisp</u> provides the analogue functions in its menus for display interactive work):

- 1. <u>st generate <further parameters></u> Generates new shapes
- 2. <u>st manipulate <further parameters></u>

Manipulates existing shapes

- 3. <u>st analyze <further parameters></u> Analyze existing shapes
- 4. <u>st add <further parameters></u> Add 2 existing shapes to form a new shape
- 5. <u>st merge <further parameters></u> Merge 2 existing shapes to form a new shape

3.2 Generate a new shape

• Syntax: <u>st generate <shape type> <further parameters></u>

The result is stored the file

\$XWINNMRHOME/exp/stan/nmr/lists/wave/<shape type>.

3.2.1 Burp Shapes

• Syntax:

```
st generate <shape type> <size>
or
st generate <shape type> <size> false
```

In the second case only amplitude data are generated. int <size> = shape size in number of points string <shape type> = one of: EBurp1, EBurp2, IBurp1, IBurp2, ReBurp, UBurp1, I2Burp, I3Burp.

• Example: st generate IBurp1 256

3.2.2 CosSinc Shapes

• Syntax:

st generate CosSinc <size> <cycles> <window size> or st generate CosSinc <size> false <cycles> <window size> In the second case only amplitude data are generated. int <size> = shape size in number of points int <cycles> = number of cycles to calculate double <window size> = window size in per cent

• Example: st generate CosSinc 128 4 50

3.2.3 Gauss Shapes

• Syntax:

st generate Gauss <size> <trunclevel> or st generate Gauss <size> false <trunclevel>

In the second case only amplitude data are generated. int <size> = shape size in number of points double <trunclevel> = truncation level in per cent

• Example: st generate Gauss 256 10

3.2.4 HalfGauss Shapes

• Syntax:

<u>st generate HalfGauss</u> <u><size> <trunclevel></u> or <u>st generate HalfGauss</u> <u><size> false <trunclevel></u>

In the second case only amplitude data are generated. int <size> = shape size in number of points double <trunclevel> = truncation level in per cent

• Example: st generate HalfGauss 256 10

3.2.5 GaussCascade Shapes

• Syntax:

<u>st generate <shape type> <size></u> or st generate <shape type> <size> false

In the second case only amplitude data are generated. string <shape type> = one of: GaussCascadeG3 (Inversion), GaussCascadeG4 (Exitation) GaussCascadeQ3 (180 deg.), GaussCascadeQ5 (90 deg.) int <size> = shape size in number of points

• Example: st generate GaussCascadeG4 256

3.2.6 Hermite Shapes

• Syntax:

```
<u>st generate Hermite</u> <<u>size></u> <<u>trunclevel></u> <<u>quadcoeff></u>
or
<u>st generate Hermite</u> <<u>size></u> false <<u>trunclevel></u> <<u>quadcoeff></u>
```

```
In the second case only amplitude data are generated.
int <size> = shape size in number of points
double <trunclevel> = truncation level in per cent
double <quadcoeff> = coefficient of quadratic term (-4.0 .... +6.0)
```

• Example: st generate Gauss 256 10 1.0

3.2.7 Hyperbolic Secant Shape

• Syntax:

st generate HypSec <size> <phase> <trunclevel> tru or st generate HypSec <size> <phase> <trunclevel> false

int <size> = shape size in number of points double <phase> = phase (squareness) factor ($0.0 \le p \le 16.0$) double <trunclevel> = truncation level in % bolean <full / half > = full / half full or half passage (true / false)

• Example:st generate HypSec 256 1.0 .1 true If only amplitude data are needed:

The call is: st generate HypSec 256 false 1.0.1 true

3.2.8 Rectangle Shape

• Syntax: st generate Rectangle <size> <phase> <trunclevel>

int <size> = shape size in number of points double <trunclevel> = truncation level in %

• Example: <u>st generate Rectangle 256 10</u> If only amplitude data are needed: The call is: <u>st generate Rectangle 256 false 10</u>

3.2.9 Generate SinCos Shape

 Syntax: <u>st generate SinCos <size> <trunclevel> tru</u> or <u>st generate SinCos <size> <trunclevel> false</u>

 $\begin{array}{l} \mbox{int} <\mbox{size}> =\mbox{shape size in number of points} \\ \mbox{double} <\mbox{factor}> =\mbox{factor phase amplitude factor } (0.0 <=\mbox{p} <= 16.0) \\ \mbox{bolean} <\mbox{full} /\mbox{half}> =\mbox{full} /\mbox{half full or half passage (true / false)} \end{array}$

• Example: <u>st generate SinCos 256 8 true</u> If only amplitude data are needed: The call is: <u>st generate SinCos 256 false 8 true</u>

3.2.10 Sinc Shape

 Syntax: <u>st generate Sinc <size> <cycnum></u>

int <size> = shape size in number of points
int <cycnum> = cycnum nunber of cycles to calculate

• Example: <u>st generate Sinc 256 8</u> If only amplitude data are needed: The call is: st generate Sinc 256 false 8

3.2.11 Sine * Sinc Shape

 Syntax: <u>st generate SineSinc <size> <cycnum></u>

int <size> = shape size in number of points int <cycnum> = cycnum nunber of cycles to calculate

• Example: <u>st generate SineSinc 256 8</u> If only amplitude data are needed: The call is: <u>st generate SineSinc 256 false 8</u>

3.2.12 Sinus Shape

• Syntax: st generate Sinus <size> <cycnum> <phase>

int <size> = shape size in number of points int <cycnum> = cycnum nunber of cycles to calculate double <phase> = phase phase angle (0.0 <= p <= 360.0)

• Example: <u>st generate Sinus 256 8 180.0</u> If only amplitude data are needed: The call is: <u>st generate Sinus 256 false 8 180.0</u>

3.2.13 Sneeze Shape

• Syntax: <u>st generate Sneeze <size></u>

int <size> = shape size in number of points

• Example: <u>st generate Sneeze 256</u> If only amplitude data are needed: The call is: <u>st generate Sneeze 256 false</u>

3.2.14 Snob Shapes

• Syntax: st generate DSnob <size>

> int <size> = shape size in number of points The following Snob Shapes are implemented: ESnob ISnob2 ISnob3 DSnob ESnob

• Example: <u>st generate DSnob 256</u> If only amplitude data are needed: The call is: <u>st generate IBurp1 256 false</u>

3.2.15 Triangle Shape

• Syntax: <u>st generate Triangle <size></u>

int <size> = shape size in number of points

• Example: <u>st generate Triangle 256</u> If only amplitude data are needed: The call is: <u>st generate Triangle 256 false</u>

3.3 Manipulate existing Shape

3.3.1 command cosm2

single cosine modulation (+/- offset)

• Syntax: st manipulate Gauss cosm2 <size> <pulDur> <f>

```
int <size> = shape size in number of points
double <pulDur> = pulDur = length of shaped pulse (in us)
double <f> = offset frequency (in Hz)
```

• Example: <u>st manipulate Gauss cosm2 128 100 3000</u> If only amplitude data are needed: The call is: st generate Triangle 256 false

3.3.2 command offs

phase modulation according to offset frequencies.

The manipulate command offs has following subcommands:

- b beginning at phase 0
- m phase of 0 at middle of shape
- e ending at phase 0
- Syntax: st manipulate Gauss offs b <pulDur> <n> <f1> <f2>

```
double <pulDur> = pulDur = length of shaped pulse (in us)
double <f1> = offset frequency 1 (in Hz)
double <f2> = offset frequency 2 (in Hz)
.....
```

```
..... double \langle fn \rangle = offset frequency n (in Hz)
```

• Example: <u>st manipulate Gauss offs b 100 2 2000 3000</u> Calculates phase modulation beginning at phase 0.

Optional commands are:

- f frequencies taken from frequency list.
- Syntax:

```
<u>st manipulate Gauss offs b f <pulDur> <freqlist</u>>
string <freqlist> = file with frequency-list
```

- Example: st manipulate Gauss offs b f 100 freqlist
- p additional phase setting.
- Syntax:

st manipulate Gauss offs m p <pulDur> <n> <f1> <p1> <f2> <p2>

double <pulDur> = pulDur = length of shaped pulse (in us) int <num> = number of offset-frequencies double <f1> = offset frequency 1 (in Hz)

```
double <p1> = initial phase 1 (in degree)
double <f2> = offset frequency 2 (in Hz)
double <p2> = initial phase 2 (in degree)
......
double <fn> = offset frequency n (in Hz)
double <pn> = initial phase n (in degree)
```

- Example: st manipulate Gauss offs m p 100 2 2000 90 3000 3000
- s with additional scaling.
- Syntax:

st manipulate Gauss offs e s <pulDur> <n> <f1> <s1>> <f2> <s2>

```
\begin{array}{l} \mbox{double <pulDur> = pulDur = length of shaped pulse (in us)} \\ \mbox{int <num> = number of offset-frequencies} \\ \mbox{double <fl> = offset frequency 1 (in Hz)} \\ \mbox{double <sl> = scaling factor 1 (in \%)} \\ \mbox{double <fl> = offset frequency 2 (in Hz)} \\ \mbox{double <sl> = scaling factor 2 (in \%)} \\ \mbox{.....} \\ \mbox{.....} \\ \mbox{double <fn> = offset frequency n (in Hz)} \\ \mbox{double <sn> = scaling factor n (in \%)} \end{array}
```

• Example: st manipulate Gauss offs e s 100 2 2000 50 3000 75

The options f, p, and s, respectively, may be combined:

- Example: st manipulate Gauss offs m f s 100 50 75 freqlist
- st manipulate Gauss offs e f p 100 90 180 freqlist

The manipulation command offs reads the shape file from *XWINNMRHOME/exp/ stan/nmr/lists/wave* and writes the manipulated file back to XWINNMRHOME/exp/stan/nmr/lists/wave

3.3.3 command power

modulation according to frequency sweep

• Syntax: st manipulate Gauss power <exponent> double <exponent> = exponential factor

• Example: <u>st manipulate Gauss power 2</u>

3.3.4 command scale

modulation according to frequency sweep

• Syntax: <u>st manipulate Gauss scale <scalfac></u>

double <scalfac> = scaling factor

• Example: st manipulate Gauss scale 0.5

3.3.5 command sinm2

single cosine modulation (+/- offset)

• Syntax: <u>st manipulate Gauss sinm2 <scalfac></u>

int <nPoints> = shape size in number of points double <pulDur> = length of shaped pulse (in us) double <offset> = offset frequency (in Hz)

• Example: st manipulate Gauss sinm2 128 100 3000

3.3.6 command sweep

modulation according to frequency sweep

• Syntax: st manipulate Gauss sweep <scalfac>

int <nPoints> = shape size in number of points double <pulDur> = length of shaped pulse (in us) double <sw> = total sweep-width (in Hz)

• Example: st manipulate Gauss sweep 128 100 5000

3.3.7 command trev

time reversal

trev needs no additional parameters.

• Example: st manipulate HalfGauss trev

3.4 Analyze existing Shape

• Syntax:

st analyze <shape type> <analyze command> <further parameter>

The shape is loaded from XWINNMRHOME/exp/stan/nmr/lists/wave

The following analyzing commands are implemented:

- <u>integr</u> integrate shape and compare to square.
- <u>integr2</u> calculate power level compared to hard pulse.
- <u>calcb1m</u> calculate gammaB1max on resonance.
- <u>calcb1mo</u> calculate gammaB1max off resonance.
- <u>calcpav</u> calculate average power level.
- <u>bandw2</u> calculate bandwidth for excitation.
- <u>bandw2i</u> calculate bandwidth for inversion.
- <u>simProf</u> simulate excitation Profile.

3.4.1 command bandw2

calculate bandwidth for excitation

• Syntax: st analyze Gauss bandw2 <rotation>

double <rotation> = total rotation (in degree)

- Example: st analyze Gauss bandw2 90
- Result: DeltaOmega-DeltaT factor.

3.4.2 command bandw2i

calculate bandwidth for inversion

• Syntax: st analyze Gauss bandw2i <rotation>

double <rotation> = total rotation (in degree)

- Example: st analyze Gauss bandw2i 180
- Result: DeltaOmega-DeltaT factor.

3.4.3 command calcb1m

calculate gammaB1max on resonance

• Syntax: <u>st analyze Gauss bandw2i <rotation></u>

double <pulDurShape> = length of shaped pulse (in us)
double <rotation> = total rotation (in degree)

- Example: st analyze Gauss calcb1m 100 180
- Result:

-maximum gammaB1 (on resonance). -corresponding 90 degree square pulse.

3.4.4 command calcb1mo

calculate gammaB1max off resonance

• Syntax: st analyze Gauss calcb1mo <pulDurShape> <rotation> <offset>

double <pulDurShape> = length of shaped pulse (in us)
double <rotation> = total rotation (in degree)
double <offset> = offset (in Hz

• Example: st analyze Gauss calcb1mo 100 90 3000

 Result: -maximum gammaB1 (on resonance).
 -corresponding 90 degree square pulse.

3.4.5 command calcpav

calculate average power

calcpav needs no additional parameters.

- Example: st analyze Gauss calcpav
- Result:

-mplitude relativ to a square pulse of 100% -corresponding difference in dB. -power relativ to a square pulse of 100%

3.4.6 command integr

integrate shape and compare to square.

integr needs no additional parameters.

- Example: st analyze Gauss integr
- Result:

-integral ratio compared to a square pulse of 100% -corresponding difference in dB.

3.4.7 command integr2

calculate power level compared to hard pulse.

• Syntax: st analyze Gauss integr2 <pulDurShape> <pulDurHard>

double <pulDurShape> = length of shaped pulse (in us)
double <pulDurHard> = length of reference hard pulse (in us)

- Example: st analyze Gauss integr2 100 50
- Result:

-integral ratio compared to a square pulse of 100% -change of power level compared to level of hard pulse in dB.

3.4.8 command simProf

simulate excitation profile

 Syntax: <u>st analyze Gauss prof <from> <to> <rfFielf> <pulDur></u>

double <from> = start point (in Hz) double <to> = end point (in Hz) double <rfField> = rf-field (in Hz) double <pulDur> = length of shaped pulse (in us)

• Example: st analyze Gauss simProf 0 5000 1000 250

The result is written in the following files:

XWINNMRHOME/exp/stan/nmr/lists/wave/Gauss.x XWINNMRHOME/exp/stan/nmr/lists/wave/Gauss.y XWINNMRHOME/exp/stan/nmr/lists/wave/Gauss.z XWINNMRHOME/exp/stan/nmr/lists/wave/Gauss.prof

3.5 Add 2 existing shapes to form a new shape

If the two input files are different in size, the smaller one is taken.

• Example: st add Gauss.1 Gauss.2 Gauss.result

Input files are taken from: XWINNMRHOME/exp/stan/nmr/lists/wave Output file is written back to: XWINNMRHOME/exp/stan/nmr/lists/wave

3.6 Merge 2 existing shapes to form a new shape

If the two input files are different in size, the smaller one is taken.

• Example: st merge Gauss.1 Gauss.2 Gauss.result

Input files are taken from:

XWINNMRHOME/exp/stan/nmr/lists/wave Output file is written back to: XWINNMRHOME/exp/stan/nmr/lists/wave

3.7 The Interactive Display Command stdisp

The command <u>stdisp</u> opens a new window that allows you to work work interactively with the shape tool.

3.7.1 The Shapes Menu

The commands described in the previous chapters may also be invoked from menues. To generate a new shape (cf. Chapter *Generate a new shape*) use the *Shapes*

📥 Shapes
Sinc
Gaussian
Half Gauss
Rectangular
Triangle
Hermite
Hyperbolic Secant
Sine / Cosine
Cossinc
Sine
Sine * Sinc
Sneeze
Gaussian Pulse Cascade 🕨
Burp 🕨
Snob 🕨

Figure 3.1 The Shapes menu

menu (Figure 3.1):

To generate a new Gauss shape, click on in the *Shapes* menu. Now an editor window, (Figure 3.2) opens, which allows you to define parameters for your selected shape. In the case of a Gaussian shape two parameters are required

-size of the shape -truncation level

Click the OK button to generate a gaussian shape with 1000 points and truncation level of 1%. The <u>Apply</u> button starts the calculation of a new shape, but the editor window will stay open for further changes.

	Editor for Shape Parameters					
Parameters for Shape Gauss:						
	Size of Shape:	1000				
	1					
	OK Apply Cancel					

Figure 3.2 Editor for Shape Parameters

3.7.2 The Manipulate Menu

To manipulate a shape (cf. Chapter *Manipulate existing Shape*) on the display, use the *Manipulate* menu: (Figure 3.3)

- Manipulate
offs – Phase Modulation acc. to Offset Freq.
sinm2 – Single Sine Modulation
cosm2 – Single Sine Modulation
sweep – Modulation acc. to Freq. Sweep
power – Power of Amplitude
scale – Scale Amplitude
trev – Time Reversal
region – Calc. Shape from Excitation Region

Figure 3.3 The Manipulate menu

As an example for maniplating shapes the <u>offs</u> command (Phase Modulation according to Offset frequency) will be described. To define the parameters needed by the command, an editor is available (Figure 3.4).

In the first *radiobuttonbox* the alignment for the phase is defined (phase 0 at the beginning, middle or end of the shape).

In the second *radiobuttonbox* you select, whether the reference frequency is taken from the first entry of the frequency list or from O1 of the current data set.

The *checkbuttonbox* provides options for the <u>offs</u> command. If the frequencies should be taken from a frequency list, please use XWIN-NMR to define a frequency list. The shape tool reads the frequency list from the directory:

XWINNMRHOME/exp/stan/nmr/lists/f1

Parameters for Manipulate Command offs:			
Reference Frequency:			
frequencies taken from frequency list with additional phase setting with additional scaling			
Loweth of Dules (in us): 10000			
Length of Pulse (in us): 10000 Name of Frequency List: freqlist			

Figure 3.4 Editor for offs command

3.7.3 The Analyze menu

To analyze a shape (cf. Chapter *Analyze existing shape*) use the *Analyze* menu (Figure 3.5):

The results of the analyzing commands are displayed in a text window. For instance the command <u>integr</u> (Integrate Shape and compare to Square results for a Gaussian shape with 1000 points and 1% truncation in the following window (Figure 3.6).

- Analyze		
bandw2 - Calculate Bandwidth for Exitation Shape		
bandw2i – Calculate Bandwidth for Inversion Shape		
calcb1m – Calculate gammaB1max on Resonance		
calcb1mo – Calculate gammaB1max off Resonance		
calcpav – Calculate average Power Level		
integr – Integrate Shape and compare to Square		
integr2 - Integrate Shape compared to Hard Pulse		
simProf - Simulate Exitation Profile		

Figure 3.5 The Analyze menu

- Results from Analyzing Commands	
integr: Integral Ratio compared to a square Pulse of 100% on Resonamce: corresponding difference in dB:	0.411578 -7.71097
Seen	

Figure 3.6 Result of integr command

3.7.4 File handling

All file related functions of the shape tool can be found in the *File* menu. For reading a shape file from disk, use the <u>Open</u> command. The <u>Save</u> command writes the actually displayed shape including its type name to disk. The command <u>Save As</u>, allows you to specify your own file name.

All file related commands read, or write their files (by default) into the directory:

XWINNMRHOME/exp/stan/nmr/lists/wave

The default directory can be changed using <u>Set Path to Shape Directory</u> command in the *Options* menu.

The commands <u>Open</u> and <u>Save As</u> open a *file selection box* to specify a file name (Figure 3.7).

The default shape directory can also be changed within the *file selection box* by editing the Filter entry. After changing the Filter click the *Filter* button to change to the entered directory.

3.7.5 The Split menu

The shape tool allows you to split the display window in several subwindows. The active windows are surrounded by a blue frame. You can activate one subwindow by clicking on it with the *left mouse button*. To activate all subwindows use the menu entry <u>Select all</u>. To split a window horizontally or vertically, use the corresponding menu entries <u>Split horizontally</u> and <u>Split vertically</u>. To return to single window display use the menu entry <u>Remove</u>.

The splitting of the display window may also be done interactively by dragging the blue frame with the *left mouse button* depressed to any position within the window.

Shape Filename:					
Filter					
winnmr2.	winnmr2.a/exp/stan/nmr/lists/wave/*				
Directorit Files					
¥	/CosSinc /EBurp1 /Gauss /GaussCascadeG3 /GaussCascadeG4 /GaussCascadeQ3 /GaussCascadeQ3 /GaussCascadeQ5				
Selection					
mr2.a/exp/stan/nmr/lists/wave/Gauss					
OK Filter Cancel					

Figure 3.7 File selection box

Chapter 4 Writing Pulse Programs

4.1 Introduction

A pulse program is an ASCII text consisting of a number of lines. Each line may contain one or more pulse program commands which specify actions to be performed by the acquisition hardware and software. You set up a pulse program by means of the XWIN-NMR commands <u>edpul</u> or <u>edcpul</u>. The chapter *The File Menu* describes their syntax, and in which disk directories pulse programs are stored. The XWIN-NMR acquisition commands <u>gs</u>, <u>go</u>, and <u>zg</u> execute the pulse program defined by the acquisition parameter PULPROG (to be set up with <u>eda</u> or <u>pulprog</u>). Pulse program execution is a two-step process: When typing <u>gs</u>, <u>go</u>, or <u>zg</u>, the *pulse program compiler* is invoked and translates the pulse program text into an internal binary form suitable to be executed. During this stage, syntax errors are reported (acquisition will not be started if there are any). If compilation is successful, the compiled pulse program is loaded into the acquisition hardware, and the measurement begins.

This chapter describes the pulse program language for AVANCE type instruments. Many of the features are also valid for AMX/ARX/ASX type spectrometers, but we recommend that you cross-check with the *UXNMR User's Guide 910901* for differences.

Please note:

A few features described in this chapter are not implemented in XWIN-NMR 2.0, but are planned for revision 2.1. These features are printed in grey rather than black letters, e.g. the section 4.9.4 and the bottom of Table 4.13.

4.2 Pulse program library

Routine users of a spectrometer usually do not need to write their own pulse programs. Instead, they can make use of the Bruker pulse program library provided with XWIN-NMR, covering the most useful experiments. The <u>edpul</u> command displays a list of these experiments and allows you to view the pulse program text, which includes a description of each. Please refer to the description of <u>edpul</u> for details. Viewing the library pulse programs requires that the <u>expinstall</u> command was executed by the administrator, which copies the pulse programs suitable for your spectrometer into the appropriate working directory.

When writing your own pulse programs, it is often helpful to start with a Bruker pulse program and apply modifications, rather than beginning from scratch.

4.3 Pulse program display

A graphical representation of a pulse program for Avance type spectrometers may be obtained using the command <u>pulsdisp</u>, which is described in its own section of the XWIN-NMR manual. After writing your own pulse program, <u>pulsdisp</u> will not only check its syntax, but will also allow you to visualize any fine timing detail before you try a real experiment.

4.4 Basic syntax rules

Table 4.1 shows a simple example from the Bruker pulse program library. It is used to demonstrate some of the most important *programming rules*.

- **6.** Pulse programs are *line oriented*. Each line specifies an *action* to be executed by the acquisition hardware or software.
- **7.** A semicolon (;) indicates the beginning of a comment. You can put it anywhere in a line. Following its occurrence, the rest of the line will be treated as a com-

;zgcw30 ;avance-version ;1D sequence with CW decoupling ;using 30 degree flip angle #include <Avance.incl> 1 ze d11 pl14:f2 d11 cw:f2 2 d1 p1*0.33 ph1 go=2 ph31 wr #0 d11 do:f2 exit ph1=0 2 2 0 1 3 3 1 ph31=0 2 2 0 1 3 3 1 ;pl1: f1 channel - power level for pulse (default) ;pl14: f2 channel - power level for cw/hd decoupling ;p1 : f1 channel - 90 degree high power pulse ;d1 : relaxation delay; 1-5 * T1 ;d11: delay for disk I/O [30 msec]

 Table 4.1 Pulse program example

ment.

8. #include <filename> or #include "filename"

This command allows you to use pulse program text, which is stored in another file, in your pulse program. It helps you to keep your pulse program reasonably sized, and to re-use the same code in other pulse programs. By convention, if the filename is given in angle brackets (<>), the file must be stored in the working directory *\$XWINNMRHOME/exp/stan/nmr/lists/pp/*. Alternatively, double quotes (" ") are normally used to specify the entire path name of the file to be included.

9. 1 ze

Any pulse program line may be preceded by a label ("1" in this case). You need

to label only those lines which must be reached by loop or branch commands such as <u>go=label</u>, <u>lo to label times n</u>, <u>goto label</u>. You may also use labels for numbering the lines of a pulse program. Labels need not necessarily be numbers. You can also use an alphanumeric string, followed by a comma, e.g. <u>first-label, ze</u>.

The command \underline{ze} serves the following purpose:

- To reset the scan counter (displayed during acquisition) to 0

- To enable the execution of *dummy scans*. This will cause the pulse program command $\underline{go=label}$ to perform DS dummy scans before accumulating NS data acquisition scans. If you replace \underline{ze} with \underline{zd} , $\underline{go=label}$ will omit the dummy scans

- To reset phases to the beginning of phase programs (phase lists)

10.d11 pl14:f2

Execute a delay whose duration is given by the acquisition parameter D11. Behind any delay command, you may specify further actions to be executed during the delay (provided the duration chosen was sufficiently large). In this example, the power level of channel f2 is switched to the value given by the acquisition parameter PL14.

11.d11 cw:f2

Execute a delay whose duration is given by the acquisition parameter D11, and, at the same time, turn on continuous wave (cw) decoupling on frequency channel f2. Decoupling will remain active until explicitly switched off with the command do:f2. This delay and cw decoupling will begin immediately after the end of the delay specified in item 5 above.

Items 5 and 6 illustrate a general feature of pulse programs: The actions specified in two adjacent lines are executed sequentially. Actions specified within the same line are started simultaneously and execute in parallel.

12.2 d1

Execute a delay whose duration is given by the acquisition parameter D1. This line is preceded by the label "2", which is where the command $\underline{go=2}$ loops back to.

13.p1*0.33 ph1

Execute a pulse on frequency channel f1 whose duration is given by the acquisition parameter P1, multiplied by 0.33. For convenience, P1 is often used to hold the pulse width for a 90 degree flip angle. The command p1*0.33 would
then execute a 30 degree pulse. In general, you may specify the operator * behind (not before!) a pulse or delay command, followed by a floating point number. Please note that the channel f1 does not occur explicitly: If no channel is given, f1 is assumed, and the command p1*0.33 is identical to p1*0.33:f1.

The pulse is executed with a power (amplitude) defined by the acquisition parameter PL1. PL1 is the default power level for channel f1, but you may also use a different parameter. For example, the command <u>pl7:f1</u> would set the channel f1 power from the parameter PL7. You would not specify, however, this command in the same line as <u>p1*0.33</u>, but previously, along with a delay, in order to give the transmitter time to settle before the pulse is executed.

The phase of this pulse is selected according to ph1 with ph1 being the name of a *phase program* or *phase list*, to be specified behind the pulse. It is defined at the end of the pulse program, in our example $ph1=0.2\ 2.0\ 1.3\ 3.1$. The phase of the pulse varies according to the current data acquisition scan. For scan 1, p1 will get the phase 0*90 degrees, for scan 2 2*90, for scan 3 2*90, for scan 4 0*90, etc. After 8 scans, the list is exhausted. With scan 9 the pulse phase is *cycled*, i.e. it restarts with the first phase in the list. Phase cycling is a method of artefact suppression in the spectrum to be acquired. Cycling the pulse phase requires the receiver phase to be cycled accordingly to ensure that coherent signals of subsequent scans will accumulate rather than cancel. This is achieved by the receiver phase program ph31 in our example.

14.go=2 ph31

Execute 1 data acquisition scan, then loop to the pulse program line with label "2" until NS scans have been accumulated, NS being an acquisition parameter. The NS scans are preceded by DS *dummy scans* (because the command <u>ze</u> is used at the beginning of the pulse sequency rather than <u>zd</u>). A *dummy scan* does not acquire any data, but requires the same time (given by the acquisition parameter AQ) as a true scan. Dummy scans are used to put the spin system of the sample into a steady state before acquisition starts.

The receiver phase is changed after each scan as described above for the pulse phase. Phase cycling is based on the phase program ph31. Phase cycling is also employed during the execution of dummy scans. DS and NS should therefore be a multiple of the number of phases in the list.

The <u>go=label</u> command executes a delay, the so-called *pre-scan delay* to avoid pulse feed through before it starts digitizing the NMR signal. During this time

the receiver gate is opened, and for certain instruments the frequency is switched from *transmit* to *receive*. You can read more about DE and its components DE1 and DE2 in the chapter *The Acquire Menu*. Normally, you can accept the default DE value suggested by the program. The total time the <u>go=label</u> command requires to execute a scan is DE+AQ+3 milliseconds. The last duration is required for internal reasons, namely the preparation of the next scan. The 3 msec are valid for Avance type instruments. Older instruments require 6 milliseconds.

15.wr #0

Writes the accumulated data as file *fid* into the EXPNO directory of the current data set. Please note that the way the *zgcw30* pulse program is built data are not stored on disk before all NS scans have been accumulated. However, while acquisition is in progress, you can force the program at any time to store the data acquired so far, by entering the command <u>tr</u> on the keyboard (or by calling it from the *Acquire* menu). You may process and plot this data while the acquisition continues. If you want to protect your data against power failures during long term experiments, we recommend that you write the data on disk in regular intervals, for example every 1000 scans. To accomplish this, you would set NS=1000, and add the line <u>lo to 1 times 30</u> before the <u>exit</u> command. The pulse program would then accumulate 30.000 scans totally, but store the result every 1000 scans. Please note that the loop must branch to the <u>ze</u> command. The reason for this is that <u>wr #0</u> adds the last acquired data to the data already present in the file. The real time fid display will only show the data currently present in the acquisition processor's memory.

16.exit

Signals the end of the pulse program.

4.5 Pulse generation commands

Table 4.2 shows the available types of commands for the generation of high frequency pulses.

A high frequency pulse is described by its

- duration (= pulse width)
- frequency
- phase

<u>p0, p1, , p31</u>	Generate a pulse whose duration is taken from the acquisition parameter P0,, P31, respectively.
<u>3.5up</u> , <u>10mp</u> , <u>0.1sp</u>	Generate a pulse of fixed length: up = a microsecond pulse (mp = millisec, sp = sec).
P135, <u>30degree1Hpulse</u>	Generate a pulse whose name is defined by means of a <u>define pulse</u> statement, and whose duration is defined by an expres- sion.
<u>vp</u>	Generate a pulse whose duration is looked up in a pulse list.
<u>pulse(flipangle</u> , auto, <i>duration</i>) <u>pulse(flipangle</u> , <i>power</i> , auto)	Generate a pulse causing the specfied flip angle.

Table 4.2 Pulse generation commands

- power (= amplitude) and shape
- flip angle

The following paragraphs will describe these items.

4.5.1 Pulse duration

The pulse duration is selected according to the name of the pulse command.

4.5.1.1 <u>p0-p31</u>

The command

p0

would execute a pulse of width P0, where P0 is an acquisition parameter. It is set from \underline{eda} , or by typing $\underline{p0}$ on the keyboard. Likewise, the command

p1

would execute a pulse of width P1.

4.5.1.2 Fixed length pulses

The command

10mp

would execute a pulse of width 10 milliseconds (called a *fixed pulse* because its duration cannot be further manipulated, cf. below). The duration must be followed by *up*, *mp*, or *sp*. These units indicate microseconds, milliseconds, and seconds, respectively. If you omit the terminating "p", a delay is executed rather than a pulse.

4.5.1.3 User defined pulses

The command

30degree1Hpulse

would execute a pulse whose name is defined by the user, and whose duration is determined by an arithmetic expression. For example, the line

define pulse 30degree1Hpulse

would define 30degree1Hpulse to be a pulse command, and the line

"30degree1Hpulse=p1*0.33"

(which must be quoted using double-quote characters (")) would constitute the expression for its duration.

The *define* line must be placed somewhere at the beginning of the pulse program before the beginning of the actual pulse sequence.

Note: The defining expression of a user defined pulse must occur before the actual start of the pulse sequence. It is evaluated at compile time of the pulse program, not at run time.

4.5.1.4 Variable list pulses

The command

vp

executes a pulse whose duration is given by the current value of a *pulse list*. A pulse list is a text file containing one pulse duration per line. Pulse lists are set up with the command <u>edlist</u> (described in the chapter *The File Menu*). The command \underline{vp} uses the list file given by the acquisition parameter VPLIST. When the pulse program begins, the first duration in the list is valid. The pulse program command \underline{ivp} can be used to advance the list pointer to the next duration. If the end of the list is encountered, the pointer is reset to the beginning. The command \underline{ivp} must be specified behind a delay for internal reasons. Examples:

d1 ivp 0.1u ivp.

The delay length is not important, any small or large duration is legal.

It is also possible to calculate the list position by means of an equation. Example:

"vpidx=5" vp

The command \underline{vp} will execute a pulse whose duration is selected from position 5 of the pulse list. At the right of the equal sign any dimensionless expression is allowed which may contain parameters from Table 4.3.

4.5.1.5 Manipulating pulse durations: The operator *

A pulse duration can be manipulated by the operator "*" if appended to the pulse command. Examples of legal commands:

p1*1.5 30degree1Hpulse*3.33 p3*oneThird vp*3

(10mp*0.33 would be illegal, since 10mp is a fixed pulse). The operator must be placed behind the pulse command, not before. *oneThird* is the name of a macro which was defined at the beginning of the pulse program by means of a *#define* statement, e.g. *#define* oneThird 0.33.

4.5.1.6 Manipulating pulse durations: Changing <u>p0-p31</u> by a constant value

Each pulse command po-p31 has been assigned an acquisition parameter INPO-

INP31, containing a duration (in microseconds). The pulse program commands <u>ipu0-ipu31</u> add INP0-INP31 to the current value of <u>p0-p31</u>, respectively. Likewise, <u>dpu0-dpu31</u> subtract INP0-INP31 from the current value of <u>p0-p31</u>. The commands <u>rpu0-rpu31</u> reset <u>p0-p31</u> to their original value, i.e. to the values of the parameters P0-P31 set up with <u>eda</u>. For internal reasons, the commands presented in this paragraph must be specified behind a delay whose length is of no importance. Examples:

d1 ipu3 0.1u dpu0 d1 rpu0

4.5.1.7 Manipulating pulse durations: Redefining p0-p31 via an expression

The duration of the pulses $\underline{p0}$ - $\underline{p31}$ is normally given by the parameters P0-P31. However, you may overwrite these values in the pulse program using an expression. The following examples show some of the possibilities:

"p13=3s + aq - dw*10" "p13=p13 + (p1*3.5 + d2/5.7)*td"

The result of such an expression must have a time dimension. You may therefore include acquisition parameters such as pulses, fixed pulses, delays, fixed delays, acquisition time AQ, dwell time DW within the expression, but also parameters without a dimension such as the time domain size TD. The complete list is shown in Table 4.3. An expression must be quoted using double quote characters (""). It can be placed anywhere in the pulse program, but must occur before the line containing the corresponding pulse command (which would be p13 in our example). Please note that the second expression in the example above assigns a new value to p13 each time the expression is encountered, e.g. if contained in a pulse program loop.

Note that expressions in labelled lines are not supported. Give the labelled line a small duration and put the expression into the following line.

Expressions do not introduce a delay in the pulse program. Pre-evaluation is applied before the pulse program is started, and the result is stored in the available buffer memory to be accessed at run time. At run time, pre-evaluation is performed during the cycle time of the loops in which the commands are embedded. If loops are executed too fast in a particular pulse program, a run time message is printed.

```
d0-d31 [sec]

p0-p31 [microsec]

l0-l31 (loop counters)

in0-in31[sec]

inp0-inp31 [microsec]

aq [sec]

dw [microsec]

dwov [microsec]

de1, de2, de3 [microsec]

vd [sec]

vp [microsec]

nbl, ds, ns, nsdone, td, td1, td2

decim

cpdtim1-cpdtim8 [sec]

cnst0-cnst31
```

Table 4.3 Parameters which may be included in expressions

4.5.1.8 Manipulating the durations of user defined pulses

Users may define their own pulse commands using a statement such as

define pulse 30degree1Hpulse

at the beginning of the pulse program. The pulse would be executed by the command

30degree1Hpulse.

The duration of the pulse can be defined by a pulse program line such as

"30degree1Hpulse=p1*0.33".

For such an expression the same rules apply, that are valid for the manipulation of $\underline{p0}-\underline{p31}$, described in the previous section.

Note: The defining expression of a user defined pulse must occur before the actual start of the pulse sequence. It is evaluated at compile time of the pulse program,

not at run time.

4.5.2 Pulse frequency

4.5.2.1 Frequency channels

The RF frequency of a pulse is selected via the spectrometer channel number f1, ..., f8 (the actual number of channels in your instrument depends on its type and equipment). Executing a pulse on a particular channel means executing it with the frequency defined for this channel. The commands

p1:f2 p2*0.33:f2 30degree1Hpulse*3.33:f2 vp:f2

would execute pulses with the duration P1 (P2*0.33, 30degree1Hpulse*3.33, VPLIST, respectively) on channel f2. The pulse frequency is that assigned to f2, namely SFO2, which is an acquisition parameter in MHz units. If the channel is omitted, f1 is assumed as default. Generally, the frequencies of the channels f1-f8 are given by the parameters SFO1-SFO8 (cf. SFO1, NUCLEI, and edasp for more information about defining frequencies for a particular channel). These parameters are loaded into the synthesizer(s) before the pulse program starts, rather than at the time a pulse is executed. This gives the hardware time to stabilize before the experiment begins.

4.5.2.2 Changing the frequency

It is possible to change the frequency of a channel within a pulse program. XWIN-NMR provides the commands $\underline{fq1}$ - $\underline{fq8}$ for this purpose. They refer to frequency lists from where the new frequency will be taken. A frequency list is a text file whose lines are frequency values (cf. command <u>edlist</u> on the exact list format, and how to set up a list). For example, the command

d1 fq2:f3

or

d1 fq=fq2:f3

uses the frequency list whose file name is given by the acquisition parameter

FQ2LIST (<u>fq1</u> would use FQ1LIST, etc.). You may set FQ1LIST etc. from the <u>eda</u> command, and you can modify a selected list with <u>edlist</u>. The example would set the frequency of channel 3 (:f3) by taking the *current* value from the list defined by FQ2LIST. The *current* value is the first value in the list when you start the pulse program and <u>fq2</u> is executed the first time. Next time <u>fq2</u> is ecountered (e.g. because it occurs several times in the pulse program, or because it is contained in a loop) the current value will be the next one in the list, etc. At the end of the list, the current value will be reset to the first list value. In general, <u>fq1-fq8</u> not only set a frequency, but also increment the list pointer to the next frequency in the list. The <u>fq1-fq8</u> commands must be written behind a delay. The frequency change occurs at the beginning of this delay, which must not be shorter than 2 microseconds.

The list can optionally contain a frequency offset in MHz. If this is the case, the frequency list values in Hz are added to this offset. If not, the list values are added to the channel frequency (SFO1 for F1, SFO2 for F2, etc.).

The frequency can also be set from the parameters CNST0-31 or from any number:

d1 fq=cnst20 d1 fq=3000	; SFOn [MHz] + cnst20 [Hz]		
	; SFOn [MHz] + 3000 Hz		

These two commands set the frequency as specified in CNST20 or directly in the pulse program.

4.5.3 Pulse phase

4.5.3.1 Phase programs: Definition

Pulse phases are relative phases with respect to the reference phase for signal detection. A phase must be specified behind a pulse in form of a *phase program*. For example, the commands

10mp:f1 ph3 p2*0.33:f2 ph4 30degree1Hpulse*3.33:f3 ph5 vp:f4 ph6

would execute pulses on the channels f1, f2, and f3, f4, repectively. The channel frequencies would be SFO1, SFO2, SFO3, and SFO4. The channel phases would

be set according to the current value of the *phase programs* ph3, ph4, ph5, and ph6, respectively. If a phase program is omitted from a pulse command, the pulse will have the last phase assigned to the channel where the pulse is executed (the default phase is zero when a pulse program is started and no phase was defined).

The four examples above can also be written in the following form:

```
(10mp ph3):f1
(p2*0.33 ph4):f2
(30degree1Hpulse*3.33 ph5):f3
(vp ph6):f4
```

This form expresses more clearly that a phase is a property of a spectrometer channel.

4.5.3.2 Phase programs: Syntax

A phase program may be specified according to the following examples:

(1) $ph1 = 0 \ 0 \ 1 \ 1 \ 2 \ 2 \ 3 \ 3$ (2) $ph1 = (5) \ 0 \ 3 \ 2 \ 4 \ 1$ (3) $ph1 = \{0\}^*4 \ \{2\}^*4$ (4) $ph1 = \{0 \ 2\}^{n1}$ (5) $ph1 = \{0 \ 2\}^{n1}^{n2}^{n3}$ (6) $ph1 = \{1 \ 3\}^{n1}^{n2}^{n2}^{n2}$ (7) $ph1 = \{\{0 \ 2\}^*2\}^{n1}^{n2}$ (8) $ph1 = \{\{0\}^*2\}^{n3}^{n1}^{n2}^{n2}$ (9) $ph1 = (5) \ \{1 \ 2\}^*2^{n1}$ (10) $ph1 = ph2^*2 + ph3$

A phase program may contain an arbitrary number of phases. The list of phases in a phase program can be split over several lines.

In (1), the phases contained in the program are given in units of 90 degrees. The actual phase values would therefore be 0, 0, 90, 90, 180, 180, 270, 270.

In (2), the phases are given in units of 360/5 degrees, corresponding to the actual phase values 0*72, 3*72, 2*72, 4*72, 1*72 = 0, 216, 144, 288, 72 degrees. The divisor, to be specified in parentheses () in front of the phase list, may be as large as 65536 (corresponding to 16 bits), resulting in a digital phase resolution of 360/65536, which is better than 0.006 degrees.

In (3) - (9), the operators "*" and "^" are introduced, which allow you to write lengthy phase programs in a compact form. For phase programs with less than 16 phases, the explicit forms (1) and (2) are usually easier to read. The operator "*n" (with n = 2, 3, ...) must be specified behind a list of phases enclosed in braces { }. It repeats the contents of the braces (n-1) times. The operator "^m" (with n = 1, 2, 3, ...) must be specified behind a list of phases enclosed in braces { }, or behind a previous "^m" or "*" operator. Each "^m" repeats what is in the braces exactly one time. In addition, the repeated phase list will be incremented by m (module 4). The following lines display the phase programs (3) - (9) in their explicit form, after applying these rules.

(3') ph1 = 0 0 0 0 2 2 2 2
(4') ph1 = 0 2 1 3
(5') ph1 = 0 2 1 3 2 0 3 1
(6') ph1 = 1 3 2 0 3 1 1 3
(7') ph1 = 0 2 0 2 1 3 1 3 2 0 2 0
(8') ph1 = 0 0 2 2 3 3 1 1 2 2 0 0 1 1 3 3
(9') ph1 = (5) 1 2 1 2 2 0

In (10), phase programs are combined by multiplication with an integer constant, and by addition. This is illustrated by the following example. Let

ph2 = 0 2 1 3 ph3 = 1 1 1 1 3 3 3 3

We want to calculate ph5 = ph2*2 + ph3. At first, we build ph2*2:

 $ph2*2 = 0\ 0\ 2\ 2$

We have to extend ph2 to the same size as ph3 before we can evaluate the equation:

 $ph2 = 0\ 0\ 2\ 2\ 0\ 0\ 2\ 2$ $ph3 = 1\ 1\ 1\ 1\ 3\ 3\ 3$

Then we can calculate the result:

ph1 = 1 1 3 3 3 3 1 1

4.5.3.3 Position of phase programs

Phase programs must be specified at the end of the pulse program (cf. the pulse program example in Table 4.1 at the beginning of this chapter). Any pulse program

may contain up to 32 phase programs (ph0 to ph31).

4.5.3.4 Phase cycling

When a pulse program begins to execute, the first phase of each phase program will become valid as soon as the pulse sequence is followed by data acquisition scans (rather than *dummy* scans). The next phase will become valid with the next scan or dummy scan (cf. the section *Commands to start data acquisition* for details). If the end of a phase program is encountered, it will be repeated from the beginning (*phase cycling*).

4.5.3.5 Phase pointer increment

It is possible to explicitly switch to the next phase in a phase program. Consider the two commands

p1:f2 ph8^ p2:f2 ph8

p1 is executed with the currently active phase of ph8, then p2 is executed with the next phase in p8. If you had omitted the carret (^) sign in the first line, p2 would have been executed with the same phase as p1. A carret sign appended to a phase program will increment the phase pointer to the next phase in the list. This phase will become valid with the next pulse program command including this phase program (this can be the same command if included in a loop).

The following two commands have the same effect as the previous example:

p1:f2 ph8 ipp8 p2:f2 ph8

In this case, the command <u>ipp8</u> is used to increment the pointer in the phase program ph8. Please note that <u>ipp8</u> is specified on the same line as p1 and therefore does not introduce an extra delay between p1 and p2. In general, the increment commands <u>ipp0-ipp31</u> are provided for the phase programs ph0-ph31. These commands can also be specified with a delay. For example,

d1 ipp7

would advance the pointer to the next phase in ph7.

4.5.3.6 Adding a constant to a phase program

You may change all phases in a phase program by a constant amount without having to modify the phase program itself. Each phase program ph0-ph31 has a constant assigned, PHCOR0-PHCOR31 (accessable from <u>eda</u>, or from the keyboard). The phases of the pulse

(p1 ph8:r):f2

would be 2, 92, 182, 272 degrees if $ph8 = 0 \ 1 \ 2 \ 3$ and PHCOR8=2 degrees. Without the <u>ir</u> option, the phase cycle of p1 would be 0, 90, 180, 270. The ir option can be combined with the ^ sign, e.g.

(p1 ph8^:r):f2.

4.5.3.7 Phase program arithmetic

Each of the phase programs ph0-ph31 has 3 associated commands

<u>ip0-ip31, dp0-dp31, rp0-rp31</u>.

They can also be used with an integer multiplier n:

 $\underline{ip0}^{*n} - \underline{ip31}^{*n}, \underline{dp0}^{*n} - \underline{dp31}^{*n}.$

Assume we had two phase programs ph3 = 0 2 2 0 and ph4 = (5) 0 1 2 3. The pulse program command

20u ip3

would increment all phases of ph3 by 90 degrees. The next time that ph3 is encountered in the pulse program, its phase cycle will be changed to "1 3 3 1". Likewise, the pulse program command

20u ip4

would increment all phases of ph4 by 360/5 degrees. The next time that ph4 is encountered in the pulse program, its phase cycle will be changed to "(5) 1 2 3 4".

The commands <u>dp0-dp31</u> are inverse to <u>ip0-ip31</u> and decrement all phases of the associated phase program. The commands <u>rp0-rp31</u> reset all phases of a phase program to their original values, i.e. to the values they had before applying <u>ip0-ip31</u> or <u>dp0-dp31</u> the first time.

The commands

6u ip3*2 7.5u dp4*2

would increment ph3 by 2*90=180 degrees and decrement ph4 by 2*360/5=144 degrees.

An increment/decrement phase program command must always be specified behind a delay, which must be long enough for the increment/decrement to be computed. The required time depends on the number of phases in the phase program, and amounts to 1.5 microseconds per phase and channel.

4.5.3.8 Phase presetting

A pulse program command such as

(p1 ph8):f2

suggests that phase switching to the current phase of the specified phase program is executed at the same time as the pulse begins. Depending on the hardware, however, the phase may take a certain amount of time to become stable. For this reason, XWIN-NMR allows you to instruct the pulse program to set the new phase somewhat earlier. Eight parameters PHASPR1-PHASPR8 are provided (to be set from the <u>edscon</u> parameter editing command), one for each spectrometer channel. Their default value is 3 microseconds. Therefore, the phase for any pulse will be set 3 microseconds before the pulse begins. If you would change PHASPR3 to 4 microseconds, the phase of all pulses executed on channel 3 would be set 4 microseconds before the pulses begin.

Consider the following section of a pulse program, which executes two pulses in a row (i.e. without a delay in between) on the same channel, but with different phases:

(10up ph8):f2 (20up ph9):f2

Assume that PHASPR2 = 3 microseconds. Phase setting for the second pulse would be initiated before the first pulse is finished, namely 7 microseconds after its beginning.

4.5.3.9 Phase setting independent of pulse commands

XWIN-NMR allows you set the phase for a particular spectrometer channel independent of pulse commands. For this purpose, you must specify a phase program behind a delay. Examples:

(d1 ph1):f3 (0.1u ph3):f2

4.5.3.10 The 4-phase modulator

The phases for the spectrometer channels f1-f8, as presented so far, are realized in the FCU (Frequency Control Unit) hardware of the acquisition system during digital frequency generation. For special applications such as certain solid state experiments an accessory is available called 4-phase modulator. This device permanently provides four phases of the current frequency: 0, 90, 180, and 270 degrees. It therefore allows for faster switching between these phases. The following example shows the phase program syntax for the 4-phase modulator:

ph1 = +x + y - x - y

This is equivalent to the following conventional phase program:

ph1 = 0 90 180 270

Whenever the special syntax is used in a phase program, the phases will be selected from the 4-phase modulator rather than from the FCU's digital frequency generator.

The 4-phase modulator has a number of internal adjustment parameters which can be set with the command <u>ed4ph</u>.

If the spectrometer is not equipped with a HPCU the modulator must be defined in the file *\$xwinnmehome/conf/instr/<intrumentName>/hardware_list.* <u>cf</u> is required afterwards. The unit may be connected to *tty10* or *tty20* on the CCU.

4.5.4 Power and shape

4.5.4.1 Rectangular pulses

A rectangular pulse has the same constant power while it is executing. The power

is, like the pulse frequency, set via the spectrometer channel number f1, ..., f8. There are acquisition parameters PL0, ..., PL31 which hold power values (in dB). You set the amplitude for a particular channel from these parameters using the pulse program commands <u>pl0-pl31</u>. For example,

d1 pl5:f2

would set the transmitter power for channel f2 to the value given by PL5. Any pulse executed on this channel would therefore get the frequency SFO2 and the power PL5. The <u>pl0-pl31</u> commands must be written behind a delay. The power change occurs within this delay, which must not be shorter than 2 microseconds.

The default power levels of channel f1-f8 are set according to PL0-PL8.

4.5.4.2 Shaped pulses

A *shaped* pulse changes its power (and possibly phase) in regular time intervals while it is executing. The pulse *shape* is a sequence of numbers (stored in a file, see below) describing the power and phase values which are active during each time interval. The interval length is calculated by the program by dividing the pulse duration by the number of power values in the shape file. Should it become less than 200 nsec, an error message is displayed which tells you how much the pulse duration must be increased.

The next 3 examples would generate shaped pulses.

```
(10mp:sp2 ph7):f1
(p1:sp1 ph8):f2
(30degree1Hpulse*3.33:sp3 ph9):f3
(vp:sp4 ph10):f4 Illegal! Shaped pulses with <u>vp</u> are not supported.
```

The pulse durations would be 10 milliseconds, P1, and 30degree1Hpulse*3.33, respectively. The pulses would be executed on the frequency channels f1, f2, and f3 (i.e. the pulse frequencies would be SFO1, SFO2, and SFO3), respectively. The pulse shape characteristics would be described by the entries 2, 1, and 3 (corresponding to :sp2, :sp1, and :sp3) of the *shaped pulse parameter table*. This table is displayed when you click on the SPO7 label within <u>eda</u>. The table has 16 entries with index 0-15. You may use :sp0 - :sp15 in a shaped pulse command to refer to entry 0-15. The examples indicate that a phase program may be appended to a shaped pulse just as to a rectangular pulse. The current phase of the phase program is added to the phase of each component of the shaped pulse.

Each entry of the shaped pulse parameter table has 3 parameters assigned: A *power value*, an *offset*, and a *file name*.

File name

The name of a shape file. A shape file can be generated by means of the programs *shape* and *xshape*, which are part of the XWIN-NMR release. You must start these programs from the Unix level. You can also generate your own shape files in ASCII format and convert them to the binary format required by XWIN-NMR using the *shape* program. The ASCII and binary file formats are described in the Chapter *File Formats*.

The specified file must be located in the directory

\$XWINNMRHOME/exp/stan/nmr/lists/wave/.

Offset frequency

The offset frequency allows you to shift the frequency of the shaped pulse by a certain amount (in Hertz). This shift is internally realized by applying phase changes during the shaped pulse's time intervals. This method was chosen to maintain phase coherency of the frequency.

Power value [dB]

This is the absolute power value which is assigned to the 100% relative power value in a shape file. If the shape file contains a relative power less than 100, its absolute power (including that of all others in the file) will be reduced accordingly.

You can access the table entries not only from <u>eda</u>, but also from the keyboard. For example, the command <u>spnam5</u> would display the file name corresponding to table entry 5, <u>spoffs2</u> the frequency offset of entry 2, and <u>sp15</u> the power value of entry 15.

4.5.4.3 Shaped pulse presetting

Please read the description of the parameters SHAPPR[1-8] in the chapter *The Acquire Menu* to ensure that shaped pulses are executed correctly.

4.5.5 Generating pulses causing a desired flip angle

The pulse generation commands described so far do not allow one to specify a flip angle. The flip angle is implicitly defined via pulse duration and power. In fact, *flip angle, power,* and *duration* are not independent of each other: Once two of them

are known, the third one can be calculated. XWIN-NMR provides the possibility of storing calibrated 90 degree pulses (using the pulse widths and the corresponding power levels) for different probe heads and solvents (cf. commands <u>prosol</u> and <u>solvloop</u>). Based on these values, the command <u>pulse(*flipangle, power, duration*)</u> generates a pulse of the desired flip angle, provided the user specifies the flip angle and one of the two other arguments, and sets the unknown argument to *auto*. Note that this feature is not yet available in XWIN-NMR 1.0 and 1.1.

Example 1:

pulse(90 deg, 0 dB, auto):f1 ph1

Generate a 90 degree pulse using a power level of 0 dB. The program calculates the required pulse width and executes the pulse on channel f1 using the phase program ph1. Please note that a space character must be specified in front of the deg and db units.

Example 2:

pulse(30 deg, pl2, auto):f2 ph1

Generate a 30 degree pulse using the power given by the acquisition parameter PL2. The program calculates the required pulse width and executes the pulse on channel f2 using the phase program ph1. In general, you may use pl0-pl31 for the second argument, referring to the parameters PL0-PL31.

Example 3:

pulse(45 deg, auto, 5u):f2 ph1

Generate a 45 degree pulse using a pulse width of 5 microseconds. The program calculates the required power and executes the pulse on channel f2 using the phase program ph1. You may append 'u', 'm', or 's' to the number in the third argument to designate microseconds, milliseconds, or seconds.

Example 4:

pulse(45 deg, auto, p1):f2 ph1

Generate a 45 degree pulse using a pulse width defined by the acquisition parameter P1. The program calculates the required power and executes the pulse on channel f2 using the phase program ph1. Please note that you may specify P0-P31 as the third argument, but no pulses defined by means of the

define pulse statement.

Example 5:

pulse(90 deg, auto, 10m):sp1:f1 ph1

Generate a 90 degree shaped pulse using a pulse width of 10 milliseconds. The program calculates the required power and executes the pulse on channel f1 using the phase program ph1. *Please note:* Power calculation uses the same formula for rectangular and shaped pulses. To account for the deviation from a rectangular pulse, a power correction value must be specified for shaped pulses. The program expects this number (in db) in the corresponding SP entry of the shaped pulse parameter table. This example (:sp1), would fetch the correction value from entry 1 (you could also enter the command <u>sp1</u> on the keyboard to set the value).

Example 6:

pulse(auto, pl2, p2):f2 ph4

Generate a pulse using a width defined by the acquisition parameter P2, and a power defined by PL2. This command is equivalent to

p2:f2 ph4

except that the power PL2 is set and the flip angle can be inspected by the command <u>ased</u>.

4.6 Delay generation commands

Table 4.4 shows the available types of commands for the generation of delays. The duration of a delay is selected according to the name of the delay command.

4.6.1 <u>d0-d31</u>

The command

d0

would execute a delay of width D0, where D0 is an acquisition parameter. It is set from \underline{eda} , or by typing $\underline{d0}$ on the keyboard. Likewise, the command

<u>d0, d1, , d31</u>	Generate a delay whose duration is taken from the acquisition parameter D0,, D31, respectively.
<u>d0:r</u> , <u>d31:r</u>	Generate a delay whose duration is taken from the acquisition parameter D0,, D31 and which is randomly varied according to the percentage specified in the acquisition parameter V9
<u>3.5u, 10m, 0.1s</u>	Generate a delay of fixed length: $u = a$ microsecond delay (m = millisec, s = sec).
compensationTime	Generate a delay whose name is defined by means of a <u>define delay</u> statement, and whose duration is defined by an expression.
<u>vd</u>	Generate a delay whose duration is looked up in a delay list.
<u>de1, de2, de3</u>	Generate a delay of length DE1, DE2-DE1, DE-DE2, respectively.
<u>dw, dwov</u>	Generate a delay of length DW, DWOV.
aq	Generate a delay of length AQ.

 Table 4.4 Delay generation commands

d1

would execute a delay of width D1.

4.6.2 Random delays

The command

d0:r

would execute a delay of width D0 which is varied randomly within a percentage as specified in the acquisition parameter V9. It is set from <u>eda</u>, or by typing $\underline{v9}$ on the keyboard.

Please note that during gs the randomization is disabled.

4.6.3 Fixed length delays

The command

10m

would execute a delay of 10 milliseconds (called a *fixed delay* because its duration cannot be further manipulated, cf. below). The duration must be followed by u, m, or s. These units indicate microseconds, milliseconds, and seconds, respectively.

4.6.4 User defined delays

For example, the line

define delay compensationTime

would define compensationTime to be a delay command, and the line

"compensationTime=d1*0.33".

(which must be quoted using double-quote characters (")) would constitute the expression for its duration.

Then, the command

compensationTime

would execute a delay whose name is defined by the user, and whose duration is determined by an arithmetic expression. The *define* line must be placed somewhere at the beginning of the pulse program before the beginning of the actual pulse sequence.

Note: The defining expression of a user defined delay must occur before the actual start of the pulse sequence. It is evaluated at compile time of the pulse program, not at run time.

4.6.5 Variable list delays

The command

vd

executes a delay whose duration is given by the current value of a *delay list*. A delay list is a text file containing one delay per line. Delay lists are set up with the

command <u>edlist</u> (described in the chapter *The File Menu*). The command <u>vd</u> uses the list file given by the acquisition parameter VDLIST. When the pulse program is started, the first duration in the list is valid. The pulse program command <u>ivd</u> must be used to advance the list pointer to the next duration. If the end of the list is encountered, the pointer is reset to the beginning. The command <u>ivd</u> must be specified behind a delay for internal reasons. Examples:

```
d1 ivd
0.1u ivd
```

The delay length is not important, any small or large duration is legal.

It is also possible to calculate the list position by means of an equation. Example:

```
"vdidx=5"
vd
```

 \underline{vd} will execute a delay whose duration is selected from position 5 of the delay list. At the right of the equal sign any dimensionless expression is allowed which may contain parameters from Table 4.3.

4.6.6 Special purpose delays

These are the delay commands <u>de1</u>, <u>de2</u>, <u>de3</u>, <u>dw</u>, and <u>aq</u> of Table 4.4. They are provided to facilitate the construction of pulse programs which must use the command <u>adc</u> (rather than <u>go=*label*</u>) to start data acquisition.

4.6.7 Manipulating delays: The operator *

A delay can be manipulated by the operator "*" if appended to the delay command. Examples of legal commands:

```
d1*1.5
compensationTime*3.33
d3*oneThird
vd*3
```

(10m*0.33 would be illegal, since 10m is a fixed delay). The operator must be placed behind the delay command, not before. *oneThird* is the name of a macro which was defined at the beginning of the pulse program by means of a #define statement, e.g. #define oneThird 0.33.

4.6.8 Manipulating delays: Changing <u>d0-d31</u> by a constant value

Each delay command $\underline{d0}$ - $\underline{d31}$ has assigned an acquisition parameter IN0-IN31, containing a duration (in seconds). The pulse program commands $\underline{id0}$ - $\underline{id31}$ add INP0-INP31 to the current value of $\underline{d0}$ - $\underline{d31}$, respectively. Likewise, $\underline{dd0}$ - $\underline{dd31}$ subtract INP0-INP31 from the current value of $\underline{d0}$ - $\underline{d31}$. The commands $\underline{rd0}$ - $\underline{rd31}$ reset $\underline{d0}$ - $\underline{d31}$ to their original value, i.e. to the values of the parameters D0-D31 set up with \underline{eda} . For internal reasons, the commands presented in this paragraph must be specified behind a delay whose length is of no importance. Examples:

d1 id3 0.1u dd0 d1 rd0

In Bruker pulse programs, D0 and D10 are used as incrementable delays for 2D and 3D experiments, IN0 and IN10 are the respective increments which are used to calculate the sweep widths SW(F1) and SW(F2), respectively (cf. IN0, IN10 in the chapter *The Acquire Menu* for more details).

4.6.9 Manipulating delays: Redefining do-d31

The duration of the $\underline{d0}$ - $\underline{d31}$ commands is normally given by the parameters D0-D31. However, you may overwrite these values in the pulse program using an expression in C language systex. The following examples show some of the possibilities:

"d13=3s + aq - dw*10" "d13=d13 + (p1*3.5 + d2/5.7)*td"

The result of such an expression must have a time dimension. You may therefore include acquisition parameters such as pulses, fixed pulses, delays, fixed delays, acquisition time AQ, dwell time DW etc. within the expression, but also parameters without a dimension such as the time domain size TD. The complete list is shown in Table 4.3. An expression must be double-quoted (""). It can be placed anywhere in the pulse program, but must occur before the line containing the corresponding delay command (which would be d13 in our example). Please note that the second expression in the example above assigns a new value to d13 each time the expression is encountered, e.g. if contained in a pulse program loop.

4.6.10 Manipulating the durations of user defined delays

Users may define their own delay commands using a statement such as

define delay compensationTime

at the beginning of the pulse program. The delay would be executed with the command

compensationTime.

The delay length would be defined by a pulse program line such as

"compensationTime=d1*0.33".

For such an expression the same rules apply as for the manipulation of d0-d31, described in the previous section.

Note: The defining expression of a user defined delay must occur before the actual start of the pulse sequence. It is evaluated at compile time of the pulse program, not at run time.

4.7 Simultaneous pulses and delays

4.7.1 Rules

The following rules are valid in pulse programs:

- 1. Pulses and delays specified on subsequent lines are executed sequentially.
- **2.** Pulses and delays which are specified on the same line, and which are enclosed in the same set of parentheses, are executed sequentially.
- **3.** Pulses and delays which are specified on the same line, and which are enclosed in different sets of parentheses, are executed simultaneously. The first item within a set of parentheses is started at the same time as the first item in the other parentheses. You may specify an arbitrary number of sets of parentheses on a line.

4.7.2 Examples

4.7.2.1 Rule 1

The pulse program section

(p1 ph1):f1 100u (p2 ph2):f2

would execute a pulse on channel f1, followed by a delay, followed by a pulse on channel f2 (Figure 4.1).



Figure 4.1 Rules 1 and 2: An example

4.7.2.2 Rule 2

The pulse program section

(p1 ph1 100u):f1 (p2 ph2):f2

would, as shown in the example above, execute a pulse on channel f1, followed by a delay, followed by a pulse on channel f2 (Figure 4.1).

4.7.2.3 Rule 3

The pulse program section

(p1 ph1):f1 (100u) (p2 ph2):f2

would execute a pulse on channel f1. At the same time, the 100u delay would

begin, since it is enclosed in a separate set of parentheses. The pulse on channel f2 would not be executed before either p1 or 100u are finished, whichever is longer (Figure 4.1)..



Figure 4.2 Rule 3: example 1

The following example is a typical section of a DEPT pulse program.

(p4 ph2):f2 (p1 ph4 d2):f1 (p0 ph3):f2 (p2 ph5):f1

The pulses p4 and p1 begin at the same time, p4 on channel f2 and p1 on channel f1. The pulses p0 and p2 start again simultaneously, but not before the sequence with the longest duration of the previous line has terminated (Figure 4.3)..



Figure 4.3 Rule 3: DEPT example

The following 2 lines were extracted from the COLOC Bruker pulse program.

```
(d6) (d0 p4 ph2):f2 (d0 p2 ph4):f1
(p3 ph3):f2 (p1 ph5):f1
```

We have three sets of parentheses in this case. The first items in each parenthesis start at the same time, namely d6 and d0. After d0, p4 on channel f2 and p2 on

channel f1 start simultaneously. Assuming that d6 is larger than d0+p4 and d0+p2, the second line would be executed after d6 has terminated.

A final example for rule 3 is a line from the HNCO3D Bruker pulse program:

(p14:sp3 ph1):f2 (p22 ph1):f3

The shaped pulse p14 is started simultaneously with the rectangular pulse p22.

4.8 Decoupling

4.8.1 Decoupling commands

Table 4.4 shows the available types of decoupling commands. Composite pulse decoupling is discussed in more detail in the next section of this chapter in more detail.

cw	continuous wave decoupling
hd	homodecoupling
<u>cpds1</u> , , <u>cpds8</u>	composite pulse decoupling with cpd sequence 1,, 8, synchronous mode
<u>cpd1</u> , , <u>cpd8</u>	composite pulse decoupling with cpd sequence 1,, 8, asynchronous mode
do	terminate decoupling

Table 4.5 Decoupling commands

Each pulse program line may contain one (and only one) decoupling command. For example, the line

d1 cw:f1

would turn on cw-decoupling on channel f1 at the same time that d1 starts. The line

go=2 cpds1:f2

would turn on composite pulse decoupling on channel f2 at the same time that fid

detection starts.

Decoupling commands are legal in pulse program lines containing delay commands or <u>go</u>, but not in lines with pulses, <u>adc</u>, <u>rcyc</u>, <u>lo</u>, <u>if</u>, and <u>goto</u> commands or expressions.

Decoupling will remain active until explicitly turned off with do, e.g.

0.1u do:f1

(turn decoupling off on channel f1).

4.8.2 Decoupling frequency

The decoupling frequency is selected by specifying the spectrometer channel behind the decoupling command. Omitting the channel is illegal. For example,

d1 cw:f2

would turn on cw decoupling on channel f2, i.e. with the frequency SFO2. This syntax is the same as used for selecting pulse frequencies (cf. the section *Pulse frequency* in this chapter). The next example

0.1u cpds1:f3

would turn on the composite decoupling sequence 1 on channel f3, i.e. with the frequency SFO3. The final example

3m do:f3

would terminate decoupling on channel f3 with the beginning of the 3 millisecond delay.

4.8.3 Decoupling phase

The relative phase of the decoupling frequency can be controlled using a phase program. This is equivalent to controlling the phases of pulses (cf. the section *Pulse phases* in this chapter). Examples:

(d1 cpds1 ph2):f3 0.1u (cw ph1):f2

Please note that phase cycling (cf. the section Phase cycling in this chapter) is

applied to phase programs specified after decoupling commands in the same way that phase programs are specified with pulses. A simple example demonstrating this feature is the pulse program section

```
1m (cw ph1):f2
d1 do:f2
```

which is equivalent to the section

```
(1mp ph1):f2
d1
```

A 1 millisecond pulse is executed on channel f2, followed by a delay d1. Its phase is cycled according to ph1.

4.8.4 Decoupling power

The power of the spectrometer channel specified with the decoupling command is valid. The power setting commands are <u>pl0-pl31</u> (cf. the section *Power and shape* in this chapter for details). For composite pulse decoupling, additional power setting features are provided (see below).

4.9 Composite pulse decoupling (cpd)

4.9.1 General

Composite pulse decoupling, as opposed to cw and hd decoupling, offers a large degree of freedom for the users to set up their own decoupling pulse sequences. Up to 8 different cpd sequences can be used in a pulse program. For example, the command line

d1 (cpds1 ph2):f3 (cpds2 ph4):f2

would start cpd sequence 1 on channel f3 and simultaneously cpd sequence 2 on channel f2, while at the same time d1 begins. Sequence 1 is obtained from a text file whose name is given by the acquisition parameter CPDPRG1. Likewise, sequence 2 is obtained from a text file whose name is given by the acquisition parameter CPDPRG2, etc. The text file contains the user defined cpd sequence, or a standard sequence provided by Bruker (e.g. WALTZ16, GARP, BB). Cpd sequences (= cpd programs) are set up with the command <u>edcpd</u> (described in the

<u>cpds1</u> , , <u>cpds8</u>	Start decoupling using the cpd program CPDPRG1,, CPDPRG8. The decoupling sequence will start at line 1.
<u>cpd1</u> , , <u>cpd8</u>	Like <u>cpds1-cpds8</u> , however, the decoupling sequence will continue from the line where it was stopped using <u>do</u> .
:f1,, :f8	Channel selector. To be appended to the cpd commands.
cpdngs1, ,cpdngs8 cpdng1, ,cpdng8	Like the cpd commands above, however, the transmitter gate for the specified channel will not be opened. Gating is controlled by the main pulse program, and can be tailored by the user.

chapter The File Menu). Table 4.6 shows the commands available to start a cpd

Table 4.6 Available cpd commands

decoupling sequence; and Table 4.7 shows the commands available to build a cpd sequence.

4.9.2 Syntax of cpd sequences

The syntax of cpd sequences is demonstrated by examples. Table 4.8 shows the realization of Broadband and Garp decoupling by means of cpd sequences. Each sequence is an infinite loop, indicated by the last command jump to 1.

A pulse width may be specified in the form of a *fixed pulse*, (e.g. <u>850up</u>) as indicated from pulse programs, or as a command <u>p0-p31</u>, also indicated from pulse programs. The Garp sequence shows the usage of the <u>lo</u> loop command.

The Garp sequence, as well as the sequences in Table 4.9, make use of the command <u>pcpd</u> to generate pulses. This facilitates the executiion of the same sequence for different nuclei on different channels. For example, if executed on channel f2 (f3), the pulse duration of <u>pcpd</u> is given by the parameter PCPD2 (PCPD3). This allows you to specify the 90 degree pulse width for two different nuclei in PCPD2 and PCPD3, and decouple both nuclei within the same pulse program using the

p0, , p31 10up, 5mp, 2.5sp pcpd	Generate pulses with durations P0,, P31. Generate pulses in micro-, milli-, and seconds Generate a pulse with duration according to PCPD1, , PCPD8, depending on the channel where the cpd sequence is executed (use <u>eda</u> to set PCPD).
d0, , d31	Generate delays with durations D0,, D31.
10u, 5m, 2.5s	Generate delays in micro-, milli-, and seconds.
*3.5	Multiplier. Can be appended to p0-p31 or d0-d31.
:135.5	Phase in degrees. Can be appended to pulses .
:sp0, , :sp15	Shaped pulse selectors. Can be appended to pulses.
pl=	Power specifier (see example in Table 4.9):
pl=5	in dB
pl=sp13	according to shaped pulse parameters SP0-15
pl=pl25	according to PL0-15
fq=	Frequency change
fq=2357	in Hz (relative to SFO1 for channel 1, SFO2 for 2)
fq=cnst25	from the parameters CNST0-31
fq=fq2	from the frequency list specified in FQ2LIST
;	Begin of a comment (until end of line)
lo to <i>label</i> times <i>n</i>	Loop to <i>label n</i> times
jump to <i>label</i>	Branch to <i>label</i> . Usually the last statement.
#addphase, #setphase	Special phase control commands

Table 4.7 Commands available to build cpd sequences

same cpd program.

Table 4.9 shows two cpd sequences based on shaped pulses. Shapes are selected in the same way as described earlier in this chapter using the :sp0, ..., :sp15 pulse selector options. The examples demonstrate the order in which duration multiplier, shape selector and phase must be specified.

The sequences in Table 4.8 do not contain a power setting command. Therefore, the current power setting of the main pulse program for the respective channel is valid. In the right column of Table 4.9, the command <u>pl=sp15</u> switches the power of the channel where the sequence is executed to SP15, which is the power given in

1 90up:0	1 pcpd*0.339:0
160up:180	pcpd*0.613:180
240up:0	pcpd*2.864:0
	pcpd*2.981:180
570up:0	pcpd*0.770:0
680up:180	
810up:0	pcpd*0.593:0
960up:180	lo to 1 times 2
1140up:0	2 pcpd*0.339:180
1000up:180	pcpd*0.613:0
850up:0	
710up:180	pcpd*2.843:180
-	pcpd*0.729:0
200up:0	pcpd*0.593:180
110up:180	lo to 2 times 2
jump to 1	jump to 1

Table 4.8 Broadband and G	GARP cpd sequences
---------------------------	--------------------

1 pcpd*2:sp15:0	1	pcpd*14.156:sp15:60 pl=sp15
pcpd*2:sp15:0		pcpd*14.156:sp15:150
pcpd*2:sp15:180		pcpd*14.156:sp15:0
pcpd*2:sp15:180		pcpd*14.156:sp15:150
pcpd*2:sp15:180		pcpd*14.156:sp15:60
pcpd*2:sp15:0	2	pcpd*14.156:sp15:240
pcpd*2:sp15:0		pcpd*14.156:sp15:330
pcpd*2:sp15:180		pcpd*14.156:sp15:180
pcpd*2:sp15:180		pcpd*14.156:sp15:330
pcpd*2:sp15:180		pcpd*14.156:sp15:240
pcpd*2:sp15:0		lo to 2 times 2
pcpd*2:sp15:0	3	pcpd*14.156:sp15:60
pcpd*2:sp15:0		pcpd*14.156:sp15:150
pcpd*2:sp15:180		pcpd*14.156:sp15:0
pcpd*2:sp15:180		pcpd*14.156:sp15:150
pcpd*2:sp15:0		pcpd*14.156:sp15:60
jump to 1		jump to 1
	1	

 Table 4.9 MLEVSP and MPF7 cpd sequences

entry 15 of the shaped pulse parameter table (which can be displayed in eda).

The following section of a pulse program starts a cpd program on channel f2, but keeps the f2 transmitter output disabled (command <u>cpdngs2</u>) except for the periods given by p2. The p2 pulse actually serves as a gating pulse for cpd decoupling, and should not have a phase program assigned to prevent overwriting the cpd program phases.

1u cpdngs2:f2 (p1 ph1 d1):f1 1 (p2 d2):f2 lo to 1 times 10 lo to 2 times 10

4.9.3 Phase setting in cpd programs: <u>#addphase</u>, <u>#setphase</u>

In cpd programs, <u>#addphase</u> is the default mode, i.e. <u>#addphase</u> needs only be specified to switch back to default mode in the case where <u>#setphase</u> mode was previously used.

Example 1:

Commands in cpd program:

#addphase pcpd:180

Pulse program command to start the cpd sequence:

d1 cpds2:f2 ph2

Resulting phase of the pcpd pulse: 180 plus the current phase in ph2.

Example 2:

Commands in cpd program:

#addphase pcpd:sp15

Pulse program command to start the cpd sequence:

d1 cpd2:f2 ph2

Resulting phase of the pcpd shaped pulse: shaped pulse phase (according to

the phases in the shape file) plus ph2.

Example 3:

Commands in cpd program:

#setphase pcpd:sp15:180

Pulse program command to start the cpd sequence:

d1 cpd2:f2

Resulting phase of the pcpd shaped pulse: shaped pulse phase (according to the phases in the shape file) plus 180.

Please note that in this example a phase program must not be used with cpd2 since the FCU does not support the real-time addition of more than two phases.

4.9.4 Frequency Setting in CPD Programs

There are three ways to change the frequency of the channel where the CPD sequence is applied. They are similar to the frequency setting commands in the pulse program except that the channel specification after the command is not necessary in CPD programs.

4.9.4.1 Frequency Setting from Lists

The first method is to use a frequency list. The commands fq1-fq8 corresponding to the frequency list parameters FQ1LIST-FQ8LIST set the frequency from the next list item thereby advancing the list pointer to the next item. In contrast to the pulse program the list assignment is done at compile time, not at run time. In the following example the frequency setting at the first fq1 statement uses the first item of the frequency list, the next command the second item. If the program loops back to label 1 and the frequency list contains more than 2 items, nevertheless during the next execution of the loop the frequency will be set from the first 2 list items.

Like in pulse programs the frequency offset can be specified in two ways: either the offset is at the to of the list in MHz, or no offset is specified in the list. In this latter case the measure frequency of the appropriate channel (SFO1 for F1, SFO2 for F2, etc.) is used as list offset.

4.9.4.2 Frequency setting using the parameters CNST0-31

The command <u>fq=cnst25</u> will set the frequency SFO1 + CNST25 [Hz]. The parameter CNST25 can also be modified during <u>gs</u>. If used on channel F2 the basic frequency SFO2 instead of SFO1 will be used etc.

4.9.4.3 Direct Specification of Frequencies

The command $\underline{fq}=3000$ will set the frequency SFO1(2,3...) + 3000 Hz.

4.10 Loop commands

The general form of a loop command is lo to label times n

Example 1:

```
label1, d1
p1:f2
lo to label1 times 10
p2:f2
```

Remember that a label can be an arbitrary string, such as *label1*, followed by a comma, or a number, such as 2, without a comma appended. The <u>lo</u> command in this example, although specified on an extra line, does not introduce an extra delay between the last p1 pulse and p2.

Example 2:

```
label1, p1:f1
label2, d1
p1:f2
lo to label2 times 10
lo to label1 times 5
p2:f2
```

The first <u>lo</u> command in this example does not introduce an extra delay in the pulse program. However, any further <u>lo</u> command will add a delay of 2.5 microseconds. XWIN-NMR will display a respective message when the pulse program compiler is invoked, i.e. when entering one of the commands <u>gs</u>, <u>zg</u>, <u>go</u>, or <u>pulsdisp</u>.

<u>lo to <i>label</i> times 5</u>	The loop counter is a constant.
<u>lo to <i>label</i> times td</u>	The loop counter is TD, the time domain size in the acquisition dimension (to be defined with the command <u>td</u> , or in the left column in <u>eda</u>).
lo to label times td1	The loop counter is TD(F1) (command <u>1 td</u> for 2D parameter sets, right column in <u>eda</u>).
lo to label times nbl	The loop counter is the parameter NBL (cf. <u>wr</u> , <u>st</u> , <u>st0</u>)
<u>lo to <i>label</i> times 10</u> <u>lo to <i>label</i> times 131</u>	The loop counter is L0,, L31 (to be defined with the keyboard commands $\underline{10},, \underline{131}$, or L array in <u>eda</u>). The pulse program commands $\underline{iu0},, \underline{iu31}$ incre- ment the counters 10-131 by 1, <u>du0</u> ,, <u>du31</u> decre- ment them, and <u>ru0</u> ,, <u>ru31</u> reset them to L0,, L31.
<u>lo to <i>label</i> times c</u>	The loop counter is taken from the list file given by VCLIST (cf. commands <u>vclist</u> and <u>edlist</u>). The pulse program command <u>ivc</u> advances the list pointer by 1. The list pointer position can also be calculated by an equation, e.g. <u>vcidx=5</u> .
lo to label times myCounter	The loop counter is defined at the beginning of the pulse program by means of a <u>define</u> statement and an expression, e.g. <u>define loopcounter myCounter</u> <u>"myCounter=aq/10m +1"</u> The result must not have a dimension.

The lo command exists in a number of variations shown in Table 4.10.

Table 4.10 The lo commands

Example 3:

ze label1, (d1 p1):f1 lo to label1 times l2
```
1u iu2
p2:f2
go=label1
```

Assume the parameter L2 is set to 1 using the keyboard command <u>12</u>, or by setting L2=2 in <u>eda</u>. Then, before scan 1, (<u>d1 p1):f1</u> would be executed once, before scan 2 twice, etc. The <u>lo</u> command does not introduce an extra delay in the sequence. The increment command <u>iu2</u> is executed during the specified 1 microsecond delay. You could replace the loop counter <u>12</u> with <u>c</u> in this example, and <u>iu2</u> with <u>ivc</u> to use the number of loops specified in a list file.

Example 4:

```
define loopcounter myCounter
"myCounter=aq/10m +1"
ze
label1, (d1 p1):f1
lo to label1 times myCounter
go=label1
```

This example defines the variable *myCounter* to represent a loop counter. An arithmetic expression assigns a value to it: the parameter AQ, divided by 10 milliseconds, plus 1. The compiler truncates the quotient aq/10m to give an integer. The expression may include any of the parameters shown in Table 4.3 on page 149.

4.11 Conditional pulse program execution

4.11.1 Conditions evaluated at compile time

Consider the pulse program at the left part of Table 4.11. It combines two experiments in one pulse program, a simple Cosy and a Cosy with presaturation during relaxation. The required pulse program statements to select or deselect presaturation are

```
#define aFlag
#ifdef aFlag
#ifndef aFlag
```

```
#define PRESAT
                            #define PRESAT
1 ze
                            1 ze
2 d11
                            2 d11
3 0.1u
                            3 0.1u
#ifdef PRESAT
                            #include <Presat.incl>
  d12 pl9:f1
                              pl phl
                              d0
  d1 cw:f1
  d13 do:f1
                              p0 ph2
  d12 pl1:f1
                              qo=2 ph31
#endif
                              d11 wr #0 if #0 id0 zd
  pl phl
                              lo to 3 times td1
  d0
                            exit
  p0 ph2
  qo=2 ph31
  d11 wr #0 if #0 id0 zd
  lo to 3 times td1
exit
```

Table 4.11 Using #define, #ifdef, #include

<u>#endif</u>

and correspond to C language pre-processor syntax. The name of aFlag can be chosen by the user, in our example aFlag=PRESAT. The identifier aFlag is considered to be *defined* in the pulse program, if the statement <u>#define aFlag</u> is present, otherwise it is considered to be *undefined*. The command <u>#ifdef aFlag</u> will ignore all pulse program commands until the next <u>#endif</u> if aFlag is *undefined*, and accept them if aFlag is *defined*. The command <u>#ifnef aFlag</u> works conversely.

In Table 4.11, <u>#define PRESAT</u> enables the presaturation command block. Inserting a semicolon (<u>:#define PRESAT</u>), would make the PRESAT flag undefined, and the presaturation block would not be executed.

The <u>#ifdef</u> and <u>#ifndef</u> statements are evaluated by a pre-processor. The pulse program compiler will take over the pre-processed pulse program. For this reason these statements do not introduce any timing changes. It is easy to view a pre-processed pulse program where all conditional statements beginning with a '#' have been removed: Enter the <u>pulsdisp</u> command and click on the button *Show program*. The example could be extended to include double quantum filtering. For this purpose, an additional flag (e.g. <u>#define DQF</u>) could be defined.

The right part of Table 4.11 shows the same pulse program in a more condensed form. The presaturation block is now contained in a separate file, *Presat.incl*, which is invoked with the <u>#include</u> statement.

Please note:

All statements beginning with a '#' character must start at the beginning of a line. No spaces or tabs are placed in front of the '#'. You can use <u>#define</u> not only to define *aFlag*, but, as known from the C language preprocessor, to define complete macros.

Example 1:

#define macro1 (p1 d1) (p2):f2 macro1

This pulse program is equivalent to:

(p1 d1) (p2):f2

Example 2:

```
#define macro2 (p1 d1) \n(p2):f2
macro2
```

This pulse program is equivalent to:

(p1 d1) (p2):f2

The definition of macro2 is continued on a new line using the $n\$ character sequence. In example 1, p1 and p2 start at the same time, while in this example p2 start at the end of the sequence (p1 d1).

Example 3:

#define macro3 (p1 d1) n (p2):f2 macro3

This pulse program is equivalent to:

(p1 d1) (p2):f2

The definition of macro3 only requires one line. However, the \n character sequence enforces a new line when the macro is invoked. For this reason, the pulse programs of the examples 2 and 3 are identical.

4.11.2 Conditions evaluated at run time

XWIN-NMR supports branching and evaluation of conditions within a pulse program while the pulse program execution is in progress. lists the available com-

goto label	Unconditional jump to <i>label</i>
if expression goto label	Branch to <i>label</i> if the <i>expression</i> evaluates to <i>true</i> .
	Branch to <i>label</i> if the <i>trigger</i> condition is true. (not yet available in XWIN-NMR 1.0 and 1.1) Positive <i>level trigger</i> specifiers:
<u>if (trigger) goto label</u>	trigpl1, trigpl2, trigpl3, trigpl4
	Negative <u>level trigger</u> specifiers: <u>trignl1, trignl2, trignl3, trignl4</u>
	The same <i>trigger</i> specifiers as above are legal. The next pulse program command will not be executed until the <i>trigger</i> condition becomes <i>true</i> . Example: <u>1u trigpl1</u>
<u>aDelay trigger</u>	Positive <u>edge trigger</u> specifiers: <u>trigpe1</u> , <u>trigpe2</u> , <u>trigpe3</u> , <u>trigpe4</u>
	Negative <i>edge trigger</i> specifiers: trigne1, trigne2, trigne3, trigne4

Table 4.12 Conditional pulse program execution

mands. These commands do not introduce a delay in the pulse program. At run time, pre-evaluation is performed during the cycle time of the loops in which the commands are embedded. If in a particular pulse program loops are executed too fast, a run time message is printed.

Example 1:

```
ze
lab1, d1
p1
d0
if "d0*2 + 7m > 500m" goto lab2
"d0 = d0 + 10m"
p2
lab2, go=lab1
```

Assume that we will start with d0=10m. The pulse p2 will no longer be executed as soon as the expression "d0*2 + 7m > 500m" becomes true.

Example 2:

```
ze
lab1, if (trigpl2) goto lab3
lab2, d1
p1
aq
lo to lab2 times ds
goto lab1
lab3, d1
p1
go=lab1
```

The TCU has 4 trigger input channels; signals arriving at the TCU can be checked using the <u>trig</u> specifiers. This example performs DS dummy scans to maintain steady state conditions as long as no positive level is detected on input channel 2. If such a level is detected, NS data acquisition scans are executed, then the pulse program again checks the external trigger signal.

Example 3:

lab1, d1 trigpl2 p1 go=lab1

This example starts executing the pulse sequence as soon as a positive level is detected on input channel 2. After each scan, the pulse program will wait until the next trigger signal is detected.

Example 4:

```
ze
lab1, d1
p1
lo to lab1 times l2
0.1u iu1 ;count number of scans
0.1u iu2 ;increment l2
if "11 <= 3" goto lab2 ;if scancounter < 4
0.1u ru2 ;reset l2 to L2
lab2, go=lab1
```

This example repeats the sequence (d1 p1) L2 times before scan 1, L2+1 times before scan 2, and L2+2 times before scan 3. Then, l2 is reset to its initial value L2. Before all remaining scans the sequence (d1 p1) is generated L2 times. L1 must be set to 1 before starting the sequence.

4.12 Commands to suspend the pulse program execution

XWIN-NMR allows to stop the pulse program execution at specified points in the pulse program. The program execution can be continued by the command ,resume'. The suspension can be done always (,autosuspend', ,calcautosuspend') or conditionally, if the command ,suspend' has been given by the user. Since the program which is executed on the TCU is precalculated, the precalculated program parts may be invalid, if a parameter has been changed before the command ,resume' is executed. Therefore the possibility to prevent the precalculation has been implemented (,calcsuspend', ,calcautosuspend'). When program execution is resumed after such a command, enough time must be left to precalculate a large enough portion of the pulse program before the program is resumed.

suspend	stop execution if command suspend has been given
autosuspend	stop execution always
calcsuspend	stop precalculation and stop execution on command
calcautosuspend	stop precalculation and stop execution always

 Table 4.13 Commands to suspend pulse program execution

4.13 Commands to start data acquisition

XWIN-NMR provides 5 basic pulse program commands to start data acquisition:

go=label, gonp=label, gosc, goscnp, adc.

The most commonly used command is go=label. It is a macro command, i.e. it includes a number of different actions required for successful data acquisition. <u>adc</u> allows you to control any fine detail of the acquisition process. All three acquisition commands place the digitized signal into a memory buffer, but do not store it on disk. The <u>wr</u> command, described in a later section, is provided to write the buffer contents to disk.

4.13.1 The commands go=label, gonp=label, gosc, goscnp

The left column of Table 4.16 shows a simple example of how to use go=label in a pulse program. All go type commands perform the following actions (we assume here that DE1 < DE2, otherwise DE1 and DE2 must be interchanged):

- 1. A delay DE1 is executed (cf. the description of DE1/DE2 in the chapter *The Acquire Menu*). At the end of this delay, the receiver gate is opened and the intermediate frequency (if required) is added to the frequency of the observe channel. This corresponds to the execution of the command <u>syrec</u>.
- **2.** A delay DE2-DE1 is executed. At the end of this delay, the phase of the receiver channel is set to 0.
- **3.** A delay DE-DE2 is executed. At the end of this delay, the digitizer is started. Please refer to the description of the parameters DW/DWOV/DIGMOD on how the sampling rate is selected. The result will be a digitized fid signal of TD data points, where the time domain size TD is defined by the user (from <u>eda</u>, or by typing <u>td</u>). The fid will be placed into the *current memory buffer*. The contents of memory buffers can be transferred to disk by means of the <u>wr</u> pulse program

command or the <u>tr</u> keyboard/menu command. The section *Working with acquisition memory buffers* discusses the usage of memory buffers and the size restrictions of TD.

- **4.** The same time the digitizer is started, a delay AQ is executed. This is a delay until the digitization of the fid is finished.
- **5.** A delay of 3 milliseconds is executed. During this time the acquisition software prepares for the next scan:
 - a) The scan counter, visible during real time fid display, is incremented to inform the user about the number of scans performed since the last executed <u>ze</u> or <u>zd</u> statement.
 - b) The frequency of the observe channel is switched back to the frequency of the observe nucleus (for spectrometers using an intermediate frequency). This corresponds to the execution of the command <u>sytra</u> (which is inverse to <u>syrec</u>).
 - c) The phase program pointers are incremented to the next phase in the lists, corresponding to the execution of the commands <u>ipp0</u>, ..., <u>ipp31</u> for the phase programs present in the pulse program. This step is skipped by <u>gonp=*label*</u> and <u>goscnp</u>.
 - d) The commands <u>go=label</u> and <u>gonp=label</u> perform a loop to *label*, whereas <u>gosc</u> and <u>goscnp</u> do not loop. The pulse program commands between *label* and <u>go</u> or <u>gonp</u> are executed DS+NS times. During the first DS loops (*dummy scans* to achieve steady state conditions), the digitizer is not activated, otherwise the dummy scans are identical to the NS data acquisition scans. If no dummy scans are desired, DS must be set to 0.

Please note: Even if DS > 0, no dummy scans will be executed if the pulse program command <u>zd</u> (rather than <u>ze</u>) was executed before a go-loop is entered (cf. the description of <u>ze</u> and <u>zd</u>). This feature is, for example, employed in 2D experiments where dummy scans are only required before the first fid is measured.

Table 4.14 shows that the <u>go</u> commands can be specified in conjunction with other commands. PH_ref is an acquisition parameter to be defined by the user.

4.13.2 The commands <u>rcyc=label</u>, <u>rcycnp=label</u>

The command <u>rcyc</u> executes step 5 of the actions performed by <u>go=label</u> and <u>gonp=label</u> (cf. the previous section). The <u>rcycnp</u> command skips step 5c.

A-	1	87	
А-	I	87	

1	go=2 ph31	Receiver phase = ph31, realized via add/ subtract and channel A/B switching. Legal phase values: 0, 90 180, 270 degrees.
2	go=2 ph30:r	Receiver phase = ph30 + PH_ref, realized via the phase of the reference frequency of the observe channel. Legal phase values: arbitrary.
3	go=2 ph31 ph30:r	Combination of (1) and (2). The receiver phase is the sum: ph31 + ph30 + PH_ref
4	go=2 ph31 ph30:r cpd1:f2	Decoupling starts at the same time the receiver is opened, and stops automati- cally when the loop is executed.
5	go=2 ph31 ph30:r cpd1:f2 ph29	Like example 4, with a phase program for the cpd sequence.

Table 4.14 Ways to specify a go or gonp command

The commands are provided for programming acquisition loops based on <u>adc</u> rather than <u>go=*label*</u> and <u>gonp=*label*</u>. You must *not* specify phase programs behind <u>rcyc</u> and <u>rcycnp</u>. However, decoupling commands are legal, although it is not meaningful to use them here. Table 4.16 displays an example of an acquisition loop based on <u>rcyc</u>.

The <u>rcyc</u> commands may also be specified behind a delay, e.g. 100u rcyc=2. In this case, they are executed during the specified delay instead of the default 3 milliseconds. The delay must not be shorter than 100 microseconds.

4.13.3 The commands eosc, eoscnp

The command <u>eosc</u> executes steps 5a-5c of the actions performed by <u>go=*label*</u> and <u>gonp=*label*</u> (cf. the previous section). The <u>eoscnp</u> command executes only steps 5a and 5b.

These commands are provided for pulse programs with data acquisition based on adc. In contrast to <u>rcyc</u>, the user must add the appropriate loop commands.

You must not specify phase programs behind eosc and eoscnp. However, decou-

pling commands are legal, although it is not meaningful to use them here. Table 4.16 displays an example of an acquisition loop based on <u>eosc</u>.

The <u>eosc</u> commands may also be specified behind a delay, e.g. <u>100u eosc</u>. In this case, they are executed during the specified delay instead of the default 3 milliseconds. The delay must not be shorter than 100 microseconds.

4.13.4 The commands <u>ze</u> and <u>zd</u>

Both commands perform the following actions:

- **1.** They set the scan counter, visible during real time fid display, to 0 or -DS. Negative values indicate that dummy scans are in progress.
- 2. They set a flag which informs the next go, gonp, gosc, goscnp, or adc command that the fid, digitized by the respective command, should replace any existing data in the acquisition memory. All NBL memory buffers are concerned. If ze or zd are placed outside an acquisition loop, the *replace* mode will only be valid for the first scan performed by the loop. The fids of all the scans that follow will be added to the data present in the memory buffer.
- **3.** The difference between \underline{ze} and \underline{zd} is that \underline{zd} inhibits the execution of dummy scans by <u>go</u>, <u>gonp</u>, <u>gosc</u>, <u>goscnp</u>, and by <u>adc</u> (combined with <u>rcyc</u> or <u>eosc</u>), even if DS > 0.

The commands \underline{ze} and \underline{zd} should be written behind a delay command (with a delay > 10 microseconds, depending on the number of phase programs in the sequence). Then, they are executed during the delay. If they are placed on a separate line, their execution will require 3 milliseconds. The *cosy* pulse program of the Bruker library (Table 4.15) illustrates the use of \underline{ze} and \underline{zd} .

4.13.5 The command <u>adc</u>

The command <u>adc</u> starts the digitizer and, at the same time, opens the receiver. Please refer to the description of the parameters DW/DWOV/DIGMOD on how the sampling rate is calculated. The result of <u>adc</u> will be a digitized fid signal of TD data points, where the time domain size TD must be defined by the user. The fid will be placed in the current memory buffer (cf. the section *Working with acquisition memory buffers*).

It is the responsibility of the user of <u>adc</u> to account for the various switching delays provided automatically by the <u>go</u> commands (cf. description of <u>go=*label*</u>). It is also

```
1 ze ; enable dummy scans
2 d1
3 p1 ph1
d0
p0 ph2
go=2 ph31
d1 wr #0 if #0 id0 zd ; the next fid requires
; no dummy scans
lo to 3 times td1
exit
ph1=0 0 0 0 1 1 1 1 2 2 2 2 3 3 3 3
ph2=0 1 2 3
ph31=0 2 0 2 3 1 3 1 2 0 2 0 1 3 1 3
```

Table 4.15 Cosy pulse program: how to use ze and zd

the responsibility of the user to enable the intermediate frequency (command <u>syrec</u>). XWIN-NMR provides the special delay generation commands <u>de1</u>, <u>de2</u>, <u>de3</u>, <u>dw, dwov</u>, and <u>aq</u> to facilitate the handling of these items (de1=DE1, de2=DE2-DE1, de3=DE-DE2). Any pulse program employing <u>adc</u> must contain one of the commands <u>eosc</u>, <u>eoscnp</u>, <u>rcyc</u>, or <u>rcycnp</u> to ensure a correct end of scan handling. You may use several <u>adc</u> commands in conjunction with, for example, a single <u>eosc</u> command.

Table 4.16 shows the same pulse program realized via go=label, adc in conjunction with <u>rcyc</u>, and <u>adc</u> in conjunction with <u>eosc</u>. Table 4.17 shows how homodecoupling during data acquisition can be realized using <u>adc</u>. Homodecoupling requires the receiver to be turned off at regular intervals. This can be achieved with the option <u>:e</u>, appended to a pulse or delay command. It disables the receiver for the duration of the respective pulse or delay. *Please note:* If <u>:e</u> does not occur between the commands <u>adc</u> and <u>rcyc/eosc</u> but elsewhere in the pulse program, its sense is reversed, i.e. the receiver is enabled rather than disabled.

The receiver phase

In pulse programs based on the <u>adc</u> command, the receiver phase must be specified behind <u>adc</u>, e.g. <u>adc ph31</u>. This construction sends an interrupt signal to the RCU, telling it to account for the receiver phase setting. Interrupt handling on the RCU is a time consuming procedure which must be carried out during the recycle time of

ze	ze 2 d1 (p1 ph1):f1	ze 2 d1 (p1 ph1):f1 ;
2 d1 (p1 ph1):f1 ; go=2 ph31 ; wr #0 exit	; de1 de2 adc ph31 syrec de3 ph30 aq rcyc=2 ; wr #0 exit	de1 de2 adc ph31 syrec de3 ph30 aq eosc lo to 2 times ns ;

Table 4.16 The same pulse program based on go, adc/rcyc, and adc/eosc

an experiment. For certain applications (e.g. some imaging experiments) it may be necessary to keep the recycle time as short as possible. Table 4.18 shows how receiver phase setting can be moved outside the recycle loop (command <u>recph ph31</u>), allowing for a shorter <u>aq</u> or <u>rcyc</u>. However, then it is not possible to set the receiver phase differently for consecutive scans: Only after NS scans the command <u>ip31</u> in Table 4.18 (right column) will the phase, which is activated by <u>recph ph31</u>, change.

4.13.6 External dwell pulses

The <u>go</u> and <u>adc</u> commands instruct the digitizer to acquire a desired number of data points with a rate given by the *dwell time*. The dwell pulses, which activate the digitizer in regular time intervals are generated internally (on the RCU) so that the detection of a complete fid is automatically accomplished once initialized via <u>go</u> or <u>adc</u>. This is the reason for the waiting time <u>aq</u> in the two rightmost columns of Table 4.16. Certain experiments, however, require that the user can control the detection of each individual data point of an fid. The pulse program in Table 4.19 displays an example of how this can be achieved. It performs identically to the middle pulse program in Table 4.16, but now the dwell pulses are externally generated in the pulse program (i.e. by the TCU) using the <u>:x</u> pulse option. A corresponding cable connection between TCU and RCU is required. In this pulse program the waiting time <u>aq</u> has been replaced by a loop generating as many dwell pulses as required to measure TD data points.



 Table 4.17 Homodecoupling during data acquisition

Please refer to the Bruker pulse program libraries for high resolution, solids, and imaging experiments for more examples using the $\underline{:x}$ option.

1	1
2 d1	2u recph ph31
10u adc ph31	2 d1
aq	10u adc
rcyc=2	aq
10u ip31	d2 rcyc=2
lo to 1 times 11	10u ip31
	lo to 1 times 11
•••	

Table 4.18 Receiver phase setting without / with recph

```
define delay dx
define pulse px
"dx=dwov/2"
"px=dwov/2"
 ze
2 d1
 (p1 ph1):f1
:-----
 de1
 de2 adc ph31 syrec
 de3 ph30
3 dx
 px:x ;external dwell pulse
 lo to 3 times tdov
 rcyc=2
:-----
 wr #0
 exit
```

Table 4.19 Acquisition based on external dwell pulses

4.14 Working with acquisition memory buffers

The acquisition commands <u>go=label</u>, <u>gonp=label</u>, and <u>adc</u> place the acquired data points into a *memory buffer* where they reside until new data points are added, or until they are replaced by new data (*replace* mode is turned on by the commands

 \underline{ze} and \underline{zd}). A memory buffer provides space for TD data points, where TD must be set by the user.

All experiments having a single fid, and as a result, need just one memory buffer. All digitized data are accumulated in this buffer by the acquisition commands, and then stored on disk.

Applications such as multi-dimensional experiments, imaging experiments, experiments varying parameters such as the decoupling frequency or recovery time generate several fids. The developer of a pulse program has the choice of utilizing a single or several memory buffers to carry out such experiments. If a single buffer is used, the buffer content must be transferred to disk before the next fid can be measured. If several buffers are used, several fids can be measured before a disk transfer is required. The latter method is appropriate if the fids of the experiment succeed one another so quickly that no disk transfer is possible in between them.

The acquisition parameter NBL determines the number of memory buffers provided (default: NBL=1). Each buffer has a size TD. The buffer size will be rounded to the next multiple of 256 data points if the TD is not a multiple of 256. The acquisition commands will put the fid into the *current* buffer. The *current* buffer 1 by default. The pulse program command st makes the *next* buffer the current buffer, and the command st0 makes the *first* buffer the current buffer. The first buffer will also become the current buffer if st is executed more than NBL times. The commands st and st0 must be specified behind a delay which must not be shorter than 10 microseconds, e.g. 10u st. Table 4.20 shows an example, the

```
1 ze
dl1 pl14:f2
dl1 fq2:f2 st0
2 d1
3 d20 cw:f2
dl3 d0:f2
pl ph1
go=2 ph31
dl fq2:f2 st
lo to 3 times l4
dl1 wr #0 if #0
exit
```

Table 4.20 Illustrating st and st0: noedif pulse program

noedif pulse program of the Bruker library. The fids acquired with different decoupling frequencies are stored in subsequent memory buffers.

The size of NBL is limited by the constraint that NBL times TD must not exceed the available RCU memory. For example, an RCU equipped with 4 Mbytes DRAM allows for about 3.8 Mbytes fid data to be stored (the remainder is needed by the acquisition parameters). Additional memory is available as an option.

4.15 Writing memory buffers to disk

Any pulse program must contain at least one disk write command to transfer the acquired data to disk, since the data acquisition commands go=label, gonp=label, gosc, goscnp, and adc put the digitized data into a memory buffer, but do not store them persistently.

Table 4.21 shows the pulse program commands provided by XWIN-NMR to access disk files. The name of the output file is *fid* or *ser*. An *fid* file contains a single fid, whereas a *ser* file contains a sequence of fids. The appropriate name is chosen by the pulse program compiler: If a pulse program contains one of the increment, decrement, or reset file pointer commands, or <u>st/st0</u>, a *ser* file will result.

Tranferring data to disk means *adding* the data to the data contained in an existing *fid* or *ser* file, or *replacing* this data (if no such file exists, one will be created). *Addition* will take place if the pulse program is started with the command <u>go</u>, *replacement* will take place if started with <u>zg</u>. However, data replacement only occurs the first time a memory buffer is tranferred to disk. Any further invocation of <u>wr</u> will cause the buffered data to be added to the data in the file.

The command \underline{zg} will create the *ser* file before acquisition starts, and fill it with zeroes if no *ser* file with the correct file size exists. Otherwise the existing *ser* file is retained, and the <u>wr</u> command in the pulse program initially will overwrite the *ser* file section defined by the current file pointer, TD, and NBL. The contents of the other parts of the file will remain unchanged.

The <u>wr</u> commands (and all other commands in Table 4.21) can be specified behind a delay (cf. the example in Table 4.20). The delay must not be shorter than 10 microseconds. The delay is required to initiate the command, not to execute it. The only timing requirement for <u>wr</u> is that the disk transfer be complete before <u>wr</u> is called again, otherwise, a run-time error message is printed. The actual execution

<u>wr #0</u>	Transfer the acquisition buffer to the file <i>fid</i> , or transfer NBL acquisition buffers to the file <i>ser</i> of the <i>current</i> data set. For <i>ser</i> files: Start writing into the file at the current position of the <i>disk file pointer</i> , which initially is at the beginning of the file.
<u>wr #1, wr #2, wr #3,</u>	Transferring is performed to the file <i>fid</i> or <i>ser</i> belonging to the data set with the number 1, 2, 3, contained in the data set list given by the acquisition parameter DSLIST.
<u>wr ##</u>	Transferring is performed to the file <i>fid</i> or <i>ser</i> of the data set which is pointed to by the <i>data set list pointer</i> . Its initial position is one item <i>before</i> the beginning of the data set list DSLIST, where the <i>current</i> data set is made available. The list pointer can be incremented by 1, decremented by 1, or reset to list begin using the commands <u>ifp</u> , <u>dfp</u> , and <u>rfp</u> , respectively.
<u>if #0, if #1, if #2,</u>	Advance the <i>disk file pointer</i> for <i>ser</i> files by TD*NBL (remember that TD is rounded to the next multiple of 256 data points if it is not a multiple of 256).
<u>df #0</u> , <u>df #1</u> , <u>df #2</u> ,	Decrement the file pointer (inverse of <u>if</u>).
<u>rf #0, rf #1, rf #2,</u>	Reset the file pointer to the beginning of the ser file.
<u>rf #0 m, rf #1 m,</u> <u>rf #2 m</u> ,	Set the file pointer to position m*TD*NBL of the <i>ser</i> file, where m is an integer number.

Table 4.21 Writing acquisition buffers to disk

time of a disk write depends on the computer hardware, the operating system, and the system load according to simultaneously active processes and users. This situation may become particularly critical if the destination disk is not physically connected to the computer, but is accessible only via a network: Networked disks do not guarantee a defined response time.

It is legal to specify the commands \underline{if} , \underline{zd} , one of $\underline{id0-31}$, one of $\underline{ip0-31}$, and decoupling commands behind the same delay used for \underline{wr} . It is important to use either a zd or ze command after each wr before the next scan. Otherwise the data will be

added to some previously acquired data.

When a pulse program writing to a ser file is started using the command zg, XWIN-NMR will at first simulate the program to determine the required file size. A ser file of this size is then created, i.e. the ser file is not required to grow during the experiment. This method avoids the danger of running out of disk space while acquisition is in progress. File size calculation takes into account the time domain size TD of the acquisition dimension, NBL, and loop commands in the pulse program. The simulation result should be equal to TD*TD(F1) for an nD experiment executed from a 2D type data set, or to TD*TD(F1)*TD(F2) for an nD experiment executed from a 3D type data set. Otherwise, a warning message will be printed even though the experiment can still be executed. It is the responsibility of the user to set the TD values of the respective dimensions to match the *ser* file size, since they are used later for the Fourier transform and other processing routines. It is important to set the correct values *before* the experiment starts, because XWIN-NMR will update the status parameter files at the end of acquisition accordingly. If for any reason the TD values in the status parameter files are incorrect after acquisition (check with dpa!), you can still adjust them using the commands 1s td, 2s td, and 3s td.

When designing 3D pulse programs, the acquisition status parameter AQSEQ describes the order 321 or 312, in which the 1D fids of a 3D acquisition are written into the *ser* file (3 = the acquisition dimension, 1 and 2 = the orthogonal dimensions). AQSEQ will be set automatically and stored in the parameter file *acqu* if td and td1 are used consistently within the 3D pulse program. However, you may explicitly define AQSEQ in the pulse program. For this purpose, insert one of the following statements in the pulse program header: <u>aqseq 321</u> or <u>aqseq 312</u>. You can also set or modify AQSEQ using the keyboard command <u>3s aqseq</u> before starting the transform.

4.16 Multiple receivers

If a spectrometer is equipped with multiple receivers, the number of the receiver (1-8) from where data should be acquired can be appended to the following commands, e.g. go5=label (if no number is present, 1 is assumed, e.g. go is equivalent to go1):

go, gonp, gosc, goscnp, adc, rcyc, rcycnp, eosc, eoscnp, ze, zd, st, st0, aq, dw, dwov, recph, wr, if.

Parameters for the n'th RCU are taken from data set n of the data set list DSLIST. The following parameters are relevant:

AQ_mod, DECIM, DIGMOD, DIGTYP, DR, DSPFIRM, DSPFVS, FTLPGN, NBL, OVERFLW, SEOUT, SFO1, SW, SW_h, TD.

4.17 Real time outputs

The TCU provides a number of real time outputs which are used to control the various spectrometer components, such as gating and blanking the transmitters. Please refer to your hardware documentation for information on which output is connected to a particular device. The pulse program compiler will select the correct output automatically, e.g. for a command like <u>p1:f2</u>.

The file *\$xwinnmrHOME/exp/stan/nmr/lists/pp/Avance.incl* contains a number of macro definitions based on the outputs, which can be used in pulse programs.

The hardware documentation will also inform you which of the outputs are still free for special purposes, e.g. for controlling a laser from a pulse program.

4.17.1 Type 1 outputs ("RCPs")

These are 35 outputs which can be set with an accuracy of 12.5 nanoseconds and a minimum of 50 nsec (also called RCP0, ..., RCP34). Two syntactically different ways are provided to enable or disable the outputs.

- 1. The commands
 - p1:c0 5u:c25 vp:c15

would generate pulses of duration P1, 5 microseconds, and according to the current pulse list pointer on the output channels 0, 25, and 15, respectively.

2. The command

1u setnmr0 | 15

would activate output channel 15 (using active=low logic). It will remain active

until explicitly deactivated, e.g. with the command

```
1u setnmr0 ^ 15
```

The characters "|" (vertical bar) and "^" (circumflex) can be used to set and clear a bit in a register consisting of 35 bits. For this reason several outputs can be disabled or enabled simultaneously. For example, the command

```
1u setnmr0 | 14 | 13 ^ 15
```

would enable the output channels 14 and 13, and disable channel 15.

The command <u>setnmr0</u> must be specified behind a delay (in the previous examples it is 1 microsecond). The minimum delay is 200 nanoseconds.

Using the setnmr syntax, a pulse on a desired channel could be realized according to the following example:

; a 4 microsecond pulse on RCP7 4u setnmr0 | 7 1u setnmr0 ^ 7

4.17.2 Type 2 outputs ("NMR control words")

These are 128 outputs which can be set with an accuracy of 25 nanoseconds and a minimum of 50 nsec. They are organized in 8 registers of 16 bit size (called NMR control words). The commands <u>setnmr1</u>, ..., <u>setnmr8</u> are provided to enable or disable the channels 0-15 of each register. The syntax is identical to <u>setnmr0</u> described in the previous section. For example, the command

1u setnmr3 | 0 | 13 ^ 15

would enable the output channels 0 and 13 and disable channel 15 of register 3.

4.18 Gradients

A gradient in a particular spacial dimension (x, y, z) is added to the homogenous

magnetic field in the following way: The plane which is orthogonal to the gradient direction and which contains the center of the receiver coil will see no gradient field, while the two most distant parts of the receiver coil in the respective direction will see the negative minimum and the positive maximum of the gradient field. The absolute value of the maximum gradient field for either dimension is described by a parameter, the *gradient strength*, which can be set by the operator (see below).

A gradient can have a constant strength during the time it is applied, or a variable strength. Formally, this is treated by XWIN-NMR similarly to executing rectangular and shaped high frequency pulses.

4.18.1 Rectangular gradients

A *rectangular* gradient has the same constant strength while it is turned on. Example:

300u gron2 1m 100u groff

The command <u>gron2</u> turns on a gradient with the beginning of a 300 microsecond delay. The gradient remains active until explicitly disabled with the command <u>groff</u>. In this example, the gradient is on during a period of 1.3 milliseconds. The gradient strength in either dimension is given by the parameters GPX2, GPY2, and GPZ2. In general, the commands <u>gron0</u>, ..., <u>gron31</u> are provided to enable a gradient using the strength values GPX0, GPY0, GPZ0, ..., GPX31, GPY31, GPZ31, respectively. The numbers must be in the range 0-100. You can set the parameters by typing <u>gpx0</u> etc. on the keyboard, or by opening the GPX/GPY/GPZ parameter table: Enter <u>eda</u>, and select the GP031 button.

4.18.2 Shaped gradients

A *shaped* gradient changes its field in regular time intervals while it is applied. The gradient *shape* is a sequence of real numbers (between -1.0 and 1.0 and stored in a file, see below), namely the gradient strength values valid for each time interval. The interval length is calculated by the program by dividing the desired gradient duration by the number of strength values in the shape file. The gradient duration must formally be specified by means of a pulse command.

The next 3 examples would generate shaped gradients.

```
10mp:gp2
p1:gp1
gradPulse*3.33:gp3
vp:gp4 Illegal! Shaped gradients with <u>vp</u> are not supported.
```

The gradients would be applied for a duration of 10 milliseconds, P1, and grad-Pulse*3.33, respectively. The gradient shape characteristics would be described by the entries 2, 1, and 3 (corresponding to :gp2, :gp1, and :gp3) of the *gradient parameter table*. This table is displayed when you click on the GP031 button within eda. The table has 32 entries with index 0-31. You may use :gp0 - :gp31 in a shaped gradient command to refer to entry 0-31.

Each entry of the shaped gradient parameter table has 4 parameters assigned: *GPX*, *GPY*, *GPZ* (the gradient strength multipliers for the 3 spacial dimensions), and a *file name* (containing the gradient strength values).

File name

File name is the name of a gradient file. A gradient file can be generated by means of the programs *shape* and *xshape*, which are part of the XWIN-NMR release. You must start these programs from the Unix level. You can also generate your own shape files in ASCII format and convert them to the binary format required by XWIN-NMR using the *shape* program. The ASCII and binary file formats are described in the Chapter *File Formats*.

The specified file must be located in the directory

\$XWINNMRHOME/exp/stan/nmr/lists/wave/.

GPX, GPY, GPZ

These are multipliers with values from 0 to 100. They are applied to the gradient strength values (which range from -1.0 to 1.0) in the shape file to obtain the final gradient field strength.

You can access the table entries not only from <u>eda</u>, but also from the keyboard. For example, the command <u>gpnam5</u> would display the file name corresponding to table entry 5, while <u>gpx15</u> would display the strength value for the x dimension of entry 15.

4.19 Miscellaneous commands

4.19.1 Switching on/off of Presetting of Blanking Pulses: preset

The presetting of blanking pulses can be switched off by the command <u>preset off</u> at the beginning of the pulse program. The generation of blanking pulses before the pulse itself is then switched off, i.e. the program will behave as if all *preset* parameters (command <u>edscon</u>) would have been set to 0. The <u>preset off</u> command must be written at the beginning of the pulse sequence and cannot be changed afterwards.

4.19.2 Generation of Blanking Pulses: gatepulse

Blanking pulses are usually generated automatically using the <u>edscon</u> preset parameters. They can also be generated manually, which is useful if the command <u>preset off</u> has been used. The <u>gatepulse</u> command will generate the transmitter blanking pulses, the preamplifier blanking pulses and the ASU blanking pulses. The syntax is: <u>delay gatepulse 1 [| 2...]</u>.

Examples:

3u gatepulse 1 ;generate blanking pulse for f1 p1:f1 d1 2u gatepulse 1|2 ;generate blanking pulses for f1 and f2 (p1):f1 (p2):f2

4.19.3 Printing messages

The command

print "Hello World"

prints the message *Hello World* during runtime of an experiment. The timing of the printout is not necessarily correlated to the execution of the pulse program because the TCU interprets the pulse program in advance of its execution. However, for debugging complex pulse programs it could be helpful.