

Pneumatic Pressure Test Pump User's Manual

# Warnings and cautions

> Low pressure gauges can very easily be over pressured if not careful.

Please take caution when applying pressures.

- > Do not exceed the safety limit of 1500 psi (100 bar) for 917 or 2000 psi (140 bar) for 918.
- > Do not switch between pressure and vacuum mode under pressure. Open the vent valve and isolation valve before switching between positive to vacuum pressures.
- > Over tightening connectors may cause damage. Light hand tightening is sufficient.
- > Store the pump in a dry non-corrosive environment.
- > Additel is not responsible for problems or damage resulting from improper use or operation of the equipment.
- > All moisture and contaminants should be cleaned out of the liquid isolator before creating a vacuum. This is accomplished by venting the pump when at high pressure.
- ightarrow For the least risk in contamination of the reference gauge, the DUT should be installed to the pressure port which is closest to the vent valve.

### Specification

- > Pressure range: 917: 95% vacuum to 1000 psi (70 bar) 918: 95% vacuum to 1500 psi (100 bar)
- > **Temperature:** (0~50) °C
- > Humidity: < 85%RH
- > Adjustment resolution: 10 Pa (0.1 mbar)
- > Maximum pressure: 917: 1500 psi ( 100 bar ) 918: 2000 psi ( 140 bar )
- > Pressure media: Air
- > Size: Height: 5.5" ( 140 mm );
  - Base: 12.4" ( 315 mm ) x 7.8" ( 198 mm )
- > Weight: 5.7 lb ( 2.6 kg )

**Remark:** If local atmosphere pressure is 1 bar, the vacuum can reach to -0.95 bar; If local atmosphere pressure is P, the vacuum can reach to - 0.95P.

#### **Configuration & Air routine**



## Troubleshooting


1 – Hand pump

[Version number:1407V02]

- 2 Transport handle
- 3 Quick connector 4 – Vent valve

Addite 917/918

- ( clockwise to close/ counterclockwise to open ) 5 – P/V selector lock
- 6 Pressure/vacuum (P/V) selector (Positive pressure: pull it out; vacuum: push it in DO NOT switch under pressure or vacuum to
- avoid damage ) 7 – Isolation valve (Bypass valve) (Isolated the calibration volume from the hand pump,
- piston, and check valve) 8 – Fine adjust handle ( clockwise to increase pressure )



Problem	Cause	Solution
Hand pump will not generate pressure	The isolation valve is not open.	Open the isolation valve.
It is difficult to increase pressure	A. Vent valve is not closed.	Close the vent valve.
	B. The O-ring seal is loose or damaged.	Replace O-ring seal.
	C. P/V selector is in the wrong position.	Pressure: pull the P/V selector out; Vacuum: push the P/V selector in.
Hard to use the fine adjust	A. The isolation valve is not closed.	Close the isolation valve.
	B. The gauges are not tightened.	Tighten the reference gauge or the gauge under test.
	C. The O-ring seal is damaged.	Replace the O-ring seal.
	D. The thread surface is not smooth.	Use Teflon tape on the thread and turn it in tight
	E. The connector type mismatched to the gauge pressure port.	Use the correct adapter.
	F. Dust or contaminants are blocking the pressure flow	Pressurize the pump several times and then suddenly vent which will push any
		contaminants out.
It is difficult to turn the quick connector	A. Too much force was previously applied.	Secure the quick connect valves with less force.
	B. The threads have no lubrication.	Apply lubrication to the threads.


(1) clo

slowly

get the sired value, d close!

H

3) fine adjustmentt



#### Remarks:

- A: Additel has made a concerted effort to provide complete and current information for the proper use of this equipment. The product specifications and other information contained in this manual are subject to change without notice.
- B: Above pictures are just for reference.

Additel Corporation, USA www.additel.com