

For Research Use Only Not for Diagnostic Use



HEPATITIS E – HEV-IgG

Catalog #: WE7296

IgG ANTIBODY TO HEPATITIS E VIRUS ELISA KIT

Two-Step Incubation, Indirect Principle

INSTRUCTIONS FOR USE

This kit is an enzyme-linked immunosorbent assay for the qualitative detection of IgG-class antibodies to hepatitis E virus in human serum or plasma. For Research Use Only

SUMMARY

Hepatitis E virus (HEV) is a non-enveloped, single-stranded RNA virus identified in 1990. Infection with HEV induces acute or sub-clinical liver diseases similar to hepatitis A. HEV infections, endemic and frequently epidemic in developing countries, is seen also in developed countries in a sporadic form with or without a history of traveling to endemic area. The overall case-fatality is 0.5~3%, and much higher (15~25%) among pregnant women. A hypothesis that HEV infection is a zoonosis was presented in 1995. Then a swine HEV and later an avian HEV were identified and sequenced separately in 1997 and 2001. Since then, HEV infection include anti-HEV, viremia and feces excretion of HEV was seen in a wide variety of animals, i.e., swine, rodents, wild monkeys, deer, cow, goats, dogs and chicken in both the developing and developed countries. A direct testimony was reported that the consumption of uncooked deer meat contaminated with HEV led to acute hepatitis E in human, and HEV genome sequences can be detected in pork livers available in the supermarkets in Japan. With the discovery of conformational epitopes in HEV, HEV serology was further explored and understood. The phenomenon of long-lasting and protective antibodies to HEV was observed which greatly enhance the understanding to the diagnosis, epidemiology, zoonosis-related studies and vaccine development.

PRINCIPLE OF THE ASSAY

This kit employs solid phase, indirect ELISA method for detection of IgG-class antibodies to HEV (anti-HEV) in two-step incubation procedure. Polystyrene microwell strips are pre-coated with recombinant, highly immunoreactive antigens corresponding to the structural regions of HEV (ORF-2). During the first incubation step,

anti-HEV specific antibodies, if present, will be bound to the solid phase pre-coated HEV antigens. The wells are washed to remove unbound serum proteins and then, rabbit anti-human IgG antibodies (anti-IgG) conjugated to horseradish peroxidase (HRP-Conjugate) are added. During the second incubation step, these HRP-conjugated antibodies will be bound to any antigen-antibody (IgG) complexes previously formed and the unbound HRP-conjugate is then removed by washing. Chromogen solutions containing Tetramethylbenzidine (TMB) and urea peroxide are added to the wells and in presence of the antigen-antibody-anti-IgG (HRP) immunocomplex; the colorless Chromogens are hydrolyzed by the bound HRP conjugate to a blue colored product. The blue color turns yellow after stopping the reaction with sulfuric acid. The amount of color intensity can be measured and is proportional to the amount of antibody captured in the wells, and to the sample respectively. Wells containing samples negative for anti-HEV-IgG remain colorless.

COMPONENTS

96 Tests

- **MICROWELL PLATE** 1 plate
Blank microwell strips fixed on white strip holder. The plate is sealed in aluminum pouch with desiccant.
8x12/12x8-well strips wells per plate.
Each well contains recombinant HEV antigens.
The microwell strips can be broken to be used separately. Place unused wells or strips in the plastic sealable storage bag together with desiccant and return to 2~8°C.
- **NEGATIVE CONTROL** 1 vial
Blue-colored liquid filled in a vial with green screw cap. 0.5ml per vial.
Protein-stabilized buffer tested non-reactive for anti-HEV. Preservatives: 0.1% ProClin 300.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **POSITIVE CONTROL** 1 vial
Red-colored liquid filled in a vial with red screw cap. 0.5ml per vial.
anti-HEV IgG antibodies diluted in protein-stabilized buffer Preservatives: 0.1% ProClin 300.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **SPECIMEN DILUENT** 1 vial
Green-colored liquid filled in a white vial with blue screw cap. 12ml per vial.
Protein-stabilized buffer, casein, and sucrose solution. Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **HRP-CONJUGATE REAGENT** 1 vial
Red-colored liquid filled in a white vial with red screw cap. 12ml per vial.
Horseradish peroxidase-conjugated rabbit anti-human IgG antibodies.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.

- **STOCK WASH BUFFER** 1 bottle
Colorless liquid filled in a clear bottle with white screw cap. 50ml per bottle.
PH 7.4 20 x PBS. (Contains Tween-20 as a detergent)
DILUTE BEFORE USE -The concentration must be diluted **1 to 20** with distilled/deionized water before use.
Once diluted, stable for one week at room temperature or for two weeks at 2-8°C.
- **CHROMOGEN SOLUTION A** 1 vial
Colorless liquid filled in a white vial with green screw cap. 7ml per vial.
Urea peroxide solution.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **CHROMOGEN SOLUTION B** 1 vial
Colorless liquid filled in a black vial with black screw cap.
TMB solution (Tetramethyl benzidine dissolved in citric acid). 7ml per vial.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **STOP SOLUTION** 1 vial
Colorless liquid filled in a white vial with yellow screw cap. 7ml per vial.
Diluted sulfuric acid solution (2.0M H₂SO₄).
Ready to use as supplied.
- **PLASTIC SEALABLE BAG** 1 unit
For enclosing the strips not in use.
- **CARDBOARD PLATE COVER** 2 sheets
To cover the plates during incubation and prevent evaporation or contamination of the wells.
- **PACKAGE INSERTS** 1 copy

ADDITIONAL MATERIALS AND INSTRUMENTS REQUIRED BUT NOT PROVIDED

- Freshly distilled or deionized water.
- Disposable gloves and timer.
- Appropriate waste containers for potentially contaminated materials.
- Disposable V-shaped troughs.
- Dispensing system and/or pipette (single or multichannel), disposable pipette tips.
- Absorbent tissue or clean towel.
- Dry incubator or water bath, 37±0.5°C.
- Microshaker for dissolving and mixing conjugate with samples.
- Microwell plate reader, single wavelength 450nm or dual wavelength 450nm and 630nm.
- Microwell aspiration/wash system.

SPECIMEN COLLECTION, TRANSPORTATION AND STORAGE

1. **Sample Collection:** Either fresh serum or plasma samples can be used for this assay. Blood collected by venipuncture should be allowed to clot naturally and completely – the serum/plasma must be separated from the clot as early as possible as to avoid hemolysis of the RBC. Care should be taken to ensure that the

serum samples are clear and not contaminated by microorganisms. Any visible particulate matters in the sample should be removed by centrifugation at 3000 RPM for at least 20 minutes at room temperature, or by filtration on 0.22µ filters. Plasma samples collected into EDTA, sodium citrate or heparin may be tested, but highly lipaemic, icteric, or hemolized samples should not be used as they could give erroneous results in the assay. Do not heat inactivate samples. This can cause sample deterioration.

2. **Transportation and Storage:** Store samples at 2-8°C. Samples not required for assaying within 3 days should be stored frozen (-20°C or lower). Multiple freeze-thaw cycles should be avoided. For shipment, samples should be packaged and labeled in accordance with the existing local and international regulations for transport of clinical samples and ethological agents.

SPECIAL INSTRUCTIONS FOR WASHING

1. A good washing procedure is essential to obtain correct and precise analytical data.
2. It is therefore recommended to use a good quality ELISA microplate washer, maintained at the best level of washing performances. In general, no less than 5 automatic washing cycles of 350-400µl/well are sufficient to avoid false positive reactions and high background.
3. To avoid cross-contaminations of the plate with sample or HRP-conjugate, after incubation do not discard the content of the wells but allow the plate washer to aspirate it automatically.
4. Anyway, we recommend calibrating the washing system on the kit itself in order to match the declared analytical performances. Assure that the microplate washer liquid dispensing channels are not blocked or contaminated and sufficient volume of Wash buffer is dispensed each time into the wells.
5. In case of manual washing, we suggest to carry out 5 cycles, dispensing 350-400µl/well and aspirating the liquid for 5 times. If poor results (high background) are observed, increase the washing cycles or soaking time per well.
6. In any case, the liquid aspirated out the strips should be treated with a sodium hypochlorite solution at a final concentration of 2.5% for 24 hours, before liquids are wasted in an appropriate way.

The concentrated Washing solution should be diluted **1 to 20** before use. For one plate, mix 50ml of the concentrate with 950ml of water for a final volume of 1000ml diluted Wash Buffer. If less than a whole plate is used, prepare the proportional volume of solution.

STORAGE AND STABILITY

The components of the kit will remain stable through the expiration date indicated on the label and package when stored between 2-8 °C, **do not freeze**. To assure maximum performance of this HEV-IgG ELISA kit, during

storage protect the reagents from contamination with microorganism or chemicals.

PRECAUTIONS AND SAFETY

This kit is intended **FOR RESEARCH USE ONLY**

The ELISA assay is a time and temperature sensitive method. To avoid incorrect result, strictly follow the test procedure steps and do not modify them.

1. Do not exchange reagents from different lots, or use reagents from other commercially available kits. The components of the kit are precisely matched as to achieve optimal performance during testing.
2. Make sure that all reagents are within the validity indicated on the kit box and are of the same lot. Never use reagents beyond the expiry date stated on reagents labels or on the kit box.
3. **CAUTION - CRITICAL STEP:** Allow the reagents and samples to stabilize at room temperature (18-30°C) before use.
4. Shake reagent gently before, and return to 2-8°C immediately after use.
5. Use only sufficient volume of sample as indicated in the procedure steps. Failure to do so may cause in low sensitivity of the assay.
6. Do not touch the bottom exterior of the wells; fingerprints or scratches may interfere with microwell reading.
7. When reading the results, ensure that the plate bottom is dry and there are no air-bubbles inside the wells.
8. Never allow the microplate wells to dry after the washing step. Immediately proceed to the next step. Avoid the formation of air-bubbles when adding the reagents.
9. Avoid assay steps long time interruptions. Assure same working conditions for all wells.
10. Calibrate the pipette frequently to assure the accuracy of samples/reagents dispensing. Always use different disposal pipette tips for each specimen and reagents as to avoid cross-contaminations. Never pipette solutions by mouth. The use of automatic pipettes is recommended.
11. Assure that the incubation temperature is 37°C inside the incubator.
12. When adding samples, avoid touching the well's bottom with the pipette tip.
13. When reading the results with a plate reader, it is recommended to determine the absorbance at 450nm or at 450nm with reference at 630nm.
14. All specimens from animal origin should be considered as potentially infectious.
15. Therefore, handle reagents and specimens with extreme caution as if capable of transmitting infectious diseases. Strict adherence to GLP (Good Laboratory Practice) regulations can ensure the personal safety. Never eat, drink, smoke, or apply cosmetics in the assay laboratory.
16. Bovine derived sera may have been used in this kit. Bovine serum albumin (BSA) and fetal calf sera (FCS) are derived from animals from BSE/TSE free-

geographical areas.

17. The pipette tips, vials, strips and sample containers should be collected and autoclaved for 1hour at 121°C or treated with 10% sodium hypochlorite for 30 minutes to decontaminate before any further steps for disposal.
18. The Stop solution (2M H₂SO₄) is a strong acid. Corrosive. Use it with appropriate care. Wipe up spills immediately or wash with water if come into contact with the skin or eyes. ProClin 300 used as a preservative can cause sensation of the skin.
19. The enzymatic activity of the HRP-conjugate might be affected from dust, reactive chemical, and substances like sodium hypochlorite, acids, alkalis etc. Do not perform the assay in the presence of such substances.
20. Materials Safety Data Sheet (MSDS) available upon request.
21. If using fully automated microplate processing system, during incubation, do not cover the plates with the plate cover. The tapping out of the remainders inside the plate after washing, can also be omitted.

ASSAY PROCEDURE

Step 1 Reagents preparation: Allow the reagents and samples to reach room temperature (**18-30°C**) for at least 15-30 minutes. Check the Wash buffer concentrate for the presence of salt crystals. If crystals have formed in the solution, resolubilize by warming at 37°C until crystals dissolve. Dilute the stock Wash Buffer **1 to 20** with distilled or deionized water. Use only clean vessels to dilute the Wash Buffer.

Step 2 Numbering Wells: Set the strips needed in strip-holder and number sufficient number of wells including three Negative control (**e.g. B1, C1, D1**), two Positive control (**e.g. E1, F1**) and one Blank (**e.g. A1**, neither samples nor HRP-Conjugate should be added into the Blank well). If the results will be determined by using dual wavelength plate reader, the requirement for use of Blank well could be omitted. Use only number of strips required for the test.

Step 3 Adding Diluent: Add **100µl** Specimen Diluent into each well.

Step 4 Adding Sample: Add **10µl** of Positive control, Negative control, and Specimen into their respective wells. **Note: Use a separate disposal pipette tip for each specimen, Negative Control, Positive Control to avoid cross-contamination.** Mix by tapping the plate gently.

Step 5 Incubating (1): Cover the plate with the plate cover and incubate for **30minutes at 37°C**. It is recommended to use thermostat-controlled water tank to assure the temperature stability and humidity during the incubation. If dry incubator is used, do not open the door frequently.

Step 6 Washing (1): At the end of the incubation, remove and discard the plate cover. Wash each well **5 times** with diluted Wash Buffer. Each time allow the microwells to soak for 30-60 seconds.

After the final washing cycle, turn the plate onto blotting paper or clean towel, and tap it to remove any remainders.

Step 7 Adding HRP-Conjugate: Add **100µl** HRP-Conjugate to each well except the Blank.

Step 8 Incubating (2): Cover the plate with the plate cover and incubate the plate for **30 minutes at 37°C**

Step 9 Washing (2): After the end of the incubation, remove and discard the plate cover. Wash each well **5times** with diluted Wash buffer as in **Step6**.

Step 10 Coloring: Dispense **50µl** of Chromogen A and **50µl** Chromogen B solution into each well including the **Blank** and mix by tapping the plate gently. Incubate the plate at **37°C for 10minutes avoiding light**. The enzymatic reaction between the Chromogen A/B solutions produces blue color in Positive control and anti-HEV/IgG positive sample wells.

Step 11 Stopping Reaction: Using a multichannel pipette or manually, add **50µl** Stop Solution into each well and mix gently by tapping the plate. Intensive yellow color develops in Positive control and anti-HEV/IgG positive sample wells.

Step 12 Measuring the Absorbance: Calibrate the plate reader with the Blank well and read the absorbance at **450nm**. If a dual filter instrument is used, set the reference wavelength at **630nm**. Calculate the **Cut-off** value and evaluate the results. (**Note:** read the absorbance within **5minutes** after stopping the reaction)

INTERPRETATION OF RESULTS AND QUALITY CONTROL

Each microplate should be considered separately when calculating and interpreting results of the assay, regardless of the number of plates concurrently processed. The results are calculated by relating each sample's optical density (OD) value to the Cut-off value (C.O.) of the plate. If the Cut-off reading is based on single filter plate reader, the results should be calculated by subtracting the Blank well OD value from the print report values of samples and controls. In case the reading is based on Dual filter plate reader, do not subtract the Blank well OD from the print report values of samples and controls.

1. Calculation of Cut-off value (C.O.) = *NC + 0.16

*NC = the mean absorbance value for three negative controls.

Important: If the mean OD value of the negative control is lower than **0.03**, take it as **0.03**. If higher than **0.03** see the Quality Control Range.

Example:

1. Calculation of NC:

| | | | |
|----------------------------|------|-------|-------|
| Well No | B1 | C1 | D1 |
| Negative controls OD value | 0.02 | 0.012 | 0.016 |

NC=0.016 (NC is lower than 0.03 so take it as 0.03)

2. Calculation of Cut-off (C.O.)= 0.03 + 0.16= 0.190

If one of the Negative control values does not meet the Quality control range specifications, it should be discarded and the mean value is calculated again using the remaining two values. If more than one negative control OD value does not meet the Quality control range specifications, the test is invalid and must be repeated.

2. Quality control range:

The test results are valid if the Quality Control criteria are verified. It is recommended that each laboratory must establish appropriate quality control system with quality control material similar to or identical with the patient sample being analyzed.

1. The OD value of the Blank well, which contains only Chromogens and Stop solution, is less than 0.080 at 450 nm.
2. The OD value of the Positive control must be equal to or greater than 0.800 at 450/630nm or at 450nm after blanking.
3. The OD value of the Negative control must be less than 0.100 at 450/630nm or at 450nm after blanking.

3. Interpretations of the results:

(S = the individual absorbance (OD) of each specimen)

Negative Results (S/C.O. < 1) : samples giving absorbance less than Cut-off value are negative for this assay, which indicates that no IgG-class antibodies to hepatitis E virus have been detected with this anti-HEV/IgG ELISA kit

Positive Results (S/C.O. ≥1) : samples giving an absorbance greater than or equal to the Cut-off value are considered initially reactive, which indicates that IgG-class antibodies to hepatitis E virus have probably been detected using this ELISA kit. Retesting in duplicates of any reactive sample is recommended. Repeatedly reactive samples could be considered positive for IgG-class antibodies to HEV

Borderline (S/C.O. =0.9-1.1) : Samples with absorbance to Cut-off ratio between 0.9 and 1.1 are considered borderline and retesting of these samples in duplicates is recommended to confirm the results. Repeatedly positive samples could be considered positive for HEV.

LIMITATIONS

1. Non-repeatable positive result may occur due to the general biological characteristics of ELISA assays. The assay is designed to achieve very high performance characteristics of sensitivity and specificity and the "indirect model" minimizes the unspecific reactions, which may occur due to interference between unknown meters in sample and the rabbit anti-human IgG used as a conjugate. Antibodies may be undetectable during the early stages of the disease and in some immunosuppressed individuals.

2. If, after retesting of the initially reactive samples, the assay results are negative, these samples should be considered as non-repeatable (false positive) and interpreted as negative. As with many very sensitive ELISA assays, false positive results can occur due to the several reasons, most of which are related but not limited

to inadequate washing step.

3. Common sources for mistakes: Kits beyond the expiry date, bad washing procedures, contaminated reagents, incorrect assay procedure steps, insufficient aspiration during washing, failure to add samples or reagents, equipment, timing, volumes, sample nature and quality.

4. The prevalence of the marker will affect the assay's predictive values.

5. This is a qualitative assay and the results cannot be used to measure antibodies concentrations.

INDICATIONS OF INSTABILITY OR DETERIORATION OF THE REAGENTS

1. Values of the Positive or Negative controls, which are out of the indicated Quality control range, are indicator of possible deterioration of the reagents and/or operator or equipment errors. In such case, the results should be considered as invalid and the samples must be retested. In case of constant erroneous results classified as due to deterioration or instability of the reagents, immediately substitute the reagents with new ones.
2. If after mixing of the Chromogen A and B solutions into the wells, the color of the mixture turns blue within few minutes, do not continue carrying out the testing and replace the reagents with fresh ones.

VALIDITY

Please do not use this kit beyond the expiry date indicated on the kit box and reagent labels.

REFERENCES:

1. Reyes GR, Purdy MA, Kim JP, et al. Isolation of a cDNA from the virus responsible for enterically transmitted non-A, non-B hepatitis. *Science* 1990; 247: 1335–1339.
2. Clayson E, Innis B, Myint K, et al. Detection of hepatitis E virus infections among domestic swine in the Kathmandu Valley of Nepal. *Am J Trop Med Hyg*, 1995, 53:228–232.
3. Meng XJ, Purcell RH, Halbur PG, et al. A novel virus in swine is closely related to the human hepatitis E virus. *Proc Natl Acad Sci USA*, 1997, 94: 9860–9865.
4. Tei S, Kitajima N, Takahashi K, et al. Zoonotic transmission of hepatitis E virus from deer to human beings. *Lancet* 2003; 362(9381):371
5. Zheng YJ, Zhang J, Xia NS. A debate about that hepatitis E is a zoonosis. *Chinese J Zoonosis* (in press)
6. Wang YC, Zhang HY, Xia NS, et al. Prevalence, Isolation, and Partial Sequence Analysis of Hepatitis E Virus From Domestic Animals in China. *J Med Virol* 2002, 67:516–521

Express Biotech International
P.O. BOX 458
Thurmont, MD 21788 USA
Tel: 301-228-2444 Fax: 301-560-6570
Toll Free: 888-562-8914
www.xpressbio.com info@xpressbio.com

