

Ansur ESA612

Plug-In

Users Manual

Table of Contents

Chapter	Title	Page
1	Introduction	1-1
	About This Manual	1-3
	Ansur Software	1-3
	Ansur Plug-Ins	1-3
	ESA612 Plug-In	1-3
	Test Elements	1-4
	Additional References	1-6
	Software Updates	1-6
	Terms and Abbreviations	1-7
2	Getting Started	2-1
	Introduction	2-3
	System Requirements	2-3
	Installing the ESA612 Plug-In	2-3
	Entering License Key	2-4
	UnInstalling the Plug-In	2-5
	Configuring the Plug-In for AAMI Safety Tests	2-5
	Selecting AAMI Nomenclature	2-6
	Selecting AAMI Terms	2-6
3	ESA612 Tests.....	3-1
	Introduction.....	3-3
	Performing a Safety Test	3-3
	Load a Test Template	3-3
	Start the Template.....	3-4
	Define Patient Modules	3-6
	Connect the ESA612	3-8
	Perform the Electrical Safety Tests	3-9
	Storing a Test Record or Template	3-12
	Printing a Test Report	3-12
	Creating an ESA612 Safety Test Template	3-14
	Adding Limits to a Safety Test	3-15

4	ESA612 Test Templates	4-1
	Introduction.....	4-3
	Power Change Delay	4-3
	Power Change Pause.....	4-5
	Multiple Earth Resistance Tests.....	4-6
	Testing Ultrasound Probes Example.....	4-6
	Create a Test Template	4-6
	Testing the Ultrasound Probes.....	4-8
	Basic Test Reports	4-9
5	Reference	5-1
	Introduction.....	5-3
	Electrical Safety Test Elements	5-3
	Custom Setup.....	5-3
	Expected Results	5-3
	Test Groups.....	5-3
	Custom Setup.....	5-4
	Test Options	5-4
	Test Each Module Separately.....	5-4
	Test Guide Settings	5-4
	Patient Lead Tests	5-4
	MAP Settings	5-5
	Patient Lead Configuration – Module Setup.....	5-5
	Value Filter	5-5
	Modular Filter	5-5
	Expected Results	5-5
	Auto Sequence	5-5
	Custom Setup.....	5-6
	Test Options	5-6
	Delay Times	5-6
	Test Each Module Separately.....	5-6
	Test Guide Settings	5-6
	Patient Lead Tests	5-6
	Patient Lead Configuration – Module Setup.....	5-7
	Value Filter	5-7
	Module Filter.....	5-7
	Expected Results	5-7
	Test Guide.....	5-7
	Module Setup.....	5-8
	Module Groups.....	5-9
	Module Nomenclature.....	5-10
	Lead Calibration	5-10
	Halt on Test Failure	5-10
	Plug-In Preferences.....	5-11

List of Tables

Table	Title	Page
1-1.	ESA612 Plug-In Electrical Safety Tests	1-4
1-2.	Terms and Abbreviations	1-7
3-1.	ESA612 International Standards	3-16
5-1.	Toolbar Buttons.....	5-7

List of Figures

Figure	Title	Page
1-1.	Electrical Safety Tests in Ansur's Test Explorer	1-4
2-1.	ESA612 Plug-In Installation File Location	2-4
2-2.	Ansur Registration Screen - License Key	2-5
2-3.	Removing ESA612 Plug-In.....	2-5
2-4.	Selecting AAMI Nomenclature.....	2-6
2-5.	Selecting AAMI Terms	2-7
3-1.	Open Dialog Box.....	3-3
3-2.	Browsing the Ansur Test Library.....	3-4
3-3.	Selecting a Ready-Made Test Template.....	3-4
3-4.	Starting a Test Template	3-5
3-5.	Test Initialization Window.....	3-5
3-6.	IEC 60601-1 Test Guide	3-6
3-7.	Module Setup Window	3-6
3-8.	Sample Module Setup Screen	3-7
3-9.	1210 Adapter Connections.....	3-7
3-10.	Module Setup for 1210 Adapter.....	3-8
3-11.	Setup Description with 1210 Adapter	3-8
3-12.	Connecting Instructions for Applied Parts	3-9
3-13.	Completed Safety Test Window	3-10
3-14.	Test Summary Window.....	3-10
3-15.	ESA612 Test Record.....	3-11
3-16.	Save As Dialog Box	3-12
3-17.	Print Report Window	3-13
3-18.	Print Preview of Test Report.....	3-13
3-19.	Test Explorer Window	3-14
3-20.	Auto Sequence added to Template.....	3-14
3-21.	Adding Electrical Safety Tests to the Auto Sequence.....	3-15
3-22.	Complete IEC 60601-1 Safety Test	3-15
3-23.	Apply When Tab	3-16
3-24.	Selecting Standards	3-17
4-1.	Custom Setup Tab	4-3
4-2.	Editing Power On/Off Delay Times.....	4-3
4-3.	Power On Delay Message	4-4
4-4.	Power-Off Delay Message with After Power Off Selected	4-4
4-5.	Power-Off Delay Message with Before Power Off Selected	4-5

4-6.	Stop for Power Configuration Change Message	4-5
4-7.	Multiple Protective Earth Resistance Message	4-6
4-8.	Configuring Module Testing	4-7
4-9.	Creating Module Groups	4-7
4-10.	Module Group with Three Ultrasound Probes	4-8
4-11.	Connect Applied Parts Window	4-9
4-12.	Add an Earth Leakage Current Test	4-9
4-13.	Assign Maximum Value to a Test Group	4-10
4-14.	Remove Test from a Report	4-10
4-15.	Maximum Value Assigned to the Test Group	4-11
4-16.	Test Record Showing Value Filtered Report	4-11
5-1.	Expected Results with No Applied Parts	5-3
5-2.	Expected Results with Applied Parts	5-3
5-3.	Test Group Custom Setup Window	5-4
5-4.	Module Information	5-5
5-5.	Auto Sequence Custom Setup Page	5-6
5-6.	ESA612 Test Guide Window	5-7
5-7.	Module Setup Window	5-8
5-8.	Modules in a Module Setup Window	5-9
5-9.	Applied Parts Connection Diagram	5-9
5-10.	Module Groups Window	5-9
5-11.	Selecting Module Nomenclature	5-10
5-12.	Test Lead Calibration Window	5-10
5-13.	Halt on Test Failure Window	5-11
5-14.	ESA612 Plug-In Preferences: Module Nomenclature	5-12

Chapter 1

Introduction

Title	Page
About This Manual	1-3
Ansur Software	1-3
Ansur Plug-Ins	1-3
ESA612 Plug-In	1-3
Test Elements	1-4
Additional References	1-6
Software Updates.....	1-6
Terms and Abbreviations	1-7

About This Manual

This Users Manual is designed to assist the reader in using the Ansur ESA612 Plug-In with Ansur software. The manual covers all features specific to the Plug-In. Familiarity with both Ansur software and Microsoft Windows and their features will help in the design and use of tests for the Ansur ESA612 Analyzer. The manual is divided into the following chapters:

Chapter 1 “Introduction” provides information on Ansur software and the ESA612 Plug-In.

Chapter 2 “Getting Started” provides information on how to install and configure the ESA612 Plug-In.

Chapter 3 “ESA612 Tests” provides step-by-step descriptions on how to perform the basic tasks of the ESA612 Plug-In.

Chapter 4 “ESA612 Test Templates” contains information on creating highly efficient safety test procedures using the ESA612 Plug-In features.

Chapter 5 “Reference” contains details about every ESA612 Plug-In feature.

Ansur Software

Ansur Test Automation software is the foundation for all Fluke Biomedical test systems. Ansur manages test procedures by allowing both manual and visual automated test sequences.

The software works hand-in-hand with Fluke Biomedical analyzers and simulators, creating a seamless integration for:

- Visual inspections
- Preventive maintenance
- Work procedures
- Performance tests
- Electrical safety tests

Ansur Plug-Ins

Ansur Test Executive software utilizes Plug-In modules that work with a wide array of Fluke Biomedical instruments. The Plug-In module is a software interface that provides test elements to the Ansur Test Executive program. This scheme allows the use of a similar user interface for all analyzers and simulators supported by Ansur.

With the purchase of a new Fluke Biomedical analyzer or simulator, it is possible to update existing Ansur software by installing a new Plug-In. Each Plug-In module allows users to work with only the options and capabilities needed for the instrument under test.

ESA612 Plug-In

The Ansur ESA612 Plug-In provides remote access to the ESA612 Electrical Safety Analyzer, referred to throughout this document as the “Analyzer.” In addition to the general test plug-ins, specialized plug-ins address all test requirements for specific instruments.

Note

The ESA612 Electrical Safety Analyzer Users Manual explains the Analyzer’s capabilities and use.

Create and use Ansur test procedures with Ansur ESA612 test elements to incorporate the capabilities of an Analyzer into automated testing. Users can customize tests to analyze specific performance requirements. There are unique test elements for each of the tests, and simulations typically run on the Analyzer.

Test Elements

The ESA612 electrical safety tests shown in Figure 1-1 are installed to Ansur’s test explorer during the ESA612 Plug-In installation.




Figure 1-1. Electrical Safety Tests in Ansur's Test Explorer

gbv01.bmp

Table 1-1 lists all the ESA612 Plug-In electrical safety tests.

Table 1-1. ESA612 Plug-In Electrical Safety Tests

Electrical Safety Test	Configuration	Unit	Applied Parts
Mains Voltage	Live to Neutral	V	No
	Neutral to Earth	V	No
	Live to Earth	V	No
Protective Earth Resistance		Ω	No
Equipment Current		A	No
Point to Point Voltage		V	No
Point to Point Resistance		Ω	No
Point to Point Leakage		μA	No

Table 1-1. ESA612 Plug-In Electrical Safety Tests (cont.)

Electrical Safety Test	Configuration	Unit	Applied Parts
Insulation Resistance	Mains to Protective Earth	MΩ	No
	Applied Parts to Protective Earth	MΩ	All
	Mains to Applied Parts	MΩ	All
	Mains to Non-Earth Accessible Conductive Part	MΩ	No
	Applied Parts to Non-Earth Accessible Conductive Part	MΩ	All
Earth Leakage Current	Normal Condition	μA	All or No
	Open Neutral	μA	All or No
	Normal Condition, Reversed Mains	μA	All or No
	Open Neutral, Reversed Mains	μA	All or No
Enclosure Leakage Current	Normal Condition	μA	All or No
	Open Neutral	μA	All or No
	Open Earth	μA	All or No
	Normal Condition, Reversed Mains	μA	All or No
	Open Neutral, Reversed Mains	μA	All or No
	Open Earth, Reversed Mains	μA	All or No
Patient Leakage Current	Normal Condition	μA	Yes
	Open Neutral	μA	Yes
	Open Earth	μA	Yes
	Normal Condition, Reversed Mains	μA	Yes
	Open Neutral, Reversed Mains	μA	Yes
	Open Earth, Reversed Mains	μA	Yes
Mains on Applied Parts	Shorted Mains	μA	Yes
	Reversed Mains	μA	Yes
Patient Auxiliary Current	Normal Condition	μA	Yes
	Open Neutral	μA	Yes
	Open Earth	μA	Yes
	Normal Condition, Reversed Mains	μA	Yes
	Open Neutral, Reversed Mains	μA	Yes
	Open Earth, Reversed Mains	μA	Yes
Alternative Equipment Leakage	Closed Earth	μA	All
	Open Earth	μA	All

Table 1-1. ESA612 Plug-In Electrical Safety Tests (cont.)

Electrical Safety Test	Configuration	Unit	Applied Parts
Differential Leakage	Normal Condition	μA	All
	Open Neutral	μA	All
	Open Earth	μA	All
	Normal Condition, Reversed Mains	μA	All
	Open Neutral, Reversed Mains	μA	All
Direct Equipment Leakage	Normal Condition	μA	All
	Open Earth	μA	All
	Normal Condition, Reversed Mains	μA	All
	Open Earth, Reversed Mains	μA	All
Direct Applied Part Leakage	Normal Condition	μA	Yes
	Normal Condition, Reversed Mains	μA	Yes

The ESA612 Plug-In provides three types of test elements: Auto Sequence, Test Groups, and Electrical Safety Tests.

Auto Sequence is a test container containing any of the ESA612 test elements. An Auto Sequence will execute all tests contained within it. An Auto Sequence also provides additional test setup such as power on/off delay time and test optimization.

Test Groups are specialized Auto Sequences, one for each group of safety test (e.g. Earth Leakage Current).

Electrical safety tests are indicated in the Test Explorer window with a light-blue icon. Each safety test element corresponds to a measurement available in the ESA612 Safety Analyzer.

Additional References

In addition to this manual, answers to questions using the Analyzer or PC may be found in the following sources:

- *Fluke Biomedical ESA612 Users Manual*
- *Fluke Biomedical Ansur Test Executive Users Manual*
- Microsoft Windows Help and Support Center

Software Updates

Updates for Ansur are published on the Fluke Biomedical website, <http://www.flukebiomedical.com>.

Terms and Abbreviations

Table 1-2 lists terms and abbreviations used in this manual.

Table 1-2. Terms and Abbreviations

Term	Description
Ansur	Ansur is a software suite using plug-ins to perform test and inspection procedures in conjunction with several Fluke Biomedical test instruments.
DUT	Device Under Test—the equipment subjected to a test using the Analyzer
DUT Info	Information used to identify one particular DUT. DUT info usually consists of a serial number, manufacturer, device type and model. Ansur also adds a few extra data fields such as location and status.
ESA612	Electrical Safety Analyzer from Fluke Biomedical
Field User	The person using Ansur to perform a test template on a DUT.
Plug-In	Add-on software program that extends Ansur so that it can interface with a specific Fluke Biomedical test instrument to configure it for a specific test and to automatically collect the measured data (if applicable)
Test Container	A test container is a test element that can contain other test elements. The ESA612 Auto Sequence is a test container
Test Element	An Ansur construct that encapsulates test configuration and results A test template is built of several test elements.
Test Guide	A window displayed by Ansur or any of its plug-ins when a test element is being performed
Test Record	An Ansur file containing the results of a performed test template The test record can be printed as a test report.
Test Template	An Ansur file containing a set of test elements that define how a particular DUT is to be tested A test template can also contain instructions on how to perform service, preventive maintenance, repair, and other tasks on a DUT.

Chapter 2

Getting Started

Title	Page
Introduction.....	2-3
System Requirements	2-3
Installing the ESA612 Plug-In.....	2-3
Entering License Key	2-4
UnInstalling the Plug-In	2-5
Configuring the Plug-In for AAMI Safety Tests	2-5
Selecting AAMI Nomenclature	2-6
Selecting AAMI Terms	2-6

Introduction

This chapter describes installation of the ESA612 Plug-In and its use together with the Ansur Test Automation software and the Analyzer.

Note

An Analyzer is not necessary to create test templates and experiment with the functionality available in Ansur and the ESA612 Plug-In. However, actual tests cannot be performed unless the Analyzer is connected to the computer.

System Requirements

The following are recommended minimum requirements for installation:

- IBM PC/XT-compatible Pentium II 350 MHz or faster processor
- Microsoft Windows 2000 or Windows XP or Vista operating system
- Fluke Biomedical Ansur V2.8.2 or newer
- 50 MB of available hard drive for software
- Hard drive space (from 100 k to several megabytes) for result and template files

Installing the ESA612 Plug-In

The ESA612 Plug-In must be installed on the computer before the features described in this user manual can be implemented. For information on obtaining the Ansur software and the ESA612 Plug-In, contact the local Fluke Biomedical representative or visit the Fluke Biomedical website (<http://www.flukebiomedical.com>).

Note

Ansur version 2.7.1 or newer must be installed before the ESA612 Plug-In can be installed and used.

Download the ESA612 Plug-In from the Fluke Biomedical website and follow the steps below:


Note

When downloading the ESA612 Plug-In from the Fluke Biomedical web site, it is possible to run the installation without first downloading.

Note

When installing Ansur or any of its components/plugin on computers running Microsoft Vista, it is important to perform the installation as the Administrator for that computer. Otherwise the registry will not be properly updated and Ansur will not work properly. For installing on Windows Vista, you must first download the file to your local computer, then locate the installation file, right click and select "Run as Administrator."

1. Open **Windows Explorer** and browse to the saved ESA612 Plug-In installation program file, usually named **Ansur ESA612 Plug-In Vn.n.n.exe**, where *n.n.n* is the Plug-In version number. See Figure 2-1.



gbv02.bmp

Figure 2-1. ESA612 Plug-In Installation File Location

2. Double-click the installation program. The installation extracts the Plug-In elements and displays the **Welcome** dialog box.
3. Click **Next** to display the license agreement.
4. Select the checkbox for “**I accept the terms in the license agreement,**” and click **Next** to display the default destination folder.
5. Choose one of the following options:
 - Click **Next** to accept the default destination folder in which Ansur was installed.
 - Click **Change** if Ansur has been installed in a different folder. In this case, the destination folder for the Plug-In is changed so that it resides in the same directory as the Ansur program.

Note

If Ansur has been installed in a different destination folder from the default, be sure to use the same folder for the ESA612 Plug-In.

6. Click **Install** to begin the installation. A progress bar indicates the status of the Plug-In installation.
After a few minutes, the installation concludes, and the window displays the dialog box and the **Finish** button.
7. Click **Finish**. The Plug-In will load when Ansur is started.

Entering License Key

When using the Plug-In for the first time, the user is prompted to enter a software license key provided by Fluke Biomedical at the time of purchase.

Note

Test templates can be created without a license key by using the demonstration mode. Demonstration mode allows many of the tasks described in this user manual. However, a user may not save or print without licensing the Plug-In.

1. Start Ansur by doing one of the following:
 - Double-click the **Ansur** icon on the desktop.
 - From the **Start** menu, select **Start | Programs | Fluke | Ansur**.

Note

The license key dialog box shown in Figure 2-2 appears at startup if a license key has not yet been entered for the Plug-In.




Figure 2-2. Ansur Registration Screen - License Key

gbv04.bmp

2. Enter the **Establishment Name** and the Plug-In **license key**. If a license key is not available, click the **Demo** button to start Ansur in demonstration mode.

Note

Because the license key is derived from the establishment name, both strings must match the license information provided by Fluke Biomedical. This information is case sensitive and space sensitive. If the establishment name has been entered in the past, this field is already filled in

3. Click **OK** to start Ansur.
4. Click **Cancel** to prevent the Plug-In from being loaded.

Uninstalling the Plug-In

To uninstall the ESA612 Plug-In:

1. Select **Start | Control Panel** and double-click **Add or Remove Programs**.
2. Locate and select the entry named **Ansur ESA612 Plug-in**.




Figure 2-3. Removing ESA612 Plug-In

gbv08.bmp

3. With the entry highlighted, click the **Remove** button as shown in Figure 2-3.
4. When asked to verify the removal, click **Yes**. A dialog box with a progress bar displays while the ESA612 Plug-In is being removed from the computer.

When the Plug-In is no longer listed in the **Add or Remove Programs** window, it has been completely removed.

Configuring the Plug-In for AAMI Safety Tests

At installation, the ESA612 Plug-In is configured for IEC nomenclature patient lead classification names. Configuring the Plug-In for AAMI testing requires changing two settings: Module Nomenclature and User Interface Language.

Module nomenclature determines patient lead classification and how they are tested. IEC uses protection classes Body, Body Float, and Cardiac Float for patient lead while AAMI

uses protection classes Non-Isolated and Isolated.

IEC terms are used for the English (default) user interface language. AAMI terms are used for the English US language.

Note

This manual uses the IEC graphics and terms. Changing the configuration to AAMI will cause some displays to appear on your PC that are different than those in this manual.

Selecting AAMI Nomenclature

To change the Plug-In configuration to AAMI:

1. Start Ansur and click **Tools | Options** to display the preferences window.
2. Click **ESA612 Safety Analyzer** to display the ESA612 preferences.
3. Click on the **Nomenclature** tab to display the Module Nomenclature page.
4. Click **AAMI Terms** to highlight AMMI terms and the description. See Figure 2-4.




Figure 2-4. Selecting AAMI Nomenclature

gbv05.bmp

5. Click on **OK** to close the preference window.

Now AAMI is the default nomenclature when defining patient leads.


Note

Changing module nomenclature will not be affect the limits applied to the electrical safety tests.

Selecting AAMI Terms

To select AAMI terms for tests:

1. Start Ansur and click **Tools | Options** to display the preferences window.
2. Click Ansur Preferences and then click on the Language tab to display the User Interface Language page.
3. Select English US from the list. See Figure 2-5.



gbv06.bmp

Figure 2-5. Selecting AAMI Terms

4. Click on OK to close the preferences window.

Note

Ansur must be restarted for the language changes to take effect.

After restarting Ansur, the ESA612 Plug-In will be using AAMI terms for electrical safety tests and all test related windows.

Chapter 3

ESA612 Tests

Title	Page
Introduction.....	3-3
Performing a Safety Test	3-3
Load a Test Template	3-3
Start the Template.....	3-4
Define Patient Modules	3-6
Connect the ESA612	3-8
Perform the Electrical Safety Tests	3-9
Storing a Test Record or Template	3-12
Printing a Test Report	3-12
Creating an ESA612 Safety Test Template	3-14
Adding Limits to a Safety Test	3-15

Introduction

This chapter describes the operation of the Ansur ESA612 Plug-In program.

Performing a Safety Test

Several ready-to-use test templates are copied to the Ansur Test Library folder when the Plug-In was installed. These templates contain complete electrical safety tests for the most common ESA612 supported standards.

Note

*The supplied test templates can be used as a starting point when creating new customized test templates. Load a template, alter it by adding visual inspections and editing the safety test for the specific need and then use **Save As** to give it another name.*

Load a Test Template

To load a test template:

1. Click **File** → **Open** on the menu bar or click on the Open Template toolbar button (📁) to open the dialog box shown in Figure 3-1.





Figure 3-1. Open Dialog Box

fcz09.bmp


2. Browse to the folder where Ansur was installed and double-click the Ansur Test Library folder. See Figure 3-2. The normal location is C:\Program Files\Fluke but may have a different name depending on the operating system language.



gbv10.bmp

Figure 3-2. Browsing the Ansur Test Library

3. Double-click the ESA612 folder to open a dialog box that lists the folder names of the most common medical standards and regulator boards as shown in Figure 3-3.



gbv11.bmp

Figure 3-3. Selecting a Ready-Made Test Template

4. Double-click the folder containing the desired standard and safety test.
5. Click a file name and then click the **Open** button, or double-click the file name to load the selected template and display its contents.

Start the Template

After loading the template, start a test by doing the following:

1. Click **Test** → **Start Test** or click the Start Test toolbar button (▶▶) to display test initialization window as shown in Figure 3-4.




Figure 3-4. Starting a Test Template

gbv13.bmp

Note

When using Ansur version 2.3.2 or newer, the test initialization window will not be displayed. Instead Ansur will automatically select the standards used in the template and go directly to the first test guide window. Continue to the Define Patient Modules section.

2. Select the standard from the list in the Test Initialization window shown in Figure 3-5 that corresponds with the selected test template. For example, if IEC 60601 test template is opened, then select 60601 standard and deselect the Ansur User Defined standard that is selected by default.





Figure 3-5. Test Initialization Window

gbv14.bmp

3. Click the Next toolbar button to go to the first Test Guide window shown in Figure 3-6.




gbv15.bmp

Figure 3-6. IEC 60601-1 Test Guide

Define Patient Modules

The ESA612 Safety Analyzer can test applied parts from different modules in the same test run. A patient module can be ECG electrodes, SpO2 sensors, temperature sensors, and others. To define a patient module:

1. Click the **Module Setup** tab on the left-hand side of the Test Guide to open the Module Setup page as shown in Figure 3-7.



gbv16.bmp


Figure 3-7. Module Setup Window

Note

If AAMI module nomenclature is selected, the module class options will be different than those shown in Figure 3-7.

2. To define a new module, click the plus button (+) next to the page title. A module labeled New is added to the list of modules and a table with module information fields is displayed in the right-hand frame.
3. Select the appropriate module class, the number of patient leads, and enter a module


code for each module tested. Figure 3-8 shows a sample module for a patient monitor.



gbv17.bmp

Figure 3-8. Sample Module Setup Screen

When the ESA612 Safety Analyzer is used with the 1210 adapter to expand the number of possible connections to fourteen, the Module Setup screen is used to define these connections. Figure 3-9 shows the Safety Analyzer connected to a Defibrillator/Pacer unit with the ten ECG leads connected through the 1210 adapter to the RA input.



fis120.eps

Figure 3-9. 1210 Adapter Connections

Figure 3-10 shows how three modules are used to define these connections in the Ansur Plug-In. The first module labeled 12-Lead ECG defines the 10 connections to the RA input of the Safety Analyzer. The other two modules for the paddles and pacer connections are defined separately.




Figure 3-10. Module Setup for 1210 Adapter

gbv101.bmp

Figure 3-11 shows a description of the test set-up using the 1210 adapter as part of an auto sequence template.





Figure 3-11. Setup Description with 1210 Adapter

gbv102.bmp

Connect the ESA612

The ESA612 Plug-In gives instructions on how to connect the ESA612 to the DUT's applied parts. To see the connection instructions:

1. Click the Procedure tab on the left-hand side of the test guide to display the connection procedure. An applied parts connection guide is shown in Figure 3-12.



gbv18.bmp

Figure 3-12. Connecting Instructions for Applied Parts

2. Connect the DUT to the safety analyzer as shown in Figure 3-12. Ensure the connections are made per the connection instructions.
3. Connect a USB printer cable between a USB port on your computer and the computer port on the ESA612.
4. Power on the ESA612 Safety Analyzer and the DUT.

Perform the Electrical Safety Tests

To run the safety test:

1. Click the Start Test button (🔵) on the toolbar of the Test Guide.

A progress bar will display in the lower-right corner of the Test Guide window while the ESA612 Plug-In searches for the ESA612 Safety Analyzer. The safety test will start when a connection is established. As each test is performed, it is highlighted in the results frame. When the safety analyzer determines the test result, the results are displayed next to the test.

Note

If the ESA612 is not connected to COM1 on the PC, Ansur displays an Instrument Not Found window where the port the ESA612 is connected to can be entered. The correct port number is displayed in the Windows Device Manager. Look for the USB Serial Port entry under "Ports."

Once the PC locates which port the ESA612 is connected to, Ansur remembers the port number and uses it as the default port for electrical safety tests.

Depending on how the loaded template is configured, different instructions may be displayed while the safety test is running. During a normal test run, instructions to calibrate the test leads or change the applied part connections may appear. If a safety test fails, a message appears asking how the failure is to be handled. See "Halt on Test Failure" in Chapter 5 (Reference).

When the test is completed, the status bar in the lower-right corner of the Test Guide is updated, the progress bar disappears, and the Start Test button is no longer

highlighted. Figure 3-13 is a sample of a completed safety test window.




Figure 3-13. Completed Safety Test Window

gbv20.bmp

2. Click the Next button (➡) to open the Test Summary window as shown in Figure 3-14.




Figure 3-14. Test Summary Window

gbv21.bmp

3. To create a Test Record, click the Next button (➡). The test summary window closes and a test record is opened in the main Ansur window as shown in Figure 3-15.




Figure 3-15. ESA612 Test Record

gbv22.bmp

Note

Plug-In Test Setup is displayed in the test record when the user selects the test element.

Storing a Test Record or Template

Each test record and test template created with the Plug-In can be saved to a file on the PC. To save a test record or template:

1. Click File Save on the main menu or click the save toolbar button (📁). The Save As dialog box opens as shown in Figure 3-16.




Figure 3-16. Save As Dialog Box

gbv24.bmp

2. Browse to the correct location and enter a filename in the file name field and click the Save button or press the Enter key. The file is saved and the Ansur title bar is updated with the new name.

Printing a Test Report

Test reports can be printed in three different formats from the test record:

- **Summary** – A report containing DUT information and the overall status of the safety test. This one-page report is good for a simple Pass or Fail report.
 - **Condensed** – A report containing the summary report information as well as each test element on a single line showing whether the test passed or failed.
 - **Detailed** – A report containing the summary and condensed information along with all configuration data, test result data, and module setup data.
1. With the Test Report open, click File Print or click the toolbar Print button to open the Print Report window shown in Figure 3-17.




Figure 3-17. Print Report Window

gbv26.bmp

2. Select the desired print type and click the Print button to display the report in the preview window as shown in Figure 3-18.




Figure 3-18. Print Preview of Test Report

gbv27.bmp

3. Print the report by clicking on the Print button (🖨️) in the left-hand window border.

Note

To skip the preview before printing, uncheck the Show Preview checkbox in the print report window shown in Figure 3-17.

Creating an ESA612 Safety Test Template

A test template is an Ansur file containing a set of test elements that define how a particular DUT is to be tested. Once created, the test template can be used to test other DUTs of the same type. A test template can also contain instructions on how to perform service, preventive maintenance, repair, and other tasks on a DUT.

To create a new test template:

1. Open a blank test template by clicking on **File** → **New** → **Template** or click the New Template button (📄) on the main toolbar.
2. In the test explorer window, click on the plus button next to the ESA612 Safety Analyzer entry to expand the electrical safety test entries for this analyzer. See Figure 3-19.




Figure 3-19. Test Explorer Window

gbv29.eps

To create a safety test made up of several measurements that are executed sequentially, an auto sequence test element must be created to hold the tests.

3. Click and hold on the Auto Sequence entry in the test explorer window and drag and release it in the template window. A copy of the Auto Sequence should appear in the template window as shown in Figure 3-20. The Auto Sequence entry is a holder of tests that are executed when called.




Figure 3-20. Auto Sequence added to Template

gbv29.eps

- To add tests to the Auto Sequence, click and hold on the desired test in the Test Explorer window and drag it to directly over the Auto Sequence entry in the Template Window and release it. The Mains Voltage test is added to the Auto Sequence shown in Figure 3-21.




Figure 3-21. Adding Electrical Safety Tests to the Auto Sequence

gbv30.bmp

Note

If a test is dragged and released UNDER the Auto Sequence entry and not on it, the test will not be performed as a part of the Auto Sequence. It will be performed when the Auto Sequence is completed displaying its own Test Guide where the operator will have to click the Play button in order to perform the test.

Repeat step 4 for each test that should be included in the Auto Sequence. A complete IEC 60601-1 safety test will look similar to the test template shown in Figure 3-22.




Figure 3-22. Complete IEC 60601-1 Safety Test


gbv31.bmp

Adding Limits to a Safety Test

The ESA612 Safety Analyzer performs electrical safety tests in accordance with several major medical standards. A Test Template allows the selection of limits for the different electrical safety tests. It is also possible to define your own test limits. For example, check that a specific DUT maintains its factory specifications.

To select an appropriate standard for a test:

- Select a safety test in the template window.
- Click the **Apply When** tab below the list of test elements as shown in Figure 3-23.



gbv32.bmp

Figure 3-23. Apply When Tab

3. Click on the standard(s) to which you want to add a limit value. The list of standards shown in Figure 3-24 may be different depending on the currently installed Plug-In. The International Standards that are used by the ESA612 Plug-In are listed in Table 3-1.



gbv33.bmp

Figure 3-24. Selecting Standards

Note

When a safety test is a part of a test container, for instance an Auto Sequence, the standards listed on the safety test will include only those selected on the parent test container.

Table 3-1. ESA612 International Standards

Standard	Use
AAMI/NFPA-99 (H)	Hospital
AAMI/NFPA-99 (M)	Manufacturer
AAMI ES60601-1	
AS/NZS 3551	
IEC 60601	
IEC 62353 (CL1)	Class 1 equipment
IEC 62353 (CL2)	Class 2 equipment
MDA DB9801S1A	
MDB DB9801S1B	
VDE 751.1 (CL1)	Class 1 equipment
VDE 751.1 (CL2)	Class 2 equipment

When adding a test element to a test template, Ansur selects the Ansur User Defined standard by default. If a test requires a particular resistance load, then only one standard should be selected. The ESA612 will be configured to use the test load of the first standard that was selected when creating the test template. Ansur filters the tests on the standards. Therefore a test marked for IEC601.1 only, will not be performed if you perform the test selecting AAMI/NFPA-99 as the only standard.

To select one standard for all tests within a test container, select the test container entry and then select the standard in the apply when performing list. For example, select the Auto Sequence entry in the template window and then select a standard. All tests in the Auto Sequence will now have the same standard applied to each test.

4. Enter a limit value by clicking on the Expected Results tab.
5. Click the High or Low cell under the desired standard. The selected row is highlighted and a cursor is displayed in the cell that was checked.
6. Enter the limit value for the selected safety test.

Note

If an international standard has statutory limits, then these limits are always enforced. Statutory limits cannot be altered by the user and are indicated with a padlock (🔒).

Chapter 4

ESA612 Test Templates

Title	Page
Introduction.....	4-3
Power Change Delay	4-3
Power Change Pause.....	4-5
Multiple Earth Resistance Tests.....	4-6
Testing Ultrasound Probes Example.....	4-6
Create a Test Template	4-6
Testing the Ultrasound Probes.....	4-8
Basic Test Reports	4-9

Introduction

This chapter describes the advanced features of the ESA612 Plug-In. This chapter assumes the reader is already familiar with the basic tasks described in chapter 3.

Power Change Delay

Some devices require time to stabilize after cycling power on or off. It is important to delay running electrical safety tests until the device stabilizes. The Auto Sequence can be delayed for a specified time to allow a device to stabilize after its power has been turned off or on. To enter a time delay:

1. Select **Auto Sequence** in the template window and click the **Custom setup** tab to open the custom setup window.

As shown in Figure 4-1, there are two delay settings under Delay Times in the General Test Group Settings section of the custom setup window: Power Up and Power Off delay times.




Figure 4-1. Custom Setup Tab

gbv35.bmp

2. To set a delay in seconds after the ESA612 turns on the DUT but before the safety test continues, click in the text box to the right of **Power on delay time (s)** and enter a delay between 0 and 9,999. See Figure 4-2.
3. To set a delay in seconds after the ESA612 turns off the DUT, click in the text box to the right of **Power off delay time (s)** and enter a delay between 0 and 9,999. See Figure 4-2.





Figure 4-2. Editing Power On/Off Delay Times

gbv36.bmp

Note

The Power off delay time must be greater than zero for the **After Power Off** and **Before Power Off** options to be visible. **After Power Off** option is selected by default.


When performing an Auto Sequence configured as the example in Figure 4-2, a 20 second delay is inserted each time the DUT outlet on the ESA612 is powered on. The dialog box shown in Figure 4-3 is displayed to indicate the progress while the Plug-In is waiting to continue the test.



gbv37.bmp

Figure 4-3. Power On Delay Message

A delay of 7 seconds is inserted each time the ESA612 DUT outlet is powered off depending on the selection of **After Power Off** or **Before Power Off**. For the **After Power Off** option, the dialog box shown in Figure 4-4 is displayed to indicate the progress while the ESA612 Plug-In is waiting to continue the safety test.



gbv71.bmp

Figure 4-4. Power-Off Delay Message with After Power Off Selected

The dialog box shown in Figure 4-5 is displayed when the **Before Power Off** option is selected.




Figure 4-5. Power-Off Delay Message with Before Power Off Selected

gbv72.bmp

Power Change Pause

Some devices require more than a delay each time DUT power is turned on or off. One example is a device that does not turn on after a power failure. In this case, personnel must turn the device back on. To work with this device, the Auto Sequence can be programmed to stop or pause whenever the power configuration changes on the ESA612 equipment outlet. There two pause options: Stop before and stop after new power configuration.

Stop before new power configuration – Causes the Auto Sequence to display a message before power on the ESA612 equipment outlet is changed.

Stop after new power configuration – Causes the Auto Sequence to display a message after power on the ESA612 equipment outlet is changed.

1. Select **Auto Sequence** in the template window and click the **Custom setup** tab to open the custom setup window.

As shown in Figure 4-1, there are several settings under Test Options in the General Test Group Settings section of the custom setup window. Two settings are labeled Stop before new power configuration and stop after new power configuration.

2. To pause the safety test before a new power configuration is applied to the equipment outlet, click **Stop before new power configuration**.
3. To pause the safety test after a new power configuration is applied to the equipment outlet, click **Stop after new power configuration**.

The dialog box shown in Figure 4-6 is displayed whenever a pause is executed.

Continue must be clicked to continue the safety test.




Figure 4-6. Stop for Power Configuration Change Message

gbv40.bmp

Note

When a stop before or stop after new power configuration option is enabled, the power on and power off delay times are ignored.

Multiple Earth Resistance Tests

To allow taking Protective Earth (PE) Resistance measurements at several test points on a DUT during one safety test, the ESA612 Plug-In can be set to repeat the PE test. To set up the plug-In for multiple PE tests:

1. Select **Auto Sequence** in the template window and click the **Custom setup** tab to open the custom setup window.

As shown in Figure 4-1, there are several settings under Test Options in the General Test Group Settings section of the custom setup window.

2. To perform multiple PE tests, click **Perform multiple PE tests**.

The dialog box shown in Figure 4-7 is displayed whenever multiple protective earth resistance tests is enabled. Clicking the **Repeat** button repeats the test while clicking **Continue** moves the Auto Sequence on to the next test. Checking the **Save multiple results** checkbox will cause each PE test to be saved.




Figure 4-7. Multiple Protective Earth Resistance Message

gbv43.bmp

Note

If the Save multiple results checkbox is not checked, only the last Protective Earth Resistance test will be saved with the test results. Otherwise each Protective Earth Resistance test result is saved with the test results.

Testing Ultrasound Probes Example

Ultrasound machines may have more than one probe. Only one probe can be attached to the ESA612 at any given time while a safety test is performed. However, the ESA612 Plug-In supports such situations by testing each probe separately through module groups. Module groups allow the testing of several modules connected to the same patient lead input.

Create a Test Template

To create a test template for ultrasound probes:

1. Create a test template with an Auto Sequence test containing a Patient Leakage Current and a Patient Auxiliary current test group. See the “Creating an ESA612 Safety Test Template” section in Chapter 3 for more information.
2. Click the Auto Sequence in the template window and click **Custom setup**.

- Click the Test Each Module Separately, Expand module groups, and Stop at new module checkboxes in the General Test Group Settings area as shown in Figure 4-8.




Figure 4-8. Configuring Module Testing

gbv44.bmp

- Define the patient modules with ultrasound probes by starting the template and proceed to the Test Guide for the Auto Sequence test.
- Click the Module Setup tab and add a new module of type CF with one lead. There is no need to enter module information at this point.
- Click the Group checkbox.

The currently selected module is displayed as Group in the module list as shown in Figure 4-9. A new module is inserted below the group, slightly indented and the **Module Type** and **No of Leads** fields are disabled.





Figure 4-9. Creating Module Groups

gbv45.bmp

- Enter Abdominal as the **Module code** for the new module.

8. Click the plus button (+) to add a new module within the group. Ensure the module currently within the group is selected, otherwise the new module will be placed outside the group.
9. Enter Endovaginal as Module code for the new module, and add another module named Vascular.

As shown in Figure 4-10, the module setup now contains a group with three modules inside it.



gbv46.bmp

Figure 4-10. Module Group with Three Ultrasound Probes

Testing the Ultrasound Probes

Ultrasound probes normally don't have an electrical connection. To perform a safety test on them, fill a container with an electrical conducting liquid (e.g. a salt water solution) in which the probe and test lead will be placed.


1. Connect one end of a test lead to the ESA612 as shown in the test guide. Put the other end in the conducting liquid.

Note

To determine where to connect the test lead, return to the procedure page by clicking the Procedure tab in the Test Guide. The module group just created will be displayed as Group in the patient lead setup. To see the contents of the group, hover the mouse pointer above the group title and a tool tip will pop up and display the Module code of every module inside the group.

2. Click the Start Test button (▶) on the Test Guide's toolbar. Alternately push F9 to start the test.

The safety test starts and a progress bar in the lower-left corner starts moving. If the test does not start, see the "Connect the ESA612" section in Chapter 3. After a short while, the Connect Applied Parts window shown in Figure 4-11 appears, instructing you to connect the first probe to patient lead 1.



gbv47.bmp


Figure 4-11. Connect Applied Parts Window

3. When instructed to connect the first probe (Abdominal), connect it to the ultrasound machine and submerge it in the liquid with the test lead.
4. Click **Continue** to continue performing the electrical safety tests on the abdominal probe.
5. When the Connect Applied Parts window reappears instructing you to connect the next probe, disconnect the first probe and connect the second probe to the ESA612. Submerge the probe in the liquid with the test lead.
6. Repeat steps 4 and 5 for the vascular probe.

Basic Test Reports

The amount of detail contained in a test report is adjustable with the value filter built into the Auto Sequence and Test Groups. The example below shows the steps to get a report with only the worst case value from a test.


1. Create a new test template and add an Earth Leakage Current test group as shown in Figure 4-12.



gbv48.bmp

Figure 4-12. Add an Earth Leakage Current Test


2. Click the test group in the template window and click the Custom setup tab to display the options shown in Figure 4-13.



gbv49.bmp

Figure 4-13. Assign Maximum Value to a Test Group

3. Under Value Assigned to Test Group click Maximum value option.
4. Expand the test group and click the first safety test (Normal Condition), then click the **Apply When** tab.
5. Uncheck **Include in report** which is found in the Report Level frame shown in Figure 4-14.




gbv50.bmp

Figure 4-14. Remove Test from a Report

6. Repeat step 4 and 5 for every safety test inside the Earth Leakage Current test group.
7. Perform the safety test.

When the safety test is complete, the Earth Leakage Current test group (the row with caption written in bold face font) is updated with the maximum value recorded by any of the Earth Leakage Current tests. See Figure 4-15.




gbv51.bmp

Figure 4-15. Maximum Value Assigned to the Test Group

8. Continue to the test record window.

As shown in Figure 4-16 only the test group appears in the test record. When a test report is printed it will also show only the test group with the maximum (worst case) value of all the performed Earth Leakage Current tests.



gbv52.bmp

Figure 4-16. Test Record Showing Value Filtered Report

Chapter 5 Reference

	Title	Page
Introduction.....		5-3
Electrical Safety Test Elements		5-3
Custom Setup.....		5-3
Expected Results		5-3
Test Groups.....		5-3
Custom Setup.....		5-4
Test Options		5-4
Test Each Module Separately.....		5-4
Test Guide Settings		5-4
Patient Lead Tests		5-4
MAP Settings		5-5
Patient Lead Configuration – Module Setup.....		5-5
Value Filter		5-5
Modular Filter		5-5
Expected Results		5-5
Auto Sequence		5-5
Custom Setup.....		5-6
Test Options		5-6
Delay Times		5-6
Test Each Module Separately.....		5-6
Test Guide Settings		5-6
Patient Lead Tests		5-6
Patient Lead Configuration – Module Setup.....		5-7
Value Filter		5-7
Module Filter.....		5-7
Expected Results		5-7
Test Guide.....		5-7
Module Setup.....		5-8
Module Groups.....		5-9
Module Nomenclature.....		5-10
Lead Calibration		5-10
Halt on Test Failure		5-10
Plug-In Preferences.....		5-11

Introduction

This chapter contains additional information on some of the ESA612 Plug-In features.

Electrical Safety Test Elements

Each electrical safety test corresponds to a unique measurement available on the ESA612 Safety Analyzer. When the electrical safety test element is executed within an Ansur test template, the returned measurement is compared to defined limits. This comparison determines if the test element passed or failed.


The most noticeable difference is between electrical safety tests measured on applied parts and those measured on other parts on the DUT (e.g. its enclosure). Before measurements are taken, the applied parts configuration for the DUT is grouped into a number of modules. Electrical safety tests measured on applied parts will be run for each module and so the measurements returned are displayed against the list of modules.

Custom Setup

There is no limitation placed on how the measurement is performed for a electrical safety test element.

Expected Results

Electrical safety tests that are not performed on applied parts will contain only one limit entry for each selected standard as shown in Figure 5-1.




Limit	High	Low	Unit	Reference
User defined				
Single test			mA	

Figure 5-1. Expected Results with No Applied Parts

gbv53.bmp

Electrical safety tests performed on applied parts will contain two or three limit entries for each selected standard. This is because applied parts can be one of three IEC classes (Body, Body Float, or Cardiac Float) or one of two AAMI classes (Non-Isolated or Isolated). In Ansur, it is possible to define limits against any of these classes. Figure 5-2 shows the results window on an applied part using the IEC classes.



Limit	High	Low	Unit	Reference
User defined				
Body			uA	
Body Float			uA	
Cardiac Float			uA	

Figure 5-2. Expected Results with Applied Parts

gbv54.bmp


Test Groups

Test groups come with a number of predetermined child safety test elements. These child elements are the same measurement but with varying power configurations. It is possible to drop other tests into this container but it is ill advised.

The test group is passed only if all of its child elements pass and the test group's value satisfy its expected results.

Custom Setup

The Test Group's custom setup enables the operator to configure how the group will be performed. Figure 5-3 shows the custom setup window.



gbv55.bmp

Figure 5-3. Test Group Custom Setup Window

Test Options

When **Halt on test failure** is selected, the test will stop each time a safety test violates its limits. The Plug-In will display the window described in the “Halt of Test Failure” section later in this chapter.

Test Each Module Separately

Usually the software gets a measurement for each safety test element for each module in turn before continuing to the next safety test element. However, it is possible to run all safety test elements against a single module before running all the tests against the next module. This is achieved by selecting the **Test Each Module Separately** option.

By selecting **Expand module groups** the electrical safety tests will be repeated against all modules within each module group. If unselected, but a number of identical applied parts do exist on the same lead positions, then the applied parts are measured in parallel.

If **Stop at new module** option is selected, the test run is halted whenever a new module is to be tested. The user must choose to continue to run the test.

If the Test Group is part of an Auto Sequence that has either Test Optimization or Test Each Module Separately selected, the module options will have no effect and will not be displayed.

Test Guide Settings

The **Skip**, **NA**, and **Step** button on the test guide (see Test Guide section below) can be enabled or disabled through the checkboxes under Test Guide Settings in the General Test Group Settings of the Test Group Setup window.

Note

If the test group is part of an Auto Sequence where the Step button has been disabled, then Test Guide Settings will have no effect and will not be displayed.

Patient Lead Tests

For the patient leads test, the “Worst Case” option is used to get the worst case value per module. The “All Lead” option is used to get a value for all leads for each module.

MAP Settings

The MAP settings options are used only for Mains on Applied Parts test to change the polarity.

Patient Lead Configuration – Module Setup

When the “Override Patient Lead Configuration” option is checked, the field user can modify the module setup data from the test guide. When not checked, the field user cannot modify the module setup from the test guide.

Module setup can be configured in custom setup during template creation. The setup can also be modified by the field user from the test guide, if the "Override Patient Lead Configuration" option is checked in the template. See the “Module Setup” section for more details on how to define modules.

Value Filter

The option selections under the **Value Assigned to Test Group** determine if the final value will be tested against the maximum measurement value, the minimum measurement value, or no value at all.

Modular Filter

The Module Filter restricts the test group to certain modules. This filter ensures the test is performed only for modules that contain module info in accordance with the filter.

- To create a module filter, click the plus button (+) above the column headers.
- Enter the name of a Module Info field in the column named **Module’s**. The default Ansur installation provides the fields **Module Code**, **Serial No.** and **Type**. Ensure that the name is typed correctly.
- In the **equals** column, enter the value the filter should apply.

When defining modules and before running the test, ensure the values are entered in the same way as the filter was created. Shown in Figure 5-4 is the module info for a module that would be included in the filter indicated in Figure 5-3. See the Module Setup section below for more details on how to define modules.

Module Code	FAST-SP02
Serial No.	818491
Type	Sp02

gbv56.bmp

Figure 5-4. Module Information

The Module Filter is only displayed on test groups containing tests designed for applied parts. Some examples are Patient Leakage Current, Mains on Applied Parts, and Patient Auxiliary Current. The Module Filter is also available on the Auto Sequence because it may contain any of the first three tests.

If no module filter is defined, then all modules will be tested.

Expected Results

The expected results are entered in an identical fashion as the individual safety test elements.

Auto Sequence

The Auto Sequence enables a series of automatic tests to be executed within the same test guide, i.e. with minimum user interaction. It is recommended that all the elements defining an equipment’s test and inspection procedures be included into a single Auto Sequence container.

The Auto Sequence is passed only if all of its child elements pass and the Auto Sequence's value satisfy its expected results.

Custom Setup

The Auto Sequence has the same configuration features as those found in the Test Group described above. However, there are a number additional test settings. Figure 5-5 shows the Auto Sequence custom setup page.




Figure 5-5. Auto Sequence Custom Setup Page

gbv57.bmp

Test Options

The **Test Optimization** option affects the sequence in which the tests are run within the container. Without this option, child tests would always be run sequentially as they appear in the container. By selecting this option, the child tests are run so that the fewest number of DUT power cycles are employed.

Halt on test failure sets the Plug-In to display the window shown in the Halt on Test Failure section below whenever a test fails.

The **Stop before / after new power configuration** options are in support of the optimization functionality. The operator performing the test will be notified whenever the power changes. This allows the operator to perform required operations on the DUT before continuing the safety test.

When **Perform multiple PE tests** is selected, the protective earth resistance test will repeat as many times as desired.

Delay Times

Power on / off delay time defines the number of seconds the Plug-In waits before continuing the safety test after toggling DUT power on or off.

Test Each Module Separately

See the Test Each Module Separately section under Custom Setup above.

Test Guide Settings

See the Test Guide Settings section under Custom Setup above.

Patient Lead Tests

See the Patient Lead Tests section under Custom Setup above.

Patient Lead Configuration – Module Setup

See the Patient Lead Configuration section under Custom Setup above.

Value Filter

See the Value Filter section under Custom Setup above.

Module Filter


See the Modular Filter section under Custom Setup above.

Expected Results

The expected results are entered in an identical fashion as the individual safety test elements. The only exception is that limits are only available for a single test; no account is made for module type.

Test Guide

The test guide is the user’s interface with the ESA612 while running a test template. The most critical user interaction is immediately prior to running the tests. This is the time to input the applied parts configuration that sets the Ansur ESA612 Plug-In for the right tests. Figure 5-6 shows the Test Guide window.



gbv58.bmp






Figure 5-6. ESA612 Test Guide Window

Table 5-1 lists the Test Window controls with their descriptions.

Table 5-1. Toolbar Buttons

Button	Name	Description
	Abort	Abort the test template.
	Previous	Go to the previous dialog screen.
	Next	Go to the next dialog screen.
	Start	Perform the test.

Table 5-1. Toolbar Buttons (cont.)

Button	Name	Description
	Stepwise	Perform a step-wise safety test. i.e. one step at a time. This button is available by clicking the drop down arrow next to the start button.
	Stop	Stop the test.
	Not applicable	Flag test as not applicable.
	Skip	Skip the test and go to the next test.
	Additional features	Allows access to test lead calibration window.

Module Setup

A module defines the configuration of the applied parts being tested within Ansur. Running an applied parts test without this definition will result in no measurements taken. See Appendix A for details of the applied parts tests.

The number of applied parts depends on the DUT configuration and can only be defined at the time of testing. This definition is done through the Module Setup window within the Test Guide. Click on the tab with the image of the ESA612 on it to open the Module Setup window shown in Figure 5-7.




Figure 5-7. Module Setup Window

gbv59.bmp

The list on the left displays the user-defined modules. A module can be of class B, BF or CF and contain any number of leads up to a cumulative limit of ten. The toolbar at the top enables the creation, deletion and sequencing of a module.

Figure 5-8 is an example of a module setup window.




Figure 5-8. Modules in a Module Setup Window

gbv60.bmp

The data items listed below the class and lead settings are named Module Info. These data items are important, as it is on these values that the module is filtered if Module Filters are in use.

When returning to the test procedure the diagram in the top right shows how the Ansur ESA612 plug in sees the applied parts configuration.

Figure 5-9 shows the applied parts connection diagram for the ESA612.




Figure 5-9. Applied Parts Connection Diagram

gbv61.bmp

Notice how the leads are displayed sequentially from socket 1 through socket 5 as defined in the module setup.

Module Groups

A module group allows multiple, identical sets of applied parts to be tested within the same test. Figure 5-10 shows the module groups window.




Figure 5-10. Module Groups Window

gbv62.bmp

To create a module group, click on the checkbox named Group just to the right of the number of leads.

Module Nomenclature

The module setup toolbar contains a button initially named IEC. Clicking this button opens a popup menu containing the items IEC and AAMI. This option is used to change the nomenclature used for the module classes.

When selecting AAMI from the popup menu shown in Figure 5-11, the module class changes from B, BF and CF to Non-Isolated and Isolated. All existing B and BF modules will be classified as Non-Isolated, while all CF modules will change to Isolated.




Figure 5-11. Selecting Module Nomenclature

gbv63.bmp

Lead Calibration

Before performing a protective earth resistance test, the test lead should be calibrated to remove the test lead resistance. If a calibration has not already been performed, the ESA612 Plug-In will automatically display the calibration window shown in Figure 5-12 when an earth resistance test is run. Connect the test lead to the 0/NULL jack and click the Calibrate button.




Figure 5-12. Test Lead Calibration Window


gbv65.bmp

Note

If the test is stopped and the test lead is replaced, you should select Calibrate Leads from the test guide's toolbar before continuing.

Halt on Test Failure

If Halt on test failure is enabled on a test group, the ESA612 Plug-In will display the dialog box shown in Figure 5-13 whenever a test fails according to any of the currently selected standards.



gbv66.bmp

Figure 5-13. Halt on Test Failure Window

The dialog box will implement a timer counting down from 90 seconds. When the timer reaches 0, the test will continue using the 'save and continue' option. If the user clicks within the window the timer will stop.

Save test result and continue

When selected, the failed test will be saved and the sequence will continue.

Use on all subsequent tests

If selected, this window will not be displayed again and all subsequent tests that fail will be saved.

Discard test result and continue

The failed test will be discarded and the sequence will continue.

Note

If the test's Skip button is disabled, this option will not be displayed.

Repeat the failed test

The failed test will be performed again. If the test fails, the same dialog box will be shown again.

Plug-In Preferences

The ESA612 Plug-In preferences contain default settings for the ESA612 Plug-In. The preferences are available from Ansur's **Tools** → **Options** menu.

The Plug-In preferences also contain a nomenclature setting that will be used as a default setting when an electrical safety test is started. Figure 5-14 shows the Module Nomenclature selection window.




Figure 5-14. ESA612 Plug-In Preferences: Module Nomenclature

gbv67.bmp