

User manual

Check-MDR ESBL

Rapid Molecular Detection of ESBL genes

REF 14-0040



SmartCycler®II

For use with the SmartCycler[®]II System (Cepheid, CA, USA) in combination with the SmartMix[®]HM PCR Master Mix (Cepheid, CA, USA)



Key to symbols used								
IVD	For In Vitro Diagnostic Use	REF	Catalog number					
LOT	Batch code	IFU	Instructions for use number					
Σ	Use before YYYY-MM	i	Consult instructions for use					
	Temperature limitation	$\sum_{i=1}^{n}$	Contains sufficient for < n > tests					
	Manufacturer							

INTRODUCTION	INTENDED USE	2
PRINCIPLE OF THE METHOD 2 KIT CONTENTS (FOR 96 REACTIONS) 2 SHELF LIFE, STORAGE AND HANDLING 2 MATERIALS REQUIRED BUT NOT SUPPLIED WITH THE KIT 3 GOOD LABORATORY PRACTICES 3 PROTOCOL 4 1. DNA EXTRACTION FROM BACTERIAL CELLS 4 2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 1. DATA ANALYSIS 7 1. DATA INTERPRETATION 6 1. DATA INTERPRETATION 6 2. DATA INTERPRETATION 6 3. REQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 2. IMITATIONS 6	INTRODUCTION	2
KIT CONTENTS (FOR 96 REACTIONS) 2 SHELF LIFE, STORAGE AND HANDLING 2 MATERIALS REQUIRED BUT NOT SUPPLIED WITH THE KIT 3 GOOD LABORATORY PRACTICES 3 PROTOCOL 4 1. DNA EXTRACTION FROM BACTERIAL CELLS 4 2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 TREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 LIMITATIONS 8	PRINCIPLE OF THE METHOD	2
SHELF LIFE, STORAGE AND HANDLING 2 MATERIALS REQUIRED BUT NOT SUPPLIED WITH THE KIT 3 GOOD LABORATORY PRACTICES 3 PROTOCOL 4 1. DNA EXTRACTION FROM BACTERIAL CELLS 4 2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 LIMITATIONS 8	KIT CONTENTS (FOR 96 REACTIONS)	2
MATERIALS REQUIRED BUT NOT SUPPLIED WITH THE KIT 3 GOOD LABORATORY PRACTICES 3 PROTOCOL 4 1. DNA EXTRACTION FROM BACTERIAL CELLS 4 2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 LIMITATIONS 8	SHELF LIFE, STORAGE AND HANDLING	2
GOOD LABORATORY PRACTICES 3 PROTOCOL 4 1. DNA EXTRACTION FROM BACTERIAL CELLS 4 2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 LIMITATIONS 8	MATERIALS REQUIRED BUT NOT SUPPLIED WITH THE KIT	
PROTOCOL 4 1. DNA EXTRACTION FROM BACTERIAL CELLS 4 2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 LIMITATIONS 8	GOOD LABORATORY PRACTICES	
1. DNA EXTRACTION FROM BACTERIAL CELLS 4 2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 LIMITATIONS 8	PROTOCOL	4
2. DNA RECOGNITION STEP A 4 3. REAL-TIME PROBE AMPLIFICATION STEP B 5 DATA ANALYSIS AND INTERPRETATION 6 1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING 7 LIMITATIONS 8	1. DNA EXTRACTION FROM BACTERIAL CELLS	4
3. REAL-TIME PROBE AMPLIFICATION STEP B	2. DNA RECOGNITION STEP A	4
DATA ANALYSIS AND INTERPRETATION	3. REAL-TIME PROBE AMPLIFICATION STEP B	5
1. DATA ANALYSIS 6 2. DATA INTERPRETATION 6 FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING IMITATIONS	DATA ANALYSIS AND INTERPRETATION	e
2. Data INTERPRETATION	1. DATA ANALYSIS	6
FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING	2. DATA INTERPRETATION	6
LIMITATIONS	FREQUENTLY ASKED QUESTIONS (FAQ) & TROUBLESHOOTING	7
	LIMITATIONS	8

Intended use

Check-MDR ESBL is designed for the rapid molecular detection of Extended Spectrum β -lactamase (ESBL) genes. In most cases the presence of the genes indicates the expression of Extended Spectrum β -Lactamase activity. For TEM and SHV, ESBL and non-ESBL variants exist and the test employs highly specific DNA markers to distinguish these gene variants. In addition the most prevalent CTX-M genes are detected. Check-MDR ESBL generates definitive results within 4,5 hours, compared to the 24 hours necessary for analysis with conventional phenotypic methods.

Introduction

The TEM, SHV and CTX-M genes encode the clinically most prevalent Extended Spectrum β -Lactamases (ESBLs). These three groups of ESBLs are generally capable of hydrolyzing first, second, third and fourth generation cephalosporins, penicillins and monobactams, thereby limiting treatment options. TEM and SHV ESBL subtypes are derived from their parental sequences by point mutations leading to amino acid substitutions (www.lahey.org/studies). These amino acid substitutions may extend the substrate spectrum to hydrolyze a wider range of β -lactam antibiotics, *i.e.*, ESBL. The CTX-M genes originate from *Kluyvera* species and presently are the most prevalent ESBLs. According to their amino acid sequence they can be divided into 5 different groups: CTX-M-1, CTX-M-2, CTX-M-9, CTX-M-8 and CTX-M-25. Conventional methods for detection of β -lactamases rely on phenotypic identification which is time consuming, frequently inconclusive and not applicable to all species. Therapeutic failures associated with infections caused by bacteria containing β -lactamase genes are often due to serious problems with interpretation of phenotypic tests. Check-MDR ESBL is a rapid molecular test for the most prevalent ESBLs, reliably identifying the genes that encode for ESBLs.



Check-MDR ESBL is based on specific molecular recognition of target sequences by DNA probe ligation followed by real-time PCR detection. The recognition step uses two specific oligonucleotides (ligation probes) that are joined by a DNA ligase only when they match perfectly with the target DNA. In addition to target-specific sequences, the DNA probes contain two universal primer binding sites as well as a DNA segment that is complementary to a molecular beacon. Subsequent real-time amplification with universal primers is used to detect the connected probes, and their accumulation is visualized by fluorescence of a molecular beacon.

Kit contents (for 96 reactions)

Components (Mat. No.)	Description	Storage conditions
Solution A (9-0050)	1 tube (purple cap ●) 600 μl	- 20°C
Solution P ESBL (9-0051)	1 tube (orange cap 🗢) 250 μl	- 20°C
Solution R (9-0053)*	1tube (yellow cap 😑) 1000 μl	- 20°C, store in the dark
Internal control (9-0054)	1 tube (transparent cap 🔵) 500 μl	- 20°C
Positive control ESBL (9-0055)	1 tube (teal cap ●) 600 μl	- 20°C
Manual (9-0057)	Leaflet – download from website	Not critical

* Volume provided for 100 reactions which takes into account the preparation of excess master mix.

Shelf life, Storage and Handling

The components of the kit must be stored at -20°C. Solution R has to be stored in the dark. Reagents stored at the appropriate storage conditions can be used until the expiration date. Please visually inspect the box upon initial opening to ensure that its content is intact. Do not use when damaged. Please contact the Check-Points office at **support@checkpoints.com** if you have any questions or in case shipping has taken more than 2 days.



Check-MDR



Materials required but not supplied with the kit

	Pre-PCR	Post-PCR
Equipment	 Thermocycler* Vortex mixer Mini-centrifuge PCR plate spinner 	 Real-Time PCR instrument*
Supplies	 SmartMix®HM PCR Master Mix DNA extraction procedure** Disposable laboratory (powder-free) gloves Pipettes & disposable (filter-) tips for volumes of 1 to 1000 µl 1.5 ml tubes ("Eppendorf tubes") 10 ml tubes (for large volumes) SmartTubes PCR tubes SmartTube rack mini-centrifuge with SmartTube 	

*contact your local representative for specifications. **see Protocol, section 1

Good laboratory practices

Recommendations for best results

The quality of the results depends on strict compliance with the following good laboratory practices, especially concerning PCR:

- The test must be performed by adequately trained personnel.
- Spinning down for a few seconds is done in the various steps to ensure that all material is collected at the bottom of the tubes.
- Do not use reagents after their expiration date.
- Before use, thaw frozen reagents completely at room temperature and vortex briefly to obtain a homogeneous solution. After vortexing briefly, spin down the solution to avoid contamination when opening the lid. Avoid unnecessary freeze-thawing of the kit content.
- Periodically, verify the accuracy and precision of pipettes, as well as correct functioning of the instruments.
- Keep Solution R (yellow cap •) in the dark to avoid photo-bleaching of the dyes.

Prevention of contaminations

PCR produces a very high quantity of DNA amplification products (amplicons) even from minute quantities of starting material. Check-MDR ESBL may therefore yield unreliable results if samples become contaminated with amplicons from previous amplification reactions prior to the PCR. Preventive measures to minimize the risk of amplicon contamination must be taken. Please read carefully and follow the instructions outlined below.

Use separate rooms: a pre-PCR room and a post-PCR room.

- Sample preparation, DNA recognition (step A) and preparation of the amplification step (step B) is carried out in the pre-PCR room.
- Incubation in the real-time PCR thermocycler of step B is carried out in the post-PCR room.
- Never bring the reaction products of step B to the pre-PCR room.
- Never prepare the ligation and/or amplification steps in the post-PCR room.

To keep the laboratory free of PCR product contamination:

- Use pipettes with hydrophobic filter tips.
- Make sure to always use a new pipette tip when adding solutions or samples to a reaction tube to avoid contamination.
- Follow proper pipette-dispensing techniques to prevent aerosols.
- Use separate equipment, pipettes, thermocyclers, sample holders, lab coats, gloves, disposables and reagents, that are assigned to these rooms. Never transfer items from the post-PCR room to the pre-PCR room.
- Wear a clean lab coat and clean gloves during all steps of the test.
- Wear clean gloves and a clean lab coat not previously worn while handling amplified PCR products or during sample preparation.
- Change gloves whenever you suspect that they are contaminated.
- Keep the tubes of all kit components and samples closed as much as possible.
- Clean the lab benches and all equipment regularly with a 0,5% sodium hypochlorite solution.



Protocol

It is strongly recommended that the full protocol is read before using the test.

The protocol consists of the following steps:

- 1. DNA extraction from bacterial cells
- 2. DNA recognition step A
- 3. Real-time probe amplification step B

1. DNA extraction from bacterial cells

Important points before starting:

- DNA extraction is completely carried out in the pre-PCR room.
- Make sure to always use a new pipette tip when adding solutions or samples to a reaction tube to avoid contamination.
- Clinical specimens should first be incubated on nutrient agar plates. Use bacterial cells from these agar plates for DNA extraction. Typical growth media include blood agar, MacConkey agar and Tryptic Soy agar.

Procedure:

- 1. Check-MDR ESBL has been validated with the following extraction methods for bacterial cells:
- NucliSENS[®] easyMAG[®] (bioMérieux, France) automated DNA extraction procedure for bacterial cells. Follow the manufacturer's protocol for bacterial cells and use 200 μl cell suspension of McFarland 0,5 1,0 or OD₆₀₀ 0,08 0,12 (this may vary between spectrophotometers). Add 2,5 μl of the Internal Control solution (IC, transparent cap) to the cell suspension and start the DNA extraction. DNA is eluted in 110 μl elution buffer. DNA extracts can be stored at -20°C or +4°C for up to 6 months.
- **DNeasy Blood & Tissue Kit** (QIAGEN, CA, USA) DNA extraction procedure for bacterial cells (manual extraction or QIAcube automated system). Follow the manufacturer's protocol for gram-negative bacteria. Prepare 1 ml cell suspension of McFarland 1,2 1,8 or OD₆₀₀ of 0,16 0,24 (this may vary between spectrophotometers) and centrifue at 14000 rpm for 10 minutes. Discard the supernatant and add 5 µl of the IC solution (transparent cap •) to the pellet. Start the DNA extraction with the pelleted cells using either the manual procedure or the QIAcube automated procedure. DNA is eluted in 200 µl elution buffer. Store DNA extracts at -20°C or +4°C for up to 6 months.

- **MagNA Pure system** (32 samples) (Roche, CH) for DNA extraction procedure for clinical specimens. Use 200 μ l of cell suspension in PBS, Phosphate buffered saline, McFarland 0,5 1,0 or OD₆₀₀ 0,08 0,12 (this may vary between spectrophotometers). Add 2,5 μ l of the IC solution (transparent cap •) to the cell suspension and start the DNA extraction. DNA is eluted in 100 μ l elution buffer. DNA extracts should be stored at +4°C only, for up to 6 months.
- 2. Use the DNA solution directly and continue with "step A" or store as specified until use.

2. DNA recognition step A

Important points before starting:

- Step A is completely carried out in the pre-PCR room.
- Make sure to always use a new pipette tip when adding solutions or samples to a reaction tube to avoid contamination.
- It is advised to perform a positive and negative control reaction. The positive control (teal cap ●) is supplied with the kit; for the negative control we recommend to perform a DNA extraction as specified earlier (with IC solution) for a sample known to be negative for the test in use (*i.e.*, ESBL negative sample, elution buffer, clean water).
- Use a thermocycler located in the pre-PCR room to perform the step A reaction. Use PCR tubes that are suitable for the type of thermocycler used.

Procedure:

- 1. Determine the number of reactions. Thaw all reagents (*i.e.*, Solution A, DNA samples If kept at -20°C, and positive control), mix well and keep on ice.
- 2. Prepare the ligation reaction mix as described in table 1 and include 10% surplus to ensure that you have enough ligation reaction mix. Mix well and spin down.
- 3. To each tube add 7,5 μl of ligation reaction mix and 10 μl of sample DNA or controls (see table 2).
- 4. Close the tubes, mix well (by tapping or vortexing) and spin down briefly. The solutions should have a uniform blue color.
- 5. Place the tube(s) in the thermocycler and run the Probe ligation program, see table 3 (total sample volume ~ 18 μ l).

 Table 1: Ligation reaction mix.

Component	Volume per reaction
Solution P ESBL (orange cap $igodoldsymbol{ imes}$)	2,5 μl
Solution A (purple cap $ullet$)	5 µl
Total volume of ligation reaction mix	7,5 μl



 Table 2: Ligation reaction setup.

Reaction Type	Component	Volume per reaction
Test sample	Sample DNA	10 µl
Positive control	ESBL positive (teal cap 🔵)	10 µl
Negative control	Negative sample extracted with the IC	10 µl

Table 3: Probe ligation cycling parameters.

Step	Temperature	Time	Cycles
1.	95°C	3 min	1
2.	65°C	120 min	1
3.	98°C	2 min	1
	4°C	hold	

Note:

- It is recommended to mix Solution P and Solution A first in a separate tube for x number of samples (including controls and 10% surplus) and then add 7,5 μl of this mix to each reaction tubes. Prepare this mix shortly before step A is started and use immediately. Do not store this mix and dispose after use.
- Close the PCR tube(s) carefully: excessive pressure may distort the cap and lead the sample evaporation during step A.

3. Real-time probe amplification step B

Important points before starting:

- The preparation of the reaction mix for step B is carried out in the pre-PCR room.
- Step B was developed for the **Cepheid SmartCycler®II** real-time system, thus Cepheid SmartTubes, SmartTube rack, mini-centrifuge with SmartTube adaptators and any other related laboratory materials should be used (see SmartCycler® II Operator Manual).
- Step B requires SmartMix[®]HM PCR Master Mix (Cepheid, CA, USA) and 10 mM Tris buffer pH8.5. One kit requires 50 SmartMix[®] HM beads and 1.5 ml of 10 mM Tris buffer.

Procedure:

- 1. Briefly spin down the reaction mixtures from step A.
- 2. Take Solution R (yellow cap \bigcirc) from the freezer. Thaw completely at room temperature **protecting from exposure to light**. Mix well and spin down briefly.

- Rehydrate SmartMix[®]HM PCR Master Mix: 1 bead is diluted in 20 μl of Tris buffer (10 mM, pH8.5). Mix well (5 sec vortex) and spin down briefly. 1 bead is used to perform 2 reactions.
- 4. Prepare the real-time PCR (qPCR) reaction mix as described in table 4 and include surplus to ensure that you have enough qPCR reaction mix.
- 5. Add 20 μ l of the qPCR master mix and 5 μ l of each step A reaction to each SmartTube. Use a new pipette tip for each step A reaction sample added. Close SmartTubes.
- 6. Transfer the SmartTubes to the post-PCR room. Centrifuge SmartTubes for minimum 7 sec and check that no air bubble was accumulated in the optical part of the tube (see SmartCycler®II Operator Manual). Then carefully place the SmartTubes in the appropriate cartridges.
- Following the manufacturer's instructions, start and set up the run of the real-time PCR SmartCycler®II system, using the parameters presented in table 5 and table 6 and select the Dye set FCTC25.
- 8. Without delay, start the run with the cycling conditions presented in table 6. When the run is completed, discard the SmartTubes according to local regulations.

Table 4: qPCR-reaction mix.

Component	Volume for 2 reactions
Tris buffer (10 mM, pH8.5)	20 µl
SmartMix [®] HM PCR Master Mix	1 bead
Solution R(yellow cap))	20 µl
Total volume of master mix	40 μl

Table 5: Real-time PCR Analysis settings.*

Ch #	Dye Name	Usage	Bkgnd Sub	Bkgnd Min Cycle	Curve Analysis	Thresh setting	Manual Thresh Fluor Units	Auto Thresh #SD's	Auto Min Cycle	Auto Max Cycle	Valid Min Cycle	Valid Max Cycle
1.	FAM	Assay	ON	5	40	Primary curve	Manual	NA	5	10	3	60
2.	Cy5	Assay	ON	5	40	Primary curve	Manual	NA	5	10	3	60

*Ch: channel; Bkgnd: background; Thresh: Threshold; Avg: average

Table 6: Real-time cycling parameters.

Step	Temperature	Time	Cycles	Data Collection	Ramp Rate
1.	95°C	120 sec	Hold	Optics OFF	
2	95°C	15 sec	10	Optics OFF	2°C/s
2.	60°C	60 sec	40	Optics ON	2°C/s



Data analysis and interpretation

SmartCycler®II real-time PCR instrument

For a detailed description on how to operate the SmartCycler®II instrument, please refer to your real-time PCR instrument user guide instruction, the *Smartcycler®II Operator Manual*.

1. Data analysis

- 1. In the *Analysis Settings* set the threshold manually in the middle of the exponential phase (check on the Amplification plot). This process is described in the *Smartcycler®II Operator Manual*.
- FAM threshold can vary depending on the DNA extracts. FAM threshold is around 22 ±
 2 Fluor Units on average; Cy5 threshold is around 10 ± 2 Fluor Units on average. It is still very important to check visually if the manual threshold is adequate to your experiment. Always click "update analysis" when modifying the threshold.
- 3. Figures 1 shows typical amplification plots of ESBL positive and negative samples.



Figure 1: Typical real-time amplification plots of: **A** FAM ESBL positive&negative samples (red line: threshold was set at 22) and **B** Cy5 ESBL positive&negative samples (red line: threshold was set at ~10). The internal control curve (Cy5) may be absent or at a higher C_T in positive samples (SmartSoft 5.5.1).

2. Data interpretation

Interpret results (positive, negative or inconclusive) with the C_T values obtained for the samples following the guidelines outlined below and summarized in table 7 and 8.

- 1. The relevant columns in the table are "Site ID", "Sample ID" (if filled in), "FAM C_T " and "Cy5 C_T ". All other columns may be ignored. For each sample the FAM and Cy-5 C_T values are shown.
- 2. Verify that the real-time PCR run is valid before data analysis and interpretation of the results. Table 7 shows criteria for a valid real-time Check-MDR ESBL run.

- 4. Negative ESBL samples will show a higher FAM C_T value than the IC Cy5 signal. The IC Cy5 signal is expected at 33 ±2, and the FAM signal is expected at $C_T > 35$.
- 5. Samples with a FAM C_T between 30 and 35 and a C_T for the IC Cy5 at 33 ±2 are considered inconclusive. Please repeat the whole procedure with a new DNA extract of the bacterial culture. If still inconclusive the sample can be regarded neither positive nor negative.
- 6. In samples with a FAM $C_T > 35$ or 0,00 and an IC Cy5 > 35 or 0,00 the assay has not worked well and should be repeated with a new DNA extract of the bacterial culture.

Table 7: Criteria for a valid run with Check-MDR real-time assay.

Reaction/Sample Type*	ESBL (FAM) C _T values	Internal Control (Cy5) C _⊤ values		
Positive control	22 ± 2	0,00		
Negative sample (extracted with IC)	C _T > 35 or 0,00	33 ±2		

* IC: internal control; If C_T values are not as expected, see FAQ.

Table 8: Data interpretation guidelines.

ESBL (FAM) C _T values	Internal Control (Cy5) C _T values	Interpretation
< 29	≥ 33 ±2	Positive sample
> 35 or 0,00	33 ±2	Negative sample
32 < C _T < 35	≥ 33 ±2	Inconclusive
> 35 or 0,00	> 35 or 0,00	Sample failed*

* see FAQ.



Frequently asked questions (FAQ) & Troubleshooting

1. How important is the amount of input cells for DNA extraction?

Check-MDR ESBL has been optimized using a well-defined amount of cells. A deviation of 20% in the amount of cells will have no major consequences. Larger deviations will not give optimal results.

2. May other DNA extraction methods be used with Check-MDR ESBL?

Check-MDR ESBL test has been optimized using DNeasy Blood & Tissue Kit (QIAGEN, CA, USA), NucliSENS[®] easyMAG[®] (bioMérieux, France), and MagNA Pure system (Roche, CH) extraction methods. Check-Points does not guarantee the performance of the test with extraction methods other than those recommended in this manual.

3. The thermocycler states an error in step A.

Please contact Check-Points Technical Support: **support@check-points.com**

4. During the step A the sample(s) have (partly) evaporated.

Reaction tubes may not have been closed properly. Please restart the procedure from step A.

5. I have left Solutions (A, P, R, Internal control or Positive control) out of the -20°C (-4°F) storage.

These reagents must be stored at -20°C (-4°F) for proper performance of the test. The performance of the product cannot be fully guaranteed if these solutions were left out of -20°C (-4°F) for more than 24 hours.

6. May I change the C_T threshold when analyzing my real-time PCR data?

The Check-MDR ESBL assay was developed using a manual threshold to analyze the data. As guidelines a range of manual threshold to use for FAM and Cy5 detection channel are provided. Please refer to the tutorial *Data Analysis on the Smart Note*[®] 6.2 and to the *SmartCycler*[®]*II Operator Manual* for more information on setting thresholds.

7. What does it mean if the real-time results show no C_T values or interpretation concluded that the sample failed?

Such results may have multiple explanations:

- The sample DNA was not added to the assay in step A.
- The sample DNA tested with Check-MDR ESBL is negative and the internal control was not added prior to DNA extraction.
- The DNA extraction failed since the internal control was not detected.
- The sample DNA contains contaminants inhibiting the reactions. Please repeat the DNA extraction.
- Solution P and/or Solution A was not added in step A. Please repeat the test.
- Solution R and/or SmartMix[®]HM PCR MasterMix was not added to the assay. Please repeat the test.
- Check SmartMix[®]HM PCR MasterMix expiration date.
- 8. What does it mean if the real-time results show no C_T values for the positive control or interpretation concluded that sample is inconclusive?

Such results may have multiple explanations:

- The positive control solution was not added to its reaction tube in step A.
- Solution P and/or Solution A was not added in step A. Please repeat the test.
- Solution R or SmartMix[®]HM PCR MasterMix was not added to the assay. Please repeat the test.
- Check SmartMix[®]HM PCR MasterMix expiration date.

9. Duplicate DNA samples tested with Check-MDR ESBL test do not yield identical results.

 C_T values of identical samples may vary slightly between individual reactions. Larger variations, > 2 C_T values, suggest pipetting errors or other differences between the duplicate samples.

10. May the assay be interrupted after step A and continued at a later time?

Reaction mixtures from Step A can be kept at hold at +4°C for up to 2 hours. For further inquiries, please contact Check-Points Technical Support: **support@check-points.com**.



Limitations

Check-MDR ESBL uses a range of specific DNA markers to identify the presence or absence of ESBL genes. The test detects the presence of CTX-M 1, 2 and 9 gene families and the most prevalent ESBL mutations in TEM and SHV. ESBL mutations TEM-E104K, TEM-R164S and SHV-G238S are covered by Check-MDR ESBL. The most frequently found TEM and SHV ESBL variants contain these mutations. More rare variants may have other ESBL mutations. (For a detailed explanation see **www.lahey.org/studies** and M. Gniadkowski, Clin. Microbiol. Infect. 2008; 14 [Suppl. 1]: 11–32). In addition various minor ESBL genes exist, although these are not frequently found in clinical settings. (For more details see T. Naas, L. Poirel and P. Nordmann, Clin. Microbiol. Infect. 2008; 14 (Suppl. 1): 42–52).

Check-MDR ESBL requires DNA purified from a colony or bacterial culture. Clinical specimens cannot be tested directly. The quality of the input DNA is an important factor for obtaining reliable results from Check-MDR ESBL. DNA must be extracted from cultured bacteria using the extraction methods validated with Check-MDR ESBL and described in this manual (page 6). The assay has been tested extensively with purified DNA from gram-negative bacteria, such as Escherichia, Salmonella, Klebsiella, Enterobacter, Citrobacter and Pseudomonas, with excellent results. However, it may never be excluded that other Gram-negative bacteria or certain strains of the above species will yield poor results. Check-MDR ESBL cannot and does not make any representation or warranty that it is capable of correctly detecting the ESBL genes in all gram-negative species, subspecies or type or in any clinical sample source. Results may need to be confirmed by additional methodologies in specific cases (e.g. for regulatory samples). Due to the high variability of bacterial genomes it is possible that certain subtypes might not be detected. The test reflects the state of knowledge of Check-Points Health B.V. The presence of multiple bacterial species in a sample may hamper the interpretation of the test. As with other diagnostic assays, the results of this test may only be interpreted in combination with additional laboratory and clinical data available to the responsible person. Use of this assay is limited to appropriately gualified personnel, well trained in the execution DNA-based molecular detection methods.

Despite the utmost care in the development and preparation of the protocol Check-Points cannot take any responsibility for errors, omissions and/or future changes herein.

Literature Citation: When describing a procedure for publication using this product, please refer to it as the Check-MDR ESBL.

Notice to Purchaser:

Molecular beacons are licensed from the Public health Research Institute of the City of New York (PHRI) under PHRI's patents.

Trademarks

Nuclisens and easyMAG and are registered trademarks of bioMérieux SA. DNeasy and QIAcube are registered trademarks of the QIAGEN group. MAGNA PURE is a trademark of Roche. FAM is a trademark of Applera Corporation. Cy5 is a registered trademark of GE Healthcare. SmartCycler II and SmartMix are trademarks of Cepheid.





Check-Points Health BV Binnenhaven 5 6709 PD Wageningen The Netherlands Tel: +31 317 453 908 Fax: +31 317 210 147 info@check-points.com www.check-points.com