

For Patient and Clinician



The Personal Pain Relief System That Breaks The Pain Cycle





USER'S GUIDE For Patient and Clinician



ADDENDUM

This Manual applies to the *Ultima 3* (three mode device without timer). If you have the Ultima or Ultima 3t, please read below:

The Ultima

Is a single mode device. This manual will provide accurate instructions omitting the modulation and burst modes.

The Ultima 3t

Is a three mode device with a timer. This manual will provide accurate instructions omitting the use of the timer.

Also note,

Throughout this manual the "c" is designated as the "conventional mode", however on most devices an "n" or "normal mode" will be found to have the same function.

LIFE TIME WARRANTY

This warranty is in lieu of any other warranty expressed or implied.

Pain Management Technologies provides a life-time warranty from the date of purchase on each new ULTIMA TENS stimulator against any defects in material and workmanship. If the new ULTIMA stimulator is found to be defective within the warranty period, it will be replaced or repaired free of charge if returned to your place of purchase.

This warranty applies to the ULTIMA unit only, and does not cover charger, batteries, electrodes, lead wires or other accessories. This warranty does not cover damage caused by misuse, negligence, accident, abuse, alteration or modification of the unit.



CONTENTS

CLINICIAN MODE DIAGRAMS	2
WARNINGS & PRECAUTIONS	
PATIENT INSTRUCTIONS	
Understanding Pain	4
How ULTIMA Relieves Pain	
Introduction to The ULTIMA Unit	5
Attaching Electrodes	6
Applying Electrodes	6
Adjusting Controls	6
Personalizing Control Settings	7
Changing Batteries	7
Skincare	8
Troubleshooting	8
CLINICIAN INFORMATION	
Indications	9
Contraindications	9
Electrode Placement	9
Modes of Operation	
Control Settings	
Conventional Mode	
Modulation Mode	10
Customizing Control Settings for	
Conventional & Modulation Modes	
Burst Mode	11
Acupuncture Mode	
(High Width, Low Rate)	11
Brief Intense Mode	
(High Width, High Rate)	11
Customizing Control Settings for Burst,	
Acupuncture & Brief Intense Modes	
Technical Specifications	
Warranty	12



TO THE CLINICIAN

These diagrams were produced from 1 second sweep durations of an oscilloscope illustrating the electrical impulses of each of the 5 modes of operation. Control settings for each mode are detailed on page 10.

CONVENTIONAL MODE

MODULATION MODE











BRIEF INTENS MODE



 After this preliminary period, slowly advance the Pulse Width control (no higher than 200 uS) until the patient perceives a change.

If the perception is only that the stimulation is growing stronger, but not expanding the area of pain relief, reduce the Pulse Width to the nominal setting.

- If the perception is that the stimulation is growing deeper, wider or spreading, sustain stimulation at the higher Pulse width setting.
- 7. Relief is usually evident within 10 to 30 minutes.
- 8. Conventional and Modulation Modes may be used as long as desired.

Burst Mode

Use the Burst Mode at low Intensity settings to treat acute superficial pain.

Use at higher Intensity settings - resulting in strong muscle contractions - to treat chronic, deep pain.

PRESET CONTROLS TO:

Intensity : OFF Mode Selector : "B" Pulse Width : 200 uS Pulse Rate : Automatically controlled by the ULTIMA unit

in this mode. (9 pulses per Burst ,2 bursts per second)

Acupuncture Mode (High Width, Low Rate) The high intensity Acupuncture Mode provides merve stimulation with electrodes positioned on the traditional Chinese acupuncture points or on any painful trigger or tender points. Stimulation is thought to cause signals to be transmitted to the brainstem, which uses descending pathways to block pain through the release of endogenous opiates such as sendorphins, as well as other pain-modulating neurotransmitters such as serotonin. This mode is very effective for deep aching chronic pain. It is also useful when previous nerve damage has occurred.

PRESET CONTROLS TO: Intensity: OFF Mode Selector: "C" Pulse Width:200 uS Pulse Rate:2-4 pps

Brief Intense Mode (High Width , High Rate) The high-intensity Brief Intense Mode is recommended for immediate relief of severe pain. It is especially useful during painful therapeutic procedures such as joint mobilization.

PRESET CONTROLS TO: Intensity: OFF Mode Selector: "C" Pulse Width:240 uS Pulse Rate:120 pps











Modes of Operation

The dual-channel ULTIMA System is designed to operate in 5 different modes to provide maximum flexibility in pain treatment:

- 1. Conventional Mode 2 Modulation Mode
- Modulation N
 Burst Mode
- Burst Mode
 Acupuncture Mode
- 5. Brief Intense Mode
- 5. bhei miense moue

Control Settings

Adjusting the ULTIMA unit for optimum pain relief is a process of custom-tailoring settings to the individual patient. Pulse Intensity, Rate and Width are interactive phenomena, and require fine tuning from nominal settings, based on the patient's reactions, to achieve optimal results. The following procedures are recommended for determining

the most effective settings for optimum patient treatment. The Conventional Mode is recommended as the initial

treatment mode for patients new to transcutaneous electrical nerve stimulation treatment.

Conventional Mode

The Conventional Mode provides nerve stimulation with pulse amplitude, width and rate fully adjustable.

PRESET CONTROLS TO: Intensity : OFF

Mode Selector : "C" Pulse Width : 60 uS (microseconds) Pulse Rate : 60 pps (pulses per second)

Modulation Mode

In the Modulation Mode, the pulse width decreases and increases in a cyclic pattern over a period of 5 seconds. Because treatment alternates between superficial and deep stimulation, we recommend this mode for situations where the patient has developed tolerance to the Conventional Mode.

PRESET CONTROLS TO: Intensity : OFF Mode Selector : "M" Pulse Width : 60 uS Pulse Rate : 60 pps

Customizing Control Settings for Conventional & Modulation Modes

1. Preset controls as indicated.

- 2. Turn ON the intensity for the channel being used. Note that the channel indicator light begins to blink.
- 3. Gradually turn intensity up until the patient notices a change in sensation in the electrode area.
- 4. Slowly increase the intensity until the patient experiences slight discomfort. Then immediately decrease intensity slightly to just below the pain threshold (the sensation, however. should still be detectable). This will allow maximum excitation of the large nerve fibers without stimulation of the smaller, pain-conducting fibers (Gate Control Theory).
- 5. As the patient adapts to the sensations, typically after 5 to 10 minutes of stimulation, and perceives that the strength of stimulation has decreased, increase the intensity slightly to restore the desired perception.



WARNINGS & PRECAUTIONS

CARDIAC PACEMAKERS -

Do not use the ULTIMA T.E.N.S. unit with demand type Cardiac Pacemakers.

THROAT (CARLTID SINUS) -

Do not use over the nerves, muscles or arteries of the throat or the larynx (voice box) - strong muscle contractions could close the airway.

EYES -

Do not use on the eyelids.

🗌 HEAD -

Do not place electrodes so electrical flow is from one side of the head to the other.

EKG MONITORS & ALARMS -

Monitors and alarms for EKG units may not operate properly when the ULTIMA unit is in use.

MYOCARDIAL DISEASE OR ARRYTHMIAS -

The ULTIMA T.E.N.S. unit should be used by patients with these conditions only

under the direction and supervision of the prescribing physician. PREGNANCY -


- The safety of the ULTIMA T.E.N.S. unit during pregnancy or delivery has not been
 - established.
 - UNDIAGNOSED PAIN -
- Do not use the ULTIMA unit to treat undiagnosed pain syndromes. MACHINERY OPERATION -
- Do not adjust the control settings of the ULTIMA unit while

operating vehicles,

heavy equipment, machinery or tools. CHILDREN -

Keep out of the reach of children.

CAUTION - United States Federal Law requires a physician's prescription for the purchase and use of the ULTIMA T.E.N.S. system. Ultima is approved by the Food & Drug Administration.



PATIENT INSTRUCTIONS

The ULTIMA T.E.N.S. system should be used only under the supervision of a physician or a qualified medical practitioner to whom you are referred by your physician. Use it only for the condition for which it is prescribed.

Understanding Pain

Pain is the body's warning system. It alerts us to injury or illness. And It is absolutely necessary for life and health.

Unfortunately, however, this vital warning system is not equipped with an "off" switch. So sometimes, long after we have gotten the "message" and responded to the problem pain signals keep on being sent to the brain.

Pain signals, in the form of electrical impulses, travel to the brain through the body's nervous system from the site of the injury or ailment. In the brain, the impulses are interpreted as pain.

If these impulses can be interfered with before the brain can receive and interpreted them, the result is pain relief.

How ULTIMA Relieves Pain

ULTIMA is the advanced, pocket-sized, fully portable, Transcutaneous Electrical Nerve Stimulation (TE.N.S.) system designed to relieve both acute and persistent pain without shots, drugs or narcotics. It sends gentle electrical impulses through the skin to the nerves to suppress pain by scrambling the pain signals before they reach the brain.

ULTIMA provides five different modes of operation for pain treatment versatility.

Because ULTIMA has two channels, it can be used to relieve pain at two body locations, simultaneously, or to provide extra relief at a single body location or over a large body area.

ULTIMA may not work for everybody who suffers from pain. But for many people, ULTIMA offers comforting relief, reducing and often even eliminating pain.

Because there are many different kinds of pain, and each individual is unique, ULTMA pain relief varies from person to person. Some users experience relief only while the ULTIMA unit is turned on. For others, relief continues for a length of time even after the unit is turned off. Some users experience relief in one mode, some in others.

The ULTIMA system can help you return to normal activity. It is about the size of a pack of playing cards and battery powered, so you can wear it everywhere- at home, to work, in your car. And it is equipped with a convenient clip so you can wear it on your belt or even inside your clothing.

CLINICIAN INFORMATION

Indications

The ULTIMA transcutaneous electrical nerve stimulation system is indicated as a first line treatment modality or as part of a multi-modality treatment strategy for symptomatic relief and management of acute, persistent and chronic pain.

ULTIMA provides proven effective pain relief for back pain, myofascial pain syndrome, post-operative surgical & scar pain, post-traumatic pain, knee & shoulder pain, arthritis, strains & sprains. TMJ neuropathies, neuralgia, RSD causalgic pain, dysmenorrhea, tension headaches, carcinoma pain and any pain condition requiring medication. References available upon request.

Contraindications

ULTIMA is contraindicated for use on patients using demand-type cardiac pacemakers. IMPORTANT: See "Warnings & Precautions" section on page 3.

Recommended Electrode Placement

This chart shows typical electrode placement positions for treatment of typical pain conditions.



In-line Placement is recommended for neck, sciatica and lower-back muscle pain.



is ulder, pine

LOWER BACK





NECK/LOWER BACK

NFCK

SHOULDER OR KNEE

Skincare

- To keep your skin healthy and avoid irritation, follow these simple recommendations.
- 1. Before applying electrodes and after removing them, wash the area of skin where electrodes are placed with mild soap and water. Rinse and dry thoroughly.
- 2. Skin oils can be removed by wiping the skin with rubbing alcohol.
- 3. Do not shave electrode area; hair may be trimmed short with scissors.
- 4. Always remove electrodes by pulling in the direction of hair growth.
- 5. Avoid skin stress that can be caused by electrodes or tae holding lead wires stretching or pulling the skin.

TROUBLESHOOTING		
THE "ON" INDICATOR LIGHTS UP BUT THE UNIT DOES NOT FUNCTION PROPERLY	THE "ON" INDICATOR IS DIM	THE "ON" INDICATOR DOES NOT LIGHT UP
Control settings may not be set to prescribed values.		The battery is discharged; replace with a fresh one.
Electrodes may not be in proper position.	The battery is discharged; replace with a fresh one.	The battery may be improperly
Electrodes may be wearing out or drying out (A prickling sensation may accompany this problem)	replace with a fresh one.	positioned; follow diagram in battery compartment.
Electrodes may be defective.		The unit may be damaged;
Lead wires may not be securely connected.		contact your supplier for service.
Lead wires may be broken inside the insulation.		



RED(+)&

BLACK(-)

PLUGE (Connects

to socket

of electrodes)

ELECTRODE


OUTPUT


output socket on top of

PLUG (Connects to

ULTIMA)

Introduction to The ULTIMA Unit











Adjusting Controls



1. Each time you use your ULTIMA unit, begin with the intensity controls for both channels 1& 2 turned OFF.



Selector Switch to the "C"(Conventional), "M"(Modulated), "B"(Burst) position as recommended by your clinician.



2. Adjust the sliding Mode 3. Adjust the Pulse Rate control the position as recommended by your clinician.

Attaching Electrodes 1. Make sure the ULTIMA unit is turned OFF 2. Determine which channel-1.2 or both - you wish to use.

electrode.

the wire can break it.

skin oils are removed.

with the electrodes packaging. 5. Use ULTIMA electrodes for best results.

Applying Electrodes

3. Plug the "L" shaped lead wire output plug into the appropriate channel output socket of the ULTIMA unit.

4. Insert the red electrode plug into the socket of one electrode.

CAUTION: Always grip the plug connector when

you plug in and unplug the lead wires. Pulling


1. Your physician or clinician will determine which type of electrode is best suited for your treatment. 2. Follow your physician's or clinician's directions as to proper electrode positioning for best results. 3. Prepare your skin where electrodes will be placed by washing, rinsing and thoroughly drying it, making sure all

4. Apply electrodes following application instructions included


5. Insert the black electrode plug into the socket of other



4. Adjust the Pulse Width control the position recommended by your clinician.







Personalizing Control Settings Your clinician will work with you to achieve optimum control

settings to custom tailor your ULTIMA unit to your own personal needs. Carefully follow his or her instructions regarding electrode placement, control settings and user adjustment procedures.



Changing Batteries

Your ULTIMA unit is designed to operate on a 9 Volt rechargeable battery like those included with it, or on the disposable alkaline type available in most stores.

- 1. The ULTIMA unit will shut off automatically when the battery needs recharging. Prior to shutting off, the indicator lights will dim and the unit may perform somewhat erratically.
- 2. Open the battery compartment by pressing up on the latch at the bottom of the compartment cover and lifting the cover off.

- 4. Insert the battery according to the diagram inside the battery compartment, making sure that the terminals are properly aligned. Make sure the lift-out tape is beneath the battery for easy removal next time.
- 5. If you are using alkaline batteries, safely dispose of the old battery.
- 6. Replace the battery compartment cover.