


## Instruction Manual

---

# HI 96770C Silica High Range ISM



Dear Customer,

Thank you for choosing a **HANNA** product. This manual will provide you with the necessary information for the correct use of the instrument. Please read it carefully before using the meter. If you need additional technical information, do not hesitate to e-mail us at [tech@hannainst.com](mailto:tech@hannainst.com).

This instrument is in compliance with **CE** directives.

## TABLE OF CONTENTS

---

PRELIMINARY EXAMINATION .....	3
GENERAL DESCRIPTION .....	4
ABBREVIATIONS .....	4
SPECIFICATIONS .....	5
PRECISION AND ACCURACY .....	5
PRINCIPLE OF OPERATION .....	6
FUNCTIONAL DESCRIPTION .....	7
ERRORS AND WARNINGS .....	9
GENERAL TIPS FOR AN ACCURATE MEASUREMENT .....	11
STARTUP .....	12
MEASUREMENT PROCEDURE .....	12
VALIDATION PROCEDURE .....	15
CALIBRATION PROCEDURE .....	17
GLP .....	20
BATTERY MANAGEMENT .....	21
BATTERY REPLACEMENT .....	22
ACCESSORIES .....	22
WARRANTY .....	23


*All rights are reserved. Reproduction in whole or in part is prohibited without the written consent of the copyright owner, Hanna Instruments Inc., Woonsocket, Rhode Island, 02895, USA.*

## PRELIMINARY EXAMINATION

Please examine this product carefully. Make sure that the instrument is not damaged. If any damage occurred during shipment, please notify your Dealer.

Each **HI 96770** Ion Selective Meter is supplied complete with:

- Two Sample Cuvetts and Caps
- Two CAL CHECK™ standard cuvetts **HI 96770-11** with Quality Certificate
- 9V Battery
- Scissors
- Tissues for wiping cuvetts
- Instrument Quality Certificate
- Instruction Manual and Quick Reference Guide
- Rigid carrying case



**Note:** Save all packing material until you are sure that the instrument works correctly. Any defective item must be returned in its original packing.

## GENERAL DESCRIPTION

---

The **HI 96770** is an auto diagnostic portable microprocessor meter that benefits from **HANNA's** years of experience as a manufacturer of analytical instruments. It has the advanced optical system based on a Light Emitting Diode (LED) and a narrow band interference filter that allows most accurate and repeatable readings. All instruments are factory calibrated and the electronic and optical design minimizes the need of frequent calibration.

With the powerful **CAL CHECK™ validation** function, you are able to validate good performance of your instrument at any time. The validation procedure is extremely user friendly. Just use the exclusive **HANNA** ready-made, NIST traceable standards to verify the performance of the instrument and recalibrate if necessary.

The instrument is splash waterproof and the LED and filter units are protected from dust or dirt by a transparent cup. This makes the instruments fulfill field applications. Display messages aid the user in routine operation. The meter has an auto-shut off feature that will turn off the instrument after 10 minutes of non use in *measurement mode* or after 1 hour if left in *calibration mode*.

The meter uses an exclusive positive-locking system to ensure that the cuvet is in the same position every time it is placed into the measurement cell. It is designed to fit a cuvet with a larger neck making it easier to add both sample and reagents. The cuvet is made from special optical glass to obtain best results.

The **HI 96770** meter measures the molybdate reactive silica ( $\text{SiO}_2$ ) content in water samples in the 0 to 200 mg/L (ppm) range. The method is an adaption of the USEPA METHOD 370.1 for drinking, surface and saline waters, domestic and industrial wastes and Standard Method 4500- $\text{SiO}_2$  C.

The reagents are in powder form and are supplied in packets. The amount of reagent is precisely dosed to ensure the maximum repeatability.

## ABBREVIATIONS

---

°C: degree Celsius

EPA: US Environmental Protection Agency

°F: degree Fahrenheit

LCD: Liquid Crystal Display

mg/L: milligrams per liter. mg/L is equivalent to ppm (part per million)

mL: milliliter

mV: millivolts

# SPECIFICATIONS

Range	0 to 200 mg/L
Resolution	1 mg/L
Precision	± 5 mg/L @ 100 mg/L
Typical EMC Deviation	±1 mg/L
Light Source	Light Emitting Diode
Light Detector	Silicon Photocell with narrow band interference filter @ 466 nm.
Method	Adaptation of the USEPA METHOD 370.1 for to drinking, surface and saline waters, domestic and industrial wastes and Standard Method 4500-SiO <sub>2</sub> C. The reaction between silica and the reagents causes a yellow tint in the sample.
Environment	0 to 50°C (32 to 122°F); max 95% RH non-condensing
Battery Type	1 x 9 volt
Auto-Shut off	After 10' of non-use in <i>measurement mode</i> , after 1 hour of non-use in <i>calibration mode</i> , with last reading reminder.
Dimensions	192 x 102 x 67 mm (7.6 x 4 x 2.6")
Weight	290 g (10 oz.).

## REQUIRED REAGENTS


Code	Description	Quantity/test
HI 96770A-0	Silica HR Reagent A	1 packet
HI 96770B-0	Silica HR Reagent B	1 packet
HI 96770C-0	Silica HR Reagent C	1 packet

# PRECISION AND ACCURACY

**Precision** is how closely repeated measurements agree with each other. Precision is usually expressed as standard deviation (SD).

**Accuracy** is defined as the nearness of a test result to the true value.

Although good precision suggests good accuracy, precise results can be inaccurate. The figure explains these definitions. In a laboratory using a standard solution of 100 mg/L silica and a representative lot of reagent, an operator obtained with a single instrument a standard deviation of 5 mg/L.



## PRINCIPLE OF OPERATION

---

Absorption of Light is a typical phenomenon of interaction between electromagnetic radiation and matter. When a light beam crosses a substance, some of the radiation may be absorbed by atoms, molecules or crystal lattices.

If pure absorption occurs, the fraction of light absorbed depends both on the optical path length through the matter and on the physical-chemical characteristics of the substance according to the Lambert-Beer Law:

$$-\log I/I_0 = \epsilon_{\lambda} c d$$

or

$$A = \epsilon_{\lambda} c d$$


Where:

$-\log I/I_0$	=	Absorbance (A)
$I_0$	=	intensity of incident light beam
$I$	=	intensity of light beam after absorption
$\epsilon_{\lambda}$	=	molar extinction coefficient at wavelength $\lambda$
$c$	=	molar concentration of the substance
$d$	=	optical path through the substance

Therefore, the concentration "c" can be calculated from the absorbance of the substance as the other factors are known.

Photometric chemical analysis is based on the possibility to develop an absorbing compound from a specific chemical reaction between sample and reagents. Given that the absorption of a compound strictly depends on the wavelength of the incident light beam, a narrow spectral bandwidth should be selected as well as a proper central wavelength to optimize measurements.

The optical system of **HANNA's HI 96770** colorimeter is based on a Light Emitting Diode (LED) combined with a narrow-band interference filter to guarantee both high performance and reliable results.



HI 96770 block diagram (optical layout)

A microprocessor controlled LED emits radiation which is first optically conditioned and beamed to the sample contained in the cuvet. The optical path is fixed by the diameter of the cuvet. Then the light is spectrally filtered to a narrow spectral bandwidth, to obtain a light beam of intensity  $I_0$  or  $I$ .

The photoelectric cell collects the radiation  $I$  that is not absorbed by the sample and converts it into an electric current, producing a potential in the mV range.

The microprocessor uses this potential to convert the incoming value into the desired measuring unit and to display it on the LCD.


The measurement process is carried out in two phases: first the meter is zeroed and then the actual measurement is performed.

The cuvet has a very important role because it is an optical element and thus requires particular attention. It is important that both, the measurement and the calibration (zeroing) cuvetts, are optically identical to provide the same measurement conditions. Whenever possible use the same cuvet for both. It is necessary that the surface of the cuvet is clean and not scratched. This to avoid measurement interference due to unwanted reflection and absorption of light. It is recommended not to touch the cuvet walls with hands.

Furthermore, in order to maintain the same conditions during the zeroing and the measuring phases, it is necessary to close the cuvet to prevent any contamination.

## FUNCTIONAL DESCRIPTION

### INSTRUMENT DESCRIPTION



- 1) GLP/▲ key
- 2) CAL CHECK key
- 3) ZERO/CFM key
- 4) READ ►/TIMER key
- 5) ON/OFF key
- 6) Liquid Cristal Display (LCD)
- 7) Cuvet alignment indicator
- 8) Cuvet holder


## KEYPAD DESCRIPTION

- **ON/OFF**: to turn the meter on and off.
- **ZERO/CFM**: to zero the meter prior to measurement, to confirm edited values or to confirm factory calibration restore.
- **READ ►/TIMER**: this is a multi-functional key. In *measurement mode*, press to make a measurement, or press and hold for three seconds to start a pre-programmed countdown.  
In *GLP mode* press to view the next screen.
- **CAL CHECK**: this is a bi-functional key. Just press to perform the validation of the meter, or press and hold for three seconds to enter *calibration mode*.
- **GLP/▲**: this is a bi-functional key. Just press to enter *GLP mode*. In *calibration mode* press to edit the date and time.

## OPERATING MODES

- *Measurement mode*: default operation mode, enables both **validation** and **measurement**.
- *Calibration mode*: may be entered by keeping **CAL CHECK** pressed for three seconds (the "CAL" tag appears), it enables **calibration** of the instrument.
- *GLP mode* may be entered by pressing **GLP/▲** ("GLP" appears), it enables consulting of user calibration date or restore factory calibration.

## DISPLAY ELEMENTS DESCRIPTION




- 1) Instrument status indicator appears during different phases of zero or reading measurement
- 2) Icons displayed in this area are warning or action icons, signaling next operation
- 3) The battery icon shows the charging level of the battery
- 4) The hourglass appears when the instrument is performing an internal check up
- 5) CAL: appears when the instrument operates in the *calibration mode*
- 6) The month, day and date icons appear when a date is displayed
- 7) The chronometer appears when the reaction timer is running
- 8) Four digit main display
- 9) Measuring units
- 10) Four digit secondary display




## ERRORS AND WARNINGS

The instrument shows clear messages when erroneous condition appears. Messages are also displayed when the obtained values are outside expected range. The beeper is playing a beep on errors.


### a) on zero reading



**Light High:** There is too much light to perform a measurement. Please check the preparation of the zero cuvet.




**Light Low:** There is not enough light to perform a measurement. Please check the preparation of the zero cuvet.




**No Light:** The instrument cannot adjust the light level. Please check that the samples does not contain any debris.


### b) on sample reading



**Inverted cuvet:** The sample and the zero cuvet are inverted.



**Zero:** A zero reading was not taken. Follow the instructions of the measurement procedure for zeroing the meter.




**Under range:** A blinking "0" indicates that the silica content in the sample is less than 0 mg/L.




**Over Range:** A flashing value of the maximum concentration indicates an over range condition. The concentration of the sample is beyond the programmed range: dilute the sample and re-run the test.

#### c) during calibration procedure




**Standard Low:** The standard reading is out of range.




**Standard High:** The standard reading is out of range.

#### d) other errors and warnings



**Cap error:** the cuvet is not inserted correctly and an excess ambient light is reaching the detector. If the cover is properly installed, then contact your dealer or the nearest Hanna Customer Service Center.



**Battery low:** The battery voltage is getting low and the battery needs to be replaced.




**Dead battery:** This indicates that the battery is dead and must be replaced. Once this indication is displayed, the meter will lock up. Change the battery and restart the meter.


## GENERAL TIPS FOR AN ACCURATE MEASUREMENT

The instructions listed below should be carefully followed during testing to ensure best accuracy.

- For a correct filling of the cuvet: the liquid in the cuvet forms a concavity on the top; the bottom of this concavity must be at the same level of the 10 mL mark.



- Proper use of the powder reagent packet:
  - (a) use scissors to open the powder packet;
  - (b) push the edges of the packet to form a spout;
  - (c) pour out the content of the packet.



- It is important that the sample does not contain any debris. This would corrupt the reading.
- Each time the cuvet is used, the cap must be tightened to the same degree.
- Whenever the cuvet is placed into the measurement cell, it must be dry outside, and completely free of fingerprints, oil or dirt. Wipe it thoroughly with **HI 731318** or a lint-free cloth prior to insertion.
- Shaking the cuvet can generate bubbles in the sample, causing higher readings. To obtain accurate measurements, remove such bubbles by swirling or by gently tapping the cuvet.
- It is possible to take multiple readings in a row, but it is recommended to take a new zero reading for each sample and to use the same cuvet for zeroing and measurement.
- All the reaction times reported in this manual are referred to 20°C (68°F). As a general rule of thumb, they should be doubled at 10°C (50°F) and halved at 30°C (86°F).
- In order to maximize accuracy, prior to a measurement follow the **validation procedure** to be sure that the instrument is properly calibrated. If necessary, calibrate the instrument.



## START UP


Prepare the instrument for measurement as follows:

- Unpack the instrument by removing the cuvet holder protection from the instrument cuvet holder.
- Place the battery in the instrument as described in the "BATTERY REPLACEMENT" chapter (see pag.22).
- Place the instrument on a flat table.
- Do not place the instrument in direct sun light.


## MEASUREMENT PROCEDURE

Turn the meter on by pressing ON/OFF and follow instruction manual to perform either the MEASUREMENT, VALIDATION or CALIBRATION PROCEDURE.

- Turn the meter on by pressing **ON/OFF**.  
The display briefly shows all tags on.




- When the beeper sounds and the LCD displays dashes, the meter is ready. The blinking "ZERO" indicates that the instrument needs to be zeroed first.




- Fill the cuvet with 10 mL of unreacted sample, up to the mark, and replace the cap.




- Place the cuvet into the cuvet holder and ensure that the notch on the cap is positioned securely into the groove.



- Press **ZERO/CFM** and the instrument status indicator will appear, depending on the measurement phase.




- After a few seconds, the display will show “-0.0-”. The meter is now zeroed and ready for measurement.




- Remove the cuvet.



- Add the content of one packet of HI 96770A-0 Silica HR reagent A and shake until complete dissolution.




- Remove the cap and add the content of one packet of HI 96770B-0 Silica HR reagent B. Replace the cap and shake **vigorously until complete dissolution.**




- Press and hold **READ ►/TIMER** for three seconds. The display will show the countdown. The beeper indicates the end of countdown period. Alternatively, wait for 10 minutes.




- Remove the cap.
- Add the content of one packet of HI 96770C-0 Silica HR reagent C and shake vigorously until complete dissolution.




- Replace the cuvet into the cuvet holder and ensure that the notch on the cap is positioned securely into the groove.



- Wait for 2 minutes and then press **READ ►/TIMER**.



- The instrument status indicator will appear depending on the measurement phase.



- The instrument directly displays concentration in mg/L of silica ( $\text{SiO}_2$ ) on the LCD.
- To convert the reading to mg/L of silicon (Si) multiply the reading by 0.467.

## INTERFERENCES

- Iron ( $\text{Fe}^{2+}$  and  $\text{Fe}^{3+}$ ) interferes at high levels.
- Sulfide ( $\text{S}^{2-}$ ) interferes at high levels.
- Phosphates does not interfere up to  $50 \text{ mg/L PO}_4^{3-}$ .

## VALIDATION PROCEDURE

Use the validation procedure to ensure that the instrument is properly calibrated.


**Warning:** do not validate or calibrate the instrument with standard solutions other than the **HANNA CAL CHECK™** Standards, otherwise erroneous results will be obtained.

For accurate Validation and Calibration results, please perform test at room temperature:  $18$  to  $25^\circ\text{C}$  ( $64.5$  to  $77.0^\circ\text{F}$ )


- Turn the meter on by pressing **ON/OFF**.




- When the beeper sounds and the LCD displays dashes, the meter is ready.




- Place the **CAL CHECK™** Standard Cuvet A into the cuvet holder and ensure that the notch on the cap is positioned securely into the groove.



- Press **ZERO/CFM** and the instrument status indicator icons will appear, depending on the measurement phase.




- After a few seconds, the display will show “-0.0-”. The meter is now zeroed and ready for validation.




- Remove the cuvet.

- Place the **CAL CHECK™** Standard HI 96770-11 Cuvet B into the cuvet holder and ensure that the notch on the cap is positioned securely into the groove.



- Press **CAL CHECK** and the instrument status indicator together with “CAL CHECK” will appear on the display, depending on the measurement phase.



- The display will show the validation standard value. The reading should be within specifications as reported in the **CAL CHECK™** Standard Certificate. If the value is found out of specifications, please check that the cuvetts are free of fingerprints, oil or dirt and repeat validation. If results are still found out of specifications, then recalibrate the instrument.





## CALIBRATION PROCEDURE


**Note:** It is possible to interrupt calibration procedure at any time by pressing **CAL CHECK** or **ON/OFF**.

**Warning:** Do not calibrate the instrument with standard solutions other than the **HANNA CAL CHECK™** Standards, otherwise erroneous results will be obtained. For accurate Validation and Calibration results, please perform test at room temperature: 18 to 25°C (64.5 to 77.0°F)


- Turn the meter on by pressing **ON/OFF**.




- When the beeper sounds and the LCD displays dashes, the meter is ready.




- Enter the *calibration mode* by holding **CAL CHECK** for three seconds. The display will show "CAL" during calibration.




- Place the **CAL CHECK™** Standard Cuvet A into the cuvet holder. Ensure that the notch on the cap is positioned securely into the groove.



- Press **ZERO/CFM** and the instrument status indicator will appear, depending on the measurement phase.




- After a few seconds the display will show “-0.0-”. The meter is now zeroed and ready for calibration.




- Remove the cuvet.

- Place the **CAL CHECK™ Standard HI 96770-11** Cuvet B into the cuvet holder and ensure that the notch on the cap is positioned securely into the groove.




- Press **READ ►/TIMER** and the instrument status indicator will appear, depending on the measurement phase.




- After measurement the instrument will show for three seconds the Cal Check Standard value.

**Note:** If the display shows “**STD HIGH**”, the standard value was too high. If the display shows “**STD LOW**”, the standard value was too low. Verify that both **CAL CHECK™ Standards HI 96770-11** Cuvets A and B are free from fingerprints or dirt and that they are inserted correctly.




- Then the date of the last calibration (e.g.: “01.08.2007”) appears on the display, or “01.01.2005” if the factory calibration was selected before. In both cases the year number is blinking, ready for date input.





## DATE INPUT

- Press **GLP/▲** to edit the desired year (2000-2099). If the key is kept pressed, the year number is automatically increased.



- When the correct year has been set, press **ZERO/CFM** or **READ ►/TIMER** to confirm. Now the display will show the month blinking.




- Press **GLP/▲** to edit the desired month (01-12). If the key is kept pressed, the month number is automatically increased.



- When the correct month has been set, press **ZERO/CFM** or **READ ►/TIMER** to confirm. Now the display will show the day blinking.



OR



- Press **GLP/▲** to edit the desired day (01-31). If the key is kept pressed, the day number is automatically increased.



**Note:** It is possible to change the editing from day to year and to month by pressing **READ ►/TIMER**.

- Press **ZERO/CFM** to save the calibration date.



- The instrument displays "**Stor**" for one second and the calibration is saved. The instrument will return automatically to *measurement mode* by displaying dashes on the LCD.




## GLP

In the GLP mode, the last user calibration date can be consulted and the factory calibration can be restored.

### LAST CALIBRATION DATE

To display the calibration date:

- Press **GLP/▲** to enter *GLP mode*. The calibration month and day will appear on the main display and the year on the secondary display.




- If no calibration was performed, the factory calibration message, "F.CAL" will appear on the main display and the instrument returns to *measurement mode* after three seconds.



### FACTORY CALIBRATION RESET

It is possible to restore the factory calibration.

- Press **GLP/▲** to enter *GLP mode*.




- Press **READ ►/TIMER** to enter in the factory calibration restore screen. The instrument asks for confirmation before deleting the last calibration.



- Press **ZERO/CFM** to restore the factory calibration or press **GLP/▲** again to abort factory calibration restore.

- The instrument briefly notifies “done” when restores factory calibration and returns to *measurement mode*.



## BATTERY MANAGEMENT

---

To save battery, the instrument shuts down after 10 minutes of non-use in *measurement mode* and after 1 hour of non-use in *calibration mode*.

If a valid measurement was displayed before auto-shut off, the value is displayed when the instrument is switched on. The blinking “ZERO” means that a new zero has to be performed.

One fresh battery lasts for around 750 measurements, depending on the light level.

The remaining battery capacity is evaluated at the instrument startup and after each measurement.

The instrument displays a battery indicator with three levels as follows:

- 3 lines for 100 % capacity
- 2 lines for 66 % capacity
- 1 line for 33 % capacity
- Battery icon blinking if the capacity is under 10 %.

If the battery is empty and accurate measurements can't be taken anymore, the instrument shows “dead batt” and turns off.

To restart the instrument, the battery must be replaced with a fresh one.




## BATTERY REPLACEMENT

---

To replace the instrument's battery, follow the steps:

- Turn the instrument off by pressing **ON/OFF**.
- Turn the instrument upside down and remove the battery cover by turning it counterclockwise.



- Extract the battery from its location and replace it with a fresh one.
- Insert back the battery cover and turn it clockwise to close.

## ACCESSORIES

---

### **REAGENT SET**

HI 96770-01	Reagents for 100 SILICA HIGH RANGE tests
HI 96770-03	Reagents for 300 SILICA HIGH RANGE tests

### **OTHER ACCESSORIES**

HI 96770-11	CAL CHECK™ Standard Cuvets (1 set)
HI 721310	9V battery (10 pcs.)
HI 731318	Tissue for wiping cuvetts (4 pcs.)
HI 731331	Glass cuvetts (4 pcs.)
HI 731335	Caps for cuvetts
HI 93703-50	Cuvets cleaning solution (230 mL)

## WARRANTY

---

HI 96770 is warranted for two years against defects in workmanship and materials when used for its intended purpose and maintained according to the instructions.

This warranty is limited to repair or replacement free of charge.

Damages due to accident, misuse, tampering or lack of prescribed maintenance are not covered.

If service is required, contact your dealer. If under warranty, report the model number, date of purchase, serial number and the nature of the failure. If the repair is not covered by the warranty, you will be notified of the charges incurred.

If the instrument is to be returned to Hanna Instruments, first obtain a Returned Goods Authorization Number from the Customer Service Department and then send it with shipment costs prepaid. When shipping any instrument, make sure it is properly packaged for complete protection.

To validate your warranty, fill out and return the enclosed warranty card within 14 days from the date of purchase.

### Recommendations for Users

Before using these products, make sure that they are entirely suitable for your specific application and for the environment in which they are used.

Operation of these instruments may cause unacceptable interferences to other electronic equipments, this requiring the operator to take all necessary steps to correct interferences.

Any variation introduced by the user to the supplied equipment may degrade the instruments' EMC performance.

To avoid damages or burns, do not put the instrument in microwave oven. For yours and the instrument safety do not use or store the instrument in hazardous environments.

**HANNA** instruments® reserves the right to modify the design, construction and appearance of its products without advance notice.



---

w w w . h a n n a i n s t . c o m

MAN96770CR1 01/07