


MOBI®

Ultra Digital Thermometer



INSTRUCTIONS

Item No. 70119


Instruction Manual for the MOBI Ultra Digital Thermometer

Thank you for purchasing MOBI's Ultra Digital Thermometer. For safe use of this product, be sure to read and fully understand this manual as well as the safety precautions, prior to any use. Please keep this manual safely stored for convenient reference.

Safety precautions

- It is dangerous for patients to perform a self-evaluation or self-treatment based on the measuring results. Be sure to follow your doctor's instructions at all times.
- Do not touch or blow on the infrared sensor.
- Clean a soiled infrared sensor with a soft damp cloth in a gentle manner.
- Keep the device out of children's reach.
- Do not take any measurements just after physical activity, which could affect body temperature. Wait 30 minutes for the body to adjust.
- You may need to allow 30 minutes for the thermometer to attain room temperature if stored in an area at a temperature different from the room in which you will use the device.
- Stop using this product if there is any pain.
- This product is not recommended for users suffering ear disease such as otitis or tympanitis.
- Do not use the product in a wet ear passage (such as from swimming or bathing).
- Do not throw batteries into fire.
- Do not use the product for other purposes.
- This is not a waterproof device.
- Do not expose this product to any liquids, chemical solvent, direct sunshine or high temperature.
- Do not use a mobile phone nearby when the thermometer is in use.
- Take out the batteries if you are not going to use the unit for a long period of time.

PRODUCT STRUCTURE



Basic functions:

1. Current date and time display:

Date display falls between calendar years 2001 and 2099; time displays either as 12-hour clock or 24-hour clock (user selectable). The initial date for a powered-up system is January 01, 2008.

2. Measurement and display of temperature:

Measurement ranges from 32.0°C to 42.9°C (89.6°F to 109.3°F).

3. 20 sets of measurement readings:

Stores up to 20 sets of temperature readings. Each set has a storage life of 30 days. Temperature is stored with measuring date, time and measurement mode (Ear or Forehead).

4. Measurement and display of room temperature:

Measurement ranges from 5.0°C to 59.9°C (41.0°F to 139.8°F).

5. Voice prompt:

Reports measurement results (in °C or °F), room temperature and current time.

6. Built-in white Mini Flashlight:

Press the **LIGHT BUTTON** to turn on the LCD backlight. The LCD backlight will turn off automatically after 8 seconds. Press and hold the **LIGHT BUTTON** for 2 seconds to turn on the white **Mini Flashlight**; the white **Mini Flashlight** will turn off automatically after 5 seconds. Pressing any key also turns off Mini Flashlight or LCD backlight.



Setup

Clock mode:


1. Displays time in a sequence of year, month, date, day of the week, hour and minutes with a flickering " : " stands for passing seconds. The LCD displays current time and room temperature in alternating 5-second intervals.

2. Press **TALK/OK** to hear current time and room temperature.



3. Press and hold **TALK/OK** for 2 seconds to toggle on/off the talking function – however, during measurement, you will hear " didi " sound whether or not the unit is in talking or non-talking mode.

4. When the **TALK** function is enabled, the [🔊] icon is turned on. If disabled, the [🔊] icon is off. The default setting is **TALK** enabled.



5. Press **MEM/SET** to retrieve the recorded temperature or press and hold **MEM/SET** for 2 seconds to enter into set up mode.

6. **With a pen depress C/F Key inside battery area to toggle between Celsius or Fahrenheit** (illustration on pg 2)

7. Press the **LIGHT BUTTON** to turn on the LCD backlight. The LCD backlight will turn off automatically after 8 seconds.

8. Press and hold the **LIGHT BUTTON** for 2 seconds to turn on the white **Mini Flashlight**; the white **Mini Flashlight** will turn off automatically after 5 seconds


9. Sleep mode - System enters into sleep mode 1 minute after last key pressed

Setup mode:

Press and hold **MEM/SET** for 2 seconds to enter into setup mode


1. Press **MEM/SET** to shift between a setting of 12/24Hr, HOUR, MINUTE, YEAR, MONTH, DAY and SLEEP enable/disable.
2. Press **MEM/SET** again to adjust the values, then press **TALK/OK** to confirm.
3. Press and hold **MEM/SET** to increase the values rapidly.

* Unit returns to clock mode automatically 30 seconds after last key pressed.



Memory mode:

1. Press **MEM/SET** in clock mode to enter into memory mode.
2. "-" will be displayed if no data is stored.
3. Maximum storage is 20 sets of readings with date and time shown.
4. Unit returns to clock mode automatically 5 seconds after last key pressed.



Notes:

1. Press **MEM/SET**, the **LIGHT BUTTON** or the **TEMPERATURE SCAN BUTTON** to stop the speech.
2. When either the Mini Flashlight or LCD backlight is on, press any key to turn it off.
3. In case of nearby electrostatic discharge, the device may need to be reset.
4. Please store the Thermometer on the table stand when not in use.
5. **Thermometer display will invert 180 degrees when turned upside down to place onto holding base and will read correctly for Clock Mode. When using in the hand the display will invert back 180 degrees.**

Preparations:

1. Reset the unit if there is missing segment on LCD, LCD is blurred or system is not working. The reset key is located inside the battery compartment. (See illustration on page 2.)
2. Replace any low-voltage batteries to ensure a full power supply.
3. An unexpected fluctuation of ambient temperature is likely to affect the measurement's accuracy. For example, never attempt to perform a temperature measurement in front of an air conditioner or heater, as this can result in inaccurate measurements.
4. Measurement should not be taken immediately after exercise or bathing, as this can affect body temperature. Please wait for a period for the body to return to normal temperature.


How to Take a Body Temperature Measurement

IMPORTANT: Before taking a reading allow the thermometer to acclimate to room temperature for 30 mins.

How to Take a Measurement:

NOTE: You may choose to take 2 or 3 consecutive readings, at 15-second intervals, to ensure a correct reading. The device will record the highest reading. Wait at least 10 minutes for the device to cool down before taking further readings in this case.

- 1) To measure **Ear Temperature:**



NOTE: Ear readings are the most accurate because they read from **INSIDE** the ear canal.

Make sure the ear is clean and dry. Pull the **FOREHEAD PROBE** off the unit, insert the temperature probe into the ear, then press the **TEMPERATURE SCAN BUTTON** (just above the LCD screen) once to start the measurement. You will hear a beep sound. Upon completion you will hear a "DI DI" sound, then hear the result via the speaker or see it on the LCD screen.

- 2) To measure **Forehead Temperature:**



NOTE: Though accurate, readings from **OUTSIDE** areas may be one degree lower than inside readings. The **BEST** outside reading area is actually the **TEMPLE**.

Put the **FOREHEAD PROBE** onto the unit, then gently position the probe onto the center of the forehead, midway between the eyebrow and the hairline. Press the **TEMPERATURE SCAN BUTTON** once to start the measurement & you will hear a beep sound. Upon completion you will hear a "DI DI" sound, then hear the result via the speaker or see it on the LCD.


- 3) The result will be stored into memory automatically. If 20 sets of readings are stored already, the earliest reading will be removed each time a new reading is taken.
(Note: recorded data will be removed automatically when it is over 30 days old)
- 4) If the LCD shows "HI", then the temperature reading was above the device's limit of 42.9°C / 109.3°F.
- 5) If the LCD shows "LO", then the temperature reading was below the device's limit of 32.0°C / 89.6°F.
- 6) The unit automatically returns to time display mode after 30 seconds on inactivity.
- 7) The display will be turned off after 1 minute of inactivity (if sleep mode is enabled).

Reviewing Previous Temperature Readings Stored in Memory

While in Clock Mode, press the Mem/Set button to view the last temperature reading. Recorded temperatures will be displayed with date and time stamp. Press the Mem/Set button again to view successive previously recorded temperatures.

Replace a battery

Remove the battery compartment cover. Use only new batteries (2 x AAA battery).
Replace the battery compartment cover.



Maintenance, storage and calibration

Maintenance:

1. Remove any stains on the device with a soft damp cloth.
2. Don't wash with water or detergent containing abrasive or benzene.

Storage:

1. Clean the thermometer and put the cover on to protect the lens after each use.
2. Do not expose the thermometer to direct sunshine, high temperatures, dampness, vibration or impact.
3. Remove the battery for long periods of no use.

Calibration:

This thermometer is calibrated at the time of manufacture. If the thermometer is used according to the instructions, periodic recalibration is not required. If at any time you question the accuracy of the measurement, please contact the retailer immediately

Do not attempt to modify or disassemble the thermometer.

Troubleshooting

Display on LCD	Possible issue	Solutions
Nothing	Battery is exhausted or placed in the wrong position	Check battery orientation or replace batteries
Battery symbol	Low battery	Change battery
Missing segment	Temporary malfunction	Reset the unit
"Lo"	Reading is below 32°C, 89.6°F	Refer back to this manual for proper use
"Hi"	Reading is above 42.9°C/ 109.3°F	Refer back to this manual for proper use
"ErrE"	Ambient temperature out of range	Do not use the thermometer under extreme temperature conditions
"ErrU"	Cannot measure	Refer back to this manual for proper use
"ErrH"	Hardware problem	Consult your dealer
"ErrS"	Sensor error	Do not use the thermometer under extreme conditions
"ErrR"	Hardware problem	Consult your dealer
"ErrL"	Battery low	Replace battery
"ErrP"	Hardware problem	Consult your dealer

Specification:

1. Product Name: Ultra Digital Thermometer
2. Power supply: DC3V (2 x AAA battery)
3. Power consumption: 30mA (for voice prompt)
4. Temperature sensing part: IR sensor
5. Measuring range: 32.0°C ~ 42.9°C (89.6°F ~ 109.3°F)
Room temperature: 5.0°C ~ 59.9°C (41.0°F ~ 139.8°F)
6. Measuring accuracy: $\pm 0.2^{\circ}\text{C}$ (or $\pm 0.4^{\circ}\text{F}$)
Room temperature accuracy: $\pm 1^{\circ}\text{C}$
7. Outer dimension: 200mm x 50mm x 25mm
8. Weight: 110 g
9. Working environment: Temperature: 15°C ~ 40°C (59°F ~ 104.0°F)
Relative humidity: 30%RH ~ 80%RH
10. Storage environment: Temperature: -10°C ~ 55°C (-14°F ~ 131°F)
Relative humidity: 30%RH ~ 85%RH

*Dispose of used batteries with care consult your distribution center for details.

Accessories:

- 1 x FOREHEAD PROBE
- 1 x User's Manual
- 1 x Table Stand Holder

Limited Warranty

Mobi® Technologies, Inc. warrants this product to be free from defects in material and workmanship for a period of ninety (90) days from the date of purchase. Should this product prove to be defective at any time during the warranty period, Mobi® Technologies, Inc. will, at its option, either replace or repair it without charge. After the warranty period, a service charge will be applied for replacement of parts or labor for repair. To obtain warranty service please return the product to Mobi® along with a dated sales receipt from the place of purchase. This warranty does not cover damage caused by accident, misuse, abuse, improper maintenance, unauthorized modification, or connection to an improper power supply. A charge will be made for repair of such damage. This warranty excludes all incidental or consequential damages and any liability other than what is stated above. © 2008 Mobi® Technologies, Inc. All Rights Reserved.

MOBI.

Mobi® Technologies, Inc.
7635 San Fernando Road, Suite A
Burbank, CA. 91505

www.getmobi.com

customer service:
1-877-MOBI-INC (1-877-662-4462)

MOBI and Ultra Digital Thermometer are registered trademarks of MOBI Technologies, Inc.
©2008 MOBI Technologies, Inc. All Rights Reserved.
