

MT 4080
LCR METER
OPERATING MANUAL



Contents

1.	INTRODUCTION.....	1
1.1	GENERAL.....	1
1.2	IMPEDANCE PARAMETERS	3
1.3	SPECIFICATION	6
1.4	ACCESSORIES	19
2.	OPERATION	20
2.1	PHYSICAL DESCRIPTION	20
2.2	MAKING MEASUREMENT.....	21
2.2.1	Battery Replacement.....	21
2.2.2	Battery Recharging/AC operation.....	22
2.2.3	Open and Short Calibration	23
2.2.4	Display Speed	24
2.2.5	Relative Mode	24
2.2.6	Range Hold.....	24
2.2.7	DC Resistance Measurement.....	25
2.2.8	AC Impedance Measurement.....	25
2.2.9	Capacitance Measurement	25
2.2.10	Inductance Measurement	26
2.3	ACCESSORY OPERATION	27
3.	USB OPERATION.....	29
3.1	COMMAND SYNTAX	30
3.2	COMMANDS.....	31
4.	APPLICATION	39
4.1	TEST LEADS CONNECTION	39
4.2	OPEN/SHORT COMPENSATION	44
4.3	SELECTING THE SERIES OR PARALLEL MODE	46
5.	WARRANTY INFORMATION	49
6.	SAFETY PRECAUTION.....	51

1. Introduction

1.1 General

The MT4080 is a high accuracy handheld LCR meter that can perform the inductor, capacitor and resistor measurement up to 100KHz within 0.2% basic accuracy. It is the most advanced handheld AC/DC impedance measurement instrument to date. The MT4080 can help engineers and students to understand the characteristic of electronics components. It is also of great assistance to those people who want to do the quality control of the electronics components.

The instrument is auto or manual ranging. Test frequencies of 100Hz, 120Hz, 1KHz, 10KHz or 100KHz (MT4080A only) may be selected on all applicable ranges. The test voltages of 50mVrms, 0.25Vrms, 1Vrms or 1VDC (DCR only) may also be selected on all applicable ranges. The dual display feature permits simultaneous measurements.

Components can be measured in the series or parallel mode as desired; the more standard method is automatically selected first but can be overridden.

The highly versatile MT4080 can perform virtually all the functions of most bench type LCR bridges. With a basic accuracy of 0.2%, this economical LCR meter may be adequately substituted for a more expensive LCR bridge in many situations. The meter is powered from two AA Batteries and is supplied with an AC to DC charging adapter and two AA Ni-Mh Rechargeable

Batteries.

The instrument has applications in electronic engineering labs, production facilities, service shops, and schools. It can be used to check ESR values of capacitors, sort values, select precision values, measure unmarked and unknown inductors, capacitors or resistors, and to measure capacitance, inductance, or resistance of cables, switches, circuit board foils, etc.

The key features are as following:

- Test condition:
 - 1 Frequency : 100Hz / 120Hz / 1KHz / 10KHz / 100KHz (MT4080A only)
 2. Level : 1Vrms / 0.25Vrms / 50mVrms / 1VDC (DCR only)
- Measurement Parameters : Z, Ls, Lp, Cs, Cp, DCR, ESR, D, Q and θ
- Basic Accuracy: 0.2%
- Dual Liquid Crystal Display
- Fast/Slow Measurement
- Auto Range or Range Hold
- USB Interface Communication
- Open/Short Calibration
- Primary Parameters Display:
 - Z : AC Impedance
 - DCR : DC Resistance
 - Ls : Serial Inductance
 - Lp : Parallel Inductance
 - Cs : Serial Capacitance
 - Cp : Parallel Capacitance

- Second Parameter Display:
 - θ : Phase Angle
 - ESR : Equivalence Serial Resistance
 - D : Dissipation Factor
 - Q : Quality Factor
- Combinations of Display:
 - Serial Mode : $Z - \theta$, $C_s - D$, $C_s - Q$, $C_s - \text{ESR}$, $L_s - D$, $L_s - Q$, $L_s - \text{ESR}$
 - Parallel Mode : $C_p - D$, $C_p - Q$, $L_p - D$, $L_p - Q$

1.2 Impedance Parameters

Due to the different testing signals on the impedance measurement instrument, there are DC impedance and AC impedance. The common digital multi-meter can only measure the DC impedance, but the MT4080 can do both. It is a very important issue to understand the impedance parameters of the electronic component.

When we analysis the impedance by the impedance measurement plane (Figure 1.1). It can be visualized by the real element on the X-axis and the imaginary element on the y-axis. This impedance measurement plane can also be seen as the polar coordinates. The Z is the magnitude and the θ is the phase of the impedance.




Figure 1.1

$$Z = R_s + jX_s = |Z| \angle \theta \text{ (}\Omega\text{)}$$

$$R_s = |Z| \cos \theta$$

$$|Z| = \sqrt{R_s^2 + X_s^2}$$

$$X_s = |Z| \sin \theta$$

$$\theta = \tan^{-1} \left(\frac{X_s}{R_s} \right)$$

Z = (Impedance)

R_s = (Resistance)

X_s = (Reactance)

Ω = (Ohm)

There are two different types of reactance: Inductive (X_L) and Capacitive (X_C). It can be defined as follows:

$$X_L = \omega L = 2\pi f L$$

$$L = \text{Inductance}(\text{H})$$

$$X_C = \frac{1}{\omega C} = \frac{1}{2\pi f C}$$

$$C = \text{Capacitance}(\text{F})$$

$$f = \text{Frequency}(\text{Hz})$$


Also, there are quality factor (Q) and the dissipation factor (D) that need to be discussed. For component, the quality factor serves as a measure of the reactance purity. In the real world, there is always

some associated resistance that dissipates power, decreasing the amount of energy that can be recovered. The quality factor can be defined as the ratio of the stored energy (reactance) and the dissipated energy (resistance). Q is generally used for inductors and D for capacitors.

$$\begin{aligned}
 Q &= \frac{1}{D} = \frac{1}{\tan \delta} \\
 &= \frac{|X_s|}{R_s} = \frac{\omega L_s}{R_s} = \frac{1}{\omega C_s R_s} \\
 &= \frac{|B|}{G} \\
 &= \frac{R_p}{|X_p|} = \frac{R_p}{\omega L_p} = \omega C_p R_p
 \end{aligned}$$

There are two types of the circuit mode. One is series mode, the other is parallel mode. See Figure 1.2 to find out the relation of the series and parallel mode.

Real and imaginary components are serial



Real and imaginary components are Parallel




Figure 1.2

1.3 Specification

- LCD Display Range:

Parameter	Display Range
Z	0.000 Ω to 9999 M Ω
L	0.000 μ H to 9999 H
C	0.000 pF to 9999 F
DCR	0.000 Ω to 9999 M Ω
ESR	0.000 Ω to 9999 Ω
D	0.000 to 9999
Q	0.000 to 9999
θ	-180.0 $^\circ$ to 180.0 $^\circ$

- Accuracy (Ae):

Z Accuracy:

Freq. \ Zx	20M ~ 10M (Ω)	10M ~ 1M (Ω)	1M ~ 100K (Ω)	100K ~ 10 (Ω)	10 ~ 1 (Ω)	1 ~ 0.1 (Ω)
DCR	2% ±1	1% ±1	0.5% ±	0.2% ±	0.5% ±	1% ±1
100Hz	❶		1	1	1	❶
120Hz						
1KHz						
10KHz	5% ±1 ❶	2% ±1				
100KHz (4080A)	NA	5%±1 ❶	2%±1	0.4% ± 1	2%±1	5%±1 ❶

Note : 1.The accuracy applies when the test level is set to 1Vrms.

2.Ae multiplies 1.25 when the test level is set to 250mVrms.

3.Ae multiplies 1.50 when the test level is set to 50mVrms.

4.When measuring L and C, multiply Ae by $\sqrt{1+Dx^2}$ if the Dx > 0.1.

❶ : Ae is not specified if the test level is set to 50mV.

C Accuracy :

100Hz	79.57 pF 159.1 pF	159.1 pF 1.591 nF	1.591 nF 15.91 nF	15.91 nF 159.1 uF	159.1 uF 1591 uF	1591 uF 15.91 mF
	2% ± 1 ①	1% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
120Hz	66.31 pF 132.6 pF	132.6 pF 1.326 nF	1.326 nF 13.26 nF	13.26 nF 132.6 uF	132.6 uF 1326 uF	1326 uF 13.26 mF
	2% ± 1 ①	1% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
1KHz	7.957 pF 15.91 pF	15.91 pF 159.1 pF	159.1 pF 1.591 nF	1.591 nF 15.91 uF	15.91 uF 159.1 uF	159.1 uF 1.591 mF
	2% ± 1 ①	1% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
10KHz	0.795 pF 1.591 pF	1.591 pF 15.91 pF	15.91 pF 159.1 pF	159.1 pF 1.591 uF	1.591 uF 15.91 uF	15.91 uF 159.1 uF
	5% ± 1 ①	2% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
100KHz (4080A)	NA	0.159 pF 1.591 pF	1.591 pF 15.91 pF	15.91 pF 159.1 nF	159.1 nF 1.591 uF	1.591 uF 15.91 uF
	NA	5% ± 1 ①	2% ± 1	0.4% ± 1	2% ± 1	5% ± 1 ①

L Accuracy :

100Hz	31.83 KH 15.91 KH	15.91 KH 1591 H	1591 H 159.1 H	159.1 H 15.91 mH	15.91 mH 1.591 mH	1.591 mH 159.1 uH
	2% ± 1 ①	1% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
120Hz	26.52 KH 13.26 KH	13.26 KH 1326 H	1326 H 132.6 H	132.6 H 13.26 mH	13.26 mH 1.326 mH	1.326 mH 132.6 uH
	2% ± 1 ①	1% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
1KHz	3.183 KH 1.591 KH	1.591 KH 159.1 H	159.1 H 15.91 H	15.91 H 1.591 mH	1.591 mH 159.1 uH	159.1 uH 15.91 uH
	2% ± 1 ①	1% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
10KHz	318.3 H 159.1 H	159.1 H 15.91 H	15.91 H 1.591 H	1.591 H 159.1 uH	159.1 uH 15.91 uH	15.91 uH 1.591 uH
	5% ± 1 ①	2% ± 1	0.5% ± 1	0.2% ± 1	0.5% ± 1	1% ± 1 ①
100KHz (4080A)	31.83 H 15.91 H	15.91 H 1.591 H	1.591 H 159.1 mH	159.1 mH 15.91 uH	15.91 uH 1.591 uH	1.591 uH 0.159 uH
	NA	5% ± 1 ①	2% ± 1	0.4% ± 1	2% ± 1	5% ± 1 ①

D Accuracy :

$ Z_x $ Freq.	20M ~ 10M (Ω)	10M ~ 1M (Ω)	1M ~ 100K (Ω)	100K ~ 10 (Ω)	10 ~ 1 (Ω)	1 ~ 0.1 (Ω)
100Hz	± 0.020 ①	± 0.010	± 0.005	± 0.002	± 0.005	± 0.010 ①
120Hz						
1KHz						
10KHz	± 0.050 ①	± 0.020				
100KHz (4080A)	NA	± 0.050 ①	± 0.020	± 0.004	± 0.020	± 0.050 ①

θ Accuracy :

$ Z_x $ Freq.	20M ~ 10M (Ω)	10M ~ 1M (Ω)	1M ~ 100K (Ω)	100K ~ 10 (Ω)	10 ~ 1 (Ω)	1 ~ 0.1 (Ω)
100Hz	± 1.046 ①	± 0.523	± 0.261	± 0.105	± 0.261	± 0.523 ①
120Hz						
1KHz						
10KHz	± 2.615 ①	± 1.046				
100KHz (4080A)	NA	± 2.615 ①	± 1.046	± 0.209	± 1.046	± 2.615 ①

Z Accuracy:

As shown in table 1.

C Accuracy:

$$|Z_x| = \frac{1}{2 \cdot \pi \cdot f \cdot C_x}$$

C_{Ac} = Ae of $|Z_x|$

f : Test Frequency (Hz)

C_x : Measured Capacitance Value (F)

$|Z_x|$: Measured Impedance Value (Ω)

Accuracy applies when D_x (measured D value) ≤ 0.1

When $D_x > 0.1$, multiply C_{Ac} by $\sqrt{1 + D_x^2}$

Example:

Test Condition:

Frequency : 1KHz

Level : 1Vrms

Speed : Slow

DUT : 100nF

Then

$$|Z_x| = \frac{1}{2 \cdot \pi \cdot f \cdot C_x}$$

$$= \frac{1}{2 \cdot \pi \cdot 10^3 \cdot 100 \cdot 10^{-9}} = 1590\Omega$$

Refer to the accuracy table, get $C_{Ac} = \pm 0.2\%$

L Accuracy:

$$|Z_x| = 2 \cdot \pi \cdot f \cdot L_x$$

L_{Ac} = Ae of $|Z_x|$

f : Test Frequency (Hz)

L_x : Measured Inductance Value (H)

$|Z_x|$: Measured Impedance Value (Ω)

Accuracy applies when D_x (measured D value) ≤ 0.1

When $D_x > 0.1$, multiply L_{Ac} by $\sqrt{1 + D_x^2}$

Example:

Test Condition:

Frequency : 1KHz

Level : 1Vrms

Speed : Slow

DUT : 1mH

Then

$$|Z_x| = 2 \cdot \pi \cdot f \cdot L_x$$

$$= 2 \cdot \pi \cdot 10^3 \cdot 10^{-3} = 6.283\Omega$$

Refer to the accuracy table, get $L_{Ac} = \pm 0.5\%$

ESR Accuracy:

$$ESR_{Ac} = \pm X_x \cdot \frac{Ae}{100}$$

$$X_x = 2 \cdot \pi \cdot f \cdot L_x = \frac{1}{2 \cdot \pi \cdot f \cdot C_x}$$

$$ESR_{Ac} = Ae \text{ of } |Zx|$$

f : Test Frequency (Hz)

Xx : Measured Reactance Value (Ω)

Lx : Measured Inductance Value (H)

Cx : Measured Capacitance Value (F)

Accuracy applies when Dx (measured D value) ≤ 0.1

Example:

Test Condition:

Frequency : 1KHz

Level : 1Vrms

Speed : Slow

DUT : 100nF

Then

$$|Zx| = \frac{1}{2 \cdot \pi \cdot f \cdot Cx}$$

$$= \frac{1}{2 \cdot \pi \cdot 10^3 \cdot 100 \cdot 10^{-9}} = 1590\Omega$$

Refer to the accuracy table, get

$C_{Ac} = \pm 0.2\%$,

$$ESR_{Ae} = \pm Xx \cdot \frac{Ae}{100} = \pm 3.18\Omega$$

D Accuracy:

$$D_{Ae} = \pm \frac{Ae}{100}$$

$$D_{Ac} = Ae \text{ of } |Zx|$$

Accuracy applies when Dx (measured D value) ≤ 0.1

When $Dx > 0.1$, multiply Dx by $(1+Dx)$

Example:

Test Condition:

Frequency : 1KHz

Level : 1Vrms

Speed : Slow

DUT : 100nF

Then

$$|Zx| = \frac{1}{2 \cdot \pi \cdot f \cdot Cx}$$

$$= \frac{1}{2 \cdot \pi \cdot 10^3 \cdot 100 \cdot 10^{-9}} = 1590\Omega$$

Refer to the accuracy table, get

$C_{Ac} = \pm 0.2\%$,

$$D_{Ac} = \pm \cdot \frac{Ae}{100} = \pm 0.002$$

Q Accuracy:

$$Q_{Ac} = \pm \frac{Qx^2 \cdot De}{1 \mp Qx \cdot De}$$

$$Q_{Ac} = Ae \text{ of } |Zx|$$

Qx : Measured Quality Factor Value

De : Relative D Accuracy

Accuracy applies when $Qx \cdot De < 1$

Example:

Test Condition:

Frequency : 1KHz

Level : 1Vrms

Speed : Slow

DUT : 1mH

Then

$$\begin{aligned}|Zx| &= 2 \cdot \pi \cdot f \cdot Lx \\ &= 2 \cdot \pi \cdot 10^3 \cdot 10^{-3} = 6.283\Omega\end{aligned}$$

Refer to the accuracy table, get

$L_{Ae} = \pm 0.5\%$,

$$De = \pm \cdot \frac{Ae}{100} = \pm 0.005$$

If measured $Qx = 20$

Then

$$\begin{aligned}Q_{Ae} &= \pm \frac{Qx^2 \cdot De}{1 \mp Qx \cdot De} \\ &= \pm \frac{2}{1 \mp 0.1}\end{aligned}$$

θ Accuracy:

$$\theta_e = \frac{180}{\pi} \cdot \frac{Ae}{100}$$

Example:

Test Condition:

Frequency : 1KHz

Level : 1Vrms

Speed : Slow

DUT : 100nF

Then

$$|Z_x| = \frac{1}{2 \cdot \pi \cdot f \cdot C_x}$$
$$= \frac{1}{2 \cdot \pi \cdot 10^3 \cdot 100 \cdot 10^{-9}} = 1590\Omega$$

Refer to the accuracy table, get

$Z_{Ae} = \pm 0.2\%$,

$$\theta_{Ae} = \pm \frac{180}{\pi} \cdot \frac{Ae}{100}$$
$$= \pm \frac{180}{\pi} \cdot \frac{0.2}{100} = \pm 0.115 \text{ deg}$$

- Testing Signal:

Level Accuracy : $\pm 5\%$

Frequency Accuracy : 0.1%

- Output Impedance : $100\Omega \pm 5\%$

- Measuring Speed:

Fast : 4.5 meas. / sec.

Slow : 2.5 meas. / sec.

- General:

Temperature	:	0°C to 40°C (Operating) -20°C to 70°C (Storage)
Relative Humidity	:	Up to 85%
Battery Type	:	2 AA size Ni-Mh or Alkaline
Battery Charge	:	Constant current 150mA approximately
Battery Operating Time	:	2.5 Hours typical
AC Operation	:	110/220V AC, 60/50Hz with proper adapter
Low Power Warning	:	under 2.2V
Dimensions	:	174mm x 86mm x 48mm (L x W x H) 6.9" x 3.4" x 1.9"
Weight	:	470g

Considerations

Test Frequency. The test frequency is user selectable and can be changed. Generally, a 1kHz test signal is used to measure capacitors that are 0.01uF or smaller and a 120Hz test signal is used for capacitors that are 10uF or larger. Typically a 1 kHz test signal is used to measure inductors that are used in audio and RF (radio frequency) circuits. This is because these components operate at higher frequencies and require that they be measured at a higher frequency of 1kHz. Generally, inductors below 2mH should be measured at 1 kHz and inductors above 200H should be measured at 120Hz.

It is best to check with the component manufacturer's data sheet to determine the best test frequency for the device.

Charged Capacitors

Always discharge any capacitor prior to making a measurement since a charged capacitor may seriously damage the meter.

Effect Of High D on Accuracy

A low D (Dissipation Factor) reading is desirable. Electrolytic capacitors inherently have a higher dissipation factor due to their normally high internal leakage characteristics. If the D (Dissipation Factor) is excessive, the capacitance measurement accuracy may be degraded.

It is best to check with the component manufacturers' data sheet to determine the desirable D value of a good component.

Combining Auto ranging and Manual Ranging Operation

Combining auto ranging and manual ranging is a very convenient way to gain the advantage of both modes. Starting in the auto ranging mode, insert or connect the inductor to be measured, The instrument quickly steps to the correct range for measurement. Next, press the **RANGE key to switch to the manual** ranging mode. The instrument will be on the correct range. The display will indicate whether a calibration needs to be performed to obtain optimum accuracy. If not, take the reading. If so, perform the calibration then take the reading, This method combines the speed of auto ranging and the accuracy of manual ranging and is very easy and simple to perform.

Series Vs Parallel Measurement (for Inductors)


The MT4080 normally measures inductance in the series equivalent mode. The series mode displays the more accurate measurement in most cases. The series equivalent mode is essential for obtaining an accurate Q reading of low Q inductors. Where ohmic losses are most significant, the series equivalent mode is preferred. However, there are cases where the parallel equivalent mode may be more appropriate. For iron core inductors operating at higher frequencies where hysteresis and eddy currents become significant, measurement in the parallel equivalent mode is preferred.

1.4 Accessories

- Operating Manual 1 pc
- 2 AA Size Ni-Mh Rechargeable Batteries 2 pcs
- Shorting Bar 1 pc
- AC to DC Adapter 1 pc
- TL08A SMD Test Probe (Optional)
- TL08B 4-Wire Test Clip (Optional)
- TL08C Kelvin Clip (Optional)
- Carrying Case (Optional)

2. Operation

2.1 Physical Description




- | | |
|--------------------------------|-------------------------------|
| 1. USB Port | 2. Primary Parameter Display |
| 3. Secondary Parameter Display | 4. Low Battery Indicator |
| 5. Model Number | 6. Power Switch |
| 7. Relative Key | 8. Measurement Level Key |
| 9. Open/Short Calibration Key | 10. Measurement Frequency Key |
| 11. Display Update Speed Key | 12. D/Q/θ/ESR Function Key |
| 13. Range Hold Key | 14. L/C/Z/DCR Function Key |
| 15. Battery Charge Indicator | 16. DC Adapter Input Jack |
| 17. Guard Terminal | 18. HPOT/HCUR Terminal |
| 19. LPOT/LCUR Terminal | 20. Battery Compartment |

2.2 Making Measurement

2.2.1 Battery Replacement

When the LOW BATTERY INDICATOR lights up during normal operation, the batteries in the MT4080 should be replaced or recharged to maintain proper operation. Please perform the following steps to change the batteries:

1. Remove the battery hatch by unscrewing the screw of the battery compartment.
2. Take out the old batteries and insert the new batteries into the battery compartment. Please watch out for battery polarity when installing new batteries.
3. Replace the battery hatch by reversing the procedure used to remove it.



Battery Replacement

Caution

Only the MT4080 standard accessory AC to DC adapter can be used with MT4080. Other battery eliminator or charger may result in damage to MT4080.

The MT4080 works on external AC power or internal batteries. To power the MT4080 with AC source, make sure that the MT4080 is off, then plug one end of the AC to DC adapter into the DC jack on the right side of the instrument and the other end into an AC outlet.

There is a small slide switch inside the battery compartment called Battery Select Switch. If the Ni-Mh or Ni-Cd rechargeable batteries are installed in MT4080, set the Battery Select Switch to "Ni-Mh" position. The Ni-Mh or Ni-Cd batteries can be recharged when the instrument is operated by AC source. The LED for indicating battery charging will light on. If the non-rechargeable batteries (such as alkaline batteries) are installed in MT4080, set the Battery Select Switch to "NORM" position for disconnecting the charging circuit to the batteries.

Warning


The Battery Select Switch must be set in the "NORM" position when using non-rechargeable batteries. Non-rechargeable batteries may explode if the AC adapter is used with non-rechargeable batteries. Warranty is voided if this happened.

2.2.3 Open and Short Calibration

The MT4080 provides open/short calibration capability so the user can get better accuracy in measuring high and low impedance. We recommend that the user performs open/short calibration if the test level or frequency has been changed.

- Open Calibration


First, remaining the measurement terminals with the open status, then, press the **CAL** key shortly (no more than two second), the LCD will display:



This calibration takes about 10 seconds. After it is finished, the MT4080 will beep to show that the calibration is done.

- Short Calibration

To perform the short calibration, insert the Shorting Bar into the measurement terminals. Press the **CAL** key for more than two second, the LCD will display:



This calibration takes about 10 seconds. After it is finished, the MT4080 will beep to show that the calibration is done.

2.2.4 Display Speed


The MT4080 provide two different display speeds (Fast/Slow). It is controlled by the **Speed** key. When the speed is set to fast, the display will update 4.5 readings every second. When the speed is set to slow, it's only 2.5 readings per second.

2.2.5 Relative Mode

The relative mode lets the user to make quick sort of a bunch of components. First, insert the standard value component to get the standard value reading. (Approximately 5 seconds in Fast Mode to get a stable reading.) Then, press the **Relative** key, the primary display will reset to zero. Remove the standard value component and insert the unknown component, the LCD will show the value that is the difference between the standard value and unknown value.


2.2.6 Range Hold

To set the range hold, insert a standard component in that measurement range. (Approximately 5 seconds in Fast Mode to get a stable reading.) Then, by pressing the **Range Hold** key it will hold the range within 0.5 to 2 times of the current measurement range. When the **Range Hold** is press the LCD display:




2.2.7 DC Resistance Measurement

The DC resistance measurement measures the resistance of an unknown component by 1VDC. Select the **L/C/Z/DCR** key to make the DCR measurement. The LCD display:



2.2.8 AC Impedance Measurement

The AC impedance measurement measures the Z of an unknown device. Select the **L/C/Z/DCR** key to make the Z measurement. The LCD display:




The testing level and frequency can be selected by pressing the **Level** key and **Frequency** key, respectively.

2.2.9 Capacitance Measurement

To measure the capacitance of a component, select the **L/C/Z/DCR** key to Cs or Cp mode. Due to the circuit structure, there are two modes can be selected (Serial Mode – Cs and Parallel Mode – Cp). If the serial mode (Cs) is selected, the D, Q and ESR can be shown on the secondary display. If the parallel mode (Cp) is selected, only


the D and Q can be shown on the secondary display. The following shows some examples of capacitance measurement:



The testing level and frequency can be selected by pressing the **Level** key and **Frequency** key, respectively.

2.2.10 Inductance Measurement


Select the **L/C/Z/DCR** key to Ls or Lp mode for measuring the inductance in serial mode or parallel mode. If the serial mode (Ls) is selected, the D, Q and ESR can be shown on the secondary display. If the parallel mode (Lp) is selected, only the D and Q can be shown on the secondary display. The following shows some examples of capacitance measurement:




The testing level and frequency can be selected by pressing the **Level** key and **Frequency** key, respectively.

2.3 Accessory Operation


Follow the figures below to attach the test probes for making measurement.




Shorting Bar



TL08A SMD Test Probe



TL08B 4-Wire Test Clip



TL08C Kelvin Clip

3. USB Operation

There are three operation modes in the USB operation of MT4080. They are **Normal**, **Remote** and **Remote Binning** modes.

- **Normal:**

The **Normal** mode is the default power on local mode. It is a local mode that the MT4080 is controlled by the keypad and the results of the measurement will be sent to LCD display and a remote USB equipped PC through the build-in USB port.

- **Remote:**

In the **Remote** mode, the MT4080 is capable of communicating to USB equipped PC or terminal through the build-in USB interface. The connection setting is as follow:

Transmission Mode	: Half Duplex
Baud Rate	: 9600
Parity Bit	: None
Data Bits	: 8
Stop Bit	: 1
Handshake	: None

In this mode, the keyboard and LCD will be locked. And, the MT4080 measurement is controlled by the external program through the USB port.

- **Remote Binning:**

In the **Remote Binning** mode, the “RMT” on the LCD will flash. The MT4080 performs as a TALK ONLY instrument. That means, the measurement of MT4080 is controlled by instrument keys, but the measured value will display on the LCD as well as output to the USB port. By this way, the user can purchase the optional application program provided by Motech to obtain the GO/NO GO comparator and the component sorting comparator.

3.1 Command Syntax

The command syntax of MT4080 is as following:

COMMAND(?) (PARAMETER)

The format of COMMAND and PARAMETER is as following:

1. There is at least one space between COMMAND and PARAMETER.
2. The PARAMETER should use only ASCII string not numerical code.
3. Value parameter can be integer, floating or exponent with the unit. For example:
50mV
0.05V
5.0e1mV
4. The question mark (?) at the end of COMMAND means a query or measure back command. For example:
"CpD" sets the measurement mode to Cp and D.
"CpD?" sets the measurement mode to Cp and D as well as measures the values and send it back.
5. The COMMAND and PARAMETER can be either upper or lower case. But the unit to describe the value in the PARAMETER should have different between milli (m) and mega (M). For example:
1mV equals 0.001V.
1MV equals 1000000V.
6. The "end of command" character should placed at the end.
There are:
ASCII CR (0DH) or
ASCII LF (0AH)

3.2 Commands

Measurement Setting (or Querying) Command

The following measurement mode-setting and the query commands are supported in the MT4080. When a mode-setting command is entered the MT4080 will return the ASCII CR (ODH) and ASCII LF (0AH) after setting is complete. When query command is entered, the MT4080 will send back the values of measurement. After a command is entered, the meter shall respond within 2.5 seconds with the return values follow the ASCII CR and ASCII LF. If an illegal command is entered, there is no response from the meter.

- DCR(?) DC resistance measurement mode setting or querying command.
- CpRp(?) Parallel capacitance and parallel resistance measurement mode setting or querying command.
- CpQ(?) Parallel capacitance and quality factor measurement mode setting or querying command.
- CpD(?) Parallel capacitance and dissipation factor measurement mode setting or querying command.
- CsRs(?) Serial capacitance and serial resistance measurement mode setting or querying command.
- CsQ(?) Serial capacitance and quality factor measurement mode setting or querying command.
- CsD(?) Serial capacitance and dissipation factor measurement mode setting or querying command.

- **LpRp(?)** Parallel inductance and parallel resistance measurement mode setting or querying command.
- **LpQ(?)** Parallel inductance and quality factor measurement mode setting or querying command.
- **LpD(?)** Parallel inductance and dissipation factor measurement mode setting or querying command.
- **LsRs(?)** Serial inductance and serial resistance measurement mode setting or querying command.
- **LsQ(?)** Serial inductance and quality factor measurement mode setting or querying command.
- **LsD(?)** Serial inductance and dissipation factor measurement mode setting or querying command.
- **RsXs(?)** Serial resistance and serial reactance measurement mode setting or querying command.
- **RpXp(?)** Parallel resistance and parallel reactance measurement mode setting or querying command.
- **ZTD(?)** Impedance and angle (Deg) measurement mode setting or querying command.
- **ZTR(?)** Impedance and angle (Rad) measurement mode setting or querying command.

Example:

CPD (*set to Cp-D measurement mode*)

CPD?

0.22724 0.12840 (*return value*)

DCR?

5.1029 (*return value*)

***IDN?**

Queries and identifies the LCR Meter. This command is used to identify the basic information of LCR Meter.

***RST**

Reset the MT4080 to the power on default status. The default status is:

1KHz 1Vrms SLOW CpD uF mH Ohm

After the MT4080 is reset, it will beep once and returns the "BEEP" string back.

ASC

Set the format of the return value. This command sets the ASCII string return or the numerical code.

PARAMETER:

ON ASCII string
OFF Numerical code

Example:

ASC ON
FREQ?
1KHz (*return value*)

ASC OFF
FREQ?
2 (*return value*)

CORR OPEN

Perform the open calibration. This command sets the MT4080 to do the open calibration. After the calibration is done, the MT4080 will beep once and returns the "BEEP" string back.

CORR SHORT

Perform the short calibration. This command sets the MT4080 to do the short calibration. After the calibration is done, the MT4080 will beep once and returns the "BEEP" string back.

FREQ(?) PARAMETER

Set (query) the measurement frequency.

- **FREQ PARAMETER**

Set the measurement frequency according to the parameter.
There is no return value.

PARAMETER:

ASCII string	Numerical code
100Hz	0
120Hz	1
1KHz	2
10KHz	3
100KHz	4

Example:

FREQ 100KHz

- **FREQ?**

Return the current measurement frequency setting.

Example:

ASC ON
FREQ?
1KHz (*return value*)

ASC OFF
FREQ?
2 (*return value*)

LEV(?) PARAMETER

Set (query) the measurement level.

- **LEV PARAMETER**

Set the measurement level according to the parameter. There is no return value.

PARAMETER:

ASCII string	Numerical code
1VDC	0
1Vrms	1
250mVrms	2
50mVrms	3

Example:

LEV 1V

- **LEV?**

Return the current measurement level setting.

Example:

ASC ON

LEV?

1Vrms (*return value*)

ASC OFF

LEV?

1 (*return value*)

MODE?

Query the measurement mode. Six fields will be returned.

1. Frequency
2. Level
3. Speed
4. Measurement mode
5. Unit of primary display
6. Unit of secondary display

The existence of field 6 depends on the measurement mode. For example, there's no field 6 if the measurement mode is DCR. The separation between fields is space (ASCII 20H).

Example:

```
ASC ON
CPD
MODE?
1KHz 1Vrms SLOW CpD uF (return value)
```

```
ASC ON
CPRP
MODE?
1KHz 1Vrms SLOW CpRp uF Ohm (return value)
```

RANG(?) PARAMETER

Set (query) the measurement unit.

- **RANG PARAMETER**

Set the measurement unit according to the parameter. There is no return value.

PARAMETER:

ASCII string	Numerical code
pF	0
nF	1
uF	2
mF	3
F	4
nH	8
uH	9
mH	10
H	11
KH	12
mOhm	17
Ohm	18
KOhm	19

MOhm 20

Example:

RANG pF

- **RANG?**

Return the current measurement unit setting.

Example:

ASC ON

RANG?

pF (*return value*)

ASC OFF

RANG?

0 (*return value*)

READ?

Return the measurement value. This command will perform a measurement according to the current measurement mode and return the measured value.

Example:

CPD

READ?

0.22724 0.12840 (*return value*)

DCR

READ?

5.1029 (*return value*)

The “DCR” measurement will send only one measured value. The other measurement modes will send two measured values separated by space (ASCII 20H).

SPEED(?) PARAMETER

Set (query) the measurement speed.

- **SPEED PARAMETER**

Set the measurement speed according to the parameter.

There is no return value.

PARAMETER:

ASCII string	Numerical code
SLOW	0
FAST	1

Example:

SPEED FAST

- **SPEED?**

Return the current measurement speed setting.

Example:

ASC ON

SPEED?

SLOW (*return value*)

ASC OFF

SPEED?

0 (*return value*)

4. Application

4.1 Test Leads Connection

Auto balancing bridge has four terminals (H_{CUR} , H_{POT} , L_{CUR} and L_{POT}) to connect to the device under test (DUT). It is important to understand what connection method will affect the measurement accuracy.

- 2-Terminal (2T)

2-Terminal is the easiest way to connect the DUT, but it contains many errors that are the inductor and resistor as well as the parasitic capacitor of the test leads (Figure 3.1). Due to these errors in measurement, the effective impedance measurement range will be limited at 100Ω to $10K\Omega$.




Figure 3.1

- 3-Terminal (3T)

3-Terminal uses coaxial cable to reduce the effect of the parasitic capacitor (Figure 3.2). The shield of the coaxial cable should connect to guard of the instrument to increase the measurement range up to $10\text{M}\Omega$.




Figure 3.2

- 4-Terminal (4T)

4-Terminal connection reduces the effect of the test lead resistance (Figure 3.3). This connection can improve the measurement range down to $10\text{m}\Omega$. However, the effect of the test lead inductance can't be eliminated.




Figure 3.3

- 5-Terminal (5T)

5-Terminal connection is the combination of 3T and 4T (Figure 3.4). It has four coaxial cables. Due to the advantage of the 3T and 4T, this connection can widely increase the measurement range for $10\text{m}\Omega$ to $10\text{M}\Omega$.




Figure 3.4

- 4-Terminal Path (4TP)

4-Terminal Path connection solves the problem that caused by the test lead inductance. 4TP uses four coaxial cables to isolate the current path and the voltage sense cable (Figure 3.5). The return current will flow through the coaxial cable as well as the shield. Therefore, the magnetic flux that generated by internal conductor will cancel out the magnetic flux generated by

external conductor (shield). The 4TP connection increases the measurement range from $1\text{m}\Omega$ to $10\text{M}\Omega$.




Figure 3.5

- **Eliminating the Effect of the Parasitic Capacitor**
When measuring the high impedance component (i.e. low capacitor), the parasitic capacitor becomes an important issue (Figure 3.6). In figure 3.6(a), the parasitic capacitor C_d is paralleled to DUT as well as the C_i and C_h . To correct this problem, add a guard plane (Figure 3.6(b)) in between H and L terminals to break the C_d . If the guard plane is connected to instrument guard, the effect of C_i and C_h will be removed.




Figure 3.6

4.2 Open/Short Compensation

For those precision impedance measuring instruments, the open and short compensation need to be used to reduce the parasitic effect of the test fixture. The parasitic effect of the test fixture can be treated like the simple passive components in figure 3.7(a). When the DUT is open, the instrument gets the conductance $Y_p = G_p + j\omega C_p$ (Figure 3.7(b)). When the DUT is short, the instrument gets the impedance $Z_s = R_s + j\omega L_s$ (Figure 3.7(c)). After the open and short compensation, Y_p and Z_s are for calculating the real Z_{dut} (Figure 3.7(d)).






Figure 3.7




Figure 3.7 (Continued)

4.3 Selecting the Series or Parallel Mode

According to different measuring requirement, there are series and parallel modes to describe the measurement result. It is depending on the high or low impedance value to decide what mode to be used.

- Capacitor
The impedance and capacitance in the capacitor are negatively proportional. Therefore, the large capacitor means the low impedance; the small capacitor means the high impedance. Figure 3.8 shows the equivalent circuit of capacitor. If the capacitor is small, the R_p is more important than the R_s . If the capacitor is large, the R_s shouldn't be avoided. Hence, uses parallel mode to measure low capacitor and series mode to measure high capacitor.




Figure 3.8

- Inductor

The impedance and inductive in the inductor are positively proportional. Therefore, the large inductor equals to the high impedance and vice versa. Figure 3.9 shows the equivalent circuit of inductor. If the inductor is small, the R_S is more important than the R_P . If the inductor is large, the R_P should be taking care of. So, uses series mode to measure low inductor and parallel mode to measure high inductor.




Figure 3.9

5. Warranty Information

ONE-YEAR-LIMITED WARRANTY

MOTECH INDUSTRIES INC. (MOTECH) warrants to the original user or purchaser that the unit is free from any defects in material or workmanship for a period of one year from the date of purchase. If any defect is discovered within the warranty period, MOTECH will repair or replace the unit, subject to verification of the defect or malfunction, upon delivery or prepaid shipment to MOTECH.

This warranty does not apply to defects or to physical damage resulting from abuse, neglect, accident, improper repair, alteration, or unreasonable use of the unit, resulting in (but not limited to) cracked or broken case or parts, or to units damaged by excessive heat. Except upon initial purchase, this warranty does not cover finish or appearance items nor does it cover items damaged in shipment to MOTECH for repair or calibration.

To receive service under this warranty, you must include proof of purchase, including date and place of purchase, (a copy of your purchase receipt) or MOTECH will not be responsible for repairs or replacement of the unit under warranty.

Any applicable implied warranties, including warranties of merchant ability and fitness for a particular use, are hereby limited to one year from the date of purchase. Consequential or incidental damages resulting from loss of use, or from a breach of any applicable express or implied warranties are hereby excluded.

The warranty is in lieu of all other agreements and warranties, general or special, express or implied no representative or person is authorized to assume for us any other liability in connection with the sale or use of this MOTECH product.

6. Safety Precaution

SAFETY CONSIDERATIONS

The MT4080 LCR Meter has been designed and tested according to Class 1A 1B or 2 according to IEC479-1 and IEC 721-3-3, Safety requirement for Electronic Measuring Apparatus.

SAFETY PRECAUTIONS

SAFETY NOTES

The following general safety precautions must be observed during all phases of operation, service, and repair of this instrument. Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards of design, manufacture, and intended use of the instrument.

The manufacturer assumes no liability for the customer's failure to comply with these requirements.

BEFORE APPLYING POWER

Verify that the product is set to match the available line voltage is installed.

SAFETY SYMBOLS



Caution, risk of electric shock



Earth ground symbol



Equipment protected throughout by double insulation or reinforced insulation



Caution (refer to accompanying documents)

DO NOT SUBSTITUTE PARTS OR MODIFY INSTRUMENT

Because of the danger of introducing additional hazards, do not install substitute parts or perform any unauthorized modification to the instrument. Return the instrument to a qualified dealer for service and repair to ensure that safety features are maintained.

INSTRUMENTS THAT APPEAR DAMAGED OR DEFECTIVE SHOULD BE MADE INOPERATIVE AND SECURED AGAINST UNINTENDED OPERATION UNTIL THEY CAN BE REPAIRED BY QUALIFIED SERVICE PERSONNEL.



MOTech INDUSTRIES INC.

8 F, No. 248, Sec. 3, Rd. Shien Fsi, Shien-Keng Hsiang, Taichai Hsien 222, Taiwan
Tel: +886-2-28626093 +886-2-28626134 Fax: +886-2-28626067
E-mail: instrument@motech.com.tw <http://www.motech.com.tw>

ZOMG-408MT-1E