

BroadLighter-S

Bench Series Broadband SLD Light Sources



User Manual for Models:

- BroadLighter S-670-G-I-5
- BroadLighter S-785-B-I-15, S-790-G-I-15
- BroadLighter S-830-G-I-20, S-840-B-I-20, S-850-G-I-20
- BroadLighter S-930-B-7, S-930-B-I-10, S-960-G-I-20
- BroadLighter S-1300-G-I-20
- BroadLighter S-1550-G-I-10

Version 1.2
Version date: September 28, 2012

Superlum, Unit 1F, Eastlink Business Park, Carrigtwohill, Co. Cork, Ireland.
Phone: +353 21 4533666, Fax: +353 21 4533026, E-mail: sales@superlum.ie.
www.superlumdiodes.com

ATTENTION:

This page is intentionally left blank.

Replace the page with the EC Declaration Conformity when printing and preparing the manual for a customer. Ensure that the EC Declaration has been filled in properly and signed.

Warranty

Warranty

Superlum warrants the material and production of the BroadLighter-S Broadband SLD Light Source for a period of 12 months from the date of shipment. During this warranty period Superlum will repair or exchange any units if they are found to be defective in material or workmanship. Repaired products are warranted for the balance of the original warranty period or at least 90 days.

For warranty repairs or service the unit must be returned to Superlum. The customer is responsible for all costs incurred to ship the unit back to Superlum.

Limitation of Warranty

Superlum does not guarantee that this manual is completely free from errors or inaccuracies, and reserves the right to change or correct operating instructions, specifications, or numerical information presented here at any time, without notice. Superlum is not liable for any incidental damage caused by failure of this product.

This warranty does not cover:

- ❑ Any defect (or damage to the unit) caused by improper user's operation, such as incorrect connection to the instrument, improper handling of the unit, etc.
- ❑ The accessories coming with the instrument—fiber optical patchcables, AC power cords, etc.

The warranty will void if the user opens the unit for disassembling, repairing, or servicing without Superlum's permission.

This product is not intended for use in medical and/or life support situations.

Statement of Calibration

Superlum certifies that this product has been inspected, tested, and calibrated to meet its published specifications.

Contents

Warranty.....	3
List of Figures.....	6
List of Tables.....	7
General Safety Information.....	8
Safety Symbols and Terms.....	10
Acronyms and Abbreviations.....	11
Chapter 1 General Information.....	12
1.1 In This Manual.....	12
1.2 Instrument Description.....	13
1.3 Product Specifications.....	23
1.4 Laser Safety Information.....	25
1.4.1 Class 1 Laser Products.....	26
1.4.2 Class 3R Laser Products.....	29
1.4.3 Class 3B Laser Products.....	33
Chapter 2 Installation.....	36
2.1 Unpacking and Initial Inspection.....	36
2.1.1 Unpacking the Instrument.....	36
2.1.2 Initial Inspection.....	36
2.2 Preparation for Use.....	37
2.3 Fuse Replacement.....	38
Chapter 3 Local Mode of Operation.....	40
3.1 Front-Panel Features.....	40
3.2 Rear-Panel Features.....	42
3.3 Operating Instructions.....	43
3.4 How to Turn Off the Instrument.....	45
3.5 Troubleshooting Information.....	45

Chapter 4	Remote Mode of Operation.....	48
4.1	Introduction to Remote Operation.....	48
4.2	Description of Data Communication Protocol.....	50
4.3	Remote Command Set.....	51
4.3.1	Identification Query.....	51
4.3.2	Local/Remote Enable Command.....	52
4.3.3	Local/Remote Check Query.....	53
4.3.4	Optical Output Enable Command.....	54
4.3.5	Optical Output Check Query.....	54
4.3.6	High-Power-Mode Enable/Disable Command.....	55
4.3.7	High-Power-Mode Check Query.....	56
4.4	The Superlum's BroadLighter Companion Software.....	56
4.4.1	About the Superlum's BroadLighter Companion Software...	56
4.4.2	Minimum System Requirements.....	57
4.4.3	Getting Started.....	57
4.4.3.1	Installing the Software on Your PC System.....	57
4.4.3.2	Connecting the BroadLighter to Your PC System.	58
4.4.3.3	Running the Program.....	58
4.4.3.4	Using the Program.....	60
4.4.3.5	Uninstalling the Program.....	64
Chapter 5	Cleaning Instructions.....	65
5.1	Safety Precautions.....	65
5.2	Cleaning Procedures.....	66
5.2.1	Cleaning a Non-Lensed Connector.....	66
5.2.2	Cleaning the Front-Panel Optical Adapter.....	68
5.2.3	Cleaning the Instrument Cabinet.....	68
Chapter 6	Service and Support.....	69
6.1	Returning the Instrument for Servicing.....	69
6.2	Preparing the Instrument for Shipment.....	69
	Revision Documentation.....	72

List of Figures

Figure	Page
FIGURE A. Illustration of an ESD Wrist Strap.....	9
FIGURE 1.1. Schematic Illustration of the Principle of the Output Protection for the BroadLighter-S.....	15
FIGURE 1.2. Fragment of the Front Panel of the BroadLighter with the HP MODE and SLD ON/OFF pushbuttons	16
FIGURE 1.3. Front-Panel Laser Safety Features.....	22
FIGURE 1.4. Rear-Panel Laser Safety Features.....	22
FIGURE 1.5. Warning Labels for Class 1 Laser Products.....	27
FIGURE 1.6. Examples of the Output Beam Divergence and the MPE.....	28
FIGURE 1.7. Warning Labels for Class 3R Laser Products.....	30
FIGURE 1.8. Warning Labels for Class 3B Laser Products.....	33
FIGURE 3.1. Front-Panel Features of the BroadLighter.....	40
FIGURE 3.2. Rear-Panel Features of the BroadLighter.....	42
FIGURE 4.1. The Wiring Configuration for RS-232 Communication.....	49
FIGURE 4.2. Pinout for the BroadLighter's RS-232 Connector.....	49
FIGURE 4.3. Example of the Program Window.....	59

List of Tables

Table	Page
TABLE A. Acronyms and Abbreviations Used.....	11
TABLE 1.1. BroadLighter Units Description.....	16
TABLE 1.2. General and Mechanical Specifications.....	23
TABLE 1.3. List of Contents.....	24
TABLE 1.4. Classification of the BroadLighters According to IEC 60825-1 (Ed. 2.0 2007-03).....	25
TABLE 2.1. AC Line Fuse Specifications.....	39

General Safety Information

The following general safety precautions must be observed during all phases of operation, service, and repair of the instrument. Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards of design, manufacture, and intended use of the instrument. Superlum assumes no liability for the customer's failure to comply with these requirements.

Before Applying Power

Verify that the product is set to match the available line voltage, and the correct fuse is installed.

Safety Earth Ground

This product is a Safety Class I instrument (provided with a protective earth terminal). To minimize shock hazard, the instrument chassis and cabinet must be connected to an electrical ground. The instrument must be connected to the AC power line through a three-conductor power cable, with the third wire firmly connected to an electrical ground (safety ground) at the power outlet. Any interruption of the protective (grounding) conductor or disconnection of the protective earth terminal will cause a potential shock hazard that could result in personal injury. If the instrument is to be energized via an external autotransformer, the ground terminal of the power connector should be connected to the neutral wire of the AC supply.

DO not operate the instrument in an explosive atmosphere or in the presence of flammable gases or fumes.

DO not use repaired fuses or short-circuited fuse holders. For continued protection against fire, replace the line fuse(s) only with fuse(s) of the same voltage, current rating and type.

DO not perform procedures involving cover or shield removal unless you are qualified to do so. Operating personnel must not remove equipment covers or shields. Procedures involving the removal of covers and shields are for use by service-trained personnel only.

DO not operate damaged equipment. Whenever it is possible that the safety protection features built into this product have been impaired, either through physical

damage, excessive moisture or any other reason, remove power and do not use the product. Return the product to Superlum to ensure the safety features are maintained.

DO not substitute parts or modify equipment. Because of the danger of introducing additional hazards, do not install substitute parts or perform any unauthorized modification to the product. Return the product to Superlum for service and repair.



Electrostatic Discharge (ESD) Precautions

There are numerous internal elements inside the instrument which are extremely susceptible to static electricity. Because of this, protection against ESD is essential while you are setting up, cleaning, or inspecting the instrument. Static electricity can build up on your body and can easily damage sensitive internal circuit elements when discharged. Static discharges even too small to be felt can cause permanent damage to the instrument. To prevent it from being damaged by an ESD event, it is necessary to follow the rules below:

- **Always** wear an ESD wrist strap (fig. A) connected to the common grounded point, having 1-MOhm resistor in series with it, when handling the ungrounded instrument.
- **Always** ground the instrument via either the AC power cord (connected to a properly grounded power outlet) or the rear-panel grounding terminal before starting to work with the instrument.




FIGURE A. Illustration of an ESD Wrist Strap.

Safety Symbols and Terms

This section deals with information on the safety symbols and terms which are used on the instrument or in the manual, and gives explanations what they stand for. These symbols and terms indicate precautions you must strictly follow to safely operate the instrument.

Before you begin using your BroadLighter, read the following explanations:

Compliance



This symbol indicates the field-wiring terminal that must be connected to earth ground before operating the equipment. It is for protection against electrical shock, in case of a fault.



The WARNING sign denotes a hazard. It calls attention to a procedure, practice, or the like, which, if not correctly performed or adhered to, could result in personal injury. Do not proceed beyond a WARNING sign until the indicated conditions are fully understood and met.



The CAUTION sign denotes a hazard. It calls attention to an operating procedure, or the like, which, if not correctly performed or adhered to, could result in damage to or destruction of the entire product or part thereof. Do not proceed beyond a CAUTION sign until the indicated conditions are fully understood and met.



The NOTE sign denotes important information. It calls attention to procedures, practice, condition, or the like, which is essential to highlight.



The CE mark shows that the product complies with all relevant European Legal Directives.



The warning label indicates visible and/or invisible laser radiation.



This label means the compliance with the requirement of the RoHS Directive (Directive 2002/95/EC).

Acronyms and Abbreviations

The necessary information on acronyms and abbreviations is summarized in the following table:

**TABLE A Acronyms and
Abbreviations Used**

Abbreviation	Definition
°C	Degree Celsius
mW	Milliwatt
nm	Nanometer
dB	Decibel
V	Volt
AC	Alternating Current
AEL	Accessible Emission Limit
APC	Automatic Power Control
BLS	Broadband Light Source
CW	Continuous Wave
DC	Direct Current
LED	Light Emitting Diode
MPE	Maximum Permissible Emission
NA	Numerical Aperture
PC	Personal Computer
SLD	Superluminescent Diode
SM	Single Mode
TEC	Thermoelectric Cooler

NOTE

In case of a term for which the definition is not given here, refer to special reference books for the necessary explanations or consult with Superlum on this issue.

1 General Information

This chapter contains general information concerning the BroadLighter Broadband SLD Light Source. Included in this chapter are a description of the instrument, a scope of this manual, general specifications, information on the laser Class to which your BroadLighter is assigned, and a description of the laser safety measures employed in this equipment.

1.1 In This Manual

This manual contains the most comprehensive information required to install, operate, and clean the Superlum's BroadLighter-S Broadband SLD Light Source.

The goal of this section is to give a short overview of the chapters and sections comprising this document. Knowing the layout of the manual helps you find answers to the vast majority of questions you may have about installing, using, programming, cleaning, and servicing your BroadLighter. To this end, the manual has been divided into 6 chapters. Detailed information about the principles of operation, installing, and using the BroadLighter is provided in four chapters which are placed at the beginning of the manual. At the end of the document will be another two chapters containing additional information (i.e. detailed instructions) about cleaning and obtaining service. Use the manual as a reference book when you are installing, operating, or suspecting problems with the instrument.

Please read carefully this manual before using your BroadLighter. You should be especially careful to *warnings* and *cautions* presented in this manual. (If necessary, refer to section "Safety Symbols and Terms" for definitions of these terms.) If, after reading this manual, some of the explanations given here still remain unclear to you or you need extra information on a particular issue that interests you, do not hesitate to contact Superlum for more details or additional explanations.

The chapters in this manual are set out as follows:

Chapter 1, "General Information", begins with outlining the scope of this manual, which comes first. After that, you will find background information

concerning the BroadLighter design—the principles of operation and design as well as the most important changes in this version of the instrument. Following this are technical specifications on the product and laser safety information, which are provided in the corresponding sections of the chapter.

Chapter 2, "Installation", takes you through unpacking and inspection procedures and explains how to prepare the BroadLighter for use. At the end of the chapter are instructions on how to replace the AC power line fuse.

Chapter 3, "Local Mode of Operation", introduces front- and rear-panel features and provides detailed operating instructions to demonstrate how easy the BroadLighter is to use.

Chapter 4, "Remote mode of Operation", explains how to control the instrument remotely by using a PC. It provides a full description of each command and query supported by the instrument and the most comprehensive information on the Superlum's BroadLighter companion software, which is supplied with the instrument.

Chapter 5, "Cleaning Instructions", deals with a description of the proper cleaning procedures that you should regularly apply to both the BroadLighter and optical accessories, such as optical patchcables, adapters, optical connectors, and the like.

Chapter 6, "Service and Support", explains what you should do if you encounter a problem with the instrument. It also gives detailed recommendations for repacking the instrument before sending it for servicing.


1.2 Instrument Description

For a number of optical applications, such as optical coherence tomography, optical metrology, characterization of optical components and the like, broadband light sources based on the use of SLD modules are considered to be best suited. The reason for choosing this type of optical modules is related to their remarkable technical parameters: Among all types of modules presented on the market, the SLD modules best combine high optical power similar to that emitted by laser diodes with broadband optical spectra characteristics of LEDs. These advantages of SLDs allow the user to avoid unwanted interference effects, like speckle or ghost signals. In contrast,

narrowband light sources as laser diodes often give rise to such effects due to high temporal coherence of the laser's output emission.

The BroadLighter is a sophisticated SLD-based broadband light source intended for bench use. Among other broadband light sources of this kind, it features excellent performance parameters, easy-to-use front-panel operation, long-term stability of its technical characteristics, high reliability, minimal maintenance, and many others. Today's state-of-the-art design of the BroadLighter allows the source to be used with equal success both in the lab and production areas. It can be controlled either manually by using its front-panel controls or remotely with a PC. For remote control, the instrument is equipped with the standard RS-232 interface. The Superlum's BroadLighter companion software, which is supplied with the device, has a user-friendly interface that permits the user to put the instrument under remote control immediately after its installation. No special skills are required for installing and using the software.

Optical performance of the BroadLighter-S is based on a Superlum's high-power, broadband SLD module emitting the radiation in the required spectral range. Because the SLDs of this type are rated as high-power optical devices, which possess high optical gain, they feature very high susceptibility to optical back reflections (usually called "optical feedback"). For this reason, the Broadlighter-S with unprotected optical output can easily be placed at risk of having its SLD damaged by such optical back reflections, which commonly exist in any experimental setup if no special measures for preventing them are provided. Catastrophic destruction of the SLD can occur immediately if the level of back reflections is far greater than -30 dB. Even if catastrophic failure does not occur immediately, the device may suffer from parametric degradation, which may lead to degraded performance. In most models of the BroadLighter-S, the optical output of the instrument is traditionally protected by an appropriate polarization-independent fiber-optic isolator, which is incorporated into the instrument design (fig. 1.1). But if the level of back reflections is expected to be well below -30dB and is insured to be unchangeable during the practical usage, the instrument can be assembled without any optical isolator. Such a version of the BroadLighter-S is also available for ordering. Please note that in this case, the use of the optical patchcable with FC-APC connectors for connecting between the BroadLighter and your measuring equipment is *absolutely necessary*.



* FC/PC-connectorization is also available

FIGURE 1.1. Schematic Illustration of the Output Protection for the BroadLighter-S.

Before starting to work with the instrument, you should clearly understand the following. Like the previous version of the BroadLighter-S, your new instrument offers the same two modes of operation:

1. Low Power mode (later in the text abbreviated as "LP")
2. High Power mode (later in the text abbreviated as "HP")

In the first operating mode, the optical output power of the instrument is far lower than the full optical power accessible in the HP mode. *Note that the main aim of the LP mode is to produce a small amount of the optical power (nearly 1 milliwatt) at the optical output of the instrument to let you safely align the optical elements in your optical scheme (or set up your measuring system) before taking the necessary measurements at the full output power.* In addition to lower power, you should bear in mind that output spectral characteristics (the -3dB bandwidth, first of all) can significantly differ from those provided in the HP mode at the full output power. When your optical scheme (or measuring system) has been aligned and ready for use, set the BroadLighter to the HP mode, in which it provides the full optical output power, and perform the necessary measurements. By toggling the HP MODE pushbutton on the front panel of the instrument you can switch between the two modes (fig. 1.2). Lighting the pushbutton LED indicates activation of the HP mode. *Please note that you must disable the optical output of the instrument before switching between the LP and HP modes.*

The internal structure of the BroadLighter is based on a modular design. The instrument consists of four plug-in units incorporated into a common main frame: (1) a DC power supply, (2) a central processing unit, (3) a current-and-temperature controller for an SLD module, (4) an optical unit, and (5) the main frame. After

installing in the main frame, all four units are covered by the front panel, which is attached to the main frame by using four hex screws.




FIGURE 1.2. Fragment of the Front Panel of the BroadLighter with the HP MODE and SLD ON/OFF pushbuttons.

Such a principle of the design for the BroadLighter greatly simplifies assembling during production and servicing as well as provides flexible "plug-and-play" exchange of any unit and the ability to easily adapt the instrument to a particular need of the user. According to the main functions of the units, they can be described as follows:

TABLE 1.1 **BroadLighter Units Description**

Unit	Description
DC Power Supply	Supplies DC power to all the units.
Central Processing Unit	Controls the units comprising the instrument for the correct operation.
Current and Temperature Controller	Controls the operating temperature and driving current of an SLD module.
Optical Unit	Contains elements that comprise the optical scheme of the instrument.

The Superlum's SLD module used in the BroadLighter-S is a fiber-coupled optical module that includes a thermoelectric cooler (TEC) and a temperature sensor for active temperature control. The Superlum's current-and-temperature controller is used to control the driving current and operating temperature of the SLD module. The current and temperature drivers are both combined on a common printed circuit board of the current-and-temperature controller. To provide safe operation of cooled SLD modules, a set of SLD protective measures has been incorporated into the controller design. Among the measures implemented, the following are considered to be the most important ones:

- **Slow start of SLD driving current**—When the current driver is activated, a slow-start circuit gradually raises the current from zero to a predetermined nominal current value during the time of 0.1 sec, allowing the current to be slowly sourced to an SLD module. With this feature in use, an overdamped switch-on response of the driver without any overshooting over the nominal current setting is guaranteed. Moreover, the time of 0.1 sec is long enough to eliminate a sharp temperature-related current impact on the TEC at start-up; therefore ensuring high reliability for the TEC operation throughout its lifetime. Another benefit of the 0.1-sec start-up time is that all turn-on transients in the driving current completely die out during this time before applying the current to an SLD, thereby not affecting the SLD operation. (It is important to note that, in most cases, applying these transients to an SLD leads to either its catastrophic destruction or parametric degradation of its performance.)
- **Clamping limit circuit for SLD current** —This circuit provides a clamping limit for the SLD driving current so that an SLD is effectively prevented from overdrive by an excessive current above the limit point (commonly called "limiting threshold"). To help the user identify the situation in which the driving current has reached the limiting threshold, a LED indicator called "Service" has been added on the front panel of the BroadLighter. Once the current has reached the limiting threshold, the LED lights up, thereby showing the user that the current is being limited.

- ❑ **Automatic power control (APC) for SLD operation**—This feature makes it possible to maintain the optical output power of an SLD module at a constant level by automatically increasing or decreasing the SLD driving current. To do this, a feedback loop with a photodiode monitor built into the SLD module is used. The function of the loop is to provide constant feedback on the photodiode current (which in turn is proportional to the output power of the SLD module) to the current driver for allowing the driver to change the current flowing through the SLD in the proper way.
- ❑ **Emergency switch-off for SLD module**—When there is any risk for an SLD module to be operated improperly (for example, the active temperature control for the SLD module has stopped to maintain the temperature within the predetermined temperature window), the module will be automatically switched off to prevent it from being damaged. In addition to this, the time response of the APC feedback loop is fast enough to rapidly switch off an SLD if unexpected changes (or fluctuations) in the photodiode current occur. Therefore, an adequate level of SLD protection from unexpected changes in the feedback loop response is also provided. When the emergency switch-off occurs, the unit will inform the user about the situation by red illumination on the front-panel SLD activation pushbutton.
- ❑ **Shunting the SLD emitter**—Each time when the operating current through the SLD is off, the SLD current driver shunts the SLD emitter with a protective short-circuit. This remains unchanged until the SLD activation pushbutton is pressed. When this happens, the shunt is opened allowing the drive current to flow through the SLD. Using this feature helps protect the SLD emitter from any damage by turn-on transients.
- ❑ **Electronic fast transient suppression**—Like laser diodes, SLD modules are very sensitive to electronic fast transients and spikes in the driving current which may lead to catastrophic degradation of an SLD module. To get an adequate level for transient suppression on apparatus level, a linear regulated DC power supply featuring a low value of the ripple-and-noise parameter has been used in the BroadLighter.

Additionally, each current-and-temperature controller for driving an SLD module contains specially designed inductive-capacitive filter. It is used to suppress transients that might reach the DC-voltage-supply rails of the controller by radiating onto connecting cables inside the instrument.

The instrument is available in two configurations for AC power line: (1) for 220V AC operating voltage and (2) for 110V AC operating voltage. While ordering the instrument, the user needs to specify the desired operating voltage. Note that the desired AC voltage configuration is preset at the factory and is inaccessible for the user to change. All essential information about AC power requirements is shown on the rear panel of the instrument. It is important to note that the BroadLighter is intended for indoor use *only*. For more information on environmental conditions, refer to section "Specifications".

This equipment is intended for fiber-optic applications. To serve this purpose, it is equipped with FC-APC fiber output. Each BroadLighter is shipped with an individual optical patchcable which is used to deliver the optical output power of the instrument to the user's measuring equipment. It is important to note that the patchcable coming with the instrument is specially selected to best match the optical output of the instrument. With the patchcable attached, a low-loss connection between the BroadLighter and the patchcable is obtained. Connecting the user's patchcables to the BroadLighter's output is permitted but not recommended, because in this case the minimum optical losses in the instrument-to-patchcable connection can hardly be ensured. When connecting the user's patchcable to the BroadLighter is the only thing to do, FC-APC connectorization of its ends is the crucial requirement for the patchcable to be used.

Regardless of the fact that you may probably have practical experience with our broadband SLD light sources of earlier versions, please pay close attention to the following: *This version of the BroadLighter S-series incorporates significant improvements in the design of the instrument.*

New to this model are laser safety features and the front panel of the instrument, which has been re-designed considerably. To meet the laser safety requirements stated in standard IEC 60825-1 "Safety of laser products" (Edition 2.0 2007-03), numerous changes in the device operation have been made. They entirely

concern some pieces of electronics used in the BroadLighter. Their operation is now based on new principles. The design of the optical unit of the instrument has been left unchanged. A dust cap on a metal chain, which will help the user not to lose the cap when it is not in use, now protects the optical output of the BroadLighter.

The BroadLighter now incorporates the following safety features required by IEC 60825-1:

- ❑ **Key-operated master control**—Located on the front panel of the instrument, this switch provides access to activation of the optical output (see Figure 1.3). The BroadLighter is inoperable when the key is removed. Using the key-operated master control prevents unauthorized activation of the optical output emission.
- ❑ **Emission indicator**—This LED, located on the SLD emission activation pushbutton, lights up green when the optical power is enabled, thereby warning the user about the output emission. The LED will switch to red when the SLD emergency switch-off occurs.
- ❑ **3-second activation delay**—For personal safety, the instrument keeps its optical output inactive during 3 seconds after pressing the SLD emission activation pushbutton. During this time, the user can cancel the SLD emission activation by pressing the button again. Consider this feature as a protective measure when activation of the instrument's output is unintentionally caused.
- ❑ **Remote interlock connector**—This feature, located on the rear panel of the instrument, allows the instrument to be connected to your local external controls for laser safety (see Figure 1.4). The instrument is inoperable unless the interlock connection is made. If there are no external controls in your local area or the use of these controls is not essential for you, an appropriate short-circuited connector must be installed to the rear-panel remote interlock jack. With this connector installed, you will be able to use the instrument independently of your local laser safety system. The necessary connector is supplied with the instrument.
- ❑ **Manual reset for the remote interlock connection**—This feature requires the user to reset the instrument each time when the remote

interlock connection is opened. Until the instrument is reset, it ignores any user's attempt to enable the optical output and warns the user about the situation by visual and audible warning signals. The reset of the instrument is performed with the front-panel key of the master control. Turning the key to the OFF position and then back to the ON position resets the BroadLighter and makes it ready for further operation.

- **Visual and audible warning about optical output emission**—These features start to operate when the user activates the optical output by pressing the SLD activation pushbutton. The features remain active during 3 sec after pressing the pushbutton, thereby warning the user that the output SLD's emission will appear at the optical output soon.
- **Optical radiation hazard labeling**—Each BroadLighter carries on its cabinet a set of warning labels informing the user about the Class of radiation hazard to which the instrument is assigned. Classification is based on the rules specified in standard IEC 60825-1 (Edition 2.0 2007-03) "Safety of laser products—Part 1: Equipment classification and requirements".

This new user manual has accommodated all these changes. Please read the manual thoroughly to feel at ease with the instrument. Sitting down with this manual and the instrument and trying all its features and functions are also highly recommended. It will not take long to become familiar with the instrument and its features.




FIGURE 1.3. Front-Panel Laser Safety Features




FIGURE 1.4. Rear-Panel Laser Safety Features

1.3 Product Specifications

Contained in this section are general specifications for the BroadLighter-S series which are summarized in the tables below.

Note that the information given here covers general and mechanical specifications *only*. For the technical information concerning optical characteristics for your BroadLighter, please refer to its Acceptance Test Report or contact Superlum on this issue.

TABLE 1.2 General and Mechanical Specifications

Parameter	Value or Description
General	
Power Requirements	100-120VAC ¹ /220-240VAC, 50/60Hz, approx. 45VA
AC Power Cord Plug	Euro or US Standards
Operating Temperature Range	+10 to +40 °C
Storage Temperature Range	-30 to +70 °C
Humidity	< 90% relative humidity, non-condensing
Warm-up Time	10 min
Mechanical	
Dimensions (WxHxD)	257×161×317 mm
Weight	7 kg
Enclosure	Metal case
Applicable Optical Connector	FC/APC ²
Applicable Optical Fibers	SM Corning types
RS-232 Interface Connector	DB-9 (D-SUB type)

¹ – the unit configured for 110VAC operation is shipped on request.

² – for the models with a built-in optical isolator, FC/PC connector is available on request.

Included in the package is one of each of the following:

TABLE 1.3 List of Contents

Item	Qty.
BroadLighter-S Broadband SLD Light Source	1 pc.
AC Power Cord	1 pc.
Master Key	2 pcs.
Optical Patch Cable	1 pc.
Short-circuited monophonic connector	1 pc.
Quick Start Guide	1 pc.
Acceptance Test Report	1 pc.
CD-ROM	1 pc.
RS-232 null-modem cable (DB9F/DB9F)	1 pc.

NOTE

- 1. Optical performance parameters, such as 3-dB bandwidth, central wavelength, output power, etc., are specified and measured for each instrument individually. For detailed information on this issue, refer to the Acceptance Test Report supplied with the instrument.**
 - 2. Specifications are subject to change without notice.**
-

1.4 Laser Safety Information

Depending on a degree of optical radiation hazard of your BroadLighter (which in turn depends on its optical output power and the emission wavelength), it can be assigned to one of three possible laser Classes as specified in Table 1.4. Refer to the table to familiarize yourself with the necessary information on the laser Class for your BroadLighter. Note that the results of classification summarized in Table 1.4 have been obtained in accordance with the requirements specified in international standard IEC 60825-1 (Edition 2.0 2007-03) "Safety of laser products—Part 1: Equipment classification and requirements". First examine the table closely to understand which Class of laser products your BroadLighter is assigned to; then select and read the information on that Class in the corresponding section which follows later.

**TABLE 1.4 Classification of the BroadLighters
According to IEC 60825-1 (Ed. 2.0 2007-03)**

Model	Power	Spectral Characteristics Wavelength/Width	Laser Class	Page
S-670-G-I-5	5mW	670nm/7nm	Class 3R	29
S-680-G-I-5	5mW	680nm/9nm	Class 3R	29
S-785-B-I-15	15mW	785nm/50nm	Class 3B	33
S-790-G-I-15	15mW	790nm/17nm	Class 3B	33
S-830-G-I-20	20mW	830nm/14nm	Class 3B	33
S-840-B-I-20	15mW	840nm/50nm	Class 3B	33
S-850-G-I-20	20mW	850nm/14nm	Class 3B	33
S-930-B-7	7mW	930nm/100nm	Class 3R	29
S-930-B-I-10	10mW	930nm/70nm	Class 3R	29
S-960-G-I-20	20mW	960nm/30nm	Class 3B	33
S-1300-G-I-20	20mW	1300nm/50nm	Class 1	26
S-1550-G-I-10	10mW	1550nm/45nm	Class 1	26
D-810-HP	12mW	810nm/100nm	Class 3B	33
D-840-HP	14mW	840nm/100nm	Class 3B	33

D-840-HP-I	10mW	840nm/100nm	Class 3B	33
D-880-HP	10mW	880nm/100nm	Class 3R	29
D-880-HP-I	7mW	880nm/100nm	Class 3R	29
D-890-HP	6mW	890nm/150nm	Class 3R	29
D-980-HP	7mW	980nm/210nm	Class 3R	29
D-1300-HP	10mW	1300nm/120nm	Class 1	26
D-1300-HP-I	9mW	1300nm/120nm	Class 1	26
D-1500-LP	0.2mW	1500nm/170nm	Class 1	26
T-860-HP	15mW	860nm/135nm	Class 3B	33
T-870-HP	8mW	870nm/180nm	Class 3R	29
Q-870-HP	5mW	870nm/200nm	Class 3R	29
Q-940-HP	4mW	940nm/300nm	Class 3R	29
Q-1350-HP	5mW	1350nm/200nm	Class 1	26
Q-1430-HP	5mW	1430nm/250nm	Class 1	26

NOTE

- 1. Each BroadLighter is assigned to a particular Class only if it completely meets the requirements specified in IEC 60825-1 (Ed. 2.0 2007-03) for that Class.**
 - 2. The results of classification listed in Table 1.4 have been obtained by Superlum in accordance with classification rules of the IEC 60825-1 standard.**
-

1.4.1 Class 1 Laser Products

If you begin to read this section, this means that your BroadLighter has been assigned to Class 1 laser products. Before you start working with your instrument, acquaint yourself with the important information on this Class by reading through the explanations below.

Each BroadLighter that is assigned to this Class of laser products carries on the instrument's cabinet a set of warning labels with the necessary information according to its laser hazard, like the ones shown in Figure 1.5.




FIGURE 1.5. Warning Labels for Class 1 Laser Products.

It should be noted that the labels above are only examples of possible labeling for a particular BroadLighter assigned to this Class of laser products; consequently, the labels shown in Figure 1.5 can be different from those really affixed to your instrument. This difference lays in numeric data for "MAX power" and "Wavelength" shown on the right-hand label. To find out the difference, refer to the same label on your BroadLighter.

General specifications for all BroadLighters assigned to this Class can be summarized in the following way:

Product Type:	220VAC-Powered Broadband SLD Light Source
Model Number:	BroadLighter S-1300-G-I-20 BroadLighter S-1550-G-I-10 BroadLighter D-1300-HP BroadLighter D-1300-HP-I BroadLighter D-1500-LP BroadLighter Q-1350-HP BroadLighter Q-1430-HP
Laser Class (according to IEC 60825-1 [Edition 2.0 2007-03]):	Class 1
Maximum Rated Output Power:	See Table 1.4
Wavelength:	See Table 1.4
Output Aperture (for all models):	SM fiber with NA=0.14 (see Figure 1.6)




FIGURE 1.6. Examples of the Output Beam Divergence and the MPE.

(*) – To experimentally determine the MPE, it is necessary to use an aperture of 7 mm in diameter for measuring the MPE at the cornea of the human eye, and an aperture of 3.5 mm in diameter, for measuring the MPE at the skin.

Light sources assigned to this Class are safe during the use, including the long-term direct intrabeam viewing, even if the exposure with optical viewing instruments (e.g. eye loupes or binoculars) occurs. Extra care must be taken to the following: *Using controls, adjustments, or operating procedures that are different from those specified in this manual may result in radiation exposure hazardous to the user.*

When the instrument is used with certain optical elements—such as optical amplifiers, lenses, collimators, and the like—which are not supplied with the product as original accessories, or the product is integrated into a high-level system, it is strictly recommended to reclassify the final optical system in order to determine its total degree of optical radiation hazard. The reason for this is that the Class may increase under these conditions. It is the responsibility of a person (or organization) who performs such integration to provide correct reclassification for the optical system at the final stage.

It must be noted that it is the *user* who is responsible for being aware of possible risk associated with using Class 1 laser products as well as being adequately trained for working with such light sources. Superlum is not liable for any personal injury, property damage, or other (direct or indirect) damage of any nature caused by personal faults, improper and/or inadequate using of the instrument coming from disregarding the necessary aspects of laser safety.

Being initially designed as a Class 3B product, the instrument has inherited most of the laser safety features used for that Class. Although they are redundant for Class 1 laser products, they have been remained unchanged in this instrument. These features are as follows:

- ❑ Key-operated master control
- ❑ Remote interlock connector
- ❑ Illuminated indicator on the pushbutton for the output emission activation
- ❑ Three-second-time delay for the optical output activation
- ❑ Audible and visible warning signals for the output emission activation
- ❑ Protective metal cap for the optical output
- ❑ Laser safety labeling

This manual contains the most comprehensive information on the laser safety controls (and measures) built into the instrument. Please read thoroughly this manual to acquaint yourself with all features and functions for laser safety before you start working with the BroadLighter.

CAUTION Using controls, adjustments, or operating procedures that are different from those specified in this manual may result in radiation exposure hazardous to the user.

NOTE According to IEC 60825-1, this product has been classified for an exposure duration of 100 ms, which is supposed to be worst-case scenario. For other practical applications in which the exposure duration is greater than 100 ms, the product must be reclassified.

1.4.2 Class 3R Laser Products

If you begin to read this section, this means that your BroadLighter has been assigned to Class 3R laser products. Before you start working with your instrument, acquaint yourself with the important information on this Class by reading through the explanations below.

Each BroadLighter that is assigned to this Class of laser products carries on the instrument's cabinet a set of warning labels with the necessary information according to its laser hazard, like the ones shown in Figure 1.7.




FIGURE 1.7. Warning Labels for Class 3R Laser Products.

It should be noted that the labels above are only examples of possible labeling for a particular BroadLighter assigned to this Class of laser products; consequently, the labels shown in Figure 1.7 can be different from those really affixed to your instrument. This difference lays in numeric data for "MAX power" and "Wavelength" shown on the right-hand label. To find out the difference, refer to the same label on your BroadLighter.

General specifications for all BroadLighters assigned to this Class can be summarized in the following way:

Product Type:	220VAC-Powered Broadband SLD Light Source
Model Number:	BroadLighter S-680-G-I-5 BroadLighter S-930-B-7 BroadLighter S-930-B-I-10 BroadLighter D-880-HP BroadLighter D-880-HP-I BroadLighter D-890-HP BroadLighter T-870-HP BroadLighter Q-870-HP BroadLighter Q-940-HP
Laser Class (according to IEC 60825-1 [Edition 2.0 2007-03]):	Class 3R
Maximum Rated Output Power:	See Table 1.4
Wavelength:	See Table 1.4
Output Aperture (for all models):	SM fiber with NA=0.14 (see Figure 1.6)

Light sources assigned to this Class emit optical radiation that can exceed the MPE (Maximum Permissible Exposure) under the direct intrabeam viewing, but in most cases the risk of injury is relatively small because the AEL (Accessible Emission Limit) for Class 3R is only 5 times greater than the AEL for Class 1, which is safe for the user. The risk of injury rises as the exposure duration increases. The exposure will be safe for the user unless it is a deliberate ocular exposure.

The risk of injury is actually minimized due to the following facts:

- ❑ Unintentional exposures would rarely reflect worst-case conditions, for example, of beam alignment with the pupil or worst-case accommodation.
- ❑ The inherent safety margin in the MPE.
- ❑ Natural aversion behavior for exposure to bright light for the case of visible radiation and by the response to heating of the cornea for far infrared radiation.

The direct intrabeam viewing is supposed to be unlikely for possible practical applications of this instrument. Otherwise, the user should reclassify the instrument so that Class 3B can be assigned to this equipment. After reclassification of the instrument, new labeling according to the requirements for Class 3B should be done as well. If you are unsure about providing the correct reclassification, contact Superlum on this issue.

Pay close attention to the following information: *Using controls, adjustments, or operating procedures that are different from those specified in this manual may result in radiation exposure hazardous to the user.*

When the instrument is used with certain optical elements—such as optical amplifiers, lenses, collimators, and the like—which are not supplied with the product as original accessories, or the product is integrated into a high-level system, it is strictly recommended to reclassify the final optical system in order to determine its total degree of optical radiation hazard. The reason for this is that the Class may increase under these conditions. It is the responsibility of a person (or organization) who performs such integration to provide correct reclassification for the optical system at the final stage.

It must be noted that it is the *user* who is responsible for being aware of possible risk associated with using Class 3R laser products as well as being adequately trained for working with such light sources. Superlum is not liable for any personal injury, property damage, or other (direct or indirect) damage of any nature caused by personal faults, improper and/or inadequate using of the instrument coming from disregarding the necessary aspects of laser safety.

Being initially designed as a Class 3B product, the instrument has inherited most of the laser safety features used for that Class. Although they are redundant for Class 3R laser products, they have been remained unchanged in this instrument. These features are as follows:

- ❑ Key-operated master control
- ❑ Remote interlock connector
- ❑ Illuminated indicator on the pushbutton for the output emission activation
- ❑ Three-second-time delay for the optical output activation
- ❑ Audible and visible warning signals for the output emission activation
- ❑ Protective metal cap for the optical output
- ❑ Laser safety labeling

This manual contains the most comprehensive information on the laser safety controls (and measures) built into the instrument. Please read thoroughly this manual to acquaint yourself with all features and functions for laser safety before you start working with the BroadLighter.

CAUTION Using controls, adjustments, or operating procedures that are different from those specified in this manual may result in radiation exposure hazardous to the user.

NOTE According to IEC 60825-1, this product has been classified for an exposure duration of 100 ms, which is supposed to be worst-case scenario. For other practical applications in which the exposure duration is greater than 100 ms, the product must be reclassified.

1.4.3 CLASS 3B Laser Products

If you begin to read this section, this means that your BroadLighter has been assigned to Class 3B laser products. Before you start working with your instrument, acquaint yourself with the important information on this Class by reading through the explanations below.

Each BroadLighter that is assigned to this Class of laser products carries on the instrument's cabinet a set of warning labels with the necessary information according to its laser hazard, like the ones shown in Figure 1.8.




FIGURE 1.8. Warning Labels for Class 3B Laser Products.

It should be noted that the labels above are only examples of possible labeling for a particular BroadLighter assigned to this Class of laser products; consequently, the labels shown in Figure 1.8 can be different from those really affixed to your instrument. This difference lays in numeric data for "MAX power" and "Wavelength" shown on the right-hand label. To find out the difference, refer to the same label on your BroadLighter.

General specifications for all BroadLighters assigned to this Class can be summarized in the following way:

Product Type:	220VAC-Powered Broadband SLD Light Source
Model Number:	BroadLighter S-790-G-I-15 BroadLighter S-830-G-I-20 BroadLighter S-840-B-I-20 BroadLighter S-850-G-I-20 BroadLighter S-960-G-I-20 BroadLighter D-810-HP BroadLighter D-840-HP BroadLighter D-840-HP-I BroadLighter T-860-HP
Laser Class (according to IEC 60825-1 [Edition 2.0 2007-03]):	Class 3B
Maximum Rated Output Power:	See Table 1.4
Wavelength:	See Table 1.4
Output Aperture (for all models):	SM fiber with NA=0.14 (see Figure 1.6)

Light sources assigned to this Class are always hazardous for the intrabeam ocular viewing even if an accidental short-time exposure occurs. Viewing them through optical elements with diffuse-reflecting properties is usually safe for the user.

Pay close attention to the following information: *Using controls, adjustments, or operating procedures that are different from those specified in this manual may result in radiation exposure hazardous to the user.*

When the instrument is used with certain optical elements—such as optical amplifiers, lenses, collimators, and the like—which are not supplied with the product as original accessories, or the product is integrated into a high-level system, it is strictly recommended to reclassify the final optical system in order to determine its total degree of optical radiation hazard. The reason for this is that the Class may change under these conditions. It is the responsibility of a person (or organization) who performs such integration to provide correct reclassification for the optical system at the final stage.

It must be noted that it is the *user* who is responsible for being aware of possible risk associated with using Class 3B laser products as well as being adequately trained for working with such lighter sources. Superlum is not liable for any personal injury,

property damage, or other (direct or indirect) damage of any nature caused by personal faults, improper and/or inadequate using of the instrument coming from disregarding the necessary aspects of laser safety.

Incorporated into this equipment are the following laser safety features:

- ❑ Key-operated master control
- ❑ Remote interlock connector
- ❑ Illuminated indicator on the pushbutton for the output emission activation
- ❑ Three-second time delay for the optical output activation
- ❑ Audible and visible warning signals for the output emission activation
- ❑ Protective metal cap for the optical output
- ❑ Laser safety labeling

This manual contains the most comprehensive information on the laser safety controls (and measures) built into the instrument. Please read thoroughly this manual to acquaint yourself with all features and functions for laser safety before you start working with the BroadLighter.

CAUTION Using controls, adjustments, or operating procedures that are different from those specified in this manual may result in radiation exposure hazardous to the user.

NOTE According to IEC 60825-1, this product has been classified for an exposure duration of 100 ms, which is supposed to be worst-case scenario. For other practical applications in which the exposure duration is greater than 100 ms, the product must be reclassified.

2 Installation

This chapter will guide you through simple procedures of installation you should first apply after receiving your BroadLighter. In addition, the chapter contains detailed instructions on the AC power line fuse replacement.

2.1 Unpacking and Initial Inspection

2.1.1 Unpacking the Instrument

To unpack the instrument, proceed as follows:

1. Place the shipping container on a flat rigid surface.
2. Cut the adhesive tapes over the top of the carton exterior with a sharp knife and open the top flap.
3. Remove the cushioning material around the instrument.
4. Lift the instrument and set it aside.
5. Lift the accessories and set them aside.

After unpacking the instrument, verify the receipt of the BroadLighter Broadband SLD Light Source. Refer to the list of contents in section "Specifications" if necessary.

2.1.2 Initial Inspection

To perform initial inspection, proceed as follows:

1. After unpacking the instrument, check the shipping container and its contents thoroughly to be certain nothing was damaged during shipment.
2. If the contents are damaged or defective, contact Superlum immediately. Keep the shipping materials for the carrier's inspection.

CAUTION After transiting or handling the container in cold weather, it must be stored at room temperature for about two hours before being opened. Avoid sharp changes in temperature that might cause water condensation on the internal parts of the instrument. Any amount of water condensation can be destructive for optical components inside the BroadLighter.

- NOTE**
1. The BroadLighter is a sophisticated optical instrument containing delicately aligned optical equipment; consequently, it should be handled with great care.
 2. Retain the original container for possible future use. If the instrument is to be sent to Superlum for servicing and the original container is not available, use an equal replacement for repacking. Refer to section "Service and Support" for the detailed instructions.
-

2.2 Preparation for Use

Your BroadLighter has been configured to operate with the AC power supply specified on the rear panel of the instrument. The instrument is supplied with a 3-conductor detachable power cord which grounds the instrument cabinet when it is connected to an appropriate receptacle. If you do not have the ground wire in your receptacle, you can ground the instrument via the rear-panel grounding terminal. The type of possible plugs acceptable for the AC power cord is specified in section "Specifications". Before connecting the instrument to your AC power source, check the rear-panel line fuse to be certain the correct fuse is installed in the instrument. Refer to Table 2.1 for specifications on the fuse and to section "Fuse Replacement" for detailed instructions on how to take out the fuse.

To perform preparation for use, proceed as follows:

1. Place the BroadLighter on a flat surface (e.g. on a table or in a rack) close to your measuring equipment in a location with sufficient ventilation. Make sure that no obstructions are provided for ventilation

openings of the instrument and the optical patchcable can be easily attached to the instrument without strain.

2. Check that the POWER pushbutton is set to OFF position. If it is not in OFF position, release the pushbutton by pressing it.
3. Connect AC power cord to the rear-panel power receptacle. Use the power cord supplied with instrument.
4. Insert the key coming with the instrument into the MASTER KEY lock and leave it in position "O".
5. Plug in the instrument.

CAUTION

1. **To prevent overheating, do not obstruct the ventilation openings of this equipment.**
2. **Before applying AC power to the BroadLighter, check the rear-panel label for the line voltage to be certain the instrument is preset to operate with the nominal line voltage in your area. If it does not meet your power source requirements, contact Superlum on this issue.**
3. **Do not operate this equipment with AC power source which is different from that specified on the rear-panel label of the instrument. Failure to obey this rule may cause permanent damage to the instrument.**

2.3 Fuse Replacement

To replace the blown line fuse, perform the following:

1. Check that the POWER pushbutton switch is set to OFF position.
2. Ensure that the plug of the AC power cord has been removed from the outlet.
3. Locate the fuse holder on the rear panel of the instrument.
4. Rotate the cap of the holder counterclockwise until the cap with the fuse protrudes from the rear panel.
5. Pull out the blown fuse and replace it.
6. Return the cap with the fuse to the rear panel.

To be perfectly safe, always use the fuses with the correct ratings as specified in Table 2.1.

TABLE 2.1 AC Line Fuse Specifications

Line Setting	Fuse Type
100-115 VAC	FS-52-G-T-2.0/250 (2A/250V, Time Delay, Glass Tube)
220-240 VAC	FS-52-G-T-1.0/250 (1A/250V, Time Delay, Glass Tube)

WARNING Any interruptions of the protective grounding conductor or disconnections of the protective earth terminal can make the instrument dangerous. It is strictly prohibited to perform an intentional interruption of the protective grounding conductor.

CAUTION Always use the specified fuses as listed in Table 2.1.

3 Local Mode of Operation

This chapter details the front- and rear-panel features of the BroadLighter and provides step-by-step instructions on how to operate the instrument in the local operating mode.

3.1 Front-Panel Features

The front and rear panels of the instrument are designed for ease of use by logically grouping features functionally related. To feel at ease with the instrument, acquaint yourself with the basic information about all controls, connectors, and indicators located on the panels of the instrument by reading through the explanations below. The explanations describe the features as they are numerically indexed in the corresponding figures.

Refer to Figure 3.1 for the illustration of the front-panel controls of the BroadLighter.




FIGURE 3.1. Front-Panel Features of the BroadLighter.

Here is a description of the features shown in Figure 3.1:

1—POWER. By pressing this pushbutton, the operator can turn the instrument on and off. When the pushbutton is in the depressed position, the LED indicator on the pushbutton lights blue, thereby indicating that the BroadLighter is enabled.

2—MASTER KEY (Laser Safety Measure). This is a key-operated master control used as a laser safety measure specified in IEC 60825-1 (Edition 2.0 2007-03). Turning the key in position "I" allows the operator to get access to SLD emission activation. With the key in position "O", the access is denied and the output emission of the SLD cannot be enabled.

3—REMOTE. This LED illuminates green when the instrument is controlled remotely by using a PC. When the instrument is set for the local operation, the LED goes blank.

4—LOCAL. When the BroadLighter is under remote control, pressing this pushbutton causes the instrument to return to the local operation.

5—SERVICE. If the instrument lights this LED, the output power degradation of the SLD has reached the maximum acceptable value and the instrument needs servicing.

6—HP MODE. By toggle this pushbutton, the user can switch between the Low Power operating mode and the High Power operating mode. When the High operating mode is selected, the LED on this pushbutton lights up green.

7—SLD ON/OFF. This is used to enable/disable the SLD emission. When the emission is activated, the pushbutton LED illuminates green.

8—PROTECTIVE CAP. This is used to protect the optical output of the BroadLighter from dust and dirty. Always recap the optical output, if the instrument is not being used and has no any patchcable attached.

9—OUTPUT APERTURE. This socket is an optical output of the instrument. The optical patchcable supplied with the instrument is connected here.

10—REMOTE INTERLOCK (Laser Safety Measure). This LED is used to inform the operator that the terminals of the remote interlock connector are open-circuited. When this happens, the LED illuminates red.

3.2 Rear-Panel Features

Refer to Figure 3.2 for the illustration of the rear-panel controls of the BroadLighter.




FIGURE 3.2. Rear-Panel Features of the BroadLighter.

Here is a description of the features shown in Figure 3.2:

1—RS-232 CONNECTOR. This is used to connect the BroadLighter to a PC for remote control.

2—REMOTE INTERLOCK (Laser Safety Measure). This is an audio monophonic jack which is used to connect the instrument to your local external controls for laser safety.

3—AC POWER INPUT RECEPTACLE. It accepts a 3-conductor power cord used to connect the instrument to your local AC power line.

4—AC LINE FUSE HOLDER. This is used to hold the line fuse.

5—GROUNDING TERMINAL. This is used to ground the instrument cabinet. This terminal accepts a male-banana-plug connector.

3.3 Operating Instructions

This section will guide you through step-by-step instructions on how to operate the instrument locally by using the front-panel controls.

To operate the BroadLighter with the front-panel controls, perform the following steps in the following order:

1. Perform preparation for use. Consult section "Preparation for Use" for more information if necessary.
2. Connect the BroadLighter to your local external controls for laser safety. Use the REMOTE INTERLOCK jack on the rear panel of the instrument for this connection. If you do not have the external controls for laser safety in your local area, use the short-circuited connector supplied with instrument.
3. Connect one end of the optical patchcable supplied with the BroadLighter to the front-panel OPTICAL APERTURE; the other, to your measuring equipment (e.g. an optical power meter). To obtain low insertion optical losses in the instrument-to-patchcable connection, always follow the instructions on the label affixed to the patchcable.
4. Press the POWER pushbutton to turn the instrument on. The pushbutton LED lights up blue. The TEC controller of the SLD is activated automatically at power-on. *Note that this controller remains always activate until the instrument is turned off.*
5. Turn the MASTER KEY control clockwise to position "I". Once the key has been turned, the instrument runs an internal test to verify whether the remote interlock connection is made or not. If the test passes, you are able to activate the optical output of the instrument (continue to step 6 to do this). If the test fails, the instrument starts to alarm continuously. In this case, the REMOTE INTERLOCK LED will steady red and the electro-magnetic sound generator built into the instrument will produce a continuous audible sound. Triggering the alarm warns of the fact that the remote interlock connection is open-circuited. *Note that after triggering the alarm, the instrument ignores any your attempt to activate the output emission until the alarm is reset.* To reset the alarm, turn the key to the OFF position. After that, check the remote interlock

- connection to be sure it is made correctly. Then turn the key to the ON position again.
6. Enable the optical output power in the Low Power mode by pressing the SLD ON/OFF pushbutton. When the pushbutton is pressed, a 3-sec activation delay runs. During the delay, the LED on the pushbutton begins blinking, and the built-in sound generator produces a beep signal. Once the delay has elapsed, the pushbutton LED switches to steady green light and the audible signal is automatically cut off; the output emission appears at the optical output of the instrument. *Please note the following: To ensure safe operation at power-on, each time after turning on the BroadLighter, it starts in the Low Power mode which is a default setting for the turn-on state. To set the instrument to the High Power mode, continue to step 8.*
 7. Disable the optical output of the BroadLighter by pressing the SLD ON/OFF pushbutton. The pushbutton LED will go blank.
 8. Press the HP MODE pushbutton to switch the instrument to the High Power mode of operation. The pushbutton LED will light up green, thereby indicating that the instrument has switched to the High Power operating mode.
 9. Enable the optical output of the BroadLighter by pressing the SLD ON/OFF pushbutton. After completing the 3-sec activation delay at start-up, the LED indicator on the SLD ON/OFF pushbutton will be steady green, indicating that the optical power has appeared at the optical output of the instrument.
 10. Verify the output power of the BroadLighter to be certain the current reading of your power meter corresponds to the value stated in the acceptance test report.

CAUTION To keep the end faces of the optical patchable clean and protected, it is strictly recommended to avoid unnecessary disconnections of the patchcable from the BroadLighter and your measuring equipment.

CAUTION NEVER connect optical connectors different from those specified for your BroadLighter, because they are not intended for the use with the instrument and can seriously damage its optical output.

NOTE

1. With no connection between your laser safety controls and the instrument, it will ignore any your attempt to activate the output power as described in step 5. If you do not have such controls in your area, install the short-circuited connector supplied with instrument.
 2. During a few minutes after turning the BroadLighter on, the optical output power of the instrument can slightly drift in time. For achieving greater stability of the output power, a 10-minute warm-up period is highly recommended.
 3. To meet personal safety requirements, the optical output of the BroadLighter remains inactive until the 3-sec activation delay completes.
-

3.4 How to Turn Off the Instrument

To turn off the instrument, perform the following steps:

1. If the optical output of the instrument is active, press the SLD ON/OFF pushbutton to disable the optical output. The pushbutton LED will go blank.
2. Turn the MASTER KEY control counterclockwise to position "O".
3. Press the POWER pushbutton. The LED on the pushbutton will go blank.

NOTE

Always disable the optical output before turning the instrument off.

3.5 Troubleshooting Information

This section deals with possible problems you can encounter when using the BroadLighter. The necessary information explaining possible instrument problems, diagnostic, and fixes is given below.

Problem 1: The POWER pushbutton is not lit when it is pressed, and the instrument does not work.

Possible Reason:

1. The AC power cord is not connected to the instrument or to your AC power source.
2. The line fuse is blown.

Fix:

1. Check that the cord is installed correctly between the BroadLighter and your AC power source, and tight enough in the sockets (in the rear-panel socket of the instrument and the outlet of your AC power source).
2. Disconnect the AC power to your BroadLighter and replace the fuse. Refer to section "Fuse Replacement" on page 38 for the necessary instructions on replacement.

Problem 2: After turning the MASTER KEY to position "I", the BroadLighter sets off the alarm and lights the REMOTE INTERLOCK LED.

Possible Reason:

1. The instrument has not been connected to your local external controls for laser safety, or the cable responsible for the necessary connection is defective.
2. Your local laser safety system has been activated.

Fix:

1. Turn the MASTER KEY to position "O" and check the remote interlock connection. Then try to start the BroadLighter again.
2. Turn the MASTER KEY to position "O" and reset your laser safety system. Then start the BroadLighter again.

Problem 3: The SERVICE LED indicator is always lit when the optical output of the instrument is activated; the output power is less than the specified value.

Possible Reason: The current value of the SLD driving current has reached the maximum acceptable value. This is most likely related to SLD performance degradation.

Fix: The BroadLighter needs servicing. Contact Superlum about the problem and send the instrument for repair. For more information on this issue, see chapter 6 "Service and Support".

Problem 4: After activation of the optical power, the LED on the SLD ON/OFF pushbutton frequently switches to red and the optical power is disabled.

Possible Reason: The circuit responsible for the emergency SLD switch-off disables the SLD. Check the quality of your AC power line for fast transients or spikes which are most often responsible for such a behavior of the instrument.

Fix:

1. Exclude any fast transients and spikes in the AC power supplied to the BroadLighter.
2. Press the SLD ON/OFF pushbutton once to reset the red illumination.
3. Press the SLD ON/OFF pushbutton one more time to re-run the SLD.

Problem 5: The current value of the optical power is lower than that specified in the Acceptance Test Report, but the SERVICE LED does not light.

Possible Reason: One of the ends of the optical patchcable (or maybe both of them) is dirty.

Fix:

1. Disconnect the patchcable and clean its connectors as described in chapter 5 "Ceaning Instructions". Then re-connect the patchcable and measure the output power again.
2. If the power still remains lower than the specified value, contact Superlum on this problem.

4 Remote Mode of Operation

Your BroadLighter Broadband SLD Light Source is capable of interfacing with a PC or terminal via the standard RS-232 interface. This chapter discusses information that explains how to program and control the BroadLighter from your computer using this interface. The chapter will show you how to write simple, one-line commands (and queries) and interpret instrument's responses. At the end of this chapter, you will find a description of the Superlum's BroadLighter companion software which is supplied with the instrument. To best serve these purposes, this chapter has been split up into four sections coming under the following headings:

- ❑ Introduction to Remote Operation
- ❑ Description of Data Communication Protocol
- ❑ Remote Command Set
- ❑ The Superlum's BroadLighter Companion Software

Section "Remote Command Set" includes five subsections that detail each command and query supported by this equipment.

4.1 Introduction to Remote Operation

The BroadLighter is equipped with the RS-232 connector located on the rear panel of the instrument. This allows the user to operate the instrument remotely by using a PC. After connecting to your PC, most of the front-panel functions of the BroadLighter accessible in local operation excepting those associated with the laser safety measures (turning on/off the MASTER KEY control, for example) can also be programmed and controlled via the RS-232 interface.

All commands supported by this equipment can be categorized as follows:

- ❑ Commands that cause the instrument to perform the desired action.
- ❑ Queries reporting either a current value of a queried operating parameter or a value corresponding to a current state of the instrument.

After addressing to the BroadLighter via the RS-232 interface, the instrument goes to remote operation automatically. The REMOTE LED on the front panel of the instrument will steady green. *Note that when the instrument is set to remote operation, you are not able to make any operational changes by using the front-panel controls since they are locked out in this mode.* To release the instrument from remote control, you should send an appropriate command from your PC or press the LOCAL pushbutton on the front panel of the instrument.

The BroadLighter contains a male 9-pin DB connector for serial communication with a PC. The RS-232 interface of the BroadLighter is configured as an RS-232 Data Terminal Equipment (see Figure 4.1). This means that data is sent from pin 3 and received on pin 2. Figure 4.2 illustrates the pinout for the RS-232 connector of the BroadLighter.





FIGURE 4.1. The Wiring Configuration for RS-232 Communication.



Pin No	Description
1	No connection
2	Receive Data RXD (Input)
3	Transmit Data (TXD) (Output)
4	No connection
5	Signal Ground (GND)
6	No connection
7	No connection
8	No connection
9	No connection

FIGURE 4.2. Pinout for the BroadLighter's RS-232 Connector.

Before operating the instrument with your PC, you need to configure the following parameters for RS-232:

- ❑ Baud rate (bits per second): 57600
- ❑ Data bits: 8
- ❑ Parity: none
- ❑ Stop bits: 1
- ❑ Flow control: none
- ❑ Data type: ASCII string

To verify functionality of the BroadLighter through the use of the RS-232 interface, you can either sent the necessary commands, which are described later in this chapter, by using the Hyper Terminal program commonly available on a PC or use the Superlum's BroadLighter companion software supplied with the instrument. In the latter case, you will need to install the software following step-by-step instructions given in this chapter. After that, you will be able to operate the instrument through the program window in the same way as you usually do from the front panel. For more information about installing and using the program, refer to section "The Superlum's BroadLighter Companion Software".

CAUTION

- 1. Use ONLY a null-modem cable to connect the BroadLighter to your PC.**
 - 2. Before connecting or disconnecting the RS-232 cable between the BroadLighter and your computer, ALWAYS disconnect all electrical power supplied to your computer and to the BroadLighter by unplugging the necessary power cords.**
-

4.2 Description of Data Communication Protocol

The instrument can interpret and execute programming messages on condition that they use the valid mnemonics placed in the correct order in a command or query. To be perfectly comfortable with programming the instrument, acquaint yourself with the following Data Communication Protocol which describes the structure of a command

(or query) for the BroadLighter. The structure of an instrument response is provided here as well.

Command/Query Structure

<command><data><CR><LF>

where,

<command> – command code (always begins with "S")

<data> – data (if required)

<CR><LF> – line termination sequence (<CR> – ASCII13, <LF> – ASCII10)

Response Structure

<command><data><CR><LF>

where,

<command> – response code (always begins with "A")

<data> – data

<CR><LF> – line termination sequence (<CR> – ASCII13, <LF> – ASCII10)

NOTE

1. Consider the triangle brackets as a punctuation mark which is used to mark off a command, a query, an integer, or a decimal digit.
2. In Windows OS, the line termination sequence is equivalent to pressing the Enter button.

4.3 Remote Command Set

4.3.1 Identification Query

Syntax: <S0>

Remarks: returned is an identification string containing numeric information about the instrument.

Response: <A0><data#1><data#2><data#3><data#4><CR><LF>

where,

<data#1> – 1-byte integer indicating the type of the instrument being used.

The value ranges from 0 to 9.

<data#2> – 1-byte integer indicating the number of active channels for the instrument being used. The value ranges from 1 to 4.

<data#3> – 1-byte integer indicating the number of the firmware version used in this instrument. The value ranges from 0 to 9.

<data#4> – 5-byte serial number of the instrument being used.

In case of error, the instrument sends the following response:

<A><E><CR><LF>

Example

Query: <S0><CR><LF>

Response: <A0><4><1><0><12345><CR><LF>

Here,

<A0> – the response code

<4> – the type of the instrument is 4

<1> – the BroadLighter being used is 1-channel device

<0> – the firmware version is 0

<12345> – the serial number of the instrument is 12345

<CR><LF> – the line termination sequence (<CR> – ASCII13, <LF> – ASCII10)

NOTE

For the BroadLighter, the number identifying the type of the instrument is 4.

4.3.2 Local/Remote Enable Command

Syntax: <S1><data><CR><LF>

Remarks: this command enables the user to select between two modes of operation: either local operation or remote operation.

In this command, the <data> is 1-byte integer indicating the mode that is going to be activated. The value of the integer can be either 1 or 2. See the table below for the description on each integer.

Data	Description
1	Sets the instrument to local operation.
2	Sets the instrument to remote operation.

NOTE

1. **By default, after switching the instrument on, it is always set to local operation.**
 2. **When the BroadLighter is operated from the front panel, it ignores any programming messages except the Remote Enable command.**
-

4.3.3 Local/Remote Check Query

Syntax: <S1><0><CR><LF>

Remarks: returned is an integer indicating the mode of operation.

Response: <A1><data><CR><LF>

Two possible values for the <data> are listed in the following table:

Data	Description
1	Means that the BroadLighter is under local control.
2	Means that the BroadLighter is under remote control.

In case of error, the instrument sends the following response:

<A><E><CR><LF>

Example

Command: <S1><2><CR><LF>

Remarks: after receiving this command, the instrument is set to remote operation.

Query: <S1><0><CR><LF>

Response: <A1><2><CR><LF>

Here,

<A1> – the response code

<2> – the integer indicating that the instrument is under remote control

<CR><LF> – the line termination sequence (<CR> – ASCII13, <LF> – ASCII10)

4.3.4 Optical Output Enable Command

Syntax: <S2><data><CR><LF>

Remarks: depending on the current status of the optical output, this command allows the user to enable or disable the optical output of the BroadLighter.

The value allowable for the <data> is listed in the following table:

Data	Description
1	Enables/disables the SLD module inside the BroadLighter. (Equivalent to pressing the SLD ON/OFF pushbutton on the front panel of the instrument.)

4.3.5 Optical Output Check Query

Syntax: <S2><0><CR><LF>

Remarks: returned is a numeric code indicating the current state of the BroadLighter.

Response: <A2><data#1><CR><LF>

where,

<data#1> – a decimal number from 000 to 255 describing the current state of the current-and-temperature controller of the instrument.

Refer to the following table for the description on each bit in the decimal numbers.

Signal	Bit	Description of the states of logical signals
TEC_GOOD	0	Set to "1" if the current temperature of an SLD is OK. Set to "0" if the current temperature of an SLD is not OK.
SLD_GOOD	1	Set to "1" if an SLD module is enabled. Set to "0" if an SLD module is disabled or faulty.
LIMIT	2	Set to "1" if the maximum operating current of an SLD module is reached.
SLD_ERROR	3	Set to "1" if an SLD module is faulty (i.e. not operational). Set to "0" if an SLD is operational.
MODE	4	Set to "1" if the instrument is used in the High Power mode of operation. Set to "0" if the instrument is used in the Low Power mode of operation.
MASTER KEY	5	Set to "1" if the MASTER KEY control is in the ON position. Set to "0" if the MASTER KEY control is in the OFF position.
INTERLOCK	6	Set to "1" if the INTERLOCK connection is open-circuited. Set to "0" if the INTERLOCK connection is closed.
SLD_WAIT	7	Set to "1" during the 3-sec activation delay.

In case of error, the instrument sends one of the following responses:

Response	Error Description
<A><E><CR><LF>	Common error
<A><L><CR><LF>	This means that the command being sent cannot be executed because the instrument is under local control.

4.3.6 High-Power-Mode Enable/Disable Command

Syntax: <S4><1><CR><LF>

Remarks: this allows the user to enable or disable the High Power mode of operation.

Please note that to enable or disable the High Power mode you must first disable the optical output of the instrument.

4.3.7 High-Power-Mode Check Query

Syntax: <S4><0><CR><LF>

Remarks: returned is a numeric code indicating the current status of the High Power mode.

Response: <A4><data><CR><LF>

where,

<data> – see item "MODE" in the table of section 4.3.5 for the necessary description of the <data> values.

4.4 The Superlum's BroadLighter Companion Software

4.4.1 About the Superlum's BroadLighter Companion Software

The interface of this program is a computer simulation of the front panel of the BroadLighter; consequently it includes the same features which you can find on the real front panel of the instrument. Moreover, to help you understand the program interface with ease, these features in the program window are grouped and arranged exactly the same as on the real front panel. Using these features of the program allows you to control the BroadLighter remotely from your PC system instead of standing in front of the instrument for pressing the necessary front-panel pushbutton.

By using this program, you can do the following:

- ❑ Activate and deactivated the SLD module.
- ❑ Monitor the current status of the front-panel controls.
- ❑ By simple clicking on the necessary button in the program window, automatically measure and save all current values of the instrument's performance parameters to a separate file on your PC which must be forwarded to Superlum when you suspect problems with the BroadLighter.

It is important to note that the last feature in the list above is accessible through the program only.

The program features very simple installation, non-severe system requirements, and a user-friendly interface. Use this program whenever you want to do the following:

- ❑ To begin to control the BroadLighter remotely as soon as you receive it and do not want to waste much time for developing the program of your own.
- ❑ *In case of a problem:* to test the instrument performance and report the results of the test to Superlum for getting the necessary assistance.

4.4.2 Minimum System Requirements

Minimum system requirements are as follows:

- ❑ **Operating system:** Windows 95, 95 OSR2, 98, or 98 SE, Windows ME, Windows NT Workstation 4.0 Service Pack 3, Windows 2000 Professional, Windows XP, Windows Vista, Windows 7,
- ❑ **Memory:** 32 MB for Windows 9x and ME, 64 MB for Windows NT and 128 MB for Windows 2000 and XP, 1 gigabyte (GB) (32-bit) or 2 GB (64-bit) for Windows Vista and Windows 7,
- ❑ **Processor:** 100-megahertz (MHz) or faster processor for Windows 9x and ME, 300-megahertz (MHz) or faster processor for Windows 2000 and XP, 1 gigahertz (GHz) or faster 32-bit (x86) or 64-bit (x64) processor for Windows Vista and Windows 7,
- ❑ **Hard disk space:** 2MB of free hard disk space.

4.4.3 Getting Started

This section will guide you through step-by-step instructions on how to get started with the Superlum's BroadLighter companion software.

To control your BroadLighter remotely by using the software supplied, you need to perform the following steps in the following order:

1. Install the software on your PC system.
2. Connect the BroadLighter to your PC system.
3. Run the software after its installation.

4.4.3.1 Installing the Software on Your PC System

To install the program on your computer, carry out the following steps:

1. Visit www.superlumdiodes.com/downloads_d.htm to download the software.

2. Download the **Companion_2_0.exe** file on your computer.
The **Companion_2_0.exe** file is a self-extracting archive that contains the following files:
 - **BCompanion.exe** – *BroadLighter Companion Software*
 - **License.txt** – *Software License Agreement*
3. Create a new folder on your hard drive.
4. Execute the **Companion_2_0.exe** file to expand the archive into the created folder.

4.4.3.2 Connecting the BroadLighter to Your PC System

To connect the BroadLighter to your PC system, proceed as follows:


1. Disconnect all electrical power to the BroadLighter and your computer. *The easiest way to do this is to unplug the power cords of the instrument and of the computer from the AC power source.*
2. Connect the BroadLighter to your computer by using a 9pin-to-9pin null modem cable. The necessary cable is supplied with the instrument. You can also use a 9pin-to-25pin cable (instead of the supplied one) if it best suits the RS-232 output of your computer.
3. Connect the BroadLighter and your computer to your AC power supply and turn on both of them.

CAUTION Always use a null-modem cable for connecting the BroadLighter to your PC.

4.4.3.3 Running the Program

To run the BroadLighter Companion Software on your PC, proceed as follows:

1. Find the **BCompanion.exe** file and execute the file.
2. When you launch the program for the first time, the following pop-up window will appear on the screen:



3. Select the name of the COM port in the drop-down list to which the BroadLighter has been connected and then click on the OK button. The program window like that shown in Figure 4.3 will appear on the screen of your computer, indicating that the BroadLighter is now under remote control. Note that in addition to the features and functions necessary to control your BroadLighter, the program also shows the current number of the firmware version (which is being used in your BroadLighter) and the serial number of the instrument.




FIGURE 4.3 Example of the Program Window.

The features shown in Figure 4.3 are used to do the following:

- **Settings**—This is used to display the pop-up window with the name of the COM port to which the BroadLighter is connected.
- **Self Test**—Clicking this button runs the self test which is used to check the BroadLighter's performance parameters when a malfunction of the instrument or any other problem with the BroadLighter occurs.
- **Close**—This is used to close the program window.

CAUTION Before installing or removing the RS-232 cable between the instrument and your PC, make sure that the BroadLighter and the PC are both completely unplugged. Failure to comply with this rule may cause permanent damage to this equipment.

NOTE Install the Superlum's BroadLighter companion software before connecting the BroadLighter to your PC.

4.4.3.4 Using the Program

Once the program has been run (see the previous section on how to do this), the window of the program as shown in Figure 4.3 will appear on the screen of your computer. As you can see from the figure, the features in the program window are grouped and arranged exactly the same as those on the front panel of the BroadLighter.

It is important to note that after running the program, the instrument is set to remote operation automatically. To make sure that the BroadLighter is under remote control, watch the REMOTE indicator either on the front panel of the instrument or in the program window—it must illuminate green, thereby indicating that remote operation has been activated. After setting the instrument to this mode, the front-panel controls—such as the SLD ON/OFF and the HP MODE pushbuttons—are locked out and remain inoperative until the LOCAL pushbutton on the front panel of the instrument or the same button in the program window is pressed. Pressing this pushbutton sets the instrument back to local operation.

Before you start to operate your BroadLighter by using the program, pay close attention to the following:

- ❑ It is not necessary to set the BroadLighter to local operating mode manually (or through the program) before exiting the program, because the instrument will switch to local operation automatically after closing the program window.
- ❑ By toggling the LOCAL button in the program window, you can switch between the remote operation and the local operation modes, whereas the front-panel LOCAL pushbutton allows switching to the local operation mode only.

- While the instrument is under local control, pressing the front-panel LOCAL pushbutton does not affect the instrument operation.


To be perfectly comfortable with the program, train yourself by following the steps below.

1. Perform preparation for use and then turn the BroadLighter on. (Refer to chapter "Installation", section "Operating Instructions", and section "Connecting the BroadLighter to Your PC System" if necessary.)
2. Set the MASTER KEY control on the front panel of the instrument to position "I".
3. Run the program on your computer. (It is supposed that the program has already been installed onto your PC. If not, refer to section "Installing the Software onto Your PC System" for the necessary instructions.)


Once the program has run, the REMOTE LED on the front panel of the instrument as well as the same indicator in the program window will light up green. The program window like the one below will be displayed on the screen of your computer.




4. Enable the SLD module in the Low Power mode by clicking on the SLD ON/OFF button in the program window. Below is how the screen will look after pressing the button.



5. Disable the SLD by clicking on the SLD ON/OFF button again. The green indicator on the button will go blank, thereby indicating that the SLD has been disabled.
6. Press the HP MODE button to activate the High Power mode. The screen will show the following:



7. Enable the SLD module in the High Power mode by clicking on the SLD ON/OFF button in the program window.




8. Disable the SLD by pressing the SLD ON/OFF button, and then press the HP MODE button to return to the Low Power mode if necessary.

NOTE

Switching of the BroadLighter to the desired mode of operation (to the High Power mode or the Low Power mode) is accessible only AFTER disabling the SLD. Any attempt to switch the operating mode of the instrument while its optical output is active will be ineffective.

By using this program, you can easily test the BroadLighter's performance parameters when you suspect a problem with the instrument operation. To do this, proceed as follows:

1. Run the program and then activate the Broadlighter in the High Power mode.
2. Click on the SELF TEST button in the program window. The following pop-up window will appear on the screen of your computer:



3. Type in your comments on the problem you have in the window below the "COMMENTS", if necessary.
4. Click on the SAVE button; then enter the file name and the location of the file in the corresponding window. The results of the test will be saved to the file on your computer.
5. Forward the file to Superlum to get the necessary assistance. Use the following link to email the file to Superlum: support@superlum.ie

4.4.3.5 Uninstalling the Program

To uninstall the Superlum BroadLighter companion software from your PC, just remove its files from the folder created during the installation.

5 Cleaning Instructions

This chapter will guide you through several cleaning procedures acceptable for the BroadLighter Broadband SLD Light Source. By following these procedures, you will be able to keep your BroadLighter in the best operating conditions, thereby ensuring its excellent performance.

5.1 Safety Precautions

This section details general safety precautions you need to observe during all phases of operating and cleaning the BroadLighter.

First of all, you should pay close attention to following common recommendations:

- ❑ Whenever possible, use physically contacting connectors and dry connections.
- ❑ Clean the connectors, interfaces, and sleeves regularly after use.
- ❑ If you are unsure of the correct cleaning procedure for your optical device, it is highly recommended that you first try cleaning a dummy or test device.

Before you begin cleaning the BroadLighter, please read the following safety rules:

1. Do not remove instrument covers when cleaning.
2. Make sure that you disable the optical output of the instrument before cleaning any optical interfaces.
3. Under no circumstances look into the end of an optical patchcable attached to the optical output of the BroadLighter when the instrument is operated. Superluminescent emission is not visible to the human eye, but it can seriously damage your eyesight.
4. When cleaning, always do the following:

- To prevent electrical shock, disconnect the instrument from AC power line before cleaning.
 - To clean external parts of the instrument cabinet, use a dry cloth or a cloth slightly dampened with water.
 - Do not attempt to clean internally.
5. Do not open the BroadLighter to perform any servicing to the instrument.
 6. Do not install parts or perform any unauthorized modification to the BroadLighter.
 7. In case of some problems with the instrument, contact Superlum for getting assistance. (Refer to the chapter 6 for more information on obtaining service.)

5.2 Cleaning Procedures

Three procedures explained in this section provide step-by-step instructions for cleaning optical patchcables, universal optical adapters, and the instrument cabinet. The procedures for cleaning optical elements may require at least two cleanings. The initial cleaning gently removes any grit and oil. If filmy deposits are still present at an optical interface, the second cleaning should be performed. It is not uncommon for a patchcable or connector to require more than one cleaning.

5.2.1 Cleaning a Non-Lensed Connector

To clean a non-lensed connector, proceed as follows:

1. Apply pure isopropyl alcohol to a new dry cotton swab or lens paper. Cotton swabs can be used as long as no cotton fibers remain on the fiber end after cleaning.
2. Clean the ferrule and other parts of the connector avoiding the end of the fiber.
3. Apply isopropyl alcohol to a new dry cotton swab or lens paper.
4. Clean the surface of the fiber end with the swab or lens paper using a small circular movement.
5. Immediately dry the fiber end with a clean dry cotton swab or lens paper.

6. Blow across the connector end face from a distance of about 20 cm using filtered, dry, compressed air. Aim the compressed air at a shallow angle to the fiber end face. Nitrogen gas or compressed dust remover can also be used.
7. As soon as the connector is dry, connect it to your equipment or cover it by the protective cap to keep it clean for further use.

After completing the initial cleaning, check performance of the optical connector. If the performance seems poor, clean the connector once again.

WARNING Never look into the end of an optical patchcable when the BroadLighter is operational.

- CAUTION**
1. To clean optical fiber ends, use natural cotton swabs.
 2. Do not shake or invert cans with the compressed air, because such an action might release particles contained in the can into the air. Always use the compressed air that is free from dust, water, and oil.
 3. Be especially careful when inserting the ferrule of a patchcable into a connector or adapter. Make sure that it does not touch the outside of the mating connector or adapter.
 4. Always disable the optical output of the instrument before cleaning.
-

- NOTE**
1. If possible, avoid using denatured alcohol because it usually contains additives. Instead, apply alcohol that is used for medical purposes.
 2. Always recap the optical output of the BroadLighter when it is not in use.
-

5.2.2 Cleaning the Front-Panel Optical Adapter

To clean the front-panel optical adapter, proceed as follows:

1. Apply isopropyl alcohol to a new dry cotton swab. Cotton swabs can be used as long as no cotton fibers remain after cleaning.
2. Clean the outside of the adapter with the swab.
3. Dry the outside of the adapter with a new clean dry cotton swab.
4. Finish the procedure by blowing the adapter with filtered, dry, compressed air.

CAUTION Never use a cotton swab to dry the adapter inside, because it may damage internal parts of the optical adapter.

NOTE Apply alcohol as little as needed for cleaning. This allows you to avoid any amount of alcohol inside the optical adapter. If, after all, the adapter contains some alcohol after cleaning, use compressed air for drying it.

5.2.3 Cleaning the Instrument Cabinet

To clean the instrument cabinet, do the following:

1. Use a dry and very soft cotton tissue for cleaning.
2. Do not apply any alcohol or solvents to the cabinet when cleaning.
3. Do not try to clean internally.
4. Do not open the instrument, as there is a danger of electric shock or electrostatic discharge.

In addition, please keep in mind the following information:

1. Opening the instrument can cause damage to ESD-sensitive components.
2. *For the instrument that is still under the warranty:* In case of opening, the warranty will be voided.

6 Service and Support

This chapter contains detailed information on what you should do when you need to have your BroadLighter serviced.

6.1 Returning the Instrument for Servicing

The instructions in this section will explain to you how to properly return the instrument for repair or calibration. Always call or email to Superlum to initiate service before returning your instrument. This ensures that the time for the repair (or calibration) is reserved and that your instrument will be returned to you as quickly as possible.

If the instrument is still under warranty, it will be repaired under the terms of the warranty (see section "Warranty" at the beginning of the manual). If the BroadLighter is no longer under warranty, Superlum will notify you of the cost of the repair after examining the device.

When the instrument is returned to Superlum for servicing, it must be adequately packaged and have a complete description of the failure symptoms attached. When describing the failure, please be as specific as possible about the nature of the problem. Include copies of additional failure information—such as the instrument failure settings, data related to instrument failure, etc—along with the instrument being returned.

6.2 Preparing the Instrument for Shipment

To properly prepare the instrument for shipment, observe the following instructions:

1. Write a complete description of the failure and attach it to the instrument.

Include any specific performance details related to the problem.

The following information should be returned with the instrument:

- Type of service required: repair or calibration
- Date of the failure
- Description of the problem:

- Whether problem is constant or intermittent
 - Whether instrument is temperature-sensitive
 - Whether instrument is vibration-sensitive
 - Instrument settings required to reproduce the problem
 - Performance data
- Company name and return address
 - Name and phone number of a technical contact person
 - Model number and serial number of the returned instrument
 - List of any accessories returned with the instrument
2. Cover all front or rear-panel connectors that were originally covered when you first received the instrument.
 3. Pack the instrument in the original shipping container, or use an adequate replacement following the directions below:
 - Wrap the instrument in antistatic plastic to reduce the possibility of damage caused by electrostatic discharge.
 - Use a double-walled, corrugated cardboard carton to pack the instrument.
 - The carton must be large enough to allow approximately 7 cm on all sides of the instrument for packing material, and strong enough to accommodate the weight of the instrument.
 - Surround the equipment with packing material to protect the instrument and prevent it from moving in the carton. The best solution is to use Air Cap™. Air Cap™ looks like a plastic sheet filled with air bubbles. Use the pink (antistatic) Air Cap™ to reduce static electricity. Wrapping the instrument several times in this material will protect the instrument and prevent it from moving in the carton.
 4. Seal the carton with strong nylon adhesive tape.
 5. Mark the carton "FRAGILE, HANDLE WITH CARE".
 6. Retain copies of all shipping papers.

NOTE

- 1. Cover optical connectors to protect them from damage caused by physical contact or dust.**
 - 2. Never use styrene pellets as packaging material. They do not adequately cushion the instrument or prevent it from shifting in the carton. They may also cause instrument damage by generating static electricity.**
-

Revision Documentation

Date	Notes
15/03/10	Version 1.0: First edition
05/12/10	Version 1.1: Corrections include the information on the laser safety labels (Chapter 1) and some typo errors (relating the number of SLD channels being used) in Chapter 4.
28/09/12	Version 1.2: New models are added to the document.