


Two-step purification with ÄKTA™ pure, or ÄKTA avant, using loop collection

Cue Card

Contents

Principles	2
Setup for multi-step purification	2
UNICORN methods	3
How to add user defined phases to a method	3
Description of the user defined phase in method one	4
Edit <i>Sample Application</i> phase in method two	5
Recommendations	6



Introduction

This cue card describes how to configure ÄKTA pure, set up methods, and perform a fully automated two-step purification using two one-step UNICORN™ methods in a method queue. ÄKTA avant can be configured applying the same principle. The fully automated two-step purification is suitable for most columns and elution combinations.

Example methods for two-step purifications can be downloaded from www.gelifesciences.com/AKTA, select **ÄKTA pure** or **ÄKTA avant** link under the heading **Featured Products**. Then click on **Related Documents** tab and find the heading **Cue Card**.

The purpose of the cue card is to help users get started, and to inspire further multi-step method development.




Principles

Two-step purification using a method queue with two one-step methods

By using one method for each purification step, column information from UNICORN can easily be used in the method. This means, for example, that pressure and flow rate limits are correct for each column and the column log book features can be utilized. Using a method queue allows full automation.

Method queue outline



Method one

The user defined phase in method one defines all functionality for peak detection and loop collection of the eluted peak. (see "Description of the user defined phase in method one")

Up to 5 loops can be connected to the Loop valve **V9-L**.

Method two

The collected peak from method one is loaded onto column two. (see "Edit Sample Application phase in method two")

Setup for multi-step purification

System configuration

Several different ÄKTA pure and ÄKTA avant configurations can be used. ÄKTA pure 25 is used in the following example. To enable multistep functionality, a Loop valve **V9-L**, a Column valve **V9-C** and a Mixer valve **V9-M** or a Versatile valve **V9-V** will be needed.

Sample loading method	Modules needed
System pump ¹	V9-L, V9-C, V9-M
Sample pump	V9-L, V9-V, V9-C, Sample pump P9-S

¹ Only applicable for ÄKTA pure as sample pump is always present in ÄKTA avant.

Flow path connections, when the Sample pump is used:


From	port	To	port
Pump P9-S	-	V9-V	3
V9-O or V9-Os	Out	V9-V	4
V9-V ¹	1	V9-Inj	SaP
V9-Inj	LoopF	V9-L	F
V9-L	E	V9-Inj	LoopE

¹ V9-V port 2 is not used.


Note: Connect a suitable loop to the Loop valve **V9-L** (see Recommendations section, Loop selection).

Flow path connections, when the System pump is used for sample loading. This is only applicable for ÄKTA pure as the sample pump is always present in ÄKTA avant.

No.	From	port	To	port
1	V9-O or V9-Os	Out	V9-M	Re-inj
2	V9-M	Out	V9-Inj	SaP
3	V9-Inj	LoopF	V9-L	F
4	V9-L	E	V9-Inj	LoopE



The illustration shows the two additional modules (blue) and the needed connections (green) installed on ÄKTA pure allowing two-step purification via re-injection (System pump configuration). Standard flow path shown with dash lines.



Important

Read ÄKTA pure Operating Instructions or ÄKTA avant Operating Instructions before using the instrument.

UNICORN methods

Method one

Objective

Execute the first purification step on column one. Use peak fractionation for the collection of the eluted peak in a loop.

Description

During elution, when the **watch** condition for peak start is fulfilled and the set delay volume has passed, valves turn into position and the content in the peak is directed to the loop for collection. The volume of the peak is counted during the backwards filling of the loop from port **F** to **E** on the loop valve.

Method two

Objective

Execute purification step two by forward loading (port **E** to **F** on the loop valve) the collected peak content from method one onto a second column, wash the column and elute the sample.

Description

Add functionality to the method so that the sample injection uses the correct loop and injects the volume of the peak collected during step one.

How to add user defined phases to a method

Create and edit phases

The UNICORN **Method Editor** software is used when creating and editing phases. Follow the steps below to create a user defined phase:

- Rename a global phase
- Add new text instructions
- A user defined phase can be saved in the **Phase Library** under **Global Phases** or **Personal Phases** for future use.

For easy identification, modified phases can be renamed starting with a **#** symbol.

The **:T** symbol is a software generated indication for a text edited phase.

Note: For a comprehensive guide to creating methods that can be run on ÄKTA pure or ÄKTA avant systems, refer to the UNICORN Method Manual.

Method one: Loop valve collect phase

- The flow path for peak collection will be washed prior to activating the watch functions for peak collection.
- During elution, the watch functions will be active and, if fulfilled, will direct the peak to the loop. Details are presented in the following pages.
- After peak collection the flow path used for peak collection will be washed.

Loop collection and injection

The volume of collected peak is counted in step one, using the **start/stop volume count** function. By inserting the instruction **hold counted volume**, the injection volume of step two can be equal to the collected peak volume. The instruction is added at the end of the **sample load** phase.


Due to the laminar flow, a larger volume than the one collected should be used during sample loading on the second column to empty the loop. Ideally, a volume that is 1.3 times larger than the collected peak volume should be used. Besides the laminar flow compensation, a compensation volume equal to the volume of the flow path from loop valve to column valve should also be added. These two volumes constitute the **chase*** volume.

The instruction **Empty loop with** is used to set preferred chase volume (laminar flow compensation volume + flow path delay volume).

Note: The desalting and size exclusion columns have limitations in the sample volumes that can be added.

* A chase volume is the volume of buffer that is used to push (i.e.

The illustration below shows an example of a UNICORN method that can be used for purification step one (**Loop valve collect phase**).



Description of the user defined phase in method one

User defined phase: Loop valve collect / Watch instructions

Note: The phase in this example has been designed for sample loading using the System pump and for collecting one peak. Multiple peaks can be collected using the same principle if additional watch commands are added. When using the Sample pump with ÄKTA pure the methods needs to be adjusted accordingly. The sample pump is always present in ÄKTA avant which is reflected in the example method for ÄKTA avant

0.00 Phase: Loop valve collect SystemP sample load

- 0.00 Base: Volume, Any
- 0.00 Block: Loop collect flowpath wash
- 0.00 Watch: UV 1, Greater than, 400.0 {mAU}, Peak start
 - 0.00 Base: Volume, Any
 - 0.20 System flow: 0.000 {ml/min}, Pre column pressure
 - 0.20 Injection valve: Sample pump load
 - 0.20 Outlet valve: Outlet 3
 - 0.20 Loop valve: 1
 - 0.20 System flow: (4.000)#Flow rate_1 {ml/min}, (Pre column pressure)#Pressure control_1
 - 0.20 Start volume count
 - 0.20 Watch: UV 1, Less than, 400.0 {mAU}, Peak end
 - 0.00 Base: Volume, Any
 - 0.54 Step volume count
- 0.54 Block: Loop collect flowpath wash
- 0.00 End_Block

Callouts:

- Volume from UV monitor to Outlet valve. In this example 0.20 ml.
- Note:** The volume of each component is found in ÄKTA pure User Manual or ÄKTA avant User Manual. The volume of tubing can be calculated using the formula:

$$\text{Volume (ml)} = \text{Length (mm)} \times (\text{i.d. (mm)})^2 \times \pi / 4.$$
- Position on Outlet valve where the tubing from Mixer valve V9-M port Re-Inj is connected.
- Note:** If a Sample pump is used to apply the sample, the instruction where the flow is directed to V9-V (Outlet valve: Outlet 3) needs to be followed by an instruction that sets V9-V to position 1-4 & 2-3.
- Position of the loop where the peak will be collected.
- Variable for the flow rate used as set in Method Settings.
- Volume from UV monitor to Loop valve. In this example 0.54 ml.

Block: Loop collect flowpath wash

0.00 Block: Loop collect flowpath wash

- 0.00 Base: Volume, Any
- 0.00 System flow: 0.000 {ml/min}, Off
- 0.00 Inlet A: (A1)#Inlet A_1
- 0.00 Inlet B: (B1)#Inlet B_1
- 0.00 Injection valve: Sample pump load
- 0.00 Column position: By-pass, Down flow
- 0.00 Outlet valve: Outlet 3
- 0.00 Mixer valve: Mixer
- 0.00 Loop valve: By-pass
- 0.00 System flow: 20.000 {ml/min}, Pre column pressure
- 5.00 System flow: 0.000 {ml/min}, Pre column pressure
- 5.00 Injection valve: Manual load
- 5.00 Outlet valve: Waste
- 5.00 Column position: (1)#Column position_1, Down flow
- 5.00 System flow: (4.000)#Flow rate_1 {ml/min}, (Pre column pressure)#Pressure control_1
- 5.00 End_Block

Callouts:


- Set Base to Volume, Any.
- Note:** If the Base is connected to column volume and used flow rate for the wash is higher than the column's max flow rate, a warning will be issued.
- Position on Outlet valve where the tubing from Mixer valve V9-M port Re-inj is connected.
- Flow rate used during Loop collect flow path wash.
- Column position variable used as set in Method Settings.
- Variable for the flow rate used as set in Method Settings.
- Note:** The use of the same variable allows the changes in Method Settings to be reflected in the Loop collect flow path wash block. For a detailed description of variables, see UNICORN Method Manual.

Edit *Sample Application* phase in method two

Phase: *Sample Application*

The *Sample Application* phase can be edited to utilize the *Hold counted volume* UNICORN instruction.

Note: When performing text editing of the *Sample Application* phase, the *Phase Properties* pane will be replaced with a list of phase variables.



Block Inject

```
0.00 Phase: Sample Application
  0.00 Base: SameAsMain
    0.00 Block: Prepare sample loading_1
      0.00 Base: SameAsMain
      0.00 Set mark: ()#Sample_ID_1
      0.00 Loop wash: By-pass. (A1)#Loop wash inlet (sample application) by-pass_1. 5 {ml}
      0.00 End_Block
    0.00 Loop valve: (1)#Loop position_1
    0.00 Block: Inject_1
      0.00 Base: Volume. ColumnSameAsMain
      0.00 Inlet A: (A1)#Inlet A
      0.00 Inlet B: (B1)#Inlet B
      0.00 Gradient: (0.0)#Percent B (Sample Appl_1 {%B}. 0.00 {base})
      0.00 Injection valve: Inject
      0.00 System flow: (5.000)#Flow rate {ml/min}. (Pre column pressure)#Pressure control
      0.00 Hold counted volume: 5.00 {ml}
      (1.00)#Empty loop with_1 End_Block
    0.00 Injection valve: Manual load
    0.00 Loop valve: By-pass
    0.00 End_Block
```

Additional instruction to *Sample Application* phase if injection volume should equal collected loop volume. It is possible to set the maximum volume to be used. In this example 5.00 ml.

Chase volume (laminar flow compensation volume + flow path delay volume). In this example 1.00 ml.

Recommendations

Peak detection

Set **watch** limits so that end peak is not triggered by start peak values. For example, set **start peak** to greater than 100 mAU and **end peak** to less than 100 mAU.

Another way is to use the instruction **Peak_start_max** before the peak end instruction (see *UNICORN Method Manual*).

The peak volume should be larger than the delay volume between the UV monitor and the Outlet valve. The delay volume is typically 0.2 to 0.5 ml.

Column selection

Take proper care in selecting the column for the second step, especially with respect to maximum load volume.

Recommended combinations are:

Purification step one	Purification step two
1 ml HiTrap™	2 × 5 ml HiTrap Desalting columns in series
1 ml HiTrap	Gel filtration/Size exclusion column, i.d. 16 mm
5 ml HiTrap	HiPrep™ Desalting
5 ml HiTrap	Gel filtration/Size exclusion column, i.d. 26 mm

For local office contact information, visit www.gelifesciences.com/contact

GE Healthcare Bio-Sciences AB
Björkgatan 30
SE-751 84 Uppsala
Sweden
www.gelifesciences.com/AKTA

GE and GE monogram are trademarks of General Electric Company.

ÄKTA, HiPrep, HiTrap, and UNICORN are trademarks of General Electric Company or one of its subsidiaries.

All other third party trademarks are the property of their respective owner.

Any use of UNICORN is subject to GE Healthcare Standard Software End-User License Agreement for Life Sciences Software Products. A copy of this Standard Software End-User License Agreement is available on request.



Loop selection

Use at least 1.3 times the expected peak volume to collect. Recommended to use Loop holder 29-0113-50.

Column CIP and Equilibration

To minimize the time the material in the eluted peak from step one spends in the loop before being re-injected during step two, you can create dedicated methods for cleaning and equilibration procedures.

- CIP and re-equilibration of column one can be methods executed after the method running step two of the purification.
- Equilibration of column two can be a method executed prior to starting method one.

Download

Example methods for two-step purifications for ÄKTA pure (either equipped with Sample pump or not) and ÄKTA avant, can be downloaded from www.gelifesciences.com/AKTA, select **ÄKTA pure** or **ÄKTA avant** link under the heading **Featured Products**. Then click on **Related Documents** tab and find the heading **Cue Card**.

Ordering information

For ordering information on columns, valves and tubing visit www.gelifesciences.com/AKTApure, or www.gelifesciences.com/AKTAavant.

© 2014-2015 General Electric Company – All rights reserved.
First published Mar. 2014

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.

GE Healthcare Europe GmbH
Munzinger Strasse 5, D-79111 Freiburg, Germany

GE Healthcare UK Limited
Amersham Place, Little Chalfont, Buckinghamshire, HP7 9NA, UK

GE Healthcare Bio-Sciences Corp.
800 Centennial Avenue, P.O. Box 1327, Piscataway, NJ 08855-1327, USA

GE Healthcare Japan Corporation
Sanken Bldg. 3-25-1, Hyakunincho Shinjuku-ku, Tokyo 169-0073, Japan