

# SystemSURE Plus ATP Cleaning Verification System



## System Implementation Guide



This guide is designed to assist users in setting up and operating an ATP cleaning verification program using the SystemSURE Plus ATP Cleaning Verification System within healthcare facilities. For instructions on how to operate the SystemSURE Plus, please refer to the operator's manual included with the system or our instructional materials online.

Watch Instructional Demos at:

[www.youtube.com/HygienaTV](http://www.youtube.com/HygienaTV)



Find downloadable resources and more information at:

[www.hygiena.com/healthcare.html](http://www.hygiena.com/healthcare.html)

## Contents

<b>Introduction</b>	2
<b>Components</b>	3
<b>Section 1: An Overview of ATP Cleaning Verification</b>	
1.1 What is ATP?	4
1.2 Measuring ATP with Bioluminescence Technology	4
1.3 Additional Uses of SystemSURE Plus	6
<b>Section 2: Implementing an ATP Cleaning Verification System</b>	
2.1 Establishing Test Locations and Limits	7
2.2 Broad Risk Categories and Limits	11
2.3 Corrective Action Procedures	12
2.4 Suggested Cleaning, Testing, and Corrective Action Procedure Flowchart	13
2.5 Programming Locations	14
2.6 Setting Up Test Plans	15
2.7 Testing Frequency	16
2.8 Daily Monitoring	19
2.9 Using SureTrend Software to Maximize a Monitoring Program	21
2.10 Additional Testing	21
2.11 Calibration	22
<b>Appendices</b>	
Appendix A: Recommended Test Locations	23
Appendix B: Proper Sampling Procedure	24

## Introduction to the SystemSURE Plus ATP Cleaning Verification System

Hygiena's SystemSURE Plus ATP Cleaning Verification System is a tool used to:

- Educate environmental services professionals and other personnel on proper cleaning technique
- Monitor and improve the cleanliness levels of surfaces in healthcare facilities
- Monitor the effects of changes within a cleaning program
- Document and track individual and overall cleaning performance

Healthcare facilities that implement an objective monitoring system experience a 42% increase in cleaning thoroughness.<sup>i</sup> The Centers for Disease Control and Prevention (CDC) encourages all hospitals to develop preventative programs to optimize and monitor the thoroughness of high-touch surface cleaning.<sup>ii</sup>

The SystemSURE Plus ATP Cleaning Verification System enables healthcare organizations to:

- Instantly assess the cleanliness of surfaces, allowing immediate corrective action to be taken
- Reduce or eliminate variation in surface cleaning performance by standardizing acceptable cleaning levels
- Improve and enhance the training of environmental services personnel
- Provide insight into whether current cleaning processes and tools are sufficient or below adequate
- Reduce the use of conventional microbiological testing methods that are slow, labor intensive, and costly
- Record and track test results to identify problem areas, make improvements, and show due diligence to auditors and compliance with regulations
- Enhance environmental cleaning programs which help to prevent the spread of harmful bacteria and viruses that are associated with healthcare associated infections
- Ensure patient safety and increase patient satisfaction

**Using Hygiena's SystemSURE Plus Cleaning Verification System, healthcare facilities are able to create a standard by which to measure cleaning effectiveness.**


---

<sup>i</sup> Carling, P.C., & Bartley, J.M. (2010). Evaluating hygienic cleaning in health care settings: What you do not know can harm your patients. American Journal of Infection Control; 38 : S41

<sup>ii</sup> <http://www.cdc.gov/HAI/toolkits/Evaluating-Environmental-Cleaning.html>

## Components of the SystemSURE Plus ATP Cleaning Verification System

The SystemSURE Plus ATP Cleaning Verification System consists of four parts:



1. **SystemSURE Plus Luminometer** – a user-friendly, handheld, light-reading unit that provides precise, on-site test results. Used with the UltraSnap testing device, extremely low levels of contamination can be detected in just 15 seconds. (Catalog # SS3H)
2. **UltraSnap Testing Device** – a convenient, all-in-one ATP test device. Simply swab, snap, and squeeze, and the test is ready to be measured in the SystemSURE Plus. Packaged 100 tests per box. (Catalog # US2020)
3. **SureTrend Data Analysis Software** – a powerful software program that allows users to upload test results to a database, analyze trends and generate reports for management and record-keeping. (Included with SystemSURE Plus Luminometer).
4. **Calibration Devices** - optional calibration rods confirm SystemSURE Plus is within specifications. (Catalog # PCD4000) *See section 2.11 Calibration for more details.*

**Hygiena's luminometer, testing devices, and software are designed to be easy to use, enabling operation by both technical and non-technical staff.**



## Section I: System Overview


The SystemSURE Plus ATP Cleaning Verification System is a rapid cleaning monitoring system used to help hospitals and other healthcare organizations achieve optimal standardized cleaning levels. The system uses bioluminescence technology to identify and measure **adenosine triphosphate**, commonly known as **ATP**.

### 1.1 What is ATP?


ATP is an energy molecule found in all living cells that allows cellular metabolism to take place. All organic matter contains ATP, including blood, saliva, and bacteria. In healthcare facilities, organic matter such as bodily fluids, blood, and bacteria left on surfaces can become a point of cross-contamination between patients and staff, leading to infections if not properly cleaned. Therefore the detection of ATP on a surface after cleaning is an indication of improper cleaning.

### 1.2 Measuring ATP with Bioluminescence Technology

UltraSnap ATP surface tests contain an enzyme called luciferase which produces a bioluminescence (light-producing) reaction when it comes into contact with ATP. The light emitted from the reaction is measured and quantified in the SystemSURE Plus luminometer. The graphic below illustrates how ATP on a surface reacts with the enzyme in UltraSnap ATP test devices to emit light.



## ATP Presence and RLU Measurement



### Higher Contamination = Higher RLU

The quantity of light generated by the bioluminescence reaction is directly proportional to the amount of ATP present in the sample. The reaction is immediate, allowing results to be processed in real-time. Results are then expressed numerically on the SystemSURE Plus screen as Relative Light Units (RLU).



### 1.3 Additional Uses of SystemSURE Plus

In addition to routine ATP cleaning verification by Environmental Services (EVS), the SystemSURE Plus ATP Cleaning Verification System can be used for:



**Central/Sterile Services** – Used for verifying the cleanliness of flexible endoscopes and other reusable medical devices. Improper cleaning of equipment before sterilization can lead to non-sterile equipment.



**Hand Hygiene Compliance** – Used to measure levels of ATP present on personnel's hands before and after hand washing to demonstrate efficacy and thoroughness of hand washing efforts.



**Food Service/Cafeteria Food Safety** – Used to verify food preparation surfaces and dining areas have been cleaned properly.

*For more information on implementing the SystemSURE Plus in these and other areas of a hospital, visit [www.hygienea.com/healthcare.html](http://www.hygienea.com/healthcare.html)*



## Section 2: Implementing an ATP Cleaning Verification System

### 2.1 Establishing Test Locations and Limits

The SystemSURE Plus comes with a preset pass limit of 25 RLU and fail limit of 50 RLU. These limits are based on studies conducted in healthcare facilities, and a starting point from which custom limits can be refined.

Pass	25 RLU
Caution	26-50 RLU
Fail	50 RLU

This section will guide users through identifying test locations and establishing appropriate pass/fail limits for those locations. To optimize an ATP cleaning verification program, many hospitals choose to set custom limits for test locations. Before testing begins, it is necessary to:

1. Identify areas within the facility that will be tested
2. Establish appropriate pass and fail limits for each location

The CDC provides a list of recommended locations to test in hospitals (*see Appendix A on page 23*). Locations typically tested are high-touch surfaces where the chance of spreading infectious bacteria is high. Monitoring of low-risk surfaces on a less frequent basis is also essential to verifying a facility is being thoroughly cleaned. Prepare the list of locations on a piece of paper or Microsoft Excel worksheet.



## Collecting Test Samples

Once locations to be tested have been identified, pass and fail limits for each location can be established by collecting samples or assigning the locations to broad risk categories (see *Section 2.2 on page 11*). **Before collecting initial samples, it is essential to master proper sampling procedure as detailed in Appendix B, page 24.**

Managers may choose to collect test samples in one of two ways:

### 1. After Routine Cleaning

*Time to implement: 1-4 weeks*

To use this method, sample each location a minimum of 10 times after routine cleaning. If locations are the same from room to room, it is acceptable to test the same location from different rooms.

### 2. Optimal cleaning

*Time to implement: 1-3 days*

To use this method, the manager overseeing the ATP cleaning verification program should clean each location that will be tested the way he or she wants the location to be cleaned each time. A minimum of 10 samples should be taken from the same locations in different rooms to ensure that no site is tested more than once. For example, the call buttons in ten different rooms could be sampled after the cleaning of these sites.

Record RLU measurements on a spreadsheet (Example 1). (For an Excel® spreadsheet that automatically calculates pass/fail limits, visit the resources section at [www.hygiene.com/healthcare.html](http://www.hygiene.com/healthcare.html))

Example 1: Collected Test Samples

Tests	1	2	3	4	5	6	7	8	9	10	Average
Location	RLU measurement										
Bedrail	48	49	50	38	38	35	73	118	36	15	
IV Pole	10	25	67	35	8	5	50	47	32	24	
Bedside Table	45	27	0	2	51	5	0	2	10	6	
Remote	112	215	78	45	89	86	95	148	62	71	

## Calculating Pass/Fail Limits with the Standard Deviation Method

Result	Calculation	Instructions
<b>Pass</b>	Average	To determine the lower (pass) limit of each location, calculate the average RLU score from the samples collected.
<b>Fail</b>	Average + (3 * Standard Deviation)	To determine the upper (fail) limit, calculate the standard deviation of the samples collected and multiply by three (3) <sup>iii</sup> . Add that value to the average.  Statistically, this calculation indicates that nearly all (99.7%) passing test results (clean surfaces) will be below the fail RLU limit. An RLU result higher than the fail limit would be a statistical outlier, indicating the location is not clean.

### Example 1 Continued: Calculated Limits

Location	Average = Pass	Avg + (3 * Standard Deviation) = Fail
Bedrail	50	134
IV Pole	30	91
Bedside Table	15	73
Remote	100	248

<sup>i</sup>. For calculations using Microsoft Excel:

- For lower limit (pass), use the function “=AVERAGE”
- For upper limit (fail) use the function “=3\*STDEV and add to “=AVERAGE”

The range between the pass and fail values is the caution range. This could give different pass, caution, and fail levels for each location. This is typical when different surface types (plastic, stainless steel, etc.) are being tested and when the age of equipment varies. Thus, for Example 1 data, the limits for the locations using the standard deviation method are as follows:

Example 1: Final Limits

Location	Pass	Caution	Fail
Bedrail	50	51-133	134
IV Pole	30	31-90	91
Bedside Table	15	16-72	73
Remote	100	101-247	248

### Eliminating the Caution Zone

Many hospitals chose to eliminate the caution zone and consider anything greater than the pass score as a fail. Refer to the table below for an example.

Example 1: Limits without Caution Range

Location	Pass	Fail
Bedrail	50	>50
IV Pole	30	>30
Bedside Table	15	>15
Remote	100	>100

## 2.2 Broad Risk Categories and Limits

While establishing custom limits for your facility would be the gold standard for setting up a monitoring program, many hospitals rely on limits for broad-risk categories that have been validated by published, peer-reviewed, and third-party studies. For information on these studies, reference *Technical Document: Establishing RLU Pass/Fail Limits* available from the resource library on [www.hygienea.com](http://www.hygienea.com) or contact your Hygienea technical representative. The general recommended limits below reflect an elimination of the caution zone. For hospitals that wish to use the caution zone, simply double the pass RLU (pass RLU x 2) to determine the fail RLU number.




Application	General Recommended Limits	
	Pass (RLU)	Fail (RLU)
<b>Public areas</b>  Examples:  Elevator call buttons  Hallway handrails  Waiting room areas	50	>50
<b>Patient rooms</b>  Examples:  Call button  Bed rails  Patient restroom  Monitor panels  IV pole	25	>25
<b>Sterile services</b>	10	>10
<b>Washer disinfectant</b>	5	>5
<b>Food preparation and catering</b>	10	>10
<b>Hand washing</b>	60	>60




## 2.3 Corrective Action Procedures

Corrective action procedures provide clear instructions for what steps should be taken following pass, caution, or fail results.

Recommended corrective action procedures are as follows:

SYMBOL	TEST RESULT	CORRECTIVE ACTION
	Pass	The surface has been adequately cleaned. No action required.
	Caution	The surface may not have been adequately cleaned. The area may be recleaned and retested, or monitored for future problems. Cleaning personnel may require retraining on proper cleaning procedures.
	Fail	The surface has not been cleaned to the cleaning standard and must be recleaned and retested. Cleaning personnel should also be retrained on proper cleaning procedures.

## 2.4 Suggested Cleaning, Testing, and Corrective Action Procedure Flowchart



## 2.5 Programming Location Pass/Fail Limits into Software

Once test locations have been identified and Pass/Fail limits determined, they must be entered into the SureTrend software and synced with the SystemSURE Plus luminometer.

For instructions on how to install SureTrend software and add locations, see the installation guide and manual included with the SureTrend CD. See the SureTrend User Manual for steps on entering locations and limits.

*For assistance programming locations and limits in the SureTrend software, please contact your Hygiena representative.*



By default the SureTrend software assigns a pass limit of 10 and fail limit of 30 when you add a new location. Assign the correct limits to each location determined in Section 2.1 or 2.2.

Group information can also be entered at this time. Group information is used for grouping locations together for reporting purposes. For example, groups can allow users to compare the performance of departments, wings, or facilities.

Below is an example of the location setup in the SureTrend software.

Prog #	Location	Group	Pass	Fail
1	Bedrail – East	Community Hospital – East	50	100
2	Remote Control - East	Community Hospital – East	25	50
3	Sink – East	Community Hospital – East	15	30
4	Bedrail - West	Community Hospital – West	50	100
5	Remote Control - West	Community Hospital – West	25	50
6	Sink - West	Community Hospital – West	15	30

## 2.6 Setting up Test Plans

Once locations and limits have been input into SureTrend software, test plans may then be set up. Test plans are helpful groups of locations that are tested one after another, grouped together, or tested on a specific day. Test plans help to keep testing and data analysis organized. See the SureTrend User Manual for directions on setting up test plans in SureTrend software.

Here are some examples of test plans with locations:

Nurses' Station
Keyboard
Phone
Countertop
File cabinet handle
Light switch

ICU Patient Room
Ventilator control panel
IV Pole
Monitor cables
Call button
Door handle

Public Areas
Handrails
Door levers
Waiting area chair
Telephone
Water cooler

ER Mobile Workstations
Crash cart
ECG cart
Laceration cart
Bedside cart
Trauma cart
IV cart
IV medication cart
Cast cart

West Wing Patient Room
Bed tray table
Patient phone
Call button
Bed rail
Main light switch
Sink handles
Toilet flush handle
Bathroom light switch
Bathroom handrail
Television remote

Monday
Bedrails
IV Pole
Bedside Table
Remote
Panels


## 2.7 Testing Frequency

Once test plans are programmed into the SureTrend software, sync the SystemSURE Plus with the software and begin testing. The frequency of testing will be determined by:

- Budget
- Size of facility
- Importance of the cleaning standard
- Logistical and staff constraints
- Compliance with CDC environmental monitoring recommendations or other auditing bodies' recommendations
- Desired statistical accuracy of reports (the greater the amount of data generated, the more representative the data will be)

The CDC has a recommended formula for sample size determination. It is divided in two segments: **Baseline Monitoring** and **Ongoing Monitoring**. Section 2.8 outlines a Daily Monitoring routine that most hospitals will chose to fully leverage an ATP Cleaning Verification program.

### 1. Baseline Monitoring – Getting Started



This is the initial testing regiment that should be conducted to accurately assess the current level of cleanliness and compliance to cleaning processes. This initial testing will become the baseline to gauge improvements or deterioration of thoroughness of cleaning and will determine the number of points which must be monitored on a regular basis.

While it would be ideal to identify small fluctuations in practice accurately (e.g., 10% relative change), such an approach takes slightly more time and testing. Instead, a meaningful change in cleaning practice (e.g., 20% relative change) can be detected without having to evaluate a substantial number of surfaces. The CDC recommends sampling all available surfaces identified in Section 2.1 or Appendix A (page 23) in a 10% - 15% sample of representative patient rooms in hospitals with over 150 beds. In hospitals with less than 150 beds, all available surfaces should be tested in a minimum of 15 rooms for baseline and ongoing monitoring.


The CDC recommends baseline monitoring to be conducted at least three times per year. This is the minimum amount of testing that can be done to show a 20% relative change in cleaning levels. If pass levels decline, then more testing should be done in order to determine what has caused deterioration. (i.e., inefficient sanitizer, insufficient cleaning process, poor employee performance, etc.).

*Source: CDC Toolkit for Evaluating Environmental Cleaning, Appendix C, <http://www.cdc.gov/HAI/toolkits/Appendices-Evaluating-Environ-Cleaning.html>*




The following is an example of a baseline evaluation (to measure levels of cleanliness):

### Baseline Evaluation Calculation for a Typical 250-Bed Hospital




For Example:					
250 Patient Rooms	10 Test Locations in Each Room	100 Operating Room Locations	100 Sterile Services Locations	2,700 Total Locations	270 – 405 Locations to Monitor



For Example:		
270 – 405 Locations to Monitor	3 Evaluations per Year	810 – 1,215 Totals Tests per Year

For a 250 bed hospital with 10 locations in each patient room, 100 operating room locations, and 100 sterile services locations, there are 2,700 total locations to be tested. To monitor 10-15% of locations, a total of 270-405 locations will need to be tested three times per year (a total of 810-1,215 tests per year).

## 2. Ongoing Monitoring – Minimalist Maintenance



When hospitals have achieved more than 80% or higher pass results from baseline monitoring, the number of surfaces to be monitored can be decreased to those available in a 5% sample of rooms per evaluation cycle unless there is deterioration in practice. Though the CDC does outline minimal testing of 5% of potential test locations, Hygiena has found that hospitals that have a high cleaning thoroughness rate usually achieve such success through Daily Monitoring - see Section 2.8 on the following page.

Below is an example of ongoing monitoring (once 80% or more of baseline test results are "Passing"):

### Ongoing Monitoring for a Typical 250-Bed Hospital



For Example:					
250 Patient Rooms	10 Test Locations in Each Room	100 Operating Room Locations	100 Sterile Services Locations	2,700 Total Locations	135 Locations to Monitor



For Example:		
135 Locations to Monitor	3 Evaluations per Year	405 Totals Tests per Year

For a 250 bed hospital with 10 locations in each room, 100 operating room locations, and 100 sterile services locations, there are 2700 total locations to be tested. To monitor 5% of locations, a total of 135 locations will need to be tested three times per year (a total of 405 tests per year).

## 2.8 Daily Monitoring – Optimal Program

Minimal






Baseline

Optimal






More and more hospitals are moving to daily or weekly ATP cleaning verification. Daily monitoring only requires a few tests per day, yet holds cleaning staff accountable for achieving an optimal level of cleanliness each and every day. Creating a reward system based on daily results can be a very powerful program that brings staff together and improves overall hospital cleanliness.

Daily monitoring does take a greater time commitment and requires more consumables, so Hygiena will help every hospital to design a testing program within its resources and budget. Below are three scenarios breaking down the time and consumable use associated with a daily ATP monitoring program using 5, 10, or 15 tests per day.






### Scenario 1: Five Tests per Day

	 Daily	 Monthly	 Yearly
 # Tests	5 tests	150 tests	1,800 tests
 Time (1 min/test)	5 minutes	2.5 hours	30 hours

## Scenario 2: Ten Tests per Day

	 Daily	 Monthly	 Yearly
# Tests	 <b>10</b> tests	<b>300</b> tests	<b>3,600</b> tests
Time (1 min/test)	 <b>10</b> minutes	<b>5</b> hours	<b>60</b> hours


## Scenario 3: Fifteen Tests per Day

	 Daily	 Monthly	 Yearly
# Tests	 <b>15</b> tests	<b>450</b> tests	<b>5,400</b> tests
Time (1 min/test)	 <b>15</b> minutes	<b>7.5</b> hours	<b>90</b> hours

## 2.9 Use SureTrend Software to Maximize a Monitoring Program

One of the greatest byproducts of a daily monitoring program is a robust database of test results to aid in decision making and management of hospital cleanliness programs. By frequently collecting test results, a hospital can build up a database of records to identify trends, assess training opportunities, and compare cleaning performance between departments, shifts, or facilities. Using these reports in regular performance feedback meetings with frontline personnel can be crucial in maintaining environmental cleaning effectiveness. SureTrend comes with dozens of preset reports so little setup is required by the user and reports can be automatically

emailed to stakeholders. Visit [www.hygiene.com](http://www.hygiene.com) for a helpful guide for using SureTrend reports in hospitals: *Quick Start Software Report Guide for the Healthcare Industry*.



### 2.10 Additional Testing

ATP cleaning verification can also be integrated into a hospital's emergency cleaning procedures as a final step to confirm thorough cleanliness. In the event of patient accidents, bio-waste spills, flood, new construction, outbreak, or other accidental contaminations involving blood, urine, or fecal waste, cleaning staff should perform an emergency cleaning of that area. To confirm thorough cleaning of the site, ATP tests can be taken. This will ensure that the spill site has been properly cleaned.



## 2.11 Calibration


To verify instrument calibration, Hygiena offers two calibration kits that are recommended for periodic use with your system.



### ***Calibration Control Rod Kit (Catalog# PCD4000)***

Though Hygiena's ATP Cleaning Verification System automatically checks calibration at startup, it is recommended that calibration is verified with the Calibration Control Kit once a month for audit record-keeping purposes. Incorporating the Calibration Control Kit into a cleaning verification program will confirm that the instrument is within specifications and operating correctly.

Each kit contains a positive rod and negative rod. The positive rod emits a very low level of constant light output that can be measured in RLU's to verify proper calibration of the unit. The negative rod produces zero (0) RLU and is used to check that background light is not entering the instrument, while ensuring that the light sensor is calibrated correctly.



### ***Positive Control Kit (Catalog # CK25)***

The Positive Control Kit is used for validating the efficacy and quality of the UltraSnap ATP Testing Device. Each kit comes with 25 sealed glass vials, containing a certain amount (approx.  $5 \times 10^{-13}$  moles) of freeze-dried ATP and sugars to provide a predictable result if UltraSnap devices are used and stored correctly.

## Appendix A: Recommended Test Locations

### Patient Areas

- Airway cart
- Anesthesia cart
- Bathroom door knobs and levers\*
- Bathroom handrails\*
- Bathroom light switch\*
- Bed control\*
- Bed rails\*
- Bedpan cleaner\*
- Bedside chair\*
- Bedside tables\*
- BP machine cuff
- Cabinet door handles
- Call button\*
- Door knobs\*
- ECG cart
- Flush handle\*
- IV cart
- IV pole\*
- IV pump control\*
- Laceration cart
- Light switch\*
- Medication cart
- Monitor controls\*
- Monitor touch screen\*
- Monitor cables\*
- Operating table base
- Operating table hand control
- Operating table mattress
- OR anesthesia machine
- OR boom
- OR keyboard
- OR door push plates
- OR spotlight
- Privacy curtain
- Phone\*
- Sink top\*
- Sink handles\*
- Supply cart
- Toilet seat\*
- Trauma cart

- Tray table\*
- TV remote
- Ventilator control panel\*

### Public Areas

- Drinking fountain buttons
- Elevator call buttons
- Hallway hand rails
- Nurses' keyboard
- Nurses' phone
- Vending machine buttons
- Visitor bathroom light switch
- Visitor bathroom door knobs
- Waiting room chairs

### Sterile Services

- Autoclave interior
- Biohazard receptacle
- Endoscope
- Sink
- Sterile tools
- Sterilizer handle

### Food Preparation:

- Cutlery
- Cutting board
- Dishwasher
- Freezer handles
- Fridge handles
- Knives
- Service trays
- Serving utensils
- Sink
- Sink handles

### Hand Hygiene:

- Preclean
- Post-clean
- Scrub sink
- Sink handles
- Soap dispenser

\* Source: CDC Environmental Checklist for Monitoring Terminal Cleaning available at <http://www.cdc.gov/HAI/toolkits/Environmental-Cleaning-Checklist-10-6-2010.pdf>

## Appendix B: Proper Sampling Procedure


Watch an instructional demo at [www.youtube.com/HygienaTV](http://www.youtube.com/HygienaTV)



The SystemSURE Plus luminometer is designed to detect contamination that cannot be seen by the naked eye. Excessive sample may interfere with the bioluminescence reaction and produce an inaccurate test result, which is why it is important to remove all visible soil from a surface before swabbing. Before collecting a sample for testing, the surface should be visibly clean. If any soiling or residue is present, reclean the area before testing.




Turn on the SystemSURE Plus luminometer. The system will run through automatic calibration. Once calibration is complete, scroll through the program numbers (PROG) using the (A) and (V) arrows to find the programmed test location that correlates to the location being tested. This action should be taken prior to swabbing.




Remove the individual UltraSnap device from the package. Next, remove the outer tube by holding onto the double ring base of the Snap-Valve while pulling down on the tube. The swab tip comes pre-moistened. Condensation may be visible on the inside of the swab tube. This is normal. **Do not touch the swab tip or shaft with fingers or anything else, as this will contaminate the test.** Discard any swabs that accidentally get tainted or activated.

*NOTE: For optimal performance, swabs that have been removed from cold storage should stand for 10 minutes at room temperature before use.*




### Incorrect swabbing technique:

- Touching the swab shaft with your finger
- Lightly touching the swab to the sample area
- Collecting sample on only one side of the swab tip




### Correct swabbing technique:

- Sufficient pressure to create flex in the swab shaft (helps to break through any biofilm on the surface)
- Rotate the swab to collect sample on all sides of swab tip
- No contact with the swab shaft wand




**Regular surfaces:** Swab a 10 x 10 cm (4 x 4 in) square on the test surface, making a criss-cross pattern as shown.




**Irregular surfaces:** Where 10 x 10 cm square sampling is not feasible, such as a bed rail, swab as much of the surface as possible to collect a representative sample.


*Note: Consistent sampling pattern on irregular surfaces is necessary to ensure reliable and repeatable results. All individuals responsible for performing swab tests should agree on similar sampling pattern.*




Re-insert the swab into the tube.




Holding the device upright, activate UltraSnap by bending the bulb at the top until the plastic Snap-Valve breaks, then bend once more in the opposite direction. Squeeze the bulb twice to expel the liquid-stable reagent contained in the bulb and allow it to flow to the bottom of the tube.



Gently shake the device with a side-to-side motion for 5-10 seconds, bathing the swab bud in the liquid-stable reagent. The test is now activated and the bioluminescence reaction is taking place. For optimal results, the reading should be taken on the SystemSURE Plus luminometer within 30 seconds of activation.



Open the lid on the SystemSURE Plus luminometer, and insert the activated UltraSnap device into the reading chamber. Close the lid.



Taking care to hold the unit upright for an accurate reading, press "OK" on the SystemSURE Plus to initiate measurement. Results are displayed on the screen in 15 seconds.



## Notes

## Notes

[illegible]



For additional information or support, please contact us:



[www.hygiena.com](http://www.hygiena.com)

#### Americas

1.805.388.8007  
[info@hygiena.com](mailto:info@hygiena.com)

#### International

+44 (0) 1923 818821  
[enquiries@hygiena.com](mailto:enquiries@hygiena.com)



1295 Morningside Avenue, Unit 16-17  
Toronto, ON M1B 4Z4 Canada  
Phone: 416-261-4865 Fax: 416-261-7879  
[www.scigiene.com](http://www.scigiene.com)

