

***Erwinia amylovora* PCR Detection Kit**

Product # 35100

Product Insert

Pathogen Information

Erwinia amylovora is a Gram negative, rod-shaped bacterium belonging to the family Enterobacteriaceae. It was the first bacterium identified as a plant pathogen and was shown to be the causative agent of the necrotic disease fireblight. The most economically important hosts are apple and pear, however infection can occur in a range of other plants including Cotoneaster, Cratagus and Pyracantha. During severe disease outbreaks, the rapid spread of bacteria through host tissue can lead to the loss of entire trees in a single growing season. *E. amylovora* infection of blossom can greatly reduce crop yield and marketability. Fireblight is therefore a highly destructive disease and is of major economic concern to fruit growers world-wide.

Principle of the Test

Norgen's *Erwinia amylovora* PCR Detection Kit constitutes a ready-to-use system for the isolation and detection of *E. amylovora* using end-point PCR. The kit first allows for the isolation of bacterial DNA using spin-column chromatography based on Norgen's proprietary resin. Bacterial DNA can be isolated from bacterial culture and plant tissue using this kit. The DNA is isolated free from inhibitors, and can then be used as the template in a PCR reaction for *E. amylovora* detection using the provided *E. amylovora* Master Mix. The *E. amylovora* 2x Detection PCR Mastermix contains reagents and enzymes for the specific amplification of a 355 bp region of the bacterial genome. In addition, Norgen's *E. amylovora* PCR Detection Kit contains a second Mastermix, the Control 2X PCR Master Mix, which can be used to identify possible PCR inhibition and/or inadequate isolation via a separate PCR reaction with the use of the internal PCR control or the provided *Isolation Control (IsoC)*, respectively. This kit is designed to allow for the testing of 24 samples.

Kit Components:

Component	Contents
Resuspension Buffer	8 mL
Lysis Solution	15 mL
Wash Solution	9 mL
Elution Buffer	3 mL
Bead Tubes	24
Spin Columns	24
Collection Tubes	24
Elution tubes (1.7 mL)	24
Ea 2x Detection PCR Master Mix	0.35 mL
Control 2x PCR Master Mix	0.35 mL
Isolation Control (IsoC)^a	0.3 mL
Ea Positive Control (PosC)^b	0.1 mL
<i>Nuclease Free-Water</i>	1.25 mL
Norgen's DNA Marker	0.1 mL
Product Insert	1

* IsoC = Isolation Control ; PosC= Positive Control
^a The isolation and PCR control are a cloned PCR product.
^b The positive control is *E. amylovora* genomic DNA

Customer-Supplied Reagents and Equipment

- Disposable powder-free gloves
- Benchtop microcentrifuge
- 1.5 mL microcentrifuge tubes
- 65°C water bath or heating block
- 96 – 100% ethanol
- 70% ethanol
- RNase A (optional)

Storage Conditions and Product Stability

All buffers should be kept tightly sealed and stored at room temperature (15-25°C). Buffers can be stored for up to 1 year without showing any reduction in performance.

The *Ea* 2x Detection PCR Master Mix, Control 2x PCR Master Mix, *Ea* Positive Control (PosC) and the Isolation Control (IsoC) should be kept tightly sealed and stored at -20°C for up to 1 year without showing any reduction in performance. Repeated thawing and freezing (> 2 x) should be avoided, as this may reduce the sensitivity. If the reagents are to be used only intermittently they should be frozen in aliquots.

General Precautions

The user should exercise the following precautions when using the kit:

- Use sterile pipette tips with filters.
- Store and extract positive material (specimens, controls and amplicons) separately from all other reagents and add it to the reaction mix in a spatially separated facility.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Work quickly on ice.

Quality Control

In accordance with Norgen's ISO 9001 and ISO 13485-certified Quality Management System, each lot of Norgen's *Erwinia amylovora* PCR Detection Kit, including the *Ea* 2x Detection PCR Master Mix, Control 2x PCR Master Mix, *Ea* Positive Control (PosC) and the Isolation Control (IsoC) are tested against predetermined specifications to ensure consistent product quality.

Product Use Limitations

Norgen's *Erwinia amylovora* PCR Detection Kit is designed for research purposes only. It is not intended for human or diagnostic use.

Product Warranty and Satisfaction Guarantee

NORGEN BIOTEK CORPORATION guarantees the performance of all products in the manner described in our product manual. The customer must determine the suitability of the product for its particular use.

Safety Information

Ensure that a suitable lab coat, disposable gloves and protective goggles are worn when working with chemicals. For more information, please consult the appropriate Material Safety Data Sheets (MSDSs). These are available as convenient PDF files online at www.norgenbiotek.com.

CAUTION: DO NOT add bleach or acidic solutions directly to the sample-preparation waste.

Protocol

A. *Erwinia amylovora* Genomic DNA Isolation

Important Notes Prior to Beginning Protocol:

- A variable speed centrifuge should be used for maximum kit performance. If a variable speed centrifuge is not available a fixed speed centrifuge can be used, however reduced yields may be observed.
- Ensure that all isolation solutions are at room temperature prior to use, and that no precipitates have formed. If necessary, warm the solutions and mix well until the solutions become clear again.
- Prepare a working concentration of the **Wash Solution** by adding 21 mL of 96 - 100 % ethanol (provided by the user) to the supplied bottle containing the concentrated **Wash Solution**. This will give a final volume of 30 mL. The label on the bottle has a box that may be checked to indicate that the ethanol has been added.
- Preheat a water bath or heating block to 65°C.
- **Isolation Control (IsoC)**
 - An Isolation Control (*IsoC*) is supplied. This allows the user to control the DNA isolation procedure. For this assay, add the Isolation Control (*IsoC*) to the lysate during the isolation procedure
 - The Isolation Control (*IsoC*) must not be added to the sample material directly.
 - Do not freeze and thaw the Isolation Control (*IsoC*) more than 2 times.
 - The Isolation Control (*IsoC*) must be kept on ice at all times during the isolation procedure.
- The PCR components of the *Erwinia amylovora* PCR Detection Kit should remain at -20°C until DNA is extracted and ready for PCR amplification.
- It is important to work quickly during this procedure.

1. Lysate Preparation

- a. **Bacteria from culture:** Transfer up to 1 mL of bacterial culture to a microcentrifuge tube (provided by user).

Bacteria from Plant Tissue or Fruit: Wash the tissue or fruit with an appropriate amount of DNase free water or **Resuspension Buffer** with vortexing. Transfer up to 1 mL of washed water to a microcentrifuge tube (provided by user).

- b. Centrifuge at 14,000 × g (~14,000 RPM) for 2 minutes to pellet the cells. Pour off the supernatant carefully so as not to disturb or dislodge the cell pellet.
- c. Add 500 µL of **Lysis Solution** to the cell pellet. Resuspend the cells by gentle vortexing.
- d. Transfer the mixture to a provided **Bead Tube** and secure the tube horizontally on a flat-bed vortex pad with tape, or in any commercially available bead beater equipment (e.g. Scientific Industries' Disruptor GenieTM).
- e. Vortex for 5 minutes at maximum speed or optimize the condition for any commercially available bead beater equipment.

Note: Foaming during the homogenization is common. This foaming is due to detergents present in the **Lysis Buffer** and will not affect the protocol.

- f. Incubate the **Bead Tube** with lysate at 65°C for 10 minutes. Occasionally mix the lysate 2 or 3 times during incubation by inverting the tube.
- g. Briefly spin the tube to remove liquid from the cap, and transfer all of the lysate, including cell debris, to a DNase-free microcentrifuge tube (provided by the user) by pipetting. Ensure that the beads are not transferred during the pipetting.
- h. Centrifuge the tube for 2 minutes at 14000 × g (~14,000 RPM).
- i. Carefully transfer clean supernatant to a new DNase-free microcentrifuge tube (provided by the user) without disturbing the pellet. Note the volume.

- j. Add an equal volume of 70% ethanol (provided by the user) to the lysate collected above (100 μL of ethanol is added to every 100 μL of lysate). Vortex to mix.
- k. Proceed to Step 2: Binding DNA to Column

2. Binding DNA to Column

- a. Assemble a spin column with one of the provided collection tubes.
- b. Add 10 μL of **Isolation Control (IsoC)** to the lysate mixture.
- c. Apply up to 600 μL of the lysate with ethanol onto the column and centrifuge for 1 minute at 14,000 \times g (\sim 14,000 RPM). Discard the flowthrough and reassemble the spin column with the collection tube.

Note: Ensure the entire lysate volume has passed through into the collection tube by inspecting the column. If the entire lysate volume has not passed, spin for an additional minute.

- d. Depending on your lysate volume, repeat step 2c if necessary.

3. Column Wash

- a. Apply 500 μL of Wash Solution to the column and centrifuge for 1 minute.
Note: Ensure the entire wash solution has passed through into the collection tube by inspecting the column. If the entire wash volume has not passed, spin for an additional minute.
- b. Discard the flowthrough and reassemble the column with its collection tube.
- c. Repeat step 3a to wash column a second time.
- d. Discard the flowthrough and reassemble the spin column with its collection tube.
- e. Spin the column for 2 minutes in order to thoroughly dry the resin. Discard the collection tube.

4. DNA Elution

- a. Place the column into a fresh 1.7 mL Elution tube provided with the kit.
- b. Add 75 μL of Elution Buffer to the column.
- c. Centrifuge for 2 minutes at 200 \times g (\sim 2,000 RPM), followed by a 1 minute spin at 14,000 \times g (\sim 14,000 RPM). Note the volume eluted from the column. If the entire volume has not been eluted, spin the column at 14,000 \times g (\sim 14,000 RPM) for 1 additional minute.

5. Storage of DNA

The purified DNA may be stored at -20°C for a few days. It is recommended that samples be placed at -70°C for long term storage.

B. *Erwinia amylovora* PCR Assay Preparation

Notes:

- Before use, suitable amounts of all PCR components should be completely thawed at room temperature, vortexed and centrifuged briefly.
 - The amount of Ea 2X Detection PCR Master Mix and Control 2X PCR Mastermix provided is enough for up to 32 PCR reactions (24 sample PCR, 4 positive control PCR and 4 no template control PCR).
 - For each sample, one PCR reaction using the Ea 2X Detection PCR Mastermix and one PCR reaction using Control 2X PCR Mastermix should be set up in order to have a proper interpretation of the results.
 - For every PCR run, one reaction containing Ea Positive Control and one reaction as no template control must be included for proper interpretation of results.
 - The recommended minimum number of DNA samples tested per PCR run is 6.
 - Using a lower volume from the sample than recommended may affect the sensitivity of Ea Limit of Detection.
1. Prepare the PCR reaction for sample detection (Set #1, using **Ea 2X Detection PCR Mastermix**) and the PCR reaction for control detection (Set #2, using **Control 2X PCR Mastermix**) as shown in Table 1 below. The recommended amount of sample DNA to be used is 2.5 μ L. However, a volume between 1 and 5 μ L of sample DNA may be used as template. Ensure that one Ea detection reaction and one control reaction is prepared for each DNA sample. Adjust the final volume of the PCR reaction to 20 μ L using the Nuclease-Free Water provided.

Table 1. PCR Assay Preparation

PCR Components	Volume Per PCR Reaction
Ea 2X Detection PCR Master Mix Or Control 2X PCR Master Mix	10 μL
Sample DNA	2.5 μL
Nuclease-Free Water	7.5 μL
Total Volume	20 μL

2. For each PCR set, prepare **one** positive control PCR as shown in Table 2 below:

Table 2. PCR Positive Control Preparation

PCR Components	Volume Per PCR Reaction
Ea 2X Detection PCR Master Mix Or Control 2X PCR Master Mix	10 μL
Ea Positive Control (PosC)	10 μL
Total Volume	20 μL

- For each PCR set, prepare **one** no template control PCR as shown in Table 3 below:

Table 3. PCR Negative Control Preparation

PCR Components	Volume Per PCR Reaction
Ea 2X Detection PCR Master Mix Or Control 2X PCR Master Mix	10 µL
<i>Nuclease-Free Water</i>	10 µL
<i>Total Volume</i>	20 µL

C. *Erwinia amylovora* PCR Assay Programming

- Program the thermocycler according to the program shown in Table 4 below.
- Run one step PCR.

Table 4. *Erwinia amylovora* Assay Program

PCR Cycle	Step	Temperature	Duration
<i>Cycle 1</i>	Step 1	95°C	3 min
<i>Cycle 2 (35x)</i>	Step 1	94°C	15 sec
	Step 2	60°C	15 sec
	Step 3	72°C	30 sec
<i>Cycle 3</i>	Step 1	72°C	5 min
<i>Cycle 4</i>	Step 1	4°C	∞

D. *Erwinia amylovora* PCR Assay Results Interpretation

- For the analysis of the PCR data, the entire 15-20 µL PCR Reaction should be loaded on a 1X TAE 1.7% Agarose DNA gel along with 10 µL of Norgen's DNA Marker (provided).
- The PCR products should be resolved on the 1X TAE 1.5% Agarose gel at 150V for 30 minutes (Gel running time will be vary depending on an electrophoresis apparatus).
- Sample results are provided below:




Figure 1. Sensitivity of Detection using the *Erwinia amylovora* PCR Detection Kit. A representative 1X TAE 1.5% agarose gel showing the amplification of *Erwinia amylovora* at different concentrations. The size of the *E. amylovora* target amplicon corresponds to 355 bp as represented by the provided DNA Marker (M). NTC = Negative Control.




Figure 2: A representative 1X TAE 1.7% agarose gel showing the amplification of the spiked **Isolation Control** and the internal **PCR Control** under different conditions using the **Control 2X PCR Mastermix**. The size of the Isolation Control amplicon and PCR Control amplicon correspond to 499 bp and 150 bp, respectively, as represented by the provided DNA Marker (M). Lanes 1 to 5 showed detection of both Isolation Control and PCR Control, suggesting that the DNA isolation as well as the PCR reaction was successful. Lane 6 showed only the detection of PCR Control suggesting that while the PCR was successful, the isolation failed to recover even the spiked-in Isolation control. **NTC** = Negative Control.

Table 5. Interpretation of PCR Assay Results

Input Type	Target reaction	Control Reaction		Interpretation
	<i>Erwinia amylovora</i> Target Band (355 bp)	IsoC Band (499 bp)	PCRC Band (150 bp)	
Positive Control	X	X	X	Valid
Negative Control			X	Valid
Sample	X	X	X	Positive
Sample		X	X	Negative
Sample			X	Re-test
Sample				Re-test
Sample		X		Negative
Sample	X		X	Positive
Sample	X	X		Positive
Sample	X			Re-test

** For results obtained that are not covered in Table 5 above, please refer to the Troubleshooting Section.

E. *Erwinia amylovora* PCR Assay Specificity and Sensitivity

- The specificity of Norgen's *Erwinia amylovora* PCR Detection Kit is first and foremost ensured by the selection of the *E. amylovora* specific primers, as well as the selection of stringent reaction conditions. The *E. amylovora* specific primers were checked for possible homologies to GenBank published sequences by sequence comparison analysis and published strains.

F. Linear Range

- The linear range of Norgen's *Erwinia amylovora* PCR Detection Kit was determined by analysing a dilution series of a *E. amylovora* quantification standards ranging from 100 ag to 1 pg.
- Each dilution has been tested in replicates (n = 4) using Norgen's *E. amylovora* PCR Detection Kit on a 1X TAE 1.7% agarose gel.
- The linear range of Norgen's *E. amylovora* PCR Detection Kit has been determined to cover concentrations from 100 ag to 1 ng
- Under the conditions of the Norgen's *Erwinia amylovora* DNA Isolation procedure, Norgen's *E. amylovora* PCR Detection Kit covers a linear range from 100 copies to 1 x 10⁶ copies.

Frequently Asked Questions

1. How many samples should be included per PCR run?

- Norgen's *Erwinia amylovora* PCR Detection Kit is designed to test 24 samples. For every 6 samples, a non-template control (Nuclease Free Water) and a Positive Control must be included. It is preferable to pool and test 6 samples at a time. If not, the provided Positive Control is enough to run 3 samples at a time.

2. How can I interpret my results if neither the *Ea* PCR control nor the Isolation Control (*IsoC*) amplifies?

- If neither the *Ea* PCR control nor the Isolation Control (*IsoC*) amplifies, the sample must be re-tested. If the positive control showed amplification, then the problem occurred during the isolation, where as if the Positive control did not amplify, therefore the problem has occurred during the setup of the PCR assay reaction.

3. How should it be interpreted if only the PCR control showed amplification but neither the *Ea* target nor the Isolation control amplified for a sample?

- This indicates a poor isolation. The isolation procedure must be repeated.

4. How should it be interpreted if only the Isolation Control (*IsoC*) was amplified in a sample?

- The sample tested can be considered as *Erwinia amylovora* negative.

5. How should it be interpreted if the PCR control and the *Ea* target showed amplification in a sample?

- The sample tested can be considered positive. It could happen when too much template was added to the reaction.

6. How should it be interpreted if only the *Ea* target and the PCR control were amplified in a sample?

- The sample tested can be considered as *Erwinia amylovora* positive.

7. How should it be interpreted if only the *Ea* target was amplified in a sample?

- It is recommended that the isolation is repeated.

8. How should it be interpreted if only the PCR control and the Isolation control showed amplification in a sample?

- The sample tested can be considered negative

9. What if I forgot to do a dry spin after my third wash?

- Your first DNA elution will be contaminated with the Wash Solution. This may dilute the DNA yield in your first elution and it may interfere with the PCR detection, as ethanol is known to be a PCR inhibitor.

10. What if I forgot to add the Isolation Control (*IsoC*) during the isolation?

- It is recommended that the isolation is repeated.

11. What if I forgot to run the Control PCR for the sample and I only ran the Detection PCR and I obtained a positive result?

- The result can be considered positive. However, any negative result must be verified by running the associated control PCR to ensure that it is a true negative and not a false negative due to problems with the RNA isolation or the PCR reactions.

Related Products	Product #
Fungi/Yeast Genomic DNA Isolation Kit	27300
Bacterial Genomic DNA Isolation Kit	17900
Plant/Fungi DNA Isolation Kit	26200

Technical Assistance

NORGEN's Technical Service Department is staffed by experienced scientists with extensive practical and theoretical expertise in sample and assay technologies and the use of NORGEN products. If you have any questions or experience any difficulties regarding Norgen's *Erwinia amylovora* PCR Detection Kit or NORGEN products in general, please do not hesitate to contact us.

NORGEN customers are a valuable source of information regarding advanced or specialized uses of our products. This information is helpful to other scientists as well as to the researchers at NORGEN. We therefore encourage you to contact us if you have any suggestions about product performance or new applications and techniques.

For technical assistance and more information, please contact our Technical Support Team between the hours of 8:30 and 5:30 (Eastern Standard Time) at (905) 227-8848 or Toll Free at 1-866-667-4362 or call one of the NORGEN local distributors (www.norgenbiotek.com) or through email at techsupport@norgenbiotek.com.

3430 Schmon Parkway, Thorold, ON Canada L2V 4Y6
Phone: (905) 227-8848
Fax: (905) 227-1061
Toll Free in North America: 1-866-667-4362