

---

**CORONAL DIAGNOSTIC SPECTROMETER**  
**SOHO**

---

**CDS SOFTWARE NOTE No. 55**

Version 5

August, 2003

---



**SOHO CDS-GIS Analysis Guide**

**August 2003**

**Carl Foley**

**Solar and Stellar Physics Group,  
Mullard Space Science Laboratory  
University College London,  
Holmbury St Mary, Dorking,  
Surrey, RH5 6NT, UK**

**[carl.foley@mssl.ucl.ac.uk](mailto:carl.foley@mssl.ucl.ac.uk)**

## Revision list

Version 1, 19 May 1999	Original
Version 2, 1 July 1999	Converted to CDS software note 55
Version 3, 23 July 1999	Added approximate wavelength coverage; added /cross_cor keyword to <b>ghost_plot_one</b> ; described ghost_info; added /plot and /hardcopy keywords to ghost_buster; ghost_buster prints wavelengths with the ghost cursor; added section on saving and restoring ghost sessions; gis_calib uses new calibration coefficients; described <b>gis_write_calib</b> ; added the new routines to Appendix A.
Version 4, 11 Jan 2000	Expanded section on wavelength calibration
Version 5, August 2003	Added automatic option to <b>ghost_buster</b> and included associated ghost information structures. Added flat field files and correction to <b>gis_calib</b> . Made photons/sec/arcsec <sup>2</sup> /cm <sup>2</sup> default output from <b>gis_calib</b> , instead of counts. Included error tag to the calibrated qlds structure from <b>gis_calib</b> Added new routine <b>gis_fit</b> which corrects, calibrates and fits the GIS data non-interactively, providing the user with a simple entry into the analysis of GIS data. Included new <b>gis_utplot</b> plot routine. Restructured and worded analysis guide throughout. Put on web in HTML format with example programs etc.

## Table of contents

1. Introduction .....	4
1.1 Related Documents.....	5
1.2 GIS Observations.....	5
2. Finding GIS Data.....	6
2.1 Browsing the CDS Catalogue .....	6
3. Reading and displaying GIS data.....	7
4. Correcting GIS Data for Instrument Effects.....	8
4.1 Fixed Patterning .....	9
4.2 Ghosts.....	10
4.2.1 Displaying GIS Ghost details.....	10
4.2.2 Finding Ghosts.....	10
4.2.3 Ghost Information files.....	12
4.3 Ghost Correction.....	13
4.3.1 Ghost free mode.....	13
4.3.2 Manual mode .....	14
4.3.3 Saving and restoring ghost_buster sessions.....	15
4.4 Edge effects.....	16
5. Calibrating GIS Data .....	16
5.1 Wavelength Calibration.....	17
5.2 Intensity Calibration.....	17
5.2.1 Calibration details.....	18
6. Using calibrated spectral data.....	19
6.1 Extracting the data from the qlds .....	19
6.2 Line fitting.....	20
6.3 GIS_FIT.....	21
6.4 GIS_UTPLOT.....	21
Appendix A. List of useful routines.....	23
Appendix B. Example Program.....	24
Appendix C. Ghost free line lists .....	25

# 1 Introduction

The objective of this document is to guide the user into the use of data obtained with the Coronal Diagnostic Spectrometer's Grazing Incidence detector, (GIS) which is part of the SOHO payload. This document should be used in combination with the instrument guide which contains more detailed information with regard to the detectors and their in-flight characteristics. It is recommended that the instrument guide is read first, and then the analysis process described here followed.

Reading and calibrating the GIS data is simplified greatly by using the Solarsoft IDL library which is maintained and distributed at Lockheed Martin (<http://www.lmsal.com/solarsoft>). It is also available at the Solar UK research Facility at MSSL (<http://surfwww.mssl.ucl.ac.uk/solarsoft>), where full instructions and support for the installation and maintenance of the software is available. Some familiarity with the IDL analysis package is assumed and Solarsoft is assumed.

To make the CDS libraries available on unix platforms and invoke IDL within SolarSoft use the system commands:

```
setssw cds
sswidl
```

If using the CDS operations version of the IDL software (available from the Rutherford Appleton Laboratory) then try the system command `cidl`, though this may vary at different sites. Many routines are available within the CDS SolarSoft library (some useful ones are summarized in **Appendix A**). To find which routines which are available for a particular task use the IDL command (`[ ]`'s enclose optional inputs):

```
IDL> tftd, 'search_string' [, lines=lines, /prog, cat=cat]
```

This searches the CDS SolarSoft IDL directories for all occurrences of the search word in routine descriptions. For more information about individual routines read the routine header documentation with:

```
xdoc, 'routine_name'
```

or

```
doc_library, 'routine_name'
```

In the following sections we describe the analysis techniques which may be used with the GIS data. It is recommended that the `'doc_library'` routine is used to investigate each of these routines further since the headers contain a lot more information, and available options which are omitted here for clarity.

## 1.1 Related Documents

The GIS software forms part of the CDS SolarSoft package put together at RAL, Oslo, Goddard, MSSL, and other places.

The associated CDS Software notes are particularly useful and they come as part of the CDS SolarSoft program library distribution. A few of them are listed below; all the notes can be obtained from the web at

*[http://orpheus.nascom.nasa.gov/cds/software\\_notes.html](http://orpheus.nascom.nasa.gov/cds/software_notes.html)*.

CDS software note 54 - The GIS Instrument Guide, a detailed description of the GIS instrument.

CDS software note 20 - CDS Quicklook Software User Manual is an introduction to the general CDS software available.

CDS software note 22 - CDS on-line help utilities. To find your way about the IDL software in the CDS SolarSoft distribution.

CDS software note 9 - The CDS Quicklook Data Structure. Gives more information about the 'quick look data structure' (qlds) and the data within it.

CDS software note 47 - The Component Fitting System for IDL. For fitting Gaussians/Voigt profiles etc. to spectra.

CDS software note 37 - Diagnostic Line Ratios (using CHIANTI) for CDS. Using the CHIANTI atomic physics databases with CDS data.

CDS software note 33 - The CHIANTI Synthetic Spectrum Program for CDS/SUMER.

CDS software note 50 - Differential Emission Measure analysis using CHIANTI.

## 1.2 GIS Observations

Unlike the NIS, observations with the GIS are astigmatic, and the whole of the spectrometer slit is used to produce the spectra. To cover an area larger than the normal GIS slits (2" \_2", 4" \_4", and for observations away from the solar disc 8" \_50") the scan mirror and slit are moved over an area of up to 4 arc minutes. Data from all four detectors are gathered simultaneously, and although the minimum integration time is 1 second per pointing, it takes about 15 seconds to transmit the data. Thus to cover an area of 20" \_20", using the 4" \_4" slit and 50 second integration time per position takes just over 20 minutes. Because the GIS electronic encoding is sensitive to the incoming count rate, different GIS set-up (GSET) parameters are needed for different expected count rates. Much of this selection process is automated when the observation is uplinked to the spacecraft, although the observer must specify the solar 'zone' beforehand. The zone is a rough description of the solar activity expected in the observation. It can be one of:

- off limb: not on the solar disk. The expected maximum detector count rates for these areas are below 5 counts/second/pixel.
- quiet sun: any quiet area of the sun, including coronal holes. The expected maximum count rates are 20 counts/second/pixel.
- active region: active regions, including those on the limb. The expected maximum count rates are above 20 counts/second/pixel.

Once the zone and other observation parameters (raster size, slit, exposure time, exact location, etc.) are determined, the CDS planner at the Experiment Operations Facility can insert the GIS observation into the current plan. The combination of zone and slit selected will prompt the CDS planner to choose an appropriate GSET. The GSETs themselves are pre-determined from special (raw) observations made with the GIS. A number of raw observations for each zone, each slit, each detector and for various high voltage settings, are examined by the GIS team at MSSL and the best selected to produce the current set of GSETs. The raw files associated with each GSET are also used to produce ghost information structures and files. These enable likely ghosting regions to be identified, and in many cases corrected for automatically, see section 4.3.

## 2 Finding GIS data

The CDS mission dataset can be browsed interactively using the powerful widget based 'xcat' routine.

```
IDL> xcat [, qlds, tstart=tstart, tend=tend]
```

A lot of progress can be made using this routine, such as displaying images and spectra. It is a good idea to play with the menus at this stage if you haven't already done so. For example try selecting 'GIS only' from the 'Detector' menu, and put the required dates in the 'Start Time' and 'Stop Time' fields, and try selecting each of the read file options.

It is also possible to investigate the databases directly using the 'list\_exper' routine. This route is particularly useful when you want to find data automatically within one of your own IDL programs.

```
IDL> LIST_EXPER,'18:00 21-Jan-1998','21:00 21-Jan-1998', OBS, N_FOUND
```

The output 'obs' is an IDL structure which contains full details of the observations which CDS performed between the selected times. You can get an idea of the contents by typing:

```
IDL> help, obs, /str
```

You can use simple logic to select particular observations. For example to select all the GIS observations performed in the selected time interval:

```
IDL> gis=where(OBS.detector eq 'G')
IDL> print, OBS.filename(gis)
```

### 3 Reading and displaying GIS data

The SOHO CDS data are stored and distributed in FITS format, which can be read into a 'quick look data structure' (qlds) using the 'readcdsfits' routine. For example, to read in the GIS data set defined as study number s4384 raster r00:

```
IDL> qlds = readcdsfits('s4384r00')
```

The qlds contains the spectral data, some wavelength calibration information, ancillary information about spacecraft pointing, and much more in a hierarchal structure similar to that obtained using the LIST\_EXPER routine (mentioned in the previous section). It is possible to browse all the information in the qlds with the following

```
IDL> help, qlds, /str
```

This will display the following:


```
QL_ID      DOUBLE  1.0570199e+009  QL_NO      INT        1
HDRTEXT    STRING  Array[481]
HEADER     STRUCT  -> HDRX Array[1]
DETDESC    STRUCT  -> GISXX3 Array[4]
BACKGROUND INT        0
WAVECAL    STRUCT  -> GIS_WAVECAL Array[1]
DEL_TIMEDATA DOUBLE  Array[20]
DEL_TIMEDESC STRUCT  -> AUX2 Array[1]
INS_XDATA  FLOAT   Array[20]
INS_XDESC  STRUCT  -> AUX2 Array[1]
INS_YDATA  FLOAT   Array[20]
INS_YDESC  STRUCT  -> AUX2 Array[1]
.....
```

The structure can be investigated further in the following fashion.

```
IDL> help, qlds.header.gset_id
```

This will display the GSET for this observation. This embedding of information with the data is a powerful method of distributing the data and greatly simplifies the analysis process, making easy to quickly display and analyse the raw data. For example to plot all four GIS data windows you can use the routine 'cds\_snapshot':

```
IDL> cds_snapshot, qlds [/spectra, /log, window=window]
```



SOHO/CDS GIS Raster, 25-Aug-1996 15:13:16

G2AL -- Activity Boundary -- s4384r00.fits  
Center = (-279",73"), Size = 81"x4"


**Figure 1:** `cds_snapshot` of GIS data, using the `/log` switch. The salt and pepper noise which can be seen over this data is due to fixed patterning (described in section 4.1). This plot displays spectra since the G2AL study is a sit and stare study. For rasters, `cds_snapshot` will display an image, unless the `/spectra` switch is used.

## 4. Correcting GIS Data for Instrument Effects

### 4.1 Fixed Patterning

Fixed patterning, is an effect caused by the interaction of the GIS digital electronics with the analogue detector read-outs; it is very pronounced in some parts of the GIS spectra. Figure 2 shows a subset of data from detector 1. A boxcar smooth of the data will reduce the fixed patterning seen in the figure but increase the line widths somewhat; the increase depending on the size of the boxcar used. Thus instead a convolution with a Hanning function is used in the program `gis_smooth`. This function preserves the total counts in the spectral lines, without introducing artifacts in the background, or increasing the line widths.

s10716r00.fits SPECT\_1 1998-03-26 GIS1



**Figure 2:** A plot of the data around an Fe XI line from study 10716, showing fixed patterning (left plot), where data shows large variations from pixel to pixel. The right plot shows the same data after running `gis_smooth`.

To smooth the data within a qlds use:

```
IDL> gis_smooth, qlds [, smoothsize=size]
```

The option `smoothsize` exists to change the size of the function used, but the default (`smoothsize=7`) works under normal circumstances. To assist with the smoothing operation, any missing data in the data set are estimated automatically, the smooth function is applied, and then the original missing data marked as missing again. It is not recommended to replace missing GIS data separately *before* using any of the GIS processing software; it is taken into account already by the programs.

## 4.2 Ghosts

An effect in the GIS detectors is the presence of ‘ghosts’. These are displaced spectral lines, or parts thereof, caused by an ambiguity in the encoding used by the electronics. The routine **ghost\_buster** is used to remove or return ghosts in the GIS data.

When the GIS is in raw data mode, it transmits the data from one detector as a series of coordinate pairs from the detectors. When these coordinates are plotted they form a spiral (see the GIS Instrument Guide) where the spectral dimension lies along the length of the spiral, and the intensity is the number of counts at each position along the spiral. To translate these pairs into spectral position, a Look up Table (LUT) is used as part of the on-board GIS processing. The parameters that define the LUT, plus the high voltage setting, form the GSET mentioned earlier. Every GSET includes a different set of LUT parameters, and the LUTs are calculated by the on-board GIS processor for each detector immediately before the first observation using that GSET. The ambiguity (or ghosting) in spectral data does not occur over the whole spectral range, but where the thin spiral arms broaden and overlap each other. If a spectral line is ghosted, it is confined to occur only at specific locations in adjacent spiral arms.

These characteristics are used in **ghost\_buster** to restore the ghosts to their original locations. The counts in a ghost must be added to the counts at the original location to produce the correct line intensity. In ghosted areas, which cover 40% to 50% of the data, it is easier to separate the original line from ghosts where there are no blends. Thus ghosted areas in quiet sun observations are easier to correct than in an active region; although the unghosted areas remain unaffected for all observation zones.

### 4.2.1 Displaying GIS Ghost details


The output from **ghost\_plot\_one** (e.g., Figure 3) displays the GIS data with an emphasis on sorting ghosts from the original data:

```
ghost_plot_one, qlds, detector_no [./sample, /nm, /angstroms,/logscale, /pixels, $  
waverange=[min, max], /cross_cor]
```

This routine allows switches to modify how the data are plotted. For instance, *waverange* can be used to zoom into a particular region of interest, by specifying a minimum and maximum wavelength to be plotted; and *sample* plots a sample theoretical spectrum.

The routine first adds all the data from an observation into one spectrum per detector, then plots the data with information to help find possible ghosts. Figure 3 was produced using:

```
IDL> qlds = readcdsfits('s10716r00')  
IDL> gis_smooth, qlds  
IDL> ghost_plot_one, qlds, 1, /sample, /angstroms
```



**Figure 3:** output from `ghost_plot_one`, (top-top) theoretical quiet sun spectrum, (top-middle) left shifted spectrum, (top-bottom) right shifted spectrum. (Bottom) Spectrum to be de-ghosted, with likely ghosting regions shaded.

The areas of the resulting plot are described from the top:

- Theoretical quiet sun spectrum: This is the spectrum requested with the `/sample` keyword. There are a number of warnings that come with the sample spectrum, see `xdoc`, `'ghost_plot_sample'` for details. The spectrum is intended only as an aid to ghost restoration; the intensities of lines are only approximate.
- left shifted spectrum and (correlation): The spectrum plotted is a copy of the observed data, but shifted to the left by one spiral arm. Any ghost that occurs can only be from an original line one arm to the left or to the right. As the ghosting scale is not linear, the amount of shift varies across the detector.
- right shifted spectrum, a copy of the observed data shifted to the right, with the optional correlation.
- SPECTRUM: The lower part of the plot contains a summary of the data, where all the exposures in the observation have been added together. The grey boxes underlying the plot are information from 'ghost information files'; which show where the ghosts are expected. The vertical scale is in arbitrary units to show

most of the lines, but to enhance the weaker lines plot logarithmically with `logscale`.

If the `cross_cor` keyword is used, the correlation between the original and shifted spectra is displayed. If the absolute cross correlation coefficient is greater than 50% it plots a thick bar. This is only a guide to where ghosts may be located, and is easily confused by line blends. By default the correlation is not plotted. To summarise the data from all four detectors in a similar way use

```
IDL> ghost_plot_all, qlds [ /angstroms, /nm]
```

## 4.2.2 Finding Ghosts

The output from `ghost_plot_one` (Figure 3) can be used to check whether a line is ghosted, or itself is a ghost. Suppose we have seen a line, and want to check it; for example there is a spectral line on the plot between 160 and 161Å (these ghosts are now corrected automatically by default, since they are located in non blending regions – see section 4.3). This line is in the ghosting region defined by the shaded region on the plot. The first check is to look at the ‘theoretical quiet sun spectrum’; it shows that there are no lines at this location. This may be a fault with the theory; however further checking shows that the ‘left shifted spectrum’ also has a line at this point. This strongly suggests that the line at about 160Å is a ghost and needs to be pushed to the right to correct the line. The manual ghost correction software explains the process of moving the line in more detail, as well as overlaying a ‘ghost cursor’ on the plot. This cursor shows exactly where lines would come from if they are ghosts. Alternatively, if we are familiar with this area on the solar spectrum we know there is a FeXIII line at 202Å (see Appendix C). To see whether this line needs correcting, first notice that it is a clear area of the plot, i.e., not in a ghosting area. Neither the left nor right shifted spectra show lines at this point, and thus this line is free from ghosts.

## 4.2.3 Ghost Information files and structures


The ghost information files contain the pixel locations of ghosting regions which aren’t corrected for automatically by `ghost_buster`. All GSETS have a ghost information file, which is stored with the other GIS calibration files located in the `$CDS_GIS_CAL_INT` directory. Also located in this directory are the ghost information structures, which are used by `ghost_buster` to automatically restore ghosts.

### 4.3 Ghost Correction

There are now three modes for ghost correction: automatic, manual and ghost free. By default `ghost_buster` works in automatic mode. This automatically corrects for most ghosts where they are not coincident on other lines and are resolvable.

```
IDL> ghost_buster, qlds [,/plot ,/hardcopy ,/nm]
```

This corrects for about half of the ghosts (or 70% of the data, see Figure 4).



**Figure 4:** `ghost_buster` automatic restoration of ghosts. (Left) uncorrected averaged smoothed spectrum. (right) The same spectrum corrected using `ghost_buster` in automatic mode. This works well for quiet sun observations, restoring almost all lines.

To correct the remaining ghosts, where the ghost lands in the location of another line, manual mode is required. This is described below. However, if from using `ghost_plot_one`, the region of interest does not need correcting then use ghost free mode. Once the ghosted areas have been removed with this mode, then the data can be further calibrated without problems. If the region does need correcting, then run in manual mode. Care must be taken with manually corrected data; it is possible to run `ghost_buster` without making any changes, yet the software still allows further calibration. Because the corrected count rates must be used for the count rate dependant gain depression calibration (see below), it is important to replace the ghosts in any region of interest before calibrating further. Only ghosts of the lines within the region of interest need to be moved, all others can remain in place without affecting the calibration. If spectral lines need to be fitted, it is recommended to correct for ghosts first, calibrate the data, and then fit the lines. In either mode, it is possible to run `ghost_buster` again in the same or a different mode, although if it is run on the same detector a warning will be

displayed. To reset all spectra to the uncorrected state it is necessary to re-read the qlds from the data file.

### 4.3.1 Ghost free mode

This mode uses the ghost information files (section 4.2.3) to return only those areas of the GIS spectra that show very little (less than about 5%) or no ghosting.

```
IDL> ghost_buster, qlds, /free [,/plot /hardcopy /nm]
```


About half of each spectrum is returned. Ghost free mode is available for all GSETs. Use the plot or hardcopy switches to see the changes applied to the qlds. After using ghost free mode, blocks of data from the known ghosting regions will be missing from the qlds. The areas remaining can be used without needing ghost correction.

### 4.3.2 Manual mode

Manual mode allows the user to move ghosted lines within a single detector. To run in manual mode supply two parameters, the qlds and a detector number.

```
IDL> ghost_buster, qlds, detector_no [,/hardcopy, /nm , /sample, /logscale, /plot,  
/angstroms, save=save_struct]
```

Manual mode allows the user to specify which GIS spectral lines to relocate. The program gives hints by using **ghost\_plot\_one** (Section 4.2.1) to show the spectrum with overlying spectra shifted both to the left and to the right, a cross correlation with these spectra, and an estimate of the amount of ghosting in each direction. It also uses a 'ghost cursor' to show exactly where a spectral line would ghost. The main cursor is a solid vertical line; either side of it is a dotted ghost cursor. The dotted lines are where any selected data will be moved to (i.e., they correspond to the next or previous spiral arms). The spacing between the dotted lines and the main cursor vary across the spectrum, as do the shifted spectra. Below the dotted lines and main cursor are printed the corresponding wavelengths for those points. To move a ghost that has been identified, position the cursor to the left of the ghost and press and drag the left mouse button over the ghost. For instance, to move the ghost previously identified at around 160Å (Section 4.2.2), press the mouse button with the cursor just to the left of 160Å, keep it depressed and move to about 162Å (see Figure 5).



**Figure 5:** Manual relocation of ghosts using **ghost\_buster** in automatic mode. (top) the region around the ghosted line located at 160\_ has been selected, and (bottom) relocated to its parent line at 175 \_.

Let go of the mouse and enter `r` in the terminal window to move the line to the right. Repeat this procedure until all the lines of interest have been corrected. When moving a ghost, remember to include enough background for line fitting on either side of the line of interest. It is possible to restore more ghosts by running the program again, on the same detector or any other. Use the `/plot` or `/hardcopy` keywords to see the final result; alternatively use `gis_plot` or `ghost_plot_one` when finished.

### 4.3.3 Saving and restoring **ghost\_buster** sessions

By adding the `save = save_struct` keyword, it is possible to save the ghost corrections made in manual mode into an IDL structure variable. It is then possible to restore these corrections to uncorrected data with

```
ghost_buster, qlds, detno, restore = save_struct [,/plot, /hardcopy]
```

For example:

```
IDL> ghost_buster, qlds1, 1, save = save_struct  
IDL> save, save_struct, filename='qlds1_det1.save'
```

and then at a later time

```
IDL> restore, 'qlds1_det1.save'  
IDL> ghost_buster, qlds2, 1, restore = save_struct
```

The first call to **ghost\_buster** is in manual mode, and the software will prompt for corrections to the data as outlined above. The corrections are stored in an IDL save file. The second call uses the restored information to reapply the corrections to a second data set. Note that both `qlds1` and `qlds2` must use the same GSET; and that the same detector (number 1 in this case) must be used.

#### 4.4 Edge effects

Most detectors show edge effects, both as spikes and a gradual increase in the background. There are many causes for these effects, notably: a compression of the wavelength scale, causing an increase in the background; end spoiling in the microchannel plates, where the strong electric fields diverge at the ends of the detectors; changes in the solar background, mainly from the HeII continuum; and some very bright (solar) lines, seen especially in detector 2. Preliminary corrections for the non solar sources are now included in **gis\_calib**. These have relatively large errors associated with them (which are passed into the output structure). These will reduce as we acquire more observations to check the corrections.

### 5. Calibrating GIS Data

The GIS calibration has recently been updated using the results of workshops which were held at the [International Space Science Institute](#) (ISSI) during 2001. These workshops were used to determine (using intercalibration observations) to find an absolute and relative intensity calibration for all the instruments on SOHO. This resulted in an overall upward shift in the sensitivity of the GIS by 2.6.

More recently it has been found that the lines in GIS 1 and 2 were becoming degraded due to the effects of Long term gain depression. For this reason new flat field files have been generated to correct for this degradation. Because of the size of the corrections (as much as 60%) and associated uncertainty, a pixel by pixel error array is now included in the calibrated `qlds` structure. This can be extracted using the **gis\_error** function

```
IDL> error=gis_error(qlds,detno)
```

## 5.1 Wavelength Calibration

Wavelength calibrations are read in automatically when the fits file is converted to a qlds. The wavelength calibration can be accessed by the routines **wave2pix**, **pix2wave**, which translate the GIS pixel co-ordinates as necessary.

```
pixel = wave2pix(spec_id, wavelength [,/limit])
```

Where, for the GIS, **spec\_id** is a string containing GIS1 GIS2 GIS3 or wavelength can be a single value or an array. See the example program B. Similarly

```
wavelength=pix2wave(spec_id, pixel)
```

Which will return the wavelengths of the pixel array for the spectrum. Again, the input can be a single value or an array. Variations of about 20% of a line width are expected between observations, caused mainly by GIS hardware temperature differences when observing different areas of the sun, or very high count-rates (more than about 40 counts/second/pixel) causing distortions in the electronic processing. Note that the wavelength calibration varies with each GIS setup (GSET) compare two GIS observations that used different GSETs it is necessary wavelength calibration. This is simplified with the routine **restore\_wavecal**. For instance, below is part of an IDL session to over plot two GIS observations:

```
IDL> qlds_1 = readcdsfits('s8965r00')
IDL> qlds_2 = readcdsfits('s8966r00')
IDL> gis_smooth, qlds_1
IDL> gis_smooth, qlds_2
IDL> print,restore_wavecal(qlds_1) ; prints "1" If restored
IDL> wave_1 = pix2wave('GIS1', findgen(2048))
IDL> data_1 = gt_windata(qlds_1, 0) ; window 0 is for detector 1
IDL> print,restore_wavecal(qlds_2)
IDL> wave_2 = pix2wave('GIS1', findgen(2048))
IDL> data_2 = gt_windata(qlds_2, 0) ; window 0 is for detector 1
IDL> plot, wave_1, data_1/max(data_1), xrange=[187,190], psym=10
IDL> oplot, wave_2, data_2/max(data_2)
```

## 5.2 Intensity Calibration

As well as the ghosts and fixed patterning, the detectors and their electronics have non-linearities that need to be corrected. On the whole these corrections are small (less than 10%, but dependent on count-rate). **gis\_calib** will correct for these (as well as correct for gain depression); and calibrate from counts into photons using the updated GIS calibration coefficients.

```
gis_calib, qlds [,errmsg=errmsg, /quiet, /steradian_m2, /counts]
```

By default the data will be returned in calibrated units of photons/sec/arcsec<sup>2</sup>/cm<sup>2</sup>. To convert to photons per second per solid angle per area use the /steradian\_m2 or (to leave in counts) /counts switches. The routine checks to make sure the routine **ghost\_buster** has been run for all the detectors - if not the error for data in the regions likely to ghost is set to 100% of the counts in the adjacent arms. Similarly, the error for regions which cannot be calibrated for any other reason is also set; this might include the occasional strong line that is too bright for the gain depression calibration (such as the HeII 304Å line), or whole detector count rates too high for the electronic dead time corrections.

The GIS calibration coefficients used are those from the results of the SOHO radiometric inter-calibration workshops which were held at the [International Space Science Institute](#) (ISSI) during 2001. The coefficients can be viewed using the program **gis\_write\_calib**.

### 5.2.1 Calibration details

The corrections and calibrations made by **gis\_calib** include:

- FIFO dead time: The dead time is a straightforward correction for an on-board First in First out (FIFO) event queuing chip. It applies to all four detectors simultaneously, and involves a constant non extending dead time allowing  $10^5$  counts per second through unhindered, with small corrections for higher rates.
- Simple correction for Quiz-show dead time: The programmed Quiz-show correction is simply an upper limit on the rates; if photon events are less than 6 microseconds apart then the data are marked as uncalibratable. The maximum rate is then  $1.0/(6.0 \cdot 10^{-6})$  or approximately  $1.7 \cdot 10^5$  over all four detectors.
- Analogue dead time: This involves an extending dead time of approximately 2 microseconds. The data used to correct for this were measured before launch using the flight electronics and are read from a data file.
- Count rate dependant gain depression (also known as short term gain depression). If more than 5% error would occur in the measured rates, then the datum (at that point only) is marked as uncalibratable.
- Long term gain depression and flat field: The correction is based on the total accumulated charge extracted from the MCP's.
- Detector / grating / telescope effective area. The calculations are based on a ground based calibration exercise for the whole of CDS. The relative calibration coefficients have since been verified in flight, but please note that there is currently a very large uncertainty in the absolute values.

## 6. Using calibrated spectral data

Various CDS and other SolarSoft routines exist to use the data once calibrated; most of these can be used with both NIS and GIS data. The generic `dsp_menu` program will work with both calibrated and un-calibrated data,

```
dsp_menu, qls [, qls2, qls3, ..., /delete, /nocheck]
```

### 6.1 Extracting the data from the qls

To extract the spectral data from a qls always use `gt_windata` or `gt_spectrum`

```
data = gt_windata(qls, window [./nocheck, /quick, /nocopy, $  
/nopadding ,errmsg=errmsg])
```

These routines will work with all storage schemes used for CDS qls data. With GIS data, specify the qls and the 'window', which is one less than the detector number. For example, to extract detector 3 data from a pre-defined qls use

```
IDL> GIS3data = gt_windata(qls, 2)
```

The returned data are a floating point array of up to 4 dimensions: spectrum (always 2048 pixels for the GIS), solar x, solar y, and time. The time dimension is only used for data that have multiple exposures at the same solar x, y point. To get information about the extracted data use

```
gt_windesc(qls, window [./nocheck, /quick, /errmsg])
```

For example:

```
IDL> GIS3desc = gt_windesc(qls, 2)
```

The returned structure tells us which units the data use (`GIS3desc.units`), the missing data flag (`GIS3desc.missing`), the wavelength range (`GIS3desc.wavemin`, `GIS3desc.wavemax`), etc. The command

```
IDL> xshow_struct, GIS3desc
```

will display all the information available in the structure. Also available is `gt_spectrum`, to get individual spectra from a raster:

```
result = gt_spectrum(qls, window=window, yix=yix, xix=xix, $  
[tix=tix, lambda=lambda, xsolar=xsolar, ysolar=ysolar, time=time])
```

Where `window`, `yix`, `xix`, `tix` are the user supplied window (i.e., the detector number-1), and for each raster position the x, y and time indices. The time index is not normally used unless the raster has repeated exposures at the same position. If supplied, `lambda` will return the wavelength for each pixel in the spectrum, `xsolar` and `ysolar`

are the locations on the sun in arc seconds from the sun centre and time is the time of the start of the exposure.

## 6.2 Line fitting

Many fitting procedures are available (see `tfid`, 'gauss' and `tfid`, 'voigt') which can be used to fit various profiles and backgrounds to ghost free or ghost corrected GIS spectra. One of these is `ezfit`:

```
ezfit, wavescale, data, waverange, k=k
```

where `wavescale` and `data` are the X and Y variables, `waverange` can be used to limit the fit to a particular line, and `k` is the polynomial order of the background. Thus to fit a Gaussian plus background to a region in the first exposure of the above GIS detector 3 data:

```
IDL> wavescale = pix2wave('GIS3', indgen(2048))
IDL> data = GIS3data[* , 0, 0, 0]
IDL> waverange = [415, 420]
IDL> ezfit, wavescale, data, waverange, k=1
```

Another routine, `cds_gauss`, has been developed for use with CDS data:

```
yfit = cds_gauss(x, y, [a, k])
```

Where `x` is the independent vector; `y` is dependent variable; `a` contains the returned coefficients and `k` defines the order of the polynomial to be fitted to the background. A more complicated but very comprehensive fitting routine is `cfit`

```
yfit = cfit(x, y, a, fit [,sigmaa] [+keywords])
```

Where `x`, `y` are the data to be fitted, `a` is an array of (nominal) parameter values – if defined then used as an initial guess to the fit; `fit` is a structure containing one tag for each component in the fit; `sigma` contains the errors for each of the parameter values in `a`. For detailed information about `cfit` see CDS software note 47 – The Component Fitting System for IDL.

## 6.3 GIS\_FIT

**GIS\_FIT** uses predefined cfit structures to automatically fit most regions of the GIS spectra.

```
IDL> gisfit=gis_fit(qlds, [/plot],gisfit=gisfit)
```


This routine works on calibrated and uncalibrated data. Un-calibrated data is corrected and calibrated automatically using the individual routines described in the previous sections of the analysis guide (**Gis\_smooth**, **Ghost\_buster**, **gis\_calib** etc). The visible identified lines are then fitted, using the **cfit** line fitting routine. The output structure contains the fitted line positions, widths, intensities, errors, and total line counts in photons/sec/arcsec<sup>2</sup> along with ancillary data such as the location and time of observation. The gisfit structure can be used to quickly produce light curves, images and DEM maps

## 6.4 GIS\_UTPLOT

To display a quick time series of data acquired with GIS you can use the routine **gis\_utplot**. This is a wrapper to the Solarsoft utplot program, which displays a time series of the data acquired. To call:

```
IDL> gis_utplot, qlds, det_no [,uttime=uttime,gisfit=gisfit ]
```

This plots a time series of the raw accumulated counts for the selected detector (see Figure 6).



**Figure 6:** The output from **gis\_utplot**, for a series of sit and stare observations in an active region located at the limb. Note the dimming periods around 17:00 and 20:00 UT. This is investigated in Figure 7 where individual lines are plotted.

This routine also works with arrays of qlds structures. So if you have a series of qlds from a particular observation you can combine these in the following way:

```
IDL> gis_utplot,[qlds, qlds1, qlds2, qlds3], det_no [, lines=lines, gisfit=gisfit]
```


Or alternatively a string array of program numbers for example:

```
IDL> gis_utplot,['26320', '26322', '26324'], 1 [, lines=lines, gisfit=gisfit]
```

This will read, correct, calibrate and then plot the data, In this case from a sit and stare observation which is displayed in Figure 6. The calibrated and fitted data is saved into the gisfit structure if this variable is passed on the command line. If it already exists it is updated. The gisfit structure can then be used to plot the corrected, calibrated and fitted lines from the observations.

```
IDL> gis_utplot, gisfit
```

Calling gis\_utplot in this fashion will result in a prompt for you to select a particular line from a list of available present in the structure to plot. Alternatively, the specific line id can be entered.



**Figure 7:** Output from gis\_utplot, for specific lines contained in the gisfit structure.

GIS specific software for determining, temperatures, emission measures, densities, and DEM's, based on the output gisfit structure, are under development. To find the latest routines, browse using

```
IDL> doc_library, 'gis*'
```

## Appendix A. List of useful routines

Some useful CDS and other SolarSoft routines, along with one-line descriptions. For detailed information examine the routine headers with `xdoc`, 'routine\_name'.

<code>cds_fill_missing</code>	Fills MISSING pixels in QLDS.
<code>cds_gauss</code>	Fits Gaussian with constant, linear or quadratic background
<code>cds_snapshot</code>	Makes a thumbnail sketch of images from a CDS study
<code>chianti_ne</code>	Calculate and plot CHIANTI density sensitive line ratios
<code>chianti_ss</code>	Calculates and plots CHIANTI synthetic spectrum
<code>chianti_te</code>	Calculate and plot CHIANTI temperature sensitive line ratios
<code>cfit</code>	Make a best fit of the sum of components to the supplied data
<code>dsp_menu</code>	Selection of display modes for CDS QL data
<code>Ezfit</code>	Easy Gauss fit to data
<code>Ghost_buster</code>	Move GIS detector ghosts.*
<code>Ghost_info</code>	Display the ghosting information associated with a QLDS*
<code>Ghost_plot_all</code>	Plots all 4 GIS detector data with likely ghost regions*
<code>Ghost_plot_sample</code>	Plots (or returns) a theoretical sample GIS spectrum*
<code>Ghost_plot_one</code>	Plots a GIS detector data with likely ghost regions*
<code>gis_calib</code>	Applies calibration factors to GIS spectra.*
<code>gis_plot</code>	Summarise data from one GIS detector*
<code>gis_smooth</code>	Smooths GIS spectra to 'remove' fixed patterning*
<code>gis_write_calib</code>	Prints calibration factors for GIS spectra*
<code>gt_bimage</code>	Return wavelength band integrated image.
<code>gt_spectrum</code>	Returns one-dimensional spectrum at specified point
<code>gt_windata</code>	Return block of detector data from one detector window.
<code>gt_windesc</code>	Return descriptor structure for one detector window
<code>integral_calc</code>	To compute the atomic data integral for use in column or volume emission measure work.
<code>mk_cds_map</code>	Make an image map from a CDS QL structure
<code>pix2wave</code>	Calculate CDS wavelength given a detector pixel location.
<code>plot_map</code>	Plot an image map
<code>readcdsfits</code>	Read and return the contents of a CDS FITS level-1 file.
<code>show_struct</code>	Display contents and breakdown of an IDL structure
<code>tftd</code>	Search for a string in header documentation
<code>utplot</code>	Plot X vs. Y with Universal time labels on bottom X axis.
<code>wave2pix</code>	Calculate detector pixel location given a CDS wavelength.
<code>xcat</code>	widget interface to CDS AS-RUN catalogue
<code>xcds_snapshot</code>	Widget interface to CDS_SNAPSHOT
<code>xdoc</code>	Front end to online documentation software.

## Appendix B. Example Program

Below is an example program that combines a series of exposures, and plots the total spectrum. The count rates within a section of the spectra are then plotted as an image.

```
pro example, study, detno, lam1, lam2
;
; study: study filename (string)
; detno: detector number, from 1 to 4
; lam1, lam2: wavelength range to integrate over (in Angstroms)
;
; e.g. to pick the 277A line in detector 2 in study s7060r00 type:
; IDL> example, 's10716r00', 2, 276, 278
;
; erb@mssl.ucl.ac.uk, 17 Feb 1999
;
print, 'Reading ' + study, + ', detector number', + detno
print, 'Plotting between wavelengths ', lam1, lam2
;
qlds= readcdsfits (study) ; read in the qlds
ghost_buster, qlds ; ghost free areas
gis_calib, qlds, /arcsec2_cm2 ; calibrate
cds_fill_missing, qlds ; guess missing data
;
; plot an average spectrum
;
window, 0
ghost_plot_one, qlds, detno, /angstroms
;
; Take a slice of the data
;
int_spec = gt_bimage(qlds, lam1, lam2, xsolar=xsolar, $
ysolar=ysolar)
;
spec_id = 'GIS' + trim(detno)
print, 'Pixels chosen:', wave2pix(spec_id, [lam1, lam2])
print, 'Bottom left position on the sun:', $
min(xsolar), min(ysolar)
;
; plot the image
;
window, 1
loadct, 0
tvsc1, congrid(int_spec, 512, 512)
;
return
end
```

## Appendix C. Ghost free line lists

Below are tables of expected count rates per second for lines, for slit 1 (2"x2"), in ghost-free areas in the four GIS detectors. The lists are originally from the CDS Scientific Report ('Blue Book') by Richard Harrison, Tables 2.1-2.6 and 3.3-3.7 with lines in ghosted areas removed. Note that this list is specifically for the ghost free areas of observations taken with GSET 41 (quiet Sun, 2" \_2" slit), although the positions of ghosting are similar for all the GSETs. A detailed check of ghosting in an observation can be made with 'ghost\_plot\_one, qlds, detno'

### DETECTOR 1

Ion	Wavelength (Å)	Quiet Sun Intensity
Ni XIV	164.13	0.4
Ar X	165.49	0.0
Fe VIII	167 - 168	2.0
Ni XIV	169.68	0.0
Fe IX	171.07	8.8
O V	172.17	0.1
O VI	172.93 - 173.08	0.7
O VI	183.94 - 184.11	0.4
Fe X	184.54	2.2
Fe XI	184.79	0.2
Fe VIII/Ni XVI	185.22	0.8
Fe VIII	186.60	0.5
Fe XII	186.88	2.4
Fe VIII	187.23	0.0
Fe XI	188.22	8.2
S XI	188.67	0.3
Fe X	190.04	1.3
Fe XIII/SXI	191.26	0.3
Fe XIII	200.02	0.7
Fe XII/XIII	201.12	2.0
Fe XIII	202.04	0.9
S VIII	202.61	0.4
Fe XIII	203.79	3.7
Fe XIII	204.26	0.0
Fe XIII	204.94	0.6
S X	208.32	0.0
Fe XIII	208.68	0.0
Fe XIII	209.62	0.0

Si VIII / Si XII	214.76 - 215.15	0.8
Si VIII	216.90	0.6
Fe XII	217.27	0.0
S XII	218.18	0.4
Fe XIV	219.12	2.0
Fe XIV	220.08	2.6

### DETECTOR 2

Ion	Wavelength (Å)	Quiet Sun Intensity	Active Sun Intensity
Si X	261.06	1.5	
Fe XVI	262.98	1.3	
Fe XIV	274.20	5.7	
Si VII	275.37 - 275.76	0.5	
Mg VII	276.15	0.2	
Si VII / Si VII	276.77 - 276.85	0.2	
Mg VII / Si X	277.04 - 277.27	2.3	
Mg VII / Si VII	278.40	0.5	
S XI	281.42	0.5	
S XI	281.83	0.0	
S XII	299.50	0.2	
Fe XI	308.54	0.0	0.0
Fe XIII	318.14	0.9	2.9
Si VIII	319.83	1.3	2.5
Fe XIII	320.80	3.5	11.5
Fe XV	321.78	2.3	7.9
Fe XV	327.02	2.4	8.0
Fe XII	338.26	1.2	4.2

Note that the He II and blended Si XI lines at 304Å are not included because the intensity of these excessively bright lines cannot be calibrated.

### DETECTOR 3

Ion	Wavelength (Å)	Quiet Sun Intensity	Active Sun Intensity
Ne V	416.20	0.2	0.8
Fe XV	417.26	1.1	34.6
S XIV	417.60	3.9	32.4
C IV / Ca X	419.50 - 419.74	0.7	3.0

Mg IX	438.60	0.2	0.8
Mg IX	444.03	0.4	2.8
Ca XV	445.00	0.0	3.6
S XIV	445.70	0.2	3.0
Si IX * / Mg VII	450.73	0.0	0.0
Ne VII	465.22	4.4	36.1
Ca IX	466.23	0.0	2.7
Ne IV	469.80	0.3	1.2
Ne III	489.50	0.2	1.2

\* Second order line blend

#### DETECTOR 4

Ion	Wavelength (Å)	Quiet Sun Intensity	Active Sun Intensity
N III	685.83	1.1	2.2
C II	687.20	0.3	0.5
Ca IX	693.80	0.3	1.6
Al III	696.00	0.0	0.1
Mg IX * / Ne I	735.90	0.4	2.6
O V	758.60	0.4	2.1
O V	759.44	0.0	0.0
O V	760.40	0.9	4.4
O V	762.00	0.0	0.0
N III	764.00	0.0	0.0
Ne VIII	780.30	1.4	16.1
S XI	782.00	0.0	0.2
Mg VIII	783.00	0.2	0.7

\* Second order line blend

#### BRIGHTEST GHOST FREE LINES VISIBLE WITH THE GIS

The Si XI 303Å and He II 304Å line are not included because the intensity of these excessively bright lines cannot be calibrated.

Ion	Wavelength (Å)	Quiet Sun Intensity(counts/second)
Fe IX	171.07	8.8

Fe XI	188.22	8.2
Fe XIII	203.79	3.7
Fe XIV	220.08	2.6
Fe XIV	264.79	4.2
Fe XIV / Si VII	274.20	5.7
Fe XV / Al IX	284.16	22.3
Fe XIII	320.80	3.5
S XIV	417.60	3.9
Ne VII	465.22	4.4

**GIS PRIME DENSITY SENSITIVE PAIRS IN GHOST-FREE REGIONS**

Note the lines marked with 'n' are NIS lines

Ion	Wavelengths (Å)
Mg VII	280.74 / 278.40
S XI	215.93 / 281.42
S XI	190.37 / 281.42
S XI	217.64 / 281.42
Fe XII	338.27 / 364.47n
Fe XIII	202.04 / 200.02
Fe XIII	201.12 / 200.02
Fe XIII	202.04 / 203.79
Fe XIII	318.12 / 320.80
Fe XIII	320.80 / 348.18n
Fe XIII	318.12 / 348.18n

**GIS GHOST-FREE TEMPERATURE SENSITIVE LINE RATIOS**

Note that 'n' is for NIS lines and 's' for SUMER lines

Ion and Ratio	Range of Sensitivity (K)
O V 172.17 / 629.73n	< 1.0 x 10 <sup>6</sup>
O VI 173 / 1032s	< few x 10 <sup>6</sup>
O VI 184 / 1032s	< few x 10 <sup>6</sup>

Ne V 416.20 / 569.20n	$10^5$ - $10^6$
Ne V 416.20 / 572.20n	$10^5$ - $10^6$

**GHOST-FREE LINES IN THE GIS RANGE SUITABLE FOR A DIFFERENTIAL EMISSION MEASURE ANALYSIS**

Ion	Wavelength (Å)	LogT(K)
Ne VII	465.22	5.7
Ne VIII	780.30	5.8
Si VII	275.37	5.8
Si VIII	319.83	5.9
Si X	261.06	6.0
Fe VIII	168.18	5.6
Fe IX	171.07	6.0
Fe XI	188.22	6.1
Fe XIII	203.79	6.2
Fe XIV	220.08	6.3
Fe XVII	200.80	6.5

**GHOST-FREE GIS LINES SUITABLE FOR HIGH TEMPERATURE STUDIES**

Ion	Wavelength (Å)
Fe XXIII	154.27
Fe XXIII	166.74
Fe XXIII	173.31
Fe XXIV	192.02
Fe XXIII	263.76
Fe XXII	217.30