# product data

# Ultimate Time & Frequency calibration & analysis

- Fast: 8000 measurements/s
- High resolution: 1ps (time) 11 digits/s (freq.), 0.001° (phase)
- Rubidium stability: 0.0001 ppm
- High trigger resolution: 1.25 mV
- Advanced arming/hold-off
- Modulation Domain Analysis SW
- EMC-immunity for noisy environments
- Ideal for fast test systems, R&D and calibration laboratoires
- 8 GHz option for microwave IRF testing

### Frequency calibration

The CNT-81 and CNT-81R can directly calibrate any application specific frequency up to 8 GHz. They are ideal for calibrating e.g. the timebase oscillator of other instruments, like frequency counters and synthesisers.

The Rubidium timebase of CNT-81R allows frequency calibration of even the highest possible specified oven oscillators. For a total uncertainty of  $10^{-10}$ , just connect the unknown frequency to the counters input and wait for a second.

Each individual 1s-measurement has a  $5x10^{-11}$  resolution. The built-in statistics averaging improves resolution further, and the std dev indicator gives added information about the stability of the unknown frequency.

### Time Interval calibration

For the calibration of time-intervals the CNT-81 provides leading performance due to the fast 50 ps single shot time resolution (1 ps averaged) and the high trigger level resolution of 1.25 mV.

The systematic start-stop channel difference is only 500 ps, which can be further reduced by calibrating the input channel difference.

### Phase calibration

With CNT-81 you can measure phase differences on signals of up to 160 MHz with a



**CNT-81 & CNT-81R** 

**Timer/Counter/Calibrators** 

With the CNT-81 series of counters and analyzers, Pendulum now offers the ultimate tools for measurement, analysis and calibration of Frequency, Time Interval or Phase, whether in test systems, on the R&D bench, in the calibration lab or out in the field (portable calibration). The series comprises 2 models; the ultra-high performance CNT-81 and the ultimate CNT-81R including a built-in Rubidium time-base reference.

Selection Chart	CNT-81	CNT-81R
Frequency, burst, time interval, phase, Vp-p	•	•
Frequency range (standard)	300 MHz	300 MHz
Frequency resolution (1s gate time)	11 digits	11 digits
Time interval resolution (single/average)	50/1 ps	50/1 ps
Vp-p (and trigger level) resolution	1.25 mV	1.25 mV
Arming/Hold-off delay by time and events	•	•
Hold-off resolution	10 ns	10 ns
Best timebase stability/month	$3x10^{-9}$	$5x10^{-11}$
No. of 10 MHz +5 MHz reference outputs	1+0	6+1
Measurement speed: GPIB to internal memory	250/s 8 k/s	250/s 8 k/s
Statistics calc .: mean, std, dev. and max/min	•	•
TimeView Documenting and Analysis SW	•	•
2.7 GHz HF-input	Option 10	Option 10
8 GHz RF-input	Option 13	Option 13

resolution better than 0.01° (below 30 MHz). This gives you outstanding resolution in measurements like laser positioning and calibration of phase meters. Calibration procedures exist that provide outstanding accuracy, with an uncertainty below 0.1°.

### Ideal for fast test systems

In manufacturing test systems two things are important; EMC-immunity and speed. CNT-81 offers excellent EMC-shielding and the highest throughput for any commercially available counter. The speed is impressive 8000/s to internal memory, and 250/s for individually triggered measurements via GPIB. Up to 20 complex measurement set-ups can be locally stored in the counter's non-volatile set-up memory and instantly recalled via a short bus command. This enables new measurement tasks to be executed one after the other at a very-high rate. A complete cycle "setup-measure-transfer" takes less than 8 ms.

The two counters comply of course to SCPI, which facilitates easy updating of new test hardware without the penalty of time-consuming SW-rewriting.

> pendulum Incorporating XL Microwave

# Modulation Domain Analysis

The analysis PC-SW TimeView converts the CNT-81/CNT-81R to a high performance modulation domain analyzer. In the modulation domain you can view rapid frequency changes vs. time, e.g. modulation, sweep, frequency setting, channel hopping etc.

The 16-bit DOS program is standardly included with all CNT-81/81R.

The 32-bit Windows program is an optional accessory (option 29).




Figure 1: The modulation domain (f vs. t) complements the time (V vs. t) and the frequency (V vs. f) domains






Figure 2: Jitter (rms and peak-peak) and noise is quantified in distribution histograms.




Figure 4: Linearity of frequency sweep can be verified in the modulation domain (frequency vs. time).

----




Figure 6: The 9 different pulse width clusters, corresponding to the 9 different pit lenghts (T3-T11) in a CD-recording.





Figure 3: The FFT-diagram reveals the modulation frequency, whether intended or unwanted.



-------

frequency switching time of approx. 10.7 us.

Figure 7: Zoom in on T3-cluster displays an rms-jitter of 13 ns, which is OK for an audio CD.

# CNT-81 & CNT-81R Specifications

## Measuring Modes

Measuring Modes	
Inputs A and B can be swappe Time.	d internally in all modes except Rise and Fall
Frequency A, B, C	
Range:	
Input A:	up to 300 MHz
Input B:	up to 100 MHz
Input C (option 20):	100 MHZ to 2,7 GHz
Input C (option 13):	300 MHz to 8 GHz
Resolution:	11 digits in 1s measuring time
Frequency Burst A, B, C	
Frequency and PRF of burst si (Ch. C) can be measured with <b>Period A</b>	gnals down to 1 $\mu$ s (CH. A and B) or 50 $\mu$ s out external control signals.
Range:	$3.3 \text{ ns to } 10^{10} \text{s}$
Resolution:	11 digits in 1s measuring time
Frequency Ratio A/B, C/B	
Range:	$10^{-9}$ to $10^{15}$
Time Interval A to B	
Range:	0 ns to $10^{10}$ s
Resolution:	
Single shot:	50 ps (1 ps average)
Pulse Width A	
Range:	3 ns to $10^{10}$ s
Rise and Fall Time A	
Range:	3 ns to 10 <sup>10</sup> s
Phase A Relative B	
Range:	-180 to +360°
Resolution:	0,01°
Duty Factor:	
Range:	0.000001 to 1.000000
Totalize A, B	12 10
Range:	0 to $10^{17}$ , 0 to $10^{10}$ in A-B modes
Modes:	A Gated by B
	A Start/Stop by B Manual gating A minus B
	Time gating A minus B
V max, V min, Vp-p A, B	5 5
Range:	-50V to + 50V
Frequency Range:	up to 100 MHz
Resolution:	1.25 mV
Inputs and Outputs	
Inputs A and B	
Coupling:	AC or DC
Impedance:	$1 \text{ M}\Omega/15 \text{ pF or } 50\Omega/(\text{VSWR} \le 2:1)$
Max. Channel Timing	500
Difference: May Sanaitivity	500 ps
Max. Sensitivity:	20 mV rms, <100 MHz x1or x10
Attenuation: Var. Hysteresis A:	30 mVp-p to 10Vp-p hup to 120 MHz
Triggerpegel:	read-out on display
Range:	(x1):-5V to $+5V$
	(x10): -50V to +50V (x10): -50V to +50V 1.25 mV
Resolution (x1):	
AUTO-Trigger Level:	Trigger level is automatically set to 50% point of input signal (10% and 90% for Rise/Fall Time, 75% and 25% for variable hysteresis A)
Min. Frequency:	Settable from 1 Hz and upwards. Default=100 Hz
Low Pass Filter A:	100 kHz
Digital LP Filter:	1 Hz to 10 MHz using trigger Hold-Off

Input C (Option 20)	
Frequency Range:	100 MHz to 2.7 GHz
Operating Input Voltage	
<i>Range:</i> 0.1 to 0.3 GHz:	20 mV rms to 12V rms
0.3 to 2.5 GHz:	10 mV rms to 12V rms
2.5 to 2.7 GHz:	20 mV rms to 12V rms
Impedance:	$50\Omega$ nominal, (VSWR<2.5:1)
Max Voltage Without	12V rms during 60s, PIN-diode protected
Damage:	
Connector:	N-type, female
Input C (Option 13)	
Operating input voltage	
<i>range:</i> 0.3 to 0.5 GHz	-21 to +30 dBm (20 mV rms to 7V rms)
0.5 to 3.0 GHz	-27  to  +30  dBm (20  mV rms to  7  V rms)
3.0 to 4.5 GHz	-21 to $+30$ dBm (20 mV rms to 7V rms)
4.5 to 6.0 GHz	-15 to $+30$ dBm ( $40$ mV rms to $7V$ rms)
6.0 to 8.0 GHz	- 9 to +30 dBm (80 mV rms to 7V rms)
Impedance:	$50\Omega$ nom. VSWR<2:1
Connector:	N-type, female
Rear Panel Inputs and O	· 1 /
Reference Input:	1, 2, 5 or 10 MHz>200mV rms
Reference Output:	, ,
CNT-81:	$1x10$ MHz>0.5V rms sinewave into $50\Omega$ load
CNT-81R:	6x10 MHz; 1x5 MHz>0.6V rms sinewave into 50Ω load
Arming Input:	Most mesuring functions can be performed using arming
Gate Output:	Gate open/gate closed signal
Trigger Level Outputs:	Outputs for channel A and B trigger levels
Probe Comp. Outputs:	Outputs for channel A and B to adjust for best pulse response when using probes for counter inputs
Analog Output:	0 to 4.98V in 20 mV steps; proportional to 3 se- lected display digits
<b>Auxiliary Functions</b>	
Trigger Hold Off Time Delay Range:	60 ng to $1.24$ g, $10$ ng resolution
Event Delay Range B:	60 ns to 1.34s, 10 ns resolution 2 to $2^{24}$ -1, max. 100 MHz
External Arming	2 to 2 -1, max. 100 WHIZ
Time Delay Range B, E:	200 ns to 1.6s, 100 ns resolution
Event Delay Range B:	$2 \text{ to } 2^{24}$ -1, max. 20 MHz
Statistics	2 to 2 1, max. 20 WHZ
Functions:	Maximum, Minimum, Mean and Standard Deviation
Sample Size:	
Mathematics	1 to $2 \times 10^{-9}$ samples
Functions:	1 to $2x10^{\circ}$ samples
runctions:	1 to 2x10 <sup>°</sup> samples (K*X+L)/M and (K/X+L)/M, X is urrent reading
	-
Other Functions	(K*X+L)/M and (K/X+L)/M, X is urrent reading and K, L and M are constants; set via keyboard or as frozen reference value ( $X_0$ ) or as value from pre- ceding measurement ( $X_{n-1}$ ).
	(K*X+L)/M and (K/X+L)/M, X is urrent reading and K, L and M are constants; set via keyboard or as frozen reference value ( $X_0$ ) or as value from pre- ceding measurement ( $X_{n-1}$ ). Single cycle, 80, 160, 320, 640, 1280 ns and 20 $\mu$ s to 20s (to 400s for some functions)
Other Functions	(K*X+L)/M and (K/X+L)/M, X is urrent reading and K, L and M are constants; set via keyboard or as frozen reference value ( $X_0$ ) or as value from pre- ceding measurement ( $X_{n-1}$ ). Single cycle, 80, 160, 320, 640, 1280 ns and 20

20 instrument setups can be saved and recalled from internal non-volatile memory. 10 can be user protected.

10-digit LCD with high-luminance back-light

Display:

# CNT-81 & CNT-81R Specifications

### **GPIB** Interface

Max Measurement Rate*	
Via GPIB:	250 readings/s
To Internal Memory:	8k readings/s
Time Stamping:	125 ns resolution
Back-to-back-Period:	Up to 40k readings/s (100 ns resolution)
Internal Memory Size*: Data Output:	Up to 6100 readings ASCII, IEEE double precision floating point

## TimeView<sup>™</sup> Time & Frequency Analyse Software

TimeView is supported on bot	h CNT-81 and CNT-81R models.
Versions:	
DOS-version:	Standardly supported
Windows (32 bit) version	: Optional accessory (option 29)
Data Capture Modes and Measurement Rate*	
Free-run sampling:	8k readings/s
Repetitive Sampling:	Up to 10 MSa/s
Back-to-back-Period:	Up to 40k readings/s
Waveform Capture:	Yes (vertical sampling)
Instrument Control:	All front panel functions and some AUX MENU functions
Data Analysis:	Measurement data vs time FFT Graph Root Allan Variance Smoothing function Zoom function Cursor measurements Distribution Histogram
File Storage:	Setup and Measurement data

\* Depending on measurement function and internal data format.

## **Time Base Options**

	1	1	1	1
Model:	CNT-81	CNT-81	CNT-81	CNT-81R
Option: Stability:	Standard UCXO	Option 30 OCXO	Option 40 OXCO	- Rubidium
Ageing:				
per month per year per 10 years	$<5x10^{-7}$ $<5x10^{-6}$ n.s.	<1x10 <sup>-8</sup> <7.5x10 <sup>-8</sup> n.s.	<3x10 <sup>-9</sup> <2x10 <sup>-8</sup> n.s.	$<5x10^{-11*}$ $<2x10^{-10}$ $<1x10^{-9}$
vs. temp:				
0°C-50°C 20°C-26°C (typ.)	<1x10 <sup>-5</sup> <3x10 <sup>-6</sup>	<5x10 <sup>-9</sup> <6x10 <sup>-10</sup>	$<2.5 x 10^{-9}$ $<4 x 10^{-10}$	$<3x10^{-10}$ $<2x10^{-11}$
Short Term:				
τ=1s(Allan Dev.)	n.s	1x10 <sup>-11</sup>	5x10 <sup>-12</sup>	$5x10^{-11}$
Warm-up Stability: after warm-up time of:	n.s. 30 min.	<1x10 <sup>-8</sup> 10 min.	<5x10 <sup>-9</sup> 10 min.	<4x10 <sup>-10</sup> 10 min.
<b>Total Uncertainty (25):</b> (20°C -26°C) 1 year after calibration 2 years after calibration	<7x10 <sup>-6</sup> <1.2x10 <sup>-5</sup>	<1x10 <sup>-7</sup> <2x10 <sup>-7</sup>	<2.5x10 <sup>-8</sup> <5x10 <sup>-8</sup>	<2.5x10 <sup>-10</sup> <5x10 <sup>-10</sup>

\* After 1 month of continuous operation.

## **General Specifications**

Environmental Data     Operating Temp:   0°C to 50°C     Storage Temp:   -40°C to 70°C     Safety:   CSA 22.2 Nr. 231, EN 61010-1, Cat. II	
Storage Temp:   -40°C to 70°C     Safety:   CSA 22.2 Nr. 231, EN 61010-1, Cat. II	
Safety: CSA 22.2 Nr. 231, EN 61010-1, Cat. II	
pollution degree 2, CE	
EMC: EN 5501 1 ISM Group 1, Class B;	
EN 50082-2; FCC Part 15J Class A, CE	
Power Line Requirements (at 25°C)	
AC Voltage:	
CNT-81: 90 to 265V rms, 45 to 440 Hz	
CNT-81R: 90 to 265V rms, 45 to 440 Hz	
Power Rating: CNT-81: Max. 35W	
CNT-81R: Max. 35W (6 min. warm-up);	
Max. 100 w (6 min. wain-up), Max. 47W (cont. operation)	
Mechanical Data	
<i>WxHxD:</i> 315x86x395 mm (12.4x3.4x15.6 in)	
Weight:	
CNT-81: Net $4 \text{ kg} (8.5 \text{ lb})$	
Shipping 7 kg (15 lb)	
CNT-81R: Net 4.8 kg (10.5 lb)	
Shipping 7.8 kg (16.8 lb)	
Ordering Information	
Basic models	
<i>CNT-81</i> Timer/Counter/Analyzer 300 MHz/50 ps.	inal
Standard timebase (5x10 <sup>-7</sup> /Month) and	mer.
Time&Frequency Software TimeView fo	r DOS
CNT-81R Timer/Counter/Calibrator 300 MHz/50 ps	s, incl.
Rubidium timebase ( $5x10^{-11}$ /Month) and	DOG
Time&Frequency Software TimeView fo	r DOS
Time&Frequency Software TimeView for Included with Instrument Power line cord	r DOS
Time&Frequency Software TimeView fo	r DOS
Time&Frequency Software TimeView for     Included with Instrument   Power line cord     Users documentation on CD-rom   Certificate of Calibration	r DOS
Included with Instrument   Time&Frequency Software TimeView for     Included with Instrument   Power line cord     Users documentation on CD-rom   Vertice	r DOS
Included with Instrument Time&Frequency Software TimeView for   Included with Instrument Power line cord   Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)*   Option 13: 8.0 GHz Input C (CNT-81/81R)	r DOS
Included with Instrument Time&Frequency Software TimeView for   Included with Instrument Power line cord   Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)*   Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)	r DOS
Included with Instrument Time&Frequency Software TimeView for Power line cord Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)* 8.0 GHz Input C (CNT-81/81R)   Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options 5.0 GHz Input C (CNT-81/81R)	r DOS
Included with Instrument Time&Frequency Software TimeView for Power line cord Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)* 0ption 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options (CNT-81)*	r DOS
Included with Instrument Time&Frequency Software TimeView for Power line cord Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)* 8.0 GHz Input C (CNT-81/81R)   Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options 5.0 GHz Input C (CNT-81/81R)	r DOS
Included with Instrument Time&Frequency Software TimeView for Power line cord Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)* Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options (CNT-81)* Very-high stability Oven Time Base (1x10*8/Monat)   Option 40: Ultra-high stability Oven Time Base	r DOS
Time&Frequency Software TimeView forIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*8.0 GHz Input C (CNT-81/81R) Coption 20:Option 20:2.7 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10-8/Monat)Option 40:Ultra-high stability Oven Time Base (5x10-9/Month)	r DOS
Included with Instrument Power line cord   Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)*   Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options (CNT-81)* Very-high stability Oven Time Base (1x10 <sup>-8</sup> /Monat)   Option 40: Ultra-high stability Oven Time Base (5x10 <sup>-9</sup> /Month)   Optional Accessories* Very-Month	r DOS
Time&Frequency Software TimeView for   Included with Instrument Power line cord   Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)*   Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options (CNT-81)* Very-high stability Oven Time Base (1x10 <sup>-8</sup> /Monat)   Option 40: Ultra-high stability Oven Time Base (5x10 <sup>-9</sup> /Month)   Option 11: Rear Panel Inputs	r DOS
Included with Instrument Frequency Software TimeView for Power line cord Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)* Option 13:   Option 10: 2.7 GHz Input C (CNT-81/81R)   Time Base Options (CNT-81)* Option 30:   Option 30: Very-high stability Oven Time Base (1x10 <sup>-8</sup> /Monat)   Option 40: Ultra-high stability Oven Time Base (5x10 <sup>-9</sup> /Month)   Option 11: Rear Panel Inputs   Option 22: Rack-Mount Kit	r DOS
Time&Frequency Software TimeView for   Included with Instrument Power line cord   Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)*   Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options (CNT-81)* Very-high stability Oven Time Base (1x10 <sup>-8</sup> /Monat)   Option 40: Ultra-high stability Oven Time Base (5x10 <sup>-9</sup> /Month)   Option 11: Rear Panel Inputs   Option 22: Rack-Mount Kit   Option 27: Carrying Case	r DOS
Time&Frequency Software TimeView forIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*8.0 GHz Input C (CNT-81/81R)Option 13:8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*/Month)Option 11:Rear Panel Inputs Option 22:Option 21:Rear Panel Inputs Carrying Case Option 27H:	
Time&Frequency Software TimeView for   Included with Instrument Power line cord   Users documentation on CD-rom Certificate of Calibration   RF Input Frequency Options (CNT-81/81R)*   Option 13: 8.0 GHz Input C (CNT-81/81R)   Option 20: 2.7 GHz Input C (CNT-81/81R)   Time Base Options (CNT-81)* Very-high stability Oven Time Base (1x10 <sup>-8</sup> /Monat)   Option 40: Ultra-high stability Oven Time Base (5x10 <sup>-9</sup> /Month)   Option 11: Rear Panel Inputs   Option 22: Rack-Mount Kit   Option 27: Carrying Case	
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*8.0 GHz Input C (CNT-81/81R)Option 13:8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*/Month)Option 11:Rear Panel Inputs Option 22:Option 27:Carrying Case Option 29:Option 29:Time View for Windows 98/2000/XP/NT	31R
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)* Option 20:8.0 GHz Input C (CNT-81/81R)Option 20:2.7 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10* <sup>8</sup> /Monat)Option 40:Ultra-high stability Oven Time Base (5x10* <sup>9</sup> /Month)Option 11:Rear Panel Inputs Option 22:Option 27:Carrying CaseOption 27:Carrying CaseOption 29:TimeView for Windows 98/2000/XP/NT OM-81:Operators Manual (printed) for CNT-81/81PM-81:Programmers Manual (printed) for CNT-81/81	31R 81/81R R
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*8.0 GHz Input C (CNT-81/81R) 2.7 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*/Monat)Option 40:Ultra-high stability Oven Time Base 	31R 81/81R R uments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*8.0 GHz Input C (CNT-81/81R)Option 13: Option 20: (CNT-81)*8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*/Monat)Option 40: Option 11: Option 22: 	31R 81/81R R iments iruments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*So GHz Input C (CNT-81/81R)Option 13:8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*/Month)Option 11:Rear Panel Inputs Option 22:Option 27:Carrying Case Option 27H:Option 29:TimeView for Windows 98/2000/XP/NT OM-81:OM-81:Operators Manual (printed) for CNT-81/81 NI 778416-01:GPIB-USB interface from National Instrum NI 778209-0:Generation GPIB-ENET interface from National Instrum	81R 81/81R R intents ruments ruments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*So GHz Input C (CNT-81/81R)Option 13:8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*8/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*9/Month)Option 11:Rear Panel Inputs Option 22:Option 27:Carrying Case Option 27H:Option 29:TimeView for Windows 98/2000/XP/NT Operators Manual (printed) for CNT-81/811NI 778416-01:GPIB-USB interface from National Instr NI 778209-0:GPIB cable type X2 (1m) from National Instr NI 763061-01:	31R 81/81R R iments ruments ruments ruments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*So GHz Input C (CNT-81/81R)Option 20:2.7 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*8/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*9/Month)Option 11:Rear Panel Inputs Option 22:Option 27:Carrying CaseOption 29:TimeView for Windows 98/2000/XP/NT Operators Manual (printed) for CNT-81/8NH 778416-01:OPERATOR Manual (printed) for CNT-81/8NI 778209-0:GPIB-USB interface from National Instrum NI 763061-01:GPIB cable type X2 (2m) from National Instrum NI 763061-02:Gene Additional Actional Instrum Option 22:	31R 81/81R R inuments rruments rruments rruments
Time&Frequency Software TimeView forIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*So GHz Input C (CNT-81/81R)Option 13:8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*8/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*9/Month)Option 11:Rear Panel Inputs Option 22:Option 27:Carrying Case Option 27:Option 29:TimeView for Windows 98/2000/XP/NT OM-81:OM-81:Operators Manual (printed) for CNT-81/81R)NI 778034-0:PCMCIA-GPIB interface from National Instruct NI 778209-0:NI 763061-01:GPIB cable type X2 (2m) from National Instruct NI 763061-03:	31R 81/81R R inuments rruments rruments rruments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)* Option 20:2.7 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10* <sup>8</sup> /Monat)Option 30:Very-high stability Oven Time Base (5x10* <sup>9</sup> /Month)Option 40:Ultra-high stability Oven Time Base (5x10* <sup>9</sup> /Month)Option 22:Rack-Mount Kit Option 27:Option 27:Carrying Case Option 29:Option 29:TimeView for Windows 98/2000/XP/NT Operators Manual (printed) for CNT-81/81NI 778416-01:GPIB-USB interface from National Instrument NI 778034-0:NI 778209-0:GPIB cable type X2 (2m) from National Instrument NI 763061-02:WarrantyGPIB cable type X2 (3m) from National Instrument Warranty	31R 81/81R R inuments rruments rruments rruments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)* Option 20:8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Yery-high stability Oven Time Base (1x10*/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*/Month)Option 11:Rear Panel Inputs Option 22:Option 22:Rack-Mount Kit Option 27:Option 27:Carrying Case Option 27:Option 29:TimeView for Windows 98/2000/XP/NT OM-81:OM-81:Operators Manual (printed) for CNT-81/81NI 778034-0:PCMCIA-GPIB interface from National Inst NI 7780361-01:NI 763061-02:GPIB-ENET interface from National Inst NI 763061-02:Warranty Standard18 months	31R 81/81R R inuments rruments rruments rruments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)*S.0 GHz Input C (CNT-81/81R)Option 13: Option 20: (CNT-81)*S.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Very-high stability Oven Time Base (1x10*/Monat)Option 30: Option 40: Option 11: Option 22: Option 27: Option 27: Option 27: Option 27: Carrying Case Option 29: Option 29: TimeView for Windows 98/2000/XP/NT 	31R 81/81R R inuments rruments rruments rruments
Time&Frequency Software TimeView foIncluded with InstrumentPower line cord Users documentation on CD-rom Certificate of CalibrationRF Input Frequency Options (CNT-81/81R)* Option 20:8.0 GHz Input C (CNT-81/81R)Time Base Options (CNT-81)*Yery-high stability Oven Time Base (1x10*/Monat)Option 40:Ultra-high stability Oven Time Base (5x10*/Month)Option 11:Rear Panel Inputs Option 22:Option 27:Carrying Case Option 27H:Option 29:TimeView for Windows 98/2000/XP/NT OM-81:OM-81:Operators Manual (printed) for CNT-81/81NI 778034-0:PCMCIA-GPIB interface from National Inst NI 778209-0:GPIB-ENET interface from National Inst NI 763061-01:GPIB cable type X2 (2m) from National Inst NI 763061-03:Warranty Standard18 months	31R 81/81R R inuments ruments ruments ruments

Specifications subject to change without notice

4031 600 80101 rev. 05 August 2006

#### US: Pendulum Instruments Inc

5811 Racine Street Oakland, CA 94609-1519, USA Voice:(510)-428-9488 Fax: (510)-428-9469 *International: Pendulum Instruments AB* PO Box 20020, SE-16102 Bromma, Sweden Voice: +46 8 598 51057 Fax:+46 8 598 51040

### Pendulum Instruments www.pendulum-instruments.com

- Experts in time & frequency calibration, measurement and analysis

> pendulum Incorporating XL Microwave