BACTRACK® B70

BREATHALYZER

TRACKS BLOOD ALCOHOL CONTENT IN SECONDS



TABLE OF CONTENTS

Congratulations on your purchase of the BACTRACK® B70 Breathalyzer! Your new alcohol tester estimates Blood Alcohol Content in seconds. For your convenience, we've provided simple instructions to get you started.

| Introduction | | | | 1 |
|--------------------------------|----|---|-------|----|
| Dose-specific Effects of Alcoh | 10 | 1 | | 4 |
| Preparation | | | | 6 |
| Components Diagram | | | | 6 |
| Operation | | | | 7 |
| Specifications | | | • | 10 |
| Precautions | | | • | 11 |
| Verification and Calibration | | | | 12 |
| One-Year Limited Warranty | | | • | 13 |

Track your B.A.C. - BACTRACK!

INTRODUCTION

The BACTRACK® B70 Breathalyzer is a professional alcohol screening device, used for the detection of alcohol in the breath. The BACTRACK provides a digital result, indicating the approximate BAC (Blood Alcohol Content) of the test subject. The BACTRACK is powered by two AA batteries and is very easy to use. Simply turn on the unit and begin testing your alcohol level in a matter of seconds. However, before you begin testing, there are a few things you should know. Please take a moment to read your owner's manual in its entirety and familiarize yourself with the BACTRACK.

Alcohol and Its Effects on the Human Body

Alcohol is absorbed from the mouth, throat, stomach and intestines into the bloodstream

Alcohol ingested by the human body can be detected in the breath because of its presence in the bloodstream.

Alcohol cannot be digested, and it cannot be chemically changed in the bloodstream. As the blood flows through the lungs, alcohol in the bloodstream moves across the membranes of the lung's air sacs (alveoli) into the air. The concentration of the alcohol in the alveolar air is directly related to the concentration of the alcohol in the blood. As the alveolar air is exhaled, the alcohol in it can be detected by the breath alcohol testing device.

The alcohol concentration in the breath is related to that in the blood, and because of this, an individual's BAC can be determined by measuring alcohol in the breath. The ratio of breath alcohol to blood alcohol is generally estimated to be 2,100:1. Therefore, 2,100 milliliters (ml) of alveolar air will contain approximately the same amount of alcohol as 1 ml of blood.

Alcohol is a Drug

Alcohol is the chemical (ethanol or ethyl alcohol) resulting from the fermentation of grapes or grain. Alcohol is absorbed directly into your bloodstream. From your bloodstream, the fluids in your body tissues absorb the alcohol. Your brain is made up of a large concentration of fluids and will absorb a large amount of any alcohol you drink. Your liver eliminates the absorbed alcohol at its own rate of time and nothing you eat or drink can speed up the process. Your liver does this by oxidizing the alcohol (converting it into water and carbon dioxide). Coffee, food or any other "cure" will not sober you up. Only time will do that.

How does alcohol affect my body?

Alcohol is a depressant. It has a relaxing effect on the muscles in your body. The muscles of your eyes relax and lose focus. Your eyesight will become fuzzy and you may experience double vision. Brain activity is slowed. Your judgment, reflexes, and coordination are all negatively affected.

Some vision impairments that occur when you have been drinking include:

- Narrowing of your field of vision
- Reduction in your depth perception
- Decreased ability to see in darkness
- Increased sensitivity to glare and a longer time for your eyes to readjust from the glare

Some mental (brain) impairment can occur when you have been drinking, including:

- Reduced awareness of danger
- Becoming overly confident and reckless
- Difficulty in making decisions
- Reduction in balance
- Slowed reflexes
- Impaired judgment

How does alcohol affect my driving?

We can't emphasize enough how dangerous it is for you to drink and drive. Your vision and your brain are the most important factors in driving safely. If you jeopardize either by drinking, and then you drive, you are likely to be involved in a serious or fatal collision.

From California Department of Motor Vehicles (http://www.dmv.ca.gov)

DO/E-/PECIFIC EFFECT/ OF ALCOHOL

| BAC | Dose-Specific Effects |
|------------|--|
| 0.02-0.03% | Slight euphoria. Loss of shyness. Depressant effects are not apparent. Impairment possible in some individuals. |
| 0.04-0.06% | Feeling of well-being, relaxation, lower inhibitions and sensation of warmth. Euphoria. Some minor impairment of reasoning and memory. Lowering of caution. Driving skills may be impaired at this level of intoxication. |
| 0.07-0.09% | Slight impairment of balance, speech, vision, reaction time, and hearing. Euphoria. Judgment and self-control are reduced. Caution, reason and memory are impaired. Driving skills are always impaired at this level of intoxication and higher. |
| 0.10-0.12% | Significant impairment of motor coordination and loss of good judgment. Speech may be slurred. Balance, vision, reaction time and hearing will be impaired. Euphoria. |

| ВАС | Dose-Specific Effects | |
|--------------|--|--|
| 0.13-0.15% | Gross motor impairment and lack of physical control. Blurred vision and major loss of balance. Euphoria is reduced and dysphoria (anxiety, restlessness) is beginning to appear. | |
| 0.16-0.20% | Dysphoria predominates. Nausea may appear. | |
| 0.25% | Needs assistance in walking. Total mental confusion. Dysphoria with nausea and some vomiting. | |
| 0.30% | Loss of consciousness. | |
| 0.40% and up | Onset of coma. Possible death due to respiratory arrest**. | |

^{*} The effects of alcohol intoxication are greatly influenced by individual variations among users. Some users will be intoxicated at a much lower BAC than shown above.

The generally accepted legal standard for alcohol intoxication in the United States is 0.08%. However, your driving skills can be impaired at a level of 0.04% or lower. It is never safe to drink any amount of alcohol and drive.

^{**}Death can occur at a lower BAC in some individuals.


PREPARATION

Installing the Battery

Apply a small amount of pressure on the indentation of the battery compartment cover located on the back of the unit. Simultaneously push down and away from the unit to release the cover.

Install two AA batteries in the battery compartment. (Batteries are not included.)

COMPONENTS DIAGRAM



OPERATION

IMPORTANT!



Users must wait 20 minutes after eating, drinking, or smoking before blowing into the BACTRACK Breathalyzer. Failure to wait 20 minutes can provide inaccurate test results and damage the sensor.



STEP 1: Press and hold power for one full second.
To prevent the BACTRACK Breathalyzer from accidentally turning on, tapping the power button for less than one second will not cause it to turn on



STEP 2: After powering on, the unit will count down quickly from 50 to zero. This countdown takes approximately 10 seconds.



STEP 3: When the countdown approaches 10, begin to inhale a deep breath.







STEP 4: When the countdown reaches zero, the unit will flash "blow" and "0". Exhale through the dual-flow breath chamber for **five seconds** while holding it as close to your mouth as possible without actually touching the product to your mouth.

IMPORTANT: Keep your lips and mouth open wide while blowing, as indicated in the diagram below.

ENSURE PROPER BREATH SAMPLE



CORRECT

Ensure mouth is open as though trying to fog a mirror or saying, "Hhhhhhhha."



NOT CORRECT

Do not pucker lips together as though blowing out candles or whistling.

Important Breath Sampling Instructions

Be sure to exhale CONTINUOUSLY for at least FIVE FULL SECONDS. You must exhale hard. Be sure to exhale as soon as the countdown reaches zero.



STEP 5: When finished taking a breath sample, the BACTRACK will circle the outer LCD fields for three seconds while analyzing the breath sample.



STEP 6: Once the BluFire® sensor technology has analyzed the breath sample, the estimated BAC (Blood Alcohol Content) value is displayed.



The test result will be displayed for 10 seconds. Then the test result will blink for another 20 seconds before shutting off.

Error Notification



If a user does not exhale a sufficient air sample through the Dual-Flow tube, the BACTRACK will display ERR. Press power twice to retest and be sure to exhale CONTINUOUSLY for at least five seconds. This feature helps achieve accurate results by ensuring a deep lung air sample is taken.

Powering Off

To turn the power off at any time, press and hold the power button for one second.

8



SPECIFICATIONS

| Dimensions | 4.25 x 1.63 x 0.81 inches (10.8 x 4.1 x 2.1 cm) |
|--------------------------|--|
| Weight | 1.4 oz (39.7 grams) |
| Sensor Technology | BluFire [®] |
| Detection Range | 0.00 – 0.40 %BAC |
| Test Count | Displays total number of tests performed |
| Power Supply | Two AA batteries, not included |
| Battery Life | At least 300 tests |
| Warm Up Time | 10 seconds |
| Response Time | 3 seconds |
| Sensor Accuracy | +/- 0.01 %BAC @ 0.02 %BAC |
| Operating Temperature | 50-104 °F (10-40 °C) |

PRECAUTIONS

- Wait at least 20 minutes after drinking, eating or smoking before testing. Not observing this waiting period can cause inaccurate readings, damage the sensor, and void your warranty.
- Do not blow smoke, food, or liquids into the Dual-Flow Breath Chamber as this will damage the sensor.
- **3.** Do not test in areas with strong winds, smoke, or in an area where a large amount of alcohol is being consumed.
- 4. Replace the two AA batteries when the screen displays "BAT".
- Send your tester in for periodic calibration service as required. (See calibration).
- 6. The BACTRACK Breathalyzer is designed to be used in a temperature range of 50-104 °F (10-40 °C). Operation of the unit in temperature ranges above or below this range may affect the accuracy of results.
- Avoid testing in the presence of any substances that contain methyl alcohol, isopropyl alcohol or acetone. These substances may interfere with the results of the test.
- 8. Diabetes, a low caloric intake diet, or certain health conditions may cause a false positive test when there is an increased keytone concentration in the breath. Consult with a licensed physician to see if this may apply to you.
- Do not use the BACTRACK as a tool to determine whether you should operate a motor vehicle or equipment, or perform any other dangerous act.
- 10. Do not drink and drive. Always have a designated driver when alcohol is being consumed.

VERIFICATION AND CALIBRATION

The BACTRACK Breathalyzer is calibrated during manufacture using advanced alcohol simulation equipment. Known alcohol concentrations are passed through the sensor to set baseline values for testing. During testing, the unit compares users' breath samples to these baseline values.

The accuracy of breath alcohol testers can fluctuate after six to twelve months of normal use, depending on the number of tests performed and operating conditions.

If this product is providing inconsistent test results, is not providing any test results, or provides unusually high or low test results, the product should be recalibrated by a trained professional.

The product should be recalibrated at least every six to twelve months. If the product is used every day, it may need to be recalibrated as frequently as every month.

If the power button is depressed for 5 seconds, the product will display the total number of tests taken since the last calibration. If the total number of tests taken is 300 or greater, "CAL" is quickly displayed as a reminder that the product may need to be recalibrated in the near future to maintain accuracy.

Calibration Instructions

Please visit www.bactrack.com or call 877-334-6876 for information on how to get your BACTRACK Breathalyzer recalibrated.

During recalibration, a professional technician will reset the sensor baseline values and also perform a full diagnostic check to ensure the accuracy and reliability of the product.

ONE-YEAR LIMITED WARRANTY

KHN Solutions' warranty obligations for this product are limited to the following:

KHN Solutions warrants this product against defects in materials and workmanship under normal use for a period of ONE (1) YEAR from the date of retail purchase by the original end-user purchaser ("Warranty Period"). Except as provided in the preceding sentence or applicable law, KHN Solutions provides this product "as is". If a defect arises and a valid claim is received by KHN Solutions within the Warranty Period, at its option, KHN Solutions will either (1) repair the product, (2) exchange the product with a product that is new or which has been manufactured from new or serviceable used parts and is at least functionally equivalent to the original product, or (3) refund the purchase price of the product. When a refund is given, the product for which the refund is provided must be returned to KHN Solutions and becomes KHN Solutions' property. This warranty is limited

to the original end-user purchaser and is not transferable to, or enforceable by, any subsequent owner.

Exclusions, Limitations, and Disclaimers

KHN Solutions makes no warranties, express or implied, as to the ability of this product to determine whether, or the extent to which, a test subject's mental or physical functioning, or judgment, may be impaired by blood alcohol including whether the test subject is intoxicated under any definition of that word. KHN Solutions expressly disclaims any liability for direct, indirect, incidental, special, or consequential damages of any nature under any legal theory. Any act or failure to act based on a reading from this product shall be at the test subject's or user's own risk. KHN Solutions assumes no responsibility for consequences to or of test subjects who use this product and later are shown to have been under the influence of alcohol or to have had their judgment or any mental or bodily function impaired by alcohol. Correlation between breath alcohol content and blood alcohol concentration depends on many variables, including environmental factors (such as air quality, wind, humidity, temperature, etc.) and health conditions of the test subject. A low BAC reading does not mean that the test subject's physical or mental performance or judgment (including reaction times) can respond to an emergency. The concentration of alcohol in the blood of a test subject cannot be exactly determined by using a breath alcohol screening product.

KHN Solutions does not warrant that the operation of the product will be error-free. KHN Solutions is not responsible for any consequences arising from failure to follow instructions or relating to the product's use. This warranty does not apply: (a) to consequences caused by accident, abuse, tampering, misuse, flood, fire, earthquake or other external causes; (b) to consequences caused by operating the product outside the permitted or intended uses described by KHN Solutions; (c) to consequences caused by service performed by anyone who is not a representative of KHN Solutions or a KHN Solutions Authorized Service Provider; or (d) to a product or part that has been modified to alter functionality or capability without the written permission of KHN Solutions.

TO THE EXTENT PERMITTED BY LAW, THIS WARRANTY AND THE REMEDIES SET FORTH ABOVE ARE EXCLUSIVE AND IN LIEU OF ALL OTHER WARRANTIES, REMEDIES AND CONDITIONS, WHETHER ORAL OR WRITTEN, STATUTORY, EXPRESS OR IMPLIED. AS PERMITTED BY APPLICABLE LAW, KHN SOLUTIONS SPECIFICALLY DISCLAIMS ANY AND ALL STATUTORY OR IMPLIED WARRANTIES, INCLUDING, WITHOUT LIMITATION, WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE AND WARRANTIES AGAINST HIDDEN OR LATENT DEFECTS. IF KHN SOLUTIONS CANNOT LAWFULLY DISCLAIM STATUTORY OR IMPLIED WARRANTIES THEN TO THE EXTENT PERMITTED BY LAW, ALL SUCH WARRANTIES SHALL BE LIMITED IN DURATION TO THE DURATION OF THE EXPRESS WARRANTY AND TO THE REPAIR

14

OR REPLACEMENT SERVICE AS DETERMINED BY KHN SOLUTIONS IN ITS SOLE DISCRETION

No KHN Solutions reseller, agent, or employee is authorized to make any modification, extension, or addition to this warranty. If any term is held to be illegal or unenforceable, the legality or enforceability of the remaining terms shall not be affected or impaired.

EXCEPT AS PROVIDED IN THIS WARRANTY AND TO THE MAXIMUM EXTENT PERMITTED BY LAW, KHN SOLUTIONS IS NOT RESPONSIBLE FOR DIRECT, SPECIAL, INCIDENTAL OR CONSEQUENTIAL DAMAGES UNDER ANY LEGAL THEORY, INCLUDING BUT NOT LIMITED TO LOSS OF USE; LOSS OF REVENUE OR INCOME; LOSS OF ACTUAL OR ANTICIPATED PROFITS (INCLUDING LOSS OF PROFITS FROM A CONTRACT); LOSS OF THE USE OF MONEY; LOSS OF ANTICIPATED SAVINGS; LOSS OF BUSINESS; LOSS OF OPPORTUNITY; LOSS OF GOODWILL, OR LOSS OF REPUTATION.

This Breath Alcohol Tester is intended to measure alcohol in human breath. Readings obtained by this device are used in the diagnosis of alcohol intoxication. The accuracy of this device has been established at a Blood Alcohol Concentration of 0.008 and 0.032. Accuracy at a Blood Alcohol Concentration greater than .032 has not been established.

Some countries, states and provinces do not allow the exclusion or limitation of incidental or consequential damages or allow

limitations on how long an implied warranty or condition may last, so the above limitations or exclusions may not apply to you. This warranty gives you specific legal rights, and you may also have other rights that vary by country, state or province. This Limited Warranty is governed by and construed under the laws of the country in which the product purchase took place.

Obtaining Warranty Service

Please be sure to read this instruction manual carefully if you believe your product is not operating properly. If you still feel that your product requires warranty service, please follow these instructions:

Obtain a Return Authorization (RA) number by calling toll-free 877-334-6876 or by emailing service@bactrack.com.

When shipping the product back to KHN Solutions, please package the product carefully and ship using a major carrier (UPS, FedEx, USPS, etc). To ensure proper credit for a returned item, be sure to obtain a delivery confirmation on the return shipment. The customer is responsible for all return shipping charges.

Include the following information with your returned product:

- Your Return Authorization number
- Name, address, and phone number as stated at the time of order
- A copy of your original sales receipt

BACTRACK® B70

WWW.BACTRACK.COM

© 2008 KHN Solutions LLC