

INSTRUCTION MANUAL



CD100 Mountable Display **with Keypad**

Revision: 1/14



Copyright © 2011-2014
Campbell Scientific, Inc.

WARRANTY AND ASSISTANCE

This equipment is warranted by CAMPBELL SCIENTIFIC (CANADA) CORP. ("CSC") to be free from defects in materials and workmanship under normal use and service for **twelve (12) months** from date of shipment unless specified otherwise. ***** **Batteries are not warranted.** ***** CSC's obligation under this warranty is limited to repairing or replacing (at CSC's option) defective products. The customer shall assume all costs of removing, reinstalling, and shipping defective products to CSC. CSC will return such products by surface carrier prepaid. This warranty shall not apply to any CSC products which have been subjected to modification, misuse, neglect, accidents of nature, or shipping damage. This warranty is in lieu of all other warranties, expressed or implied, including warranties of merchantability or fitness for a particular purpose. CSC is not liable for special, indirect, incidental, or consequential damages.

Products may not be returned without prior authorization. To obtain a Return Merchandise Authorization (RMA), contact CAMPBELL SCIENTIFIC (CANADA) CORP., at (780) 454-2505. An RMA number will be issued in order to facilitate Repair Personnel in identifying an instrument upon arrival. Please write this number clearly on the outside of the shipping container. Include description of symptoms and all pertinent details.

CAMPBELL SCIENTIFIC (CANADA) CORP. does not accept collect calls.

Non-warranty products returned for repair should be accompanied by a purchase order to cover repair costs.



Campbell Scientific (Canada) Corp.
14532 131 Avenue NW | Edmonton AB T5L 4X4
780.454.2505 | fax 780.454.2655 | campbellsci.ca

PLEASE READ FIRST

About this manual

Please note that this manual was originally produced by Campbell Scientific Inc. (CSI) primarily for the US market. Some spellings, weights and measures may reflect this origin.

Some useful conversion factors:

Area:	1 in ² (square inch) = 645 mm ²
Length:	1 in. (inch) = 25.4 mm
	1 ft (foot) = 304.8 mm
	1 yard = 0.914 m
	1 mile = 1.609 km
Mass:	1 oz. (ounce) = 28.35 g
	1 lb (pound weight) = 0.454 kg
Pressure:	1 psi (lb/in ²) = 68.95 mb
Volume:	1 US gallon = 3.785 litres

In addition, part ordering numbers may vary. For example, the CABLE5CBL is a CSI part number and known as a FIN5COND at Campbell Scientific Canada (CSC). CSC Technical Support will be pleased to assist with any questions.

About sensor wiring

Please note that certain sensor configurations may require a user supplied jumper wire. It is recommended to review the sensor configuration requirements for your application and supply the jumper wire is necessary.

Table of Contents

PDF viewers: These page numbers refer to the printed version of this document. Use the PDF reader bookmarks tab for links to specific sections.

1. Introduction	1
2. Cautionary Statements	1
3. Quickstart	2
4. Overview	2
5. Specifications.....	3
6. Installation	3
7. Operation	6
8. Maintenance	7

Figures

3-1. Connecting the CD100 to the CR1000 Datalogger	2
6-1. Template Measurements	4
6-2. Profile of Installed CD100	5
7-1. Display Menu Overview	7

Tables

7-1. Keyboard Display Key Functions	6
---	---

CD100 Mountable Display with Keypad



1. Introduction

The CD100 is an easy-to-read panel-mountable datalogger data display. It includes a large-format keypad for input. The CD100 is compatible with several Campbell Scientific dataloggers. Refer to Section 5, *Specifications*, for a complete list. When properly installed, the CD100 is weather and dust resistant. As a purchased service, Campbell Scientific offers the option of installing the CD100 into an enclosure.

2. Cautionary Statements

- The CD100 is not compatible with the CR850 datalogger, nor should it be used in tandem with the CR1000KD keyboard and display on any datalogger.
- Weather resistance of the CD100 is dependent on proper installation into the bulkhead of a weather tight enclosure. See Section 6, *Installation*, for details of proper installation.
- The CD100 is easily damaged if disassembled. If disassembly is required, refer to Section 8, *Maintenance*.

3. Quickstart

1. If the CD100 is to be mounted into an enclosure bulkhead, refer to Section 6, *Installation*, for instructions.
2. Using an SC12 cable as shown in FIGURE 3-1, connect the 9-pin CS I/O port of a CR1000 or other compatible datalogger to the 9-pin CS I/O port on the back of the CD100.



FIGURE 3-1. Connecting the CD100 to the CR1000 Datalogger

3. Power the datalogger; the display should power up and be ready for use. Keyboard commands are summarized in Section 7, *Operation*, and in the datalogger manual.

4. Overview

The CD100 is a panel-mountable display with keypad. It facilitates user input and the display of data over a wide temperature range. The vacuum fluorescent element gives the CD100 high contrast, so it is easy to read in daylight conditions.

The CD100 is compatible with the dataloggers listed in Section 5, *Specifications*. It will not operate properly with the CR850 datalogger or in tandem with the CR1000KD on any datalogger.

When properly installed, the polyvinyl face and rear gasket make the CD100 weather resistant and prevent ingress of water and dust. The back of the CD100 is not weather resistant. If used outdoors, or in a condensing environment, the CD100 should be bulkhead mounted as recommended in Section 6, *Installation*.

5. Specifications

Features:

- Operable with gloved hand
- 16 key alpha numeric keypad
- 8 line x 21 character display (64 x 128 pixel)

Compatibility: CR800 (not CR850)
CR1000
CR3000

Interface: CS I/O, SDC address 2

Temperature

Operating: –45° to 85°C

Storage: –50° to 85°C

Transit: 1 hour at –60°C

Power Requirement

Source: 12 Vdc from datalogger CS I/O port

Current Draw

Inactive: 0.02 mA

Active: 140 mA

Certification: CE Compliant

Rating

Water and dust ingress

Installed as

recommended: IP66

Uninstalled: None (vulnerable to water and dust)

Weight: 0.36 kg (0.80 lb)

Dimensions: 176 x 138 x 8 mm (6.9 x 5.4 x 1.4 in)

6. Installation

NOTE As a purchased service, Campbell Scientific offers the option of installing the CD100 into an enclosure.

Install only on a flat, non-porous surface. The supplied gasket must create a weather-tight seal when installed or permanent damage to the display and other enclosure contents may occur.

1. Use the template shipped with the CD100 or create a template using the measurements shown in FIGURE 6-1, *Template Measurements*.
2. Make the required cuts in the bulkhead.
3. Remove cutting debris from bulkhead.
4. Mount the CD100 as shown in FIGURE 6-2, *Profile of Installed CD100*.

- Before tightening screws, adjust the gasket to align it just inside the edges of the display. Uniformly tighten the screws, but do not over-tighten.




FIGURE 6-1. Template Measurements




FIGURE 6-2. Profile of Installed CD100

7. Operation

An overview of keyboard functions is presented in TABLE 7-1, *Keyboard Display Key Functions*, and FIGURE 7-1, *Display Menu Overview*. Detailed descriptions are available in the datalogger manual.

TABLE 7-1. Keyboard Display Key Functions	
Key	Special Function
[2] and [8]	To navigate up and down through the menu list one line at a time
[Enter]	Selects the line or toggles the option of the line the cursor is on
[Esc]	Back up one level in the menu
[Home]	Move cursor to top of the list
[End]	Move cursor to bottom of the list
[Pg Up]	Move cursor up one screen
[Pg Dn]	Move cursor down one screen
[BkSpc]	Delete character to the left
[Shift]	Change alpha character selected
[Num Lock]	Change to numeric entry
[Del]	Delete
[Ins]	Insert/change graph setup
[Graph]	Graph




FIGURE 7-1. Display Menu Overview

8. Maintenance

Active desiccant must be maintained in the datalogger enclosure to avoid water damage to the display electronics, the datalogger, and other contents. No other maintenance is required.

Depending on the environment and weather tightness of the installation, dust may accumulate between the surface window and the underlying display elements. If this occurs, return the display to Campbell Scientific for cleaning.

Do not disassemble the CD100. The printed circuit ribbon connecting the keypad to the display can be easily torn when not removed properly. If the display must be disassembled, take care to grasp the header and not the printed circuit itself. Sheer or torsional stresses must not be put on the printed circuit. The warranty does not cover damages caused by disassembly.

Campbell Scientific Companies

Campbell Scientific, Inc. (CSI)

815 West 1800 North
Logan, Utah 84321
UNITED STATES

www.campbellsci.com • info@campbellsci.com

Campbell Scientific Africa Pty. Ltd. (CSAf)

PO Box 2450
Somerset West 7129
SOUTH AFRICA

www.csafrica.co.za • cleroux@csafrica.co.za

Campbell Scientific Australia Pty. Ltd. (CSA)

PO Box 8108
Garbutt Post Shop QLD 4814
AUSTRALIA

www.campbellsci.com.au • info@campbellsci.com.au

Campbell Scientific do Brasil Ltda. (CSB)

Rua Apinagés, n.br. 2018 — Perdizes
CEP: 01258-00 — São Paulo — SP
BRASIL

www.campbellsci.com.br • vendas@campbellsci.com.br

Campbell Scientific Canada Corp. (CSC)

14532 — 131 Avenue NW
Edmonton AB T5L 4X4
CANADA

www.campbellsci.ca • dataloggers@campbellsci.ca

Campbell Scientific Centro Caribe S.A. (CSCC)

300 N Cementerio, Edificio Breller
Santo Domingo, Heredia 40305
COSTA RICA

www.campbellsci.cc • info@campbellsci.cc

Campbell Scientific Ltd. (CSL)

Campbell Park
80 Hathern Road
Shepshed, Loughborough LE12 9GX
UNITED KINGDOM

www.campbellsci.co.uk • sales@campbellsci.co.uk

Campbell Scientific Ltd. (CSL France)

3 Avenue de la Division Leclerc
92160 ANTONY
FRANCE

www.campbellsci.fr • info@campbellsci.fr

Campbell Scientific Ltd. (CSL Germany)

Fahrenheitstraße 13
28359 Bremen
GERMANY

www.campbellsci.de • info@campbellsci.de

Campbell Scientific Spain, S. L. (CSL Spain)

Avda. Pompeu Fabra 7-9, local 1
08024 Barcelona
SPAIN

www.campbellsci.es • info@campbellsci.es

Please visit www.campbellsci.com to obtain contact information for your local US or international representative.